

A Potential Pitfall in Detection of Diabetes Mellitus Non-invasively Using the Second-Derivative of Photoplethysmography: Coexisting Hypertension May Blunt the Distinctive Difference in Early Negative to Initial Positive Wave Ratio (b/a) Between People With and Without Diabetes

Running Head: 'Fingertip Photoplethysmography Waveforms in Hypertension and Diabetes'

Ahmet Tas*¹ BSc, Yaren Alan¹ BSc, Ilke Kara² BSc, Abdullah Savas¹ BSc, Muhammed Ikbal Bayhan¹ BSc, Diren Ekici¹, Zeynep Atay¹ BSc, Fatih Sezer¹ BSc, Cagla Kitapli¹ MD

Affiliations:

- 1) Istanbul Faculty of Medicine, Istanbul University, Istanbul, Turkey
- 2) Faculty of Medicine, Bahcesehir University, Istanbul, Turkey

Competing interests: All authors have completed the ICMJE uniform disclosure form and declare: no support from any organization for the submitted work; no financial relationships with any organizations that might have an interest in the submitted work in the previous three years; no other relationships or activities that could appear to have influenced the submitted work.

*Corresponding Author:

Ahmet Tas BSc, Intern Doctor
Istanbul University, Istanbul Faculty of Medicine, Istanbul, Turkey
ahmettas.cor@gmail.com
00905515542110
ORCID: 0000-0002-1944-2576

Figures: 1

Word Count: 1868

Keywords: Second-Derivative Photoplethysmography, Diabetes, Non-invasive Cardiovascular Screening, Fingertip Waveforms, Hypertension

Abstract

Background

Digital biomarkers have attracted increasing attention as possible non-invasive diagnostic tools for frequent cardiovascular pathologies. Alterations of features extracted from the derivatives and raw signal of photoplethysmography have been found in the presence of cardiovascular risk factors including but not limited to diabetes, hypertension, arterial stiffness, and endothelial dysfunction. In parallel, recent studies have shown promising results regarding the utilization of PPG as a non-invasive diagnostic tool for diabetes. Whether the presence of hypertension impedes the classification due to mutual endothelial insult resulting in peripheral hemodynamic perturbations remains unknown.

Methods

The previously described ratio between the amplitudes of early-negative(b) and initial-positive(a) waves (b/a) was derived from the second derivative of the PPG signal. Patients were classified according to diabetes mellitus type 2 and hypertension. Standard statistical tests were used to compare the means between unpaired subgroups.

Results

Final analysis included 132 patients. Compared to healthy control group DM2(-)HT(-) (-0.263 ± 0.115), both DM2(+)HT(-) (-0.361 ± 0.122 , $p < 0.001$, $d = 0.83$) and HT(+)DM2(-) (-0.319 ± 0.127 , $p = 0.033$, $d = 0.48$) groups had significantly lower b/a values. However, subgroups with one of two conditions had statistically indifferent mean b/a values (DM2(+)HT(-) (-0.361 ± 0.122) vs HT(+)DM2(-) (-0.319 ± 0.127 , $p = 0.212$, $d = 0.32$). Moreover, significant difference in b/a between DM2 and non-DM2 subgroups of non-HT group ($p < 0.001$) disappeared in subgroups with HT ($p = 0.665$, $d = 0.16$).

Conclusion

The utilization of the second-derivative of PPG for the detection of diabetes non-invasively may be impeded in patients with cardiovascular comorbidities, especially in the hypertensive population since both HT and DM2 induce parallel b/a ratio alterations due to the mutual undesired perturbations in the cardiovascular system, affecting peripheral flow dynamics.

Abbreviations

Diabetes Mellitus: DM ; Hypertension: HT Photoplethysmography: PPG ;Second-derivative of Photoplethysmography: SDPPG

Background

The urge to develop non-invasive screening methods for common cardiovascular diseases continues.

Photoplethysmography (PPG) is a low-cost, practical and widely-available non-invasive tool measuring the volumetric changes in blood flow[1]. Waveforms extracted from the second-derivative of the PPG (SD-PPG) signal, indicating the acceleration characteristics of flow velocity, have been shown to reflect endothelial function[2], hypertension(HT) [3], glucose intolerance[4], dyslipidemia, impaired fasting glucose [3], coronary heart disease risk, age, arterial stiffness, atherosclerosis [5] and diabetes mellitus(DM)[3, 6]. Moreover, multiple studies evaluating the PPG as a potential noninvasive tool to detect blood glucose levels as well as DM have provided encouraging results [7]. The b/a ratio, representing the ratio between the amplitudes of early-negative(b) and initial-positive(a) waves in SD-PPG, was shown to be associated with DM, impaired fasting glucose and impaired glucose challenge test response [3, 4, 6]. In parallel, 'b/a' was a successful predictor of DM in a recent classifier model [6]. However, the same ratio is also related to endothelial function [2], arterial stiffness [5], and ultimately hypertension [3]. Considering the frequent coexistence of HT and DM2, whether the b/a remains useful in the detection of DM2 in the hypertensive population, is an important question to answer. We hypothesize that, due to the mutual undesired endothelial and hemodynamic impacts of HT and DM2, the previously reported difference of b/a between DM2 and non-DM2 groups may disappear in the presence of HT.

Methods

To test our hypothesis, we used an open dataset containing PPG signal and information about DM2, HT, age and sex[8]. Subjects with cerebrovascular pathologies(n=46) were excluded. The second derivative of the PPG signal was computed using MATLAB (MathWorks,USA) and the peaks of the a and b waves were manually extracted and averaged (Figure 1A). For dichotomic assessment, patients with 'stage 1-2 hypertension' were labeled as 'HT' whereas patients originally named as 'normotensive-prehypertensive' were labeled as 'non-HT'.

Since the mean ages, which is known to affect the b/a ratio, were different between DM2(-/+ non-HT subgroups, we formed an age-matched control group of HT(-)DM2(-) prior to analysis, by including only the subjects whose ages were within the mean \pm SD age range of HT(-)DM(+). In the HT group, mean ages were indifferent between DM2 and non-DM2 subgroups, thus no adjustments were made. Data was analyzed offline using SPSS (IBM,USA). Continuous variables were expressed as mean \pm SD. Normality was assessed with the Shapiro-Wilk test. Means were compared using the Student's t and Mann-Whitney U tests accordingly. Cohen's *d* was used to interpret effect sizes. $P < 0.05$ was considered statistically significant.

The present study was performed on a previously published free-use open-dataset containing data from deidentified patients [8]. Hence, ethical approval was not sought.

Results

Figure 1B demonstrates the main findings. The final analysis included 132 patients. Three patients with inadequate signal quality were excluded. Fifty-one percent of patients were male. In the non-HT group, there were twenty-eight patients with DM2 and sixty-six patients without DM2, whereas in the HT group, there were eight patients with DM2 and thirty patients without DM2. Mean ages were indifferent between non-hypertensive patients with/without DM2 (57.3 ± 12.1 vs 53.8 ± 9.2 , $p:0.228$) and between hypertensive patients with/without DM2 (63.4 ± 10.2 vs 60.9 ± 10.7 , $p:0.555$).

In the non-HT group, patients with DM2 had significantly lower b/a ratio compared to those without DM2 (-0.361 ± 0.122 vs -0.263 ± 0.115 , $p<0.001$, $d=0.83$). In parallel, in the non-DM2 group, patients with hypertension had significantly lower b/a values compared to those without HT (-0.319 ± 0.127 vs -0.263 ± 0.115 , $p:0.033$, $d:0.48$), however, in the HT group the mean b/a ratio was indifferent between patients with and without DM2 (-0.341 ± 0.122 vs -0.319 ± 0.127 , $p:0.665$, $d:0.16$). Likewise, mean the b/a ratios were indifferent between DM2(+)HT(-) and HT(+)DM2(-) groups ($p:0.212$, $d:0.32$) meaning that subgroups with one of the two studied cardiovascular risk factors had statistically inseparable mean b/a ratios.

Discussion

The present study has two major findings. Firstly, the patients with DM2 or HT had significantly lower b/a ratios compared to healthy subjects (DM(-)HT(-)) which is in general consistent with the literature [7]. Importantly, our findings should be dealt with caution when comparing them to those of previous studies, since the reported (b/a) ratios may be absolute values as in [9]. Secondly, patients who had either DM2 or HT had similar b/a ratios, which can be justified by a miscellanea of preceding studies showing the relationships between the b/a ratio, and arterial stiffness[5], hypertension[10], impaired glucose tolerance[11] and fasting glucose as well as dyslipidemia[12]. Moreover, although the difference between means was statistically insignificant between HT(+)DM2(+) and HT(+)DM2(-) subgroups, the small sample size and numerical tendency towards lower values in the overlap group may encourage further research with the focus of overlapping cardiovascular risk factors. While the former finding is confirmatory in its nature, the latter result uniquely indicates a potential pitfall in detection of diabetes non-invasively, using digital biomarkers with the second-derivative of photoplethysmography.

Conclusion

The utilization of the second-derivative of PPG for the detection of diabetes non-invasively may be impeded in patients with cardiovascular comorbidities, especially in the hypertensive population since both HT and DM2 induce parallel b/a ratio alterations due to the mutual undesired perturbations in the cardiovascular system, affecting peripheral flow dynamics.

Competing interests

All authors have completed the ICMJE uniform disclosure form and declare: no support from any organization for the submitted work; no financial relationships with any organizations that might have an interest in the submitted work in the previous three years; no other relationships or activities that could appear to have influenced the submitted work.”

References

1. Avram R, Olgin JE, Kuhar P, et al (2020) A digital biomarker of diabetes from smartphone-based vascular signals. *Nature Medicine* 2020 26:10 26:1576–1582. <https://doi.org/10.1038/s41591-020-1010-5>
2. Scholze A, Rinder C, Beige J, et al (2004) Acetylcysteine Reduces Plasma Homocysteine Concentration and Improves Pulse Pressure and Endothelial Function in Patients With End-Stage Renal Failure. *Circulation* 109:369–374. <https://doi.org/10.1161/01.CIR.0000109492.65802.AD>
3. Otsuka T, Kawada T, Katsumata M, et al (2007) Independent determinants of second derivative of the finger photoplethysmogram among various cardiovascular risk factors in middle-aged men. *Hypertens Res* 30:1211–1218. <https://doi.org/10.1291/HYPRES.30.1211>
4. Ohshita K, Yamane K, Ishida K, et al (2004) Post-challenge hyperglycaemia is an independent risk factor for arterial stiffness in Japanese men. *Diabetic Medicine* 21:636–639
5. Park J, Seok HS, Kim SS, Shin H (2022) Photoplethysmogram Analysis and Applications: An Integrative Review. *Front Physiol* 12:2511. <https://doi.org/10.3389/FPHYS.2021.808451/BIBTEX>
6. Qawqzeh YK, Bajahzar AS, Jemmali M, et al (2020) Classification of Diabetes Using Photoplethysmogram (PPG) Waveform Analysis: Logistic Regression Modeling. *Biomed Res Int* 2020:. <https://doi.org/10.1155/2020/3764653>
7. Pablo Martínez J, Zanelli S, Ammi M, et al (2022) Diabetes Detection and Management through Photoplethysmographic and Electrocardiographic Signals Analysis: A Systematic Review. <https://doi.org/10.3390/s22134890>
8. Liang Y, Chen Z, Liu G, Elgendi M (2018) A new, short-recorded photoplethysmogram dataset for blood pressure monitoring in China. *Scientific Data* 2018 5:1 5:1–7. <https://doi.org/10.1038/sdata.2018.20>
9. Usman S, Harun N, Dziyauddin RA, Bani NA (2018) Estimation of hba1c level among diabetic patients using second derivative of photoplethysmography. *IEEE Student Conference on Research and Development: Inspiring Technology for Humanity, SCORED 2017 - Proceedings* 2018-January:89–92. <https://doi.org/10.1109/SCORED.2017.8305415>

10. Hashimoto J, Chonan K, Aoki Y, et al (2002) Pulse wave velocity and the second derivative of the finger photoplethysmogram in treated hypertensive patients: their relationship and associating factors. *J Hypertens* 20:2415–2422. <https://doi.org/10.1097/00004872-200212000-00021>
11. Ohshita K, Yamane K, Ishida K, et al (2004) Post-challenge hyperglycaemia is an independent risk factor for arterial stiffness in Japanese men. *Diabet Med* 21:636–639. <https://doi.org/10.1111/J.1464-5491.2004.01161.X>
12. Otsuka T, Kawada T, Katsumata M, et al (2007) Independent determinants of second derivative of the finger photoplethysmogram among various cardiovascular risk factors in middle-aged men. *Hypertens Res* 30:1211–1218. <https://doi.org/10.1291/HYPRES.30.1211>
13. Otsuka T, Kawada T, Katsumata M, Ibuki C (2006) Utility of second derivative of the finger photoplethysmogram for the estimation of the risk of coronary heart disease in the general population. *Circ J* 70:304–310. <https://doi.org/10.1253/CIRCJ.70.304>

Figure Legends

Fig.1 Example of PPG Waveforms (A) and Comparison of b/a between study subgroups(B). Compared to healthy control group [DM2(-)HT(-): -0.263 ± 0.115], both [DM2(+)HT(-) -0.361 ± 0.122 , , $p < 0.001$, $d = 0.83$] and [HT(+)DM2(-) -0.319 ± 0.127 , $p: 0.033$, $d = 0.48$] groups had significantly lower b/a values. However, subgroups with one of two conditions have statistically indifferent mean b/a values (DM2(+)HT(-) -0.361 ± 0.122 vs HT(+)DM2(-) -0.319 ± 0.127 , $p: 0.212$, $d: 0.32$). Moreover, significant difference in b/a between DM2 and non-DM2 subgroups of non-HT group ($p < 0.001$) disappears in subgroups with HT($p: 0.665$, $d: 0.16$).

Figures

Fig.1

