

Comparative analysis of loop-mediated isothermal amplification (LAMP)-based assays for rapid detection of SARS-CoV-2 genes

Daniel Urrutia-Cabrera^{1,2,*}, Roxanne Hsiang-Chi Liou^{1,2,*}, Jianxiong Chan^{3,4}, Sandy Shen-Chi Hung^{1,2}, Alex W Hewitt^{1,2}, Keith R Martin^{1,2}, Patrick Kwan^{3,4}, Raymond Ching-Bong Wong^{1,2,#}

¹ Centre for Eye Research Australia, Royal Victorian Eye and Ear Hospital. Australia

² Ophthalmology, Department of Surgery, University of Melbourne, Australia

³ Department of Neuroscience, Central Clinical School, Monash University, Australia

⁴ Departments of Medicine and Neurology, University of Melbourne, Royal Melbourne Hospital, Australia

* Equal contribution as first author

Corresponding author

Correspondences: Dr Raymond C.B. Wong, Centre for Eye Research Australia, Royal Victorian Eye and Ear Hospital, Level 7 Smorgon Family Wing, 32 Gisborne St, East Melbourne, VIC 3002, Australia; Phone: +613 99298480; email: wongcb@unimelb.edu.au

Abstract

The COVID-19 pandemic caused by SARS-CoV-2 has infected millions worldwide and there is an urgent need to increase our diagnostic capacity to identify infected cases. Although RT-qPCR remains the gold standard for SARS-CoV-2 detection, this method requires specialised equipment in a diagnostic laboratory and has a long turn-around time to process the samples. To address this, several groups have recently reported development of loop-mediated isothermal amplification (LAMP) as a simple, low cost and rapid method for SARS-CoV-2 detection. Herein we present a comparative analysis of three LAMP-based assays that target different regions of the SARS-CoV-2: ORF1ab RdRP, ORF1ab nsp3 and Gene N. We perform a detailed assessment of their sensitivity, kinetics and false positive rates for SARS-CoV-2 diagnostics in LAMP or RT-LAMP reactions, using colorimetric or fluorescent detection. Our results independently validate that all three assays can detect SARS-CoV-2 in 30 minutes, with robust accuracy at detecting as little as 1000 RNA copies and the results can be visualised simply by color changes. We also note the shortcomings of these LAMP-based assays, including variable results with shorter reaction time or lower load of SARS-CoV-2, and false positive results in some experimental conditions. Overall for RT-LAMP detection, the ORF1ab RdRP and ORF1ab nsp3 assays have higher sensitivity and faster kinetics for detection, whereas the Gene N assay exhibits no false positives in 30 minutes reaction time. This study provides validation of the performance of LAMP-based assays for SARS-CoV-2 detection, which have important implications in development of point-of-care diagnostic for SARS-CoV-2.

Keywords: COVID-19, SARS-CoV-2, LAMP, RT-LAMP, coronavirus diagnostics

Short title: Comparison of SARS-CoV-2 detection using LAMP

Introduction

The outbreak of Severe Acute Respiratory Syndrome Coronavirus 2 (SARS-CoV-2) has caused the catastrophic COVID-19 pandemic, which has infected >63 million and caused >1.4 million deaths worldwide to date (Johns Hopkins Coronavirus Resource Centre). Reverse transcription quantitative PCR (RT-qPCR) remained the gold standard diagnostic test for SARS-CoV-2 in many countries, with standard diagnostic protocol from the World Health Organisation. However, this procedure requires processing at a clinical diagnostic laboratory which would take hours or days, depending on the workload at the test site. Given the high infectious rate of SARS-CoV-2 with >500,000 daily new cases globally, there is an urgent need to develop scalable, low cost detection methods for SARS-CoV-2 to increase our capacity to perform daily testing.

Loop-mediated isothermal amplification (LAMP) is a one-step nucleic acid amplification method that has been applied for diagnostic of infectious diseases [1]. The LAMP reaction involves auto-cycling strand displacement DNA synthesis, using hairpin-forming LAMP primers which anneal to the target DNA template and a DNA polymerase with strand displacement activity. These annealed LAMP primers are in turn displaced by displacement primers for subsequent amplification and elongation [2]. The LAMP reaction is highly sensitive, specific and only requires one set temperature for nucleic acid amplification, providing an attractive technology for the development of a low cost, point-of-care diagnostics for SARS-CoV-2. Recently, there has been significant development on the use of LAMP for detection of SARS-CoV-2. Several LAMP assays have been developed that target different regions of SARS-CoV-2, including RNA-dependent RNA polymerase (RdRP) [3], non-structural protein 3 (nsp3) [4] and the nucleocapsid (N) gene [5]. However, how these LAMP assays perform relative to each other remain unclear and understanding this would facilitate development of a robust LAMP assay for SARS-CoV-2 detection.

In this study, we perform a comparative analysis of three of the earliest LAMP assays developed for SARS-CoV-2 detection [3–5]. We perform a head-to-head comparison of the sensitivity and kinetics to detect the RNA or cDNA of SARS-CoV-2 genes, as well as their capability to be used as colorimetric or fluorescent detection.

Materials and Methods

Molecular reagents

For positive DNA control, two regions in SARS-CoV-2 ORF1, nt.2945-3370 and nt.14971-15970, were synthesized as gBLOCK double stranded DNA fragments (Integrated DNA

Technologies), as well as the 2019-nCoV_N positive control plasmid carrying the SARS-CoV-2 Gene N (GenBank: NC_045512.2, Integrated DNA Technologies).

For positive RNA control, we used the Twist Synthetic SARS-CoV-2 RNA control which provides coverage of >99.9% of the viral genome in six non-overlapping 5kb single stranded RNA fragments (GeneBank: MT007544.1, Twist Bioscience).

RT-LAMP/LAMP colorimetric assay

Positive DNA or RNA controls were diluted to various concentrations and used in LAMP and RT-LAMP reactions respectively. The sequences of primers and controls used are listed in Supplementary table 1. The six primers (F3, B3, FIP, BIP, Loop F and Loop B) were premixed as a 10X working stock containing 2 μ M of each outer primer (F3 & B3), 16 μ M of each inner primer (FIP & BIP), and 4 μ M of each loop primer (Loop F & Loop B).

For colorimetric detection, a 20 μ l reaction was set up containing 10 μ l of WarmStart Colorimetric LAMP 2X Master Mix (New England Biolabs), 2 μ l primer mix, 1 μ l RNA/DNA control and 7 μ l nuclease-free water. The mixed reactions were incubated at 65°C using a heat block and pictures for colorimetric changes were taken from 10-45 minutes. Subsequently, the amplified samples were run on a 1% agarose gel to determine amplicon specificity.

RT-LAMP fluorescent assay

Same sets primer and positive RNA controls were used for the RT-LAMP colorimetric assay. For fluorescent detection of RT-LAMP, a 20 μ l reaction was set up containing 10 μ l of WarmStart Colorimetric LAMP 2X Master Mix (New England Biolabs), 2 μ l primer mix, 1 μ l RNA control, 1 μ l of 10X SYBR green (Thermo Fisher) and 7 μ l nuclease-free water. The mixed reactions were incubated at 65°C using a StepOne Real Time PCR machine (Thermo Fisher) for 60 minutes and fluorescence measurements were taken every 2.5 minutes.

To quantify the kinetics of the RT-LAMP assay, the ΔR_n values are extracted and plotted against the 2.5 timepoints. A threshold corresponding to half the maximum ΔR_n value is set to determine the time required to reach half the maximum fluorescence intensity, termed ' $\frac{1}{2} \Delta R_{n_{max}}$ '. If there is no detected signal after 60 minutes reaction time, the $\frac{1}{2} \Delta R_{n_{max}}$ is listed as 60 minutes.

Results

Comparative study of LAMP reactions using colorimetric detection

For this study, we used the WarmStart Colorimetric RT-LAMP 2X Master Mix (New England Biolabs) containing a warm-start reverse transcriptase RTx and an isothermal DNA polymerase Bst 2.0, which is capable of both LAMP and RT-LAMP. The reaction mix also contains a pH indicator, which allows visualisation of amplification as a result of protons produced by the LAMP reaction and lead to a red-to-yellow colour change. We selected primer sets reported by three previous studies that targeted different regions of the SARS-CoV-2 genome (Supplementary table 1). The 'ORF1ab RdRP' set [3] and the 'ORF1ab nsp3' set [4] targeting RdRP and nsp3 sequences encoded by ORF1ab respectively, as well as the 'Gene N' set which targets the nucleocapsid gene [5].

We first assessed the capability of these three primer sets in colorimetric LAMP reactions, using synthesized DNA of SARS-CoV-2 genes as control. Our results showed that the three primer sets are all capable of LAMP detection of the SARS-CoV-2 genes, albeit with some differences in performances (Figure 1). In terms of kinetics for the colorimetric LAMP reaction, visible colour changes can be seen as early as 20 minutes for ORF1ab RdRP and Gene N, compared to 25 minutes for ORF1ab nsp3. In order to determine the sensitivity of the LAMP assays, we performed LAMP reactions with a dilution series ranging from 100 to 10000 copies of DNA molecules. Overall the lowest detection limit is 100 copies of DNA molecules for both ORF1ab nsp3 and Gene N, compared to 500 copies of DNA molecules for ORF1ab RdRP (Figure 1). In addition, gel electrophoresis showed a discrete band representing the LAMP amplicons (Figure 2A-C), supporting the specificity of the LAMP amplification for all three primer sets. Collectively, all three primer sets are capable of colorimetric LAMP detection of SARS-CoV-2 within 30 minutes, with slightly faster kinetics for the ORF1ab RdRP/Gene N and slightly higher sensitivity for ORF1ab nsp3.

Comparative study of RT-LAMP reactions using colorimetric detection

Next, we compared the three primer sets in colorimetric RT-LAMP reactions, using in vitro transcribed RNA of SARS-CoV-2 genes as control. Similar to our LAMP results, the ORF1ab RdRP is the fastest out of the three with visible colour change after 20 minutes of reaction, whereas ORF1ab nsp3 and Gene N take 25 minutes and 30 minutes respectively (Figure 3). In terms of sensitivity, the lowest detection limit is observed in ORF1ab RdRP (100 copies of RNA molecules), followed by ORF1ab nsp3 and Gene N (both 500 copies of RNA molecules, Figure 3). However for the ORF1ab RdRP and nsp3 primer sets, we noticed that the no template control sometimes exhibits colour change indicative of a false-positive, while this was not observed for the Gene N set (Table 1).

Correspondingly, analysis by gel electrophoresis revealed a clear band of the correct RT-LAMP reaction product in lanes with ≥ 100 copies RNA input for ORF1ab RdRP, and ≥ 500 copies RNA input for both ORF1ab nsp3 and Gene N (Figure 4A-C). These results support the specificity of the RT-LAMP amplification using the three primer sets. Overall, the three primer sets perform similarly in LAMP or RT-LAMP for colorimetric detection, with only subtle differences in sensitivity and kinetics.

Assessment of the kinetics and sensitivity of RT-LAMP reactions using fluorescent detection

Although colorimetric detection represents an easy detection method for point-of-care diagnostic development, subtle colour changes may not be obvious with naked eye. To more accurately assess the sensitivity and kinetics of the three primer sets, we added SYBR Green I into the RT-LAMP reactions which allows fluorescent quantification using a qPCR machine. For all three primer sets, increase in the fluorescent signal starts around 12-15 minutes. The fluorescent signal reached a plateau around 30-40 minutes for ORF1ab RdRP and ORF1ab nsp3 (Figure 5A, C), indicating that 40 minutes of reaction is sufficient for these two primer sets. However, Gene N exhibits different amplification curves depending on the copy number of the RNA molecules: ~30 minutes should be sufficient for detection of ≥ 500 copies, whereas >45 minutes is needed for detection of 100 copies of RNA molecules (Figure 5E).

To compare the kinetics of RT-LAMP across the three primer sets, we measure the time required to reach half the maximum fluorescence signal ($\frac{1}{2} \Delta Rn_{max}$). We noted that $\frac{1}{2} \Delta Rn_{max}$ is generally consistent for detection of ≥ 1000 copies of RNA molecules, whereas detection of 100 or 500 copies of RNA molecules yielded variable $\frac{1}{2} \Delta Rn_{max}$ values in biological repeats. This suggested the three primer sets perform more robustly for detection of higher copies of SARS-CoV-2 RNA. For detection of 1000 copies of RNA molecules, ORF1ab nsp3 exhibit the fastest kinetics ($\frac{1}{2} \Delta Rn_{max} = 17.5$ min), followed by ORF1ab RdRP ($\frac{1}{2} \Delta Rn_{max} = 26.25$ min) and Gene N ($\frac{1}{2} \Delta Rn_{max} = 27.5$ min). However, for detection of 10000 copies of RNA molecules, all three primer sets exhibit comparable kinetics ($\frac{1}{2} \Delta Rn_{max} = 17.5 - 21.25$ min) (Figure 5B, D, F). Collectively, our results suggest that the three RT-LAMP assays have comparable kinetics for detection of high copies of SARS-CoV-2, but variable performances for detection of low copies of SARS-CoV-2.

False positive rates for RT-LAMP/LAMP reactions

In our testing, occasionally we detected positive signals in the no template controls which are indicative of false positives. Table 1 lists the false positive rate for the three primer sets in RT-LAMP and LAMP reactions. In general, we observed lower false positive rates

in shorter reaction time (15 minutes) and higher false positive rates in longer reaction time (45 minutes), thus we recommend using a reaction time of 30 minutes. Notably in a 30 minutes RT-LAMP reaction, the lowest false positive rate is observed in Gene N (0/5, 0%), followed by ORF1ab nsp3 (2/5, 40%) and ORF1ab RdRP (4/5, 80%). These results suggest that the Gene N primer set has the best accuracy for detection among the three tested assays.

Discussion

Active cases of SARS-CoV-2 can be diagnosed by detection of the viral RNA or antigens in patient samples. While antigen-based detection methods are cheap and rapid as a point-of-care diagnostics, they are inherently less sensitive compared to viral RNA detection methods given there is no amplification of protein signals. As such, the sensitivity of many antigen-based detection methods ranges from 50-90% [6]. In contrast, viral RNA detection methods, such as RT-qPCR remain the gold standard for SARS-CoV-2 diagnostics with ~98% sensitivity. However, the requirement of sample processing in a clinical diagnostic laboratory results in long turnaround time for diagnostic results and this remained the key bottleneck to upscale our capability to perform daily SARS-CoV-2 testing.

Recent development of RT-LAMP assays could address this key issue, providing a rapid and low-cost detection method for SARS-CoV-2. In this study, we compared and validated three LAMP-based assays previously developed to detect different parts of SARS-CoV-2: ORF1ab RdRP, ORF1ab nsp3 and Gene N. Our results show that the three RT-LAMP assays allow colorimetric detection of SARS-CoV-2 genes in 30 minutes, with robust accuracy at detecting 1000 RNA copies. Also, the three assays perform similarly for detection of RNA or DNA copies of SARS-CoV-2. However, we observed some variability in these RT-LAMP assays with shorter reaction time and detection of lower load of SARS-CoV-2 RNA. Also, in our testing we observed false positive results in some experimental repeats. Overall for RT-LAMP, our comparative analysis showed that the ORF1ab RdRP and ORF1ab nsp3 sets have higher sensitivity and faster kinetics for detection, whereas the Gene N primer set exhibits no false positives in 30 minutes reaction time. Future optimisation of the primer sets and modification of RT-LAMP reactions would be important to develop an accurate and sensitive detection method for SARS-CoV-2, such as the use of quenched fluorescent primers to reduce the false positive rates of RT-LAMP [7]. There are also ongoing development of new RT-LAMP-based assays to improve performance of SARS-CoV-2 detection, including the use of a two-stage isothermal amplifications (Penn-RAMP)[8], integration of CRISPR with RT-LAMP [9,10], as well as barcoded RT-LAMP reactions to allow high-throughput processing using next-generation sequencing (LAMP-seq) [11]. Future validation of these

new RT-LAMP-based assays would be important to translate these technologies into clinical testing in the real world.

Although false positives remain an issue, there are several key advantages of the RT-LAMP detection methods. The major advantage is the rapid reaction time of ~30 minutes, which is far superior compared to RT-qPCR with a turn-around time of several hours to days. Secondly, the readout of RT-LAMP is simple and can be interpreted easily by naked eye, either as a change in colour (this study) or turbidity [12]. Finally, RT-LAMP only requires one isothermal step which makes it simple to develop low cost, portable devices to process the samples. Recently, several companies have developed portable devices to carry out point-of-care RT-LAMP testing, such as DNAfit LifeSciences (UK) and Lucira Health (USA). Moreover, a recent study also demonstrated that RT-LAMP can be successfully conducted in a kitchen range oven, providing a low-cost solution to carry out RT-LAMP testing at home [13]. Further improvement to RT-LAMP would potentially provide an affordable point-of-care detection method for routine SARS-CoV-2 detection.

Conclusions

In summary, this study shows a detailed assessment of the performances of LAMP-based assays for detection of SARS-CoV-2, providing important validation data that would facilitate further development of LAMP-based diagnostics for SARS-CoV-2 and translation into real world testing.

Abbreviations: SARS-CoV-2: Severe Acute Respiratory Syndrome Coronavirus 2;
LAMP: loop-mediated isothermal amplification; RT-LAMP: Reverse transcription loop-mediated isothermal amplification; RT-qPCR: Reverse transcription quantitative PCR;
RdRP: RNA-dependent RNA polymerase; nsp3: non-structural protein 3; Gene N: Nucleocapsid gene.

Declarations

Ethics approval and consent to participate: Not applicable

Consent to publication: Not applicable

Availability of data and materials: The datasets in the current study are available from the corresponding author on reasonable request.

Competing interests: The authors declare no conflict of interest.

Funding: RCBW is supported by the National Health and Medical Research Council, the University of Melbourne and the Centre for Eye Research Australia. DU and RL are supported by the Melbourne Research Scholarship from the University of Melbourne. PK is supported by a Medical Research Future Fund Fellowship (MRF1136427). The Centre for Eye Research Australia receives operational infrastructure support from the Victorian Government.

Authors' contributions: RCBW, JC and PK designed the experiments; DUC, RHCL conducted the experiments; DUC, RHCL, RCBW, SH, AWH, KM analysed the data; SH, AH, KM, RCBW provided funding to this work; DUC, RHCL and RCBW wrote the manuscript. All authors approved the manuscript.

Acknowledgements: Not applicable.

References

1. Becherer L, Borst N, Bakheit M, Frischmann S, Zengerle R, von Stetten F. Loop-mediated isothermal amplification (LAMP) – review and classification of methods for sequence-specific detection [Internet]. *Analytical Methods*. 2020. p. 717–46. Available from: <http://dx.doi.org/10.1039/c9ay02246e>
2. Notomi T, Okayama H, Masubuchi H, Yonekawa T, Watanabe K, Amino N, et al. Loop-mediated isothermal amplification of DNA. *Nucleic Acids Res*. 2000;28:E63.

3. Yu L, Wu S, Hao X, Dong X, Mao L, Pelechano V, et al. Rapid Detection of COVID-19 Coronavirus Using a Reverse Transcriptional Loop-Mediated Isothermal Amplification (RT-LAMP) Diagnostic Platform. *Clin Chem*. 2020;66:975–7.
4. Lamb LE, Bartolone SN, Ward E, Chancellor MB. Rapid detection of novel coronavirus/Severe Acute Respiratory Syndrome Coronavirus 2 (SARS-CoV-2) by reverse transcription-loop-mediated isothermal amplification. *PLoS One*. 2020;15:e0234682.
5. Zhang Y, Odiwuor N, Xiong J, Sun L, Nyaruaba RO, Wei H, et al. Rapid Molecular Detection of SARS-CoV-2 (COVID-19) Virus RNA Using Colorimetric LAMP [Internet]. *Infectious Diseases (except HIV/AIDS)*. medRxiv; 2020. p. 1453. Available from: <https://www.medrxiv.org/content/10.1101/2020.02.26.20028373v1>
6. Service R, Robert Service. Coronavirus antigen tests: quick and cheap, but too often wrong? [Internet]. *Science*. 2020. Available from: <http://dx.doi.org/10.1126/science.abc9586>
7. Hardinge P, Murray JAH. Reduced False Positives and Improved Reporting of Loop-Mediated Isothermal Amplification using Quenched Fluorescent Primers. *Sci Rep*. 2019;9:7400.
8. El-Tholoth M, Bau HH, Song J. A Single and Two-Stage, Closed-Tube, Molecular Test for the 2019 Novel Coronavirus (COVID-19) at Home, Clinic, and Points of Entry. *ChemRxiv* [Internet]. 2020; Available from: <http://dx.doi.org/10.26434/chemrxiv.11860137>
9. Broughton JP, Deng X, Yu G, Fasching CL, Servellita V, Singh J, et al. CRISPR-Cas12-based detection of SARS-CoV-2. *Nat Biotechnol*. 2020;38:870–4.
10. Joung J, Ladha A, Saito M, Kim N-G, Woolley AE, Segel M, et al. Detection of SARS-CoV-2 with SHERLOCK One-Pot Testing. *N Engl J Med*. 2020;383:1492–4.
11. Schmid-Burgk JL, Schmithausen RM, Li D, Hollstein R, Ben-Shmuel A, Israeli O, et al. LAMP-Seq: Population-Scale COVID-19 Diagnostics Using Combinatorial Barcoding [Internet]. Available from: <http://dx.doi.org/10.1101/2020.04.06.025635>
12. Arunrut N, Suebsing R, Withyachumnarnkul B, Kiatpathomchai W. Demonstration of a very inexpensive, turbidimetric, real-time, RT-LAMP detection platform using shrimp Laem-Singh virus (LSNV) as a model. *PLoS One*. 2014;9:e108047.
13. Lei Y. Kitchen Range Oven Enabled One-tube RT-LAMP for RNA Detection at Home - A Potential Solution for Large-scale Screening of COVID-19 [Internet]. Available from: <http://dx.doi.org/10.31224/osf.io/ed85s>

Figures




Figure 1: Colorimetric detection of SARS-CoV-2 DNA using LAMP. LAMP reactions are assessed at different time points using a dilution series of positive DNA control ranging from 100 copies to 10000 copies. NTC (no template control) is used as a negative control. n=3 biological repeats.




Figure 2: Gel electrophoresis analysis of LAMP reaction. Representative gel picture is shown for A) ORF1ab RdRP, B) ORF1ab nsp3 and C) Gene N in LAMP assays, using a dilution series of positive DNA control ranging from 100 copies to 10000 copies. NTC (no template control) is used as a negative control. n=3 biological repeats.




Figure 3: Colorimetric detection of SARS-CoV-2 RNA using RT-LAMP. RT-LAMP reactions are assessed at different time points using a dilution series of positive RNA control ranging from 100 copies to 10000 copies. NTC (no template control) is used as a negative control. n = 3 biological repeats.




Figure 4: Gel electrophoresis analysis of RT-LAMP reaction. Representative gel picture is shown for A) ORF1ab RdRP, B) ORF1ab nsp3 and C) Gene N in RT-LAMP assays, using a dilution series of positive RNA control ranging from 100 copies to 10000 copies. NTC (no template control) is used as a negative control. n = 3 biological repeats.




Figure 5: Fluorescent detection of SARS-CoV-2 RNA using RT-LAMP. Real-time measurement of SARS-CoV-2 RNA using RT-LAMP for A) ORF1ab RdRP, C) ORF1ab nsp3 and E) Gene N, showing the normalised fluorescent signal (ΔR_n) against the reaction time. Quantification of the time to reach half the maximum fluorescent signal ($\Delta R_{n_{max}}$) for B) ORF1ab RdRP, D) ORF1ab nsp3 and F) Gene N. Results are displayed as the mean of 2 biological repeats, error bars represented SEM.

Table 1: Summary of false positives detected in no template control in individual biological repeats.

Primer sets	Assay	Total false positive	False positive after 15 min	False positive after 30 min	False positive after 45 min
ORF1ab RdRP	LAMP	4/6 (66%)	0/6 (0%)	3/6 (50%)	4/6 (66%)
	RT-LAMP	4/5 (80%)	1/5 (20%)	4/5 (80%)	4/5 (80%)
ORF1ab nsp3	LAMP	2/6 (33%)	0/6 (0%)	1/6 (16%)	2/6 (33%)
	RT-LAMP	3/5 (60%)	1/5 (20%)	2/5 (40%)	3/5 (60%)
Gene N	LAMP	3/6 (50%)	0/6 (0%)	0/6 (0%)	3/6 (50%)
	RT-LAMP	0/5 (0%)	0/5 (0%)	0/5 (0%)	0/5 (0%)

Supplementary table 1: Information of primers and control gBLOCK used in this study.

Primer	Sequence	Reference
RW244 nsp3 F3	TCCAGATGAGGATGAAGAAGA	[4]
RW245 nsp3 B3	AGTCTGAACAACTGGTGTAAG	[4]
RW246 nsp3 FIP(F1c+F2)	AGAGCAGCAGAAGTGGCACAGGTGATTGTGAAGAAGAAGAG	[4]
RW247 nsp3 BIP(B1c+B2)	TCAACCTGAAGAAGAGCAAGAAGCTGATTGTCCTCACTGCC	[4]
RW248 nsp3 LoopF	CTCATATTGAGTTGATGGCTCA	[4]
RW249 nsp3 LoopB	ACAAACTGTTGGTCAACAAGAC	[4]
RW250 RdRp F3	CCACTAGAGGAGCTACTGTA	[3]
RW251 RdRp B3	TGACAAGCTACAACACGT	[3]
RW252 RdRp FIP	AGGTGAGGGTTTTCTACATCACTATATTGGAACAAGCAAATTCTA TGG	[3]
RW253 RdRp BIP	ATGGGTTGGGATTATCCTAAATGTGTGCGAGCAAGAACAAGTG	[3]
RW254 RdRp LF	CAGTTTTTAACATGTTGTGCCAACC	[3]
RW255 RdRp LB-4	TAGAGCCATGCCTAACATGCT	[3]
RW256 Gene N-A-F3	TGGCTACTACCGAAGAGCT	[5]
RW257 Gene N-A-B3	TGCAGCATTGTTAGCAGGAT	[5]
RW258 Gene N-A-FIP	TCTGGCCCAGTTCCTAGGTAGTCCAGACGAATTCGTGGTGG	[5]
RW259 Gene N-A-BIP	AGACGGCATCATATGGGTTGCACGGGTGCCAATGTGATCT	[5]
RW260 Gene N-A-LF	GGACTGAGATCTTTTCATTTTACCGT	[5]
RW261 Gene N-A-LB	ACTGAGGGAGCCTTGAATACA	[5]
Positive controls	Sequence	References
RWgb80 ORF1ab RdRp 14971-15970	aaggctagactttattatgattcaatgagttatgaggatcaagatgcactttcgcataatacaaaa cgtaatgcatccctactataactcaaatgaatcttaagtatgccattagtcaaaagaatagagc tcgaccgtagctggtgtctctatctgtactatgaccaatagacagttcatcaaaaattatg aaatcaatagccgccaactagaggactactgtagtaattggaacaagcaaatctatggtggtt ggcacaacatgttaaaaactgtttatagtgatgtagaaaaccctcaccttatgggtgggattac ctaaatgtagatagccatgcctaacatgcttagaattatggcctcactgttctgctcgaaac atacaacgtgttagctgtcacaccgtttctatagattagctaatgagtgctcaagattgagt gaaatggtcatgtgtggcggtcactatattgtaaac	This study
RWgB81 ORF1ab nsp3 2945-3370	ggcattgattagatgagtgagtgatggctacataactacttattgatgagtgctggtgagttaaatt ggcttcacatatgtattgttcttctaccctccagatgaggatgaagaagaagggtattgtgaaga agaagagtttagccatcaactcaatatgagtgatggtactgaagatgattaccaaggtaaacct ttgaaattggtgccaactctgctgctctcaacctgaagaagagcaagaagaagattggttaga tgatgatagcaaaaactgtgtgcaacaagacggcagtgaggacaatcagacaactacta ttcaaaccaattgtgaggttcaacctcaattagagatggaacttacaccagttgtcagactattg aagtgaatagttttaggttatttaaaactact	This study