## From Community Acquired Pneumonia to COVID-19: A Deep Learning Based Method for Quantitative Analysis of COVID-19 on thick-section CT Scans

Zhang Li<sup>1</sup>, Zheng Zhong<sup>2</sup>, Yang Li<sup>1</sup>, Tianyu Zhang<sup>3a,3b</sup>, Liangxin Gao<sup>4</sup>, Dakai Jin<sup>5</sup>, Yue Sun<sup>6</sup>, Xianghua Ye<sup>7</sup>, Li Yu<sup>8</sup>, Zheyu Hu<sup>9</sup> Jing Xiao<sup>4</sup> Lingyun Huang <sup>4</sup>, Yuling Tang<sup>10,\*</sup>

1. College of Aerospace Science and Engineering, National University of Defense Technology, Changsha, China

Department of Radiology, The First Hospital of Changsha City, Changsha, China
 3a. GROW School for Oncology and Development Biology, Maastricht University, P. O. Box 616, 6200 MD, Maastricht, The Netherlands

3b. Department of Radiology, Netherlands Cancer Institute (NKI), Plesmanlaan 121, 1066 CX, Amsterdam, The Netherlands

4. PingAn Technology, Shenzhen, China

5. PAII Inc., Bethesda, MD, USA

6. Department of Electrical Engineering, Eindhoven University of Technology, 5600 MB Eindhoven, The Netherlands

7. Department of Radiotherapy, The First Affiliated Hospital, Zhejiang University, Hangzhou, Zhejiang, China

8. Hunan LanXi Biotechnology Ltd., Changsha, China

9. Hunan Cancer Hospital, the Affiliated Cancer Hospital of Xiangya Medical School, Central South University, Changsha, China

10. Department of Respiratory Medicine, The First Hospital of Changsha City, Changsha, China

Z. L. and Z. Z. contributed equally to this manuscript

#### Address correspondence to

Yuling Tang, Department of Respiratory Medicine, The First Hospital of Changsha City, Changsha, China (tyl71523@qq.com);

Article Type: Original Research; Thoracic Imaging

#### Abbreviations:

AUC = area under the receiver operating characteristic curve CI = confidence interval COVID-19 = coronavirus disease 2019 POI = portion of infection iHU = average infection Hounsfield unit

#### Key Results:

A deep learning based AI system was able to accurately segment the infected lung regions by COVID-19 using the thick-section CT scans (Dice coefficient  $\geq 0.74$ ).

The computed imaging bio-markers were able to distinguish between the non-severe and severe COVID-19 stages (area under the receiver operating characteristic curve 0.968).

The infection volume changes computed by the AI system was able to assess the COVID-19 progression (Cohen's kappa 0.8220).

**Summary Statement:** A deep learning based AI system built on the thick-section CT imaging can accurately quantify the COVID-19 infected lung regions, assess patients disease severity and their disease progressions.

# Abstract

**Background**: Thick-section CT scanners are more affordable for the developing countries. Considering the widely spread COVID-19, it is of great benefit to develop an automated and accurate system for quantification of COVID-19 associated lung abnormalities using thick-section chest CT images.

**Purpose**: To develop a fully automated AI system to quantitatively assess the disease severity and disease progression using thick-section chest CT images.

Materials and Methods: In this retrospective study, a deep learning based system was developed to automatically segment and quantify the COVID-19 infected lung regions on thick-section chest CT images. 531 thick-section CT scans from 204 patients diagnosed with COVID-19 were collected from one appointed COVID-19 hospital from 23 January 2020 to 12 February 2020. The lung abnormalities were first segmented by a deep learning model. To assess the disease severity (non-severe or severe) and the progression, two imaging bio-markers were automatically computed, i.e., the portion of infection (POI) and the average infection HU (iHU). The performance of lung abnormality segmentation was examined using Dice coefficient, while the assessment of disease severity and the disease progression were evaluated using the area under the receiver operating characteristic curve (AUC) and the Cohen's kappa statistic, respectively.

**Results**: Dice coefficient between the segmentation of the AI system and the manual delineations of two experienced radiologists for the COVID-19 infected lung abnormalities were  $0.74\pm0.28$  and  $0.76\pm0.29$ , respectively, which were close to the inter-observer agreement, i.e.,  $0.79\pm0.25$ . The computed two imaging bio-markers can distinguish between the severe and non-severe stages with an AUC of 0.9680 (p-value< 0.001). Very good agreement ( $\kappa = 0.8220$ ) between the AI system and the radiologists were achieved on evaluating the changes of infection volumes.

**Conclusions**: A deep learning based AI system built on the thick-section CT imaging can accurately quantify the COVID-19 associated lung abnormalities, assess the disease severity and its progressions.

# 1 Introduction

Coronavirus Disease 2019 (COVID-19) has rapidly spread all over the world since the end of 2019, and 1, 436, 198 cases have been confirmed as COVID-19 to date (9 April 2020) [1]. Reverse-transcription polymerase chain reaction (RT-PCR) is used as the standard diagnostic method. However, it suffers from low sensitivities as report in [2, 3]. Computed tomography (CT) imaging is often adopted to confirm the COVID-19 in China and some European countries, e.g., Netherlands. CT plays a key role in the diagnosis and treatment assessment of COVID-19 due to its high sensitivity [2, 4].

The explosive growing number of COVID-19 patients requires the automated AI-based computer aided diagnosis (CAD) systems that can accurately and objectively detect the disease infected lung regions, assess the severity and the progressions. Recently, several deep learning based AI systems were developed to differentiate the COVID-19 and community acquired pneumonia (CAP) [5] or other viral pneumonia [6, 7], and to quantify the infection regions[8, 9, 10, 11]. However, all these previous AI systems built upon the high resolution thin-section CT images, which have high radiation doses and require higher costs. In contrast, the thick-section CT images from affordable CT scanners has relatively low radiation doses and are popularly used in hospitals worldwide, especially in primary care. Hence, it is worthwhile to develop an AI-based CAD system using the thick-section CT images.

In this study, we developed a fully automated AI system to quantify COVID-19 associated lung 19 abnormalities, assess the disease severity and the disease progressions using thick-section chest CT 20 images. Specifically, the lung and infection regions were first segmented by a deep learning based 21 model, where the labels came from another multi-center annotated CAP CT dataset knowing that 22 COVID-19 shares similar abnormal lung patterns with other pneumonia such as ground glass opacity 23 (GGO), consolidation, bilateral infiltration, etc. Using the lung and infection segmentation masks, 24 we computed the portion of infection (POI) and the average infection HU (iHU) as two imaging 25 bio-markers, which were applied to distinguish the COVID-19 severity. Moreover, the changes of 26 POI and iHU in patient's longitudinal CT scans were calculated to evaluate the COVID-19 27 progression. For evaluation, the AI based lung abnormalities segmentation was compared to two 28 experienced radiologists manually delineations, while the AI based assessment of disease severity and 29 progression was compared to patients diagnosis status extracted from clinical and radiology reports. 30 To the best of our knowledge, this is the first AI-based study to quantitatively assess the COVID-19 31 severity and disease progression using the thick-section CT images. 32

# 2 Materials and Methods

## 2.1 Patients

This study was approved by the Ethics of Committees of the First Hospital of Changsha, Hunan, 35 China. Informed consent for this retrospective study was waived. 548 CT scans from 204 patients 36 diagnosed with COVID-19 (RT-PCR test positive) were retrospectively reviewed for the period from 37 23 January 2020 to 12 February 2020 in the First Hospital of Changsha, which is the only appointed 38 hospital healing COVID-19 patients in Changsha city, Hunan province, China. Eight patients under 39 18 years old were excluded for this study. The characteristics of the rest 196 adult patients were 40 summarized in Table 1. According to the guideline of 2019-nCoV (trial version 7) issued by the 41 China National Health Commission[12], the severity of COVID-19 includes mild, common, severe 42 and critical types. Since there were few mild and critical cases, we categorized all the CT scans into 43 severe group (including severe and critical) and non-severe group (mild and common). In total, we 44 had 79 severe CT scans from 32 patients, and 452 general CT scans from 164 patients. It should be 45 noticed that some patients were in non-severe phase when they entered the hospital, but may 46 develop into severe phase during treatment. All the COVID-19 patients were used to test the AI 47 system performance. 48

4/16

9

10

11

12

13

14

15

16

17

18

33

34

To train the lung abnormalities segmentation deep learning model, another multi-center 49 pnumonia dataset was collected consisting of 558 CT scans with manual annotations. The informed 50 consent waiver of the training data were approved by the Ethics of Committees of multiple institutes. 51

#### 2.2 CT Protocol

52

53

54

55

56

57

58

59

60

61

62

63

64

65

66

67

68

69

70

71

72

73

74

81

All COVID-19 patients underwent the CT scanning using the GE Brivo CT325 scanner (General Electric, Illinois, the United States). The scanning protocol was as follows: 120 kV; adaptive tube current (30mAs-70mAs); pitch= 0.99-1.22 mm; slice thickness= 10 mm; field of view:  $350 mm^2$ ; matrix,  $512 \times 512$ ; and breath hold at full inspiration. CT images were reconstructed with 5mm slice thickness and the soft reconstruction kernel. Note that the radiation dose (3.43mGy) from the thick-section CT imaging are reasonably lower than the conventional high resolution chest CT imaging (6.03mGy, Siemens SOMATOM go.Top).

For the multi-center pneumonia dataset, the 558 CT scans were from Siemens, Hitachi, GE, Philips and United Imaging scanners. Slice thickness ranged from 1.0 - 5.0 mm. Details of the CT imaging protocols for this multi-center pneumonia dataset is presented in the Table of the supplemental material.

## 2.3 Deep Learning Model for Lung Abnormality Segmentation

We developed a 2.5D based deep learning model to segment the pneumonia infection regions using the UNet [13] structure equipped with the Resnet 34 backbone [14]. It is able to integrate the high resolution information in the axial view with the coarse continuity information along the vertical view. We also trained a standard 2D UNet to segment the lung fields in thick-section CT scans. Details of deep learning model learning is presented in the supplemental material.

### 2.4 Imaging Bio-markers Computation

Based on the lung field and infection region segmentation masks, we computed the quantitative imaging bio-markers for COVID-19, i.e., the portion of infection (POI) and the average infection HU (iHU). Specifically, we computed the POI as the infection volume divided by the total lung volume in physical unit, and the iHU as the average HU values in the infection regions.

The computed POI and iHU are consistent with latest version (the seventh) of COVID-19 75 diagnostic guideline released by the National Health Commission of China [12]. The guideline states 76 that the POI is one of the principles to differentiate the severe and non-severe patients. It also 77 reports that lung findings in chest CT may start from small subpleural GGO to crazy paving pattern 78 and consolidation when patients conditions getting worse, which correspond to the increase in iHU 79 changes. 80

## 2.5 Statistical Analysis and Evaluation Metrics

Statistical analysis was performed by SAS (version 9.4) and Matlab (version 2018b). Sensitivity and specificity were calculated using specific cutoffs by using the Youden index generated from the receiver operating characteristic curve (ROC). Cohen's kappa statistic was used to measure of agreement between the disease progress assessment from AI and radiologists.  $\chi^2$  test was used to compare differences among different groups. A two-sided p value less than 0.05 was considered to be statistically significant. The Dice coefficient was computed to evaluate agreement between the automatic infection region segmentation and the manual infection delineations by radiologists.

## 3 Results

### 3.1 Segmentation of Lung Infection Region

Examples of the infection region segmentation for severe and non-severe patients in CT were shown 91 in Figure 1 and Figure 2. To quantitatively evaluate the accuracy of segmentation, two radiologists 92 with 20 and 15 years experiences (Z.Z and X.Y), who were blind to each other, manually delineated 93 the infection regions of interests (ROIs) to serve as the reference standard. We randomly selected 30 94 CT scans of 30 patients (3 severe and 27 non-severe) and quantitatively evaluated the accuracy of 95 the infection region segmentation on this subset. The average Dice coefficient between our method 96 and two radiologists were  $0.74\pm0.28$  (median=0.79) and  $0.76\pm0.29$  (median=0.84), respectively. The 97 inter-observer variability between the two radiologists was also assessed using Dice coefficient, which 98 is  $0.79 \pm 0.25$  (media=0.85). 99

#### 3.2 Assessment of Severe and Non-severe COVID-19

Based on the clinical diagnosis reports, 79 CT scans had been identified to belong to the severe group, while 452 scans were in the non-severe group. Figure 3 shows the box-plot of the computed POI and iHU for severe and non-severe groups. Note that both the POI and iHU show significant difference between severe and non-severe groups with p-value < 0.001.

Predictive probabilities were generated using the logistic regression model. Comparisons of 105 different imaging bio-markers for assessment of severe and non-severe exams are shown in Table 2. 106 Using the POI as input, the sensitivity and specificity for identifying the severe group are 92.4%107 90.5%, respectively. Using the iHU as input, the sensitivity and specificity for identifying the severe 108 group are 91.1% and 41.6%, respectively. When combining the POI with iHU, the sensitivity and 109 specificity for identifying the severe group are 93.7% and 88.1%, respectively. The ROC curves are 110 shown in Figure 4. The corresponding AUC values for using the iHU, POI and POI+iHU are 0.687, 111 0.968 and 0.968, respectively. The odds of severity at 1sd increase of POI was 18.762 [95% CI, 112 10.056, 35.000 (p<0.001) times higher than the baseline POI; the odds of severity at 1sd increase of 113 iHU was 1.824 [95% CI, 1.430, 2.326] (p<0.001) times higher than the odds of severity at baseline 114 iHU. The Akaike information criterion (AIC) for POI, iHU and POI+iHU are 174.877, 426.160 and 115 173.767, respectively. 116

#### 3.3 Assessment of Disease Progression

Figure 5 showed a qualitative example of the automatically segmented infection regions of a severe patient's longitudinal CT scans. We calculated the changes of the POI and iHU for each consecutive CT scan pair of the patients. The key phrases extracted from patients radiology reports were used as ground-truth reference. The correspondence of the computed bio-markers changes with radiologists assessment was described in Table 3.

To measure the agreement between the AI computed imaging bio-markers changes and the 123 radiologists assessment, we first binarize the bio-markers changes. The value 1 (or 0) represented the 124 increasing (or decreasing) of bio-markers and its corresponded phrases of radiology reports. Cohen's 125 Kappa was then used to measure the agreement, and the results were shown in Table 4. The very 126 good and moderate agreement were achieved between two AI imaging bio-markers and radiologists 127 assessment if we only consider the changes on whole lung level (ignoring the cases with phrase of 128 'partially changes'). The change of POI showed overall better agreement (very good and good) with 129 radiologists assessment than iHU (moderate and fair). 130

## 4 Discussion

In this study, we developed and evaluated an AI system for quantitative analysis of coronavirus disease 2019 (COVID-19) from thick-section chest CT scans. Our findings can be summarized as

89

90

100

117

•

6/16

131

follows: 1) The deep learning model that trained on a multi-center CAP CT dataset could be directly applied for segmenting the lung abnormalities in COVID-19 patients; 2) the portion of infection (POI) and the average infection HU (iHU), with the area under the receiver operating characteristic curve (AUC) of 0.968 [95% CI: 0.951, 0.981] and 0.687 [95% CI: 0.633, 0.737], showed significant difference (p-value<0.001) in severe and non-severe COVID-19 states; 3) POI showed *Very good* agreement ( $\kappa = 0.8220$ ) with the radiologist reports on evaluating the changes of infection volumes on the whole lung level.

Though high resolution CT is shown to have high sensitivity in detection of COVID-19, both cost and radiation doses are relatively high. In contrast, our study for the first time shows that an AI system can efficiently segment and quantify the COVID-19 lung infections in thick-section CT images (with relatively low radiation doses). This would benefit the developing or low-income countries, where the quantification of COVID-19 severity and the triage can be determined effectively using thick-section CT volumes of affordable CT scanners.

Our diagnosis system is an multi-stage AI system. The key step is to extract the infection regions. <sup>147</sup> It is interesting that this processing modules are trained using CAP cases while the detection and <sup>148</sup> segmentation accuracy is still closed to radiologist-level. Dice coefficient between the COVID-19 <sup>149</sup> infected region segmentation of the AI system and two experienced radiologists were  $0.74\pm0.28$  and <sup>150</sup>  $0.76\pm0.29$ , respectively, which were close to the inter-observer agreement, i.e.,  $0.79\pm0.25$ . <sup>151</sup>

Among our computed imaging bio-markers, only the POI shows high sensitivity and specificity for differentiating the severe from non-severe COVID-19 groups. This indicates that the POI is an effective imaging bio-marker to assess the severity of COVID-19 patients. Although the iHU value is also able to reflect infection progress, however, it is affected by several other disease irrelevant factors, such as the reconstruction slice thickness and the respiration status[15, 16]. For instance, consolidation on HRCT images might be displayed as GGO on thick-section CT images. 157

The changes of volume and density of the infection region are two key indicators that used by 158 radiologists for COVID-19 progression assessment. However, it is time consuming (or even 159 impractical) for radiologists to produce the quantitative measurements for this longitudinal analysis. 160 Our AI system provides a quantitative and objective measurement, i.e., the POI, which shows strong 161 agreement with radiologist qualitative judgements. More importantly, the AI based longitudinal 162 disease quantification is precise, reproducible and fast, which can reduce the reading time of 163 radiologists for COVID-19 each patient and improve the quality of the disease progression 164 assessment[10].165

This study has several limitations. Firstly, we only evaluated changes of imaging bio-markers at 166 the whole lung level in certain phrase. Although our model can compute the bio-markers at the lobe 167 level, the standard phrases from the radiology reports were mostly at the whole lung level. 168 Furthermore, some phrases in the reports like 'lesion absorption' might respond to either infection 169 region decreasing or HU value reduction. Thus it needs more sophisticated and precise analysis 170 evaluating our model in the future. Secondly, motion artifacts due to respiration and heart motion 171 may cause false positive segmentation in the AI system. We noticed that some false positive 172 segmentation affected the longitudinal infection evaluations 6. One possible solution is to identity the 173 motion artifacts before applying the infection segmenting. Finally, our model was only tested the 174 COVID-19 positive patients. A recent study has shown that a deep learning based AI classification 175 model can detect the COVID-19 and distinguish it from the community acquired pneumonia and 176 other non-pneumonic lung diseases using thin-section HRCT [5]. As the next step, it would be 177 interesting to see if our model can also differentiate the pneumonia caused by COVID-19 and other 178 factors using the thick-section CT imaging. 179

In conclusion, a deep learning based AI system is developed to quantify COVID-19 abnormal lung patterns, assess the disease severity and the progression using thick-section chest CT images. The imaging bio-makers computed from the AI system could be used for reproducing several findings of infection change from the reports by radiologists. These results demonstrate that the deep learning based tool has the ability to help radiologists on diagnosing and follow-up treatment for COVID-19 patients based on CT scans.

180

181

182

183

184

185

141

142

143

144

145

146

# Supporting Information

186

This work was partially supported by National Natural Science Funding of China (No.61801491), Natural Science Funding of Hunan Province (No.2019JJ50728) and The Research Program of the Hunan Health and Family Planning Commission(No.B20180393).

# References

- WHO. Coronavirus disease 2019 (covid-19)situation report-80. https://www.who.int/docs/default-source/coronaviruse/situation-reports/20200409-sitrep-80-covid-19.pdf?sfvrsn=1b685d64\_4. Accessed April 9, 2020.
- Tao Ai, Zhenlu Yang, Hongyan Hou, Chenao Zhan, Chong Chen, Wenzhi Lv, Qian Tao, Ziyong Sun, and Liming Xia. Correlation of chest ct and rt-pcr testing in coronavirus disease 2019 (covid-19) in china: a report of 1014 cases. *Radiology*, page 200642, 2020.
- Yicheng Fang, Huangqi Zhang, Jicheng Xie, Minjie Lin, Lingjun Ying, Peipei Pang, and Wenbin Ji. Sensitivity of chest ct for covid-19: comparison to rt-pcr. *Radiology*, page 200432, 2020.
- 4. Xingzhi Xie, Zheng Zhong, Wei Zhao, Chao Zheng, Fei Wang, and Jun Liu. Chest ct for typical 2019-ncov pneumonia: relationship to negative rt-pcr testing. *Radiology*, page 200343, 2020.
- Lin Li, Lixin Qin, Zeguo Xu, Youbing Yin, Xin Wang, Bin Kong, Junjie Bai, Yi Lu, Zhenghan Fang, Qi Song, et al. Artificial intelligence distinguishes covid-19 from community acquired pneumonia on chest ct. *Radiology*, page 200905, 2020.
- Jun Chen, Lianlian Wu, Jun Zhang, Liang Zhang, Dexin Gong, Yilin Zhao, Shan Hu, Yonggui Wang, Xiao Hu, Biqing Zheng, et al. Deep learning-based model for detecting 2019 novel coronavirus pneumonia on high-resolution computed tomography: a prospective study. *medRxiv*, 2020.
- Shuai Wang, Bo Kang, Jinlu Ma, Xianjun Zeng, Mingming Xiao, Jia Guo, Mengjiao Cai, Jingyi Yang, Yaodong Li, Xiangfei Meng, et al. A deep learning algorithm using ct images to screen for corona virus disease (covid-19). medRxiv, 2020.
- 8. Fei Shan, Yaozong Gao, Jun Wang, Weiya Shi, Nannan Shi, Miaofei Han, Zhong Xue, Dinggang Shen, and Yuxin Shi. Lung infection quantification of covid-19 in ct images with deep learning. *arXiv preprint arXiv:2003.04655*, 2020.
- Lei Tang, Xiaoyong Zhang, Yvquan Wang, and Xianchun Zeng. Severe covid-19 pneumonia: Assessing inflammation burden with volume-rendered chest ct. *Radiology: Cardiothoracic Imaging*, 2(2):e200044, 2020.
- Lu Huang, Rui Han, Tao Ai, Pengxin Yu, Han Kang, Qian Tao, and Liming Xia. Serial quantitative chest ct assessment of covid-19: Deep-learning approach. *Radiology: Cardiothoracic Imaging*, 2(2):e200075, 2020.
- 11. Yukun Cao, Zhanwei Xu, Jianjiang Feng, Cheng Jin, Xiaoyu Han, Hanping Wu, and Heshui Shi. Longitudinal assessment of covid-19 using a deep learning–based quantitative ct pipeline: Illustration of two cases. *Radiology: Cardiothoracic Imaging*, 2(2):e200082, 2020.
- National Health Commission of PRC. Diagnosis and treatment protocol for novel coronavirus pneumonia (trial version 7). http://en.nhc.gov.cn/2020-03/29/c\_78469.htm. Accessed April 9, 2020.

- Olaf Ronneberger, Philipp Fischer, and Thomas Brox. U-net: Convolutional networks for biomedical image segmentation. In *International Conference on Medical image computing and computer-assisted intervention*, pages 234–241. Springer, 2015.
- Kaiming He, Xiangyu Zhang, Shaoqing Ren, and Jian Sun. Deep residual learning for image recognition. In Proceedings of the IEEE conference on computer vision and pattern recognition, pages 770–778, 2016.
- 15. Stephan Achenbach, Kerstin Boehmer, Tobias Pflederer, Dieter Ropers, Martin Seltmann, Michael Lell, Katharina Anders, Axel Kuettner, Michael Uder, Werner G. Daniel, and Mohamed Marwan. Influence of slice thickness and reconstruction kernel on the computed tomographic attenuation of coronary atherosclerotic plaque. *Journal of Cardiovascular Computed Tomography*, 4(2):110–115, 2010.
- 16. L J Rosenblum, R A Mauceri, D E Wellenstein, F D Thomas, D A Bassano, B N Raasch, C C Chamberlain, and E R Heitzman. Density patterns in the normal lung as determined by computed tomography. *Radiology*, 137(2):409–416, 1980.

	All patients	Severe patients	Non-severe	p value
	F		patients	P
No.	196	32	164	-
Age	$47 \pm 15$	$56 \pm 14$	45±14	< 0.001
Male	96(49%)	14(44%)	82(50%)	0.52
Exams	531	79	452	-
Patients with multi-	162(83%)	31(97%)	131(80%)	-
ple exams				

Table 1. Characteristics of COVID-2019 patients in this study

Note: Values in parentheses are the percentage. Ages are reported as mean  $\pm$  standard deviation. COVID-2019 = coronavirus disease 2019.

 Table 2. Comparisons of different imaging bio-markers for assessment of severe and non-severe exams

	Sensitivity %	Specificity %	AUC	p value
POI	92.41 (73 of 79)	90.49 (409 of 452)	$0.9680 \ [0.9505, \ 0.9805]$	< 0.001
	[84.85, 97.50]	[87.60, 92.79]		
iHU	91.14 (72 of 79)	41.59 (188 of $452$ )	$0.6873 \ [0.6328, \ 0.7368]$	< 0.001
	[82.44, 96.15]	[36.95, 46.26]		
POI+iHU	93.67 (74 of 79)	88.05 (398 of $452$ )	$0.9677 \ [0.9437, \ 0.9794]$	< 0.001
	[86.60, 97.73]	[85.07,  90.81]		

Note: Values in parentheses are the numbers for the percentage calculation. Values in brackets are 95% confidence intervals [95%CI, %]. AUC = area under the receiver operating characteristic curve, POI = portion of infection, iHU = infection HU.

Table 3. Correspondence between the imaging bio-markers changes and radiology reports

Phrase from radiology reports	imaging bio-markers changes	No.of phrase occurrences
'infection region has (partially) expanded'	increasing of POI	86
'infection region has (partially) contracted'	decreasing of POI	98
'Density of infection region has (partially)	increasing of iHU	43
increased'		
'Density of infection region has (partially)	decreasing of iHU	39
decreased'		

Note: Opposite phrases (partially expanded and partially contracted) that exits in six patients were excluded in this table. POI = portion of infection, iHU = infection HU.

			-		
	No.of cases	Observed	Cohen's	Kappa error	Strength of
		agreement	kappa		agreement
Infection region	142	0.9155	0.8220	0.0492	very good
change					
Infection region	184	0.8533	0.7044	0.0525	good
changes (including					
partially changes)					
Intensity changes	54	0.7321	0.4643	0.1184	moderate
Intensity changes	82	0.6829	0.3718	0.1018	fair
(including partially					
changes)					

 Table 4. Correspondence between the imaging bio-markers changes and radiology reports

Note: Strength of agreement: 0-0.20,poor; 0.21-0.40, fair; 0.41-0.60, moderate; 0.61-0.80, good; 0.81-1.00, very good.



Figure 1. Lesion segmentation for three consecutive axial CTs from a severe patient. First row: original image; Second row: lesion segmentation image.



Figure 2. Lesion segmentation for three consecutive axial CTs from a non-severe patient. First row: original image; Second row: lesion segmentation image.







Figure 3. Box-plot of (a) POI and (b) iHU for the severe and non-severe patients. POI = portion of infection, iHU = infection HU.



Figure 4. Receiver operating characteristic (ROC) curves from the logistic regression model. AUC = area under the receiver operating characteristic curve, POI = portion of infection, iHU = infection HU.



**Figure 5.** The lesion segmentation of six adjacent CT scans that taken from Jan.27 to Feb.12 for a severe patient. The red dot corresponds to the time for given the 'severe' diagnosis and the green point corresponds to the time for given the 'non-severe' diagnosis. 15/16



Figure 6. The false positive segmentation from a exam with motion artifacts.