# 1 Pooled RNA sample reverse transcriptase real time PCR assay for SARS CoV -2

2	infection: a reliable, faster and economical method.
3	
4	Ekta Gupta <sup>1</sup> , Abhishek Padhi <sup>1</sup> , Arvind Khodare <sup>1</sup> , Reshu Aggarwal <sup>1</sup> , Krithiga Ramachnadran <sup>1</sup> ,
5	Vibha Mehta <sup>1</sup> , Mousumi Kilikdar <sup>1</sup> , Shantanu Dubey <sup>2</sup> , Guresh Kumar <sup>3</sup> , Shiv Kumar Sarin <sup>4*</sup> .
6	
7	1. Department of Clinical Virology, Institute of Liver and Biliary Sciences, New Delhi.
8	2. Department of Hospital Operations, Institute of Liver and Biliary Sciences, New
9	Delhi.
10	3. Department of research, Institute of Liver and Biliary Sciences, New Delhi.
11	
12	4. Department of Hepatology, Institute of Liver and Biliary Sciences, New Delhi.
	NOTE: This preprint reports new research that has not been certified by peer review and should not be used to guide clinical practice
13	. Keywords: COVID 19 laboratory diagnosis, Real time PCR, pool testing

Running title : Pooled RNA sample real time PCR assay for SARS-CoV2

## 14 Corresponding Author:

- 15 Prof. Shiv Kumar Sarin
- 16 Senior Professor, Hepatology
- 17 Institute of Liver and Biliary Sciences
- 18 Vasant Kunj, New Delhi, India
- 19 Email : <u>sksarin@ilbs.in</u>, <u>shivsarin@gmail.com</u>
- 20 Alternate corresponding author
- 21 Dr Ekta Gupta
- 22 Professor, Clinical virology
- 23 Institute of Liver and Biliary Sciences
- 24 Vasant Kunj, New Delhi, India

## 21 ABSTRACT

#### 22 Background

- 23 Corona virus disease 2019 (COVID-19) which initially started as a cluster of pneumonia
- 24 cases in the Wuhan city of China has now become a full blown pandemic. Timely diagnosis
- of COVID-19 is the key in containing the pandemic and breaking the chain of transmission.
- 26 In low and middle income countries availability of testing kits has become the major bottle
- 27 neck in testing. Novel methods like pooling of samples are the need of the hour.

## 28 Method

- 29 Extracted RNA samples were randomly placed in pools of 8 on a 96 well plate. Both
- 30 individual RNA (ID) and pooled RNA RT-qPCR for the screening E gene were done in the
- 31 same plate and the positivity for the E gene was seen.

## **32** Results

- 33 The present study demonstrated that pool testing with 8 RNA samples can easily detect even
- up to a single positive sample with Ct value as high as 38. The present study also showed that
- the results of pool testing is not affected by number of positive samples in a pool.

## **36** Conclusion

- 37 Pooling of 8 RNA samples can reduce the time and expense by one eighth, and can help
- expand diagnostic capabilities, especially during constrained supply of reagents and PCR kits
- 39 for the diagnosis of SARS-CoV-2 infection.

## 40 INTRODUCTION

- 41 Corona virus disease 2019 (COVID-19) is a severe acute respiratory infection caused by the
- 42 novel corona virus, severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) [1].
- 43 Initially started as cluster of cases from Wuhan, China [2] it has now spread to over 200
- 44 countries with 1,436,198 confirmed and 85, 522 deaths [3]. The laboratory diagnosis of

45 SARS-CoV-2 is based primarily on nucleic acid amplification test (NAAT) like real time

- 46 reverse transcriptase PCR (RT-qPCR). With an acute shortage of diagnostic kits the present
- 47 testing strategies mainly focus on testing the symptomatic individuals. But detecting the
- 48 carrier or asymptomatic individuals holds the key in containing the spread of the infection
- 49 into the community. Earlier the infected person is identified, sooner the spread of the
- 50 infection can be contained and the surveillance machinery can be activated for contact tracing
- and ultimately break in the chain of transmission of the virus. Most of the countries have

52 imposed lockdown to contain the infection but afterwards when the cases are expected to

take a vertical trajectory, scaling up the testing would be a major challenge [4]. Innovative

54 methods to conserve kits and reagents and human resources needs to be explored to enhance

the testing. Pooling the diagnostic tests has been applied in other infectious diseases and is

56 especially attractive as it requires no additional training, equipment, or materials. In this

57 method, perfected over the years [5] [6] [7], samples are mixed and tested at a single pool,

and subsequent individual tests are performed, only if the pool tests positive.

59 We undertook this study to evaluate a novel protocol of pooling of RNA samples/elutes in

60 performance of PCR for SARS CoV- 2 virus. We were extremely careful that the limits of

61 dilution, do not compromise the sensitivity and specificity of the results.

65

66

67

68

## 69 METHODOLOGY

**70** Sample Collection:

71 Combined nasopharyngeal and oropharyngeal swabs were collected by healthcare workers

72 and transported in a 3 ml viral transport media(VTM) maintaining proper cold chain and sent

73 to the virology laboratory of Institute of Liver and biliary sciences, New Delhi, India. A

volume of 200 microlitre ( $\mu$ l) of the sample was further processed for viral nucleic acid

rs extraction by Qiasymphony DSP Virus/ Pathogen mini kit (Qiagen GmbH,Germany) as per

76 the manufacturers protocol in elutes of  $60 \mu l$  each [8]. Each sample was subjected to the

addition of 10  $\mu$ l of spiked extraction control (EAC) at the time of extraction itself, to check

78 the validity of the extraction procedure.

## 79 PERFORMANCE OF RT-qPCR IN THE LABORATORY

80 The 5 µl of the extracted RNA elute/sample was subjected to RT-qPCR for the qualitative

81 detection of SARS-CoV-2 RNA utilizing with AgPath-IDTM One-Step RT-PCR Reagents

82 (Thermo Fisher Scientific) using a Applied biosystem (ABI) 7500 Real Time PCR system

83 (Thermo Fisher Scientific) and LightMix® SarbecoV E-gene (TIB MOLBIOL). Reactions

84 were heated to  $55^{\circ}$  C for 5 minutes for reverse transcription, denatured in  $95^{\circ}$  C for 5

minutes and then 45 cycles of amplification were carried out for 95°C for 5 seconds and  $60^{\circ}$ C

86 for 15 seconds using FAM parameter for E gene. This assay targets the detection of E gene

87 for SARS as well as nCoV-2. The primer details are given in Table 1. All samples that were

screened positive for E gene were confirmed by performance of RT-qPCR for the detection

89 of specific RdRP gene of SARS-CoV-2 using LightMix® Modular SARS-CoV-2 RdRP (TIB

90 MOLBIOL) using similar PCR conditions as described above.

### 91 POOLING OF SAMPLES FOR THE RT-qPCR

92 RNA samples that were obtained after extraction were randomly pooled initially in pools of

93 2, 4, 6, 8, 16 RNA elutes on a 96 well plate. The best results matching with individual RNA

test (ID) was seen in pools of 2, 4 and 8 samples. Subsequently, pools of 8 RNA elutes on a

95 96 well plate were used as shown in figure 1.Both ID and pooled RNA RT-qPCR for the

screening E gene was done in the same plate (figure 1).5 µl of the RNA sample was taken for

97 the PCR and pooling of 5  $\mu$ l (8x5= 40  $\mu$ l) was done for the pooled test. After thorough

98 mixing, 5 ul of the pooled RNA was taken for the pooled PCR. The volume of RNA sample

99 was kept similar in both the ID as well as pooled PCR so that assay sensitivity is not affected.

100 ID PCR and pooled PCR was performed in the same run, keeping the entire conditions

101 uniform.

102 Results were seen on the ABI software and each reaction was read for E gene after

103 confirmation of the performance of EAC as well as positive control and negative control

104 results. Ct value for each positive test was recorded and as per the WHO criteria, sample with

105 Ct value  $\leq 40$  were considered as positive. All initial E gene positive were confirmed as

106 positive if RdRP gene was also detected with Ct value  $\leq 40$ .

### **107 STATISTICAL ANALYSIS**

- 108 The data was compared for the ability of detection of E gene by ID test and pooled test using
- 109 one sample paired t-test (Table 1). Data of Ct value is shown as mean  $\pm$  standard deviations.
- 110 The agreement between Ct values obtained for positive sample in one positive sample pool
- 111 with ID test was done applying intra class correlation ICC followed by Bland and Altman
- 112 graph. A P value of 0.05 or less was considered statistically significant.

## **113 ETHICAL APPROVAL**

- 114 The study was approved by the Institutional ethics committee and patient consent form was
- 115 waived off because of the use of deidentified discarded RNA samples.

#### 116 **RESULTS**

#### **117 POOL TEST Vs ID TEST RESULTS**

118 Out of 280 samples that were tested, 40 were positive and 240 negative for SARS CoV-2 E

gene with a positivity rate of 16.7% (95 CI; 12.2%-22.0%). All 40 were also positive for

120 RdRP gene. Results were communicated to the patients as per the ID test results obtained.

121 All the clinical details were de-identified, delinked and kept anonymised as per the

122 Government of India data safety policy. Overall 35 pools were made from these samples. On

123 comparing the performance of pool test with ID test pools concordance in performance was

seen in 34 pools. Overall sensitivity of the pool test keeping ID test as gold standard was

125 95.4%, specificity 100%, positive predictive value 100% and negative predictive value of

123 92.86%.

124 There were 13 pools where all the ID tests were negative and pool results were also negative.

In 22 pools one or more positive sample was detected in the ID test (Table 2). Pools were

126 further classified based on the number of positive samples present in them. There were 11

pools with 1 positive sample in each pool,5 with 2 positive sample in each pool,5 with 3

positive samples in each and 1 with 4 positive samples in it.21 out of 22 (95.4%) pools could

129 correctly identify the positive test. In 1 pool which had 1 positive and 7 negative samples, the

130 ID test result showed delayed Ct at 39, therefore it was missed. The sample was however

131 positive for the RdRP gene and was given a positive result.

#### **132** EFFECT OF POOLING ON THE CT VALUE OF THE POSITIVE RESULT

133 In a positive test apart from getting a fluorescence curve the cycle number at which the

134 fluorescence starts is also important, this is measured as Ct value. WHO has defined the

135 criteria that any test which gives fluorescence after 40 cycles should not be considered as

positive [9]. We wanted to see the effect of dilution on the Ct value of the positive result. Ct

137 values of positive samples in ID test were compared with Ct values in pool test (Table 2). The

pools having only 1 sample positive in them, the Ct values were similar to the ID test (p value

139 = 0.04). In pools with more than 1 positive sample the mean Ct value of positive as ID was

140 compared with Mean Ct in pool ,the overall mean CT value of ID was 32.68 while in a

141 pooled testing it was 34.24 only an increase in Ct value of 1.56 very much with in the

142 reporting criteria of being called as positive. This clearly shows that the dilution did not alter

the Ct value of the positive result and a positive will be reported as positive despite pooling.

#### 144 EFFECT OF NUMBER OF POSITIVE SAMPLES IN A POOL

145 We tried to look at the effect of presence of number of positive samples in a pool. As the

146 number of positive samples in a pool increased the difference in the mean Ct value decreased

147 (figure 2). More the number of positives in a pool, more accurate the result of the pool

testing. When there were 4 positive samples in a pool, the mean ID Ct value were identical to

149 mean pool Ct (Table 2).

- 150 On evaluating the performance of pools with only 1 positive sample in it, a bland and altman
- 151 graph (figure 3) showed an intra class agreement of 57.3% [(-29.4% to 88.7%),p=0.024].
- 152 The reliability coefficient was found to be 74.5%. All tests were within 2 SD limit except 1.
- 153 The limit of biasness were -8.4 to 2.3.

154

#### 155 DISCUSSION

156 This novel study demonstrates that pooling of RNA samples can be done without loss of

sensitivity and specificity. The results show that pooled test and ID test for the SARS-CoV-2

158 PCR, has demonstrated that pool testing with 8 RNA samples can easily detect even upto a

single positive sample with Ct value as high as 38. The present study also showed that the

results of pool testing is not affected by number of positive samples in a pool. Pool testing

thus can be easily applied for faster PCR results. The data encourages wide spread of the

technique for the diagnosis of SARS-CoV-2 infection as it saves time and expensive PCR

163 reagents during the constrained shipping and supplies. Pooled testing has been used in many

164 infectious diseases as a simple cost effective method to enhance the speed of diagnosis

165 especially when large samples have to be tested during epidemic screening [10], pooling has

been proven to work for RT-qPCR [7] [11], a time-consuming step for which the reagents

are expected to be in short supply [12].

168 The WHO declared COVID-19 as a global pandemic on March 11<sup>th</sup>, 2020 [13]. As aptly

described by the WHO director general, the *mantra* of "detect, protect and treat" should be

170 followed to break the chain of transmission of SARS-CoV-2 [14]. Early diagnosis and

171 prompt treatment can substantially reduce the number of prospective cases. Hence, laboratory

diagnosis of SARS-CoV-2 holds the key in containing and restricting the COVID-19

pandemic. PCR remains the gold standard for the diagnosis of COVID-19 infection

174 especially in symptomatic individual, but limited supply of diagnostic kits and reagents are

175 major bottle neck in expediting testing of COVID-19 in the community [4]. Furthermore,

176 mass testing of samples should be done to escalate testing and even extend to asymptomatic

177 cases [15]. India a country of over 1.3 billion population [16] testing for COVID-19 has been

a great challenge. Government of India through the Indian Council of Medical Research

179 (ICMR) has established a robust network of laboratories in the form of Viral research and

180 diagnostic laboratories (VRDL) [17]. These VRDLs have taken a lead in the real time

diagnosis of COVID-19 in India. Still to match up to such a humongous population of India,

182 we need some fast and rapid solutions to intensify the testing. Hence a testing method where

a large number of samples can be tested and consuming minimal testing kits and reagent is

the need of the hour. Pooled testing can solve this problem. In another study on pool testing

185 of samples maximum number of pooling of samples was assessed and in this study they

186 found that pooling of samples up to a pool of 64 samples together can be done without

187 effecting the results [18].

188 In the present study, the pool size was limited to 8 based on the preliminary analysis done by

us on determination of pool size (unpublished data). Pooling of samples can be done at

190 various levels ,prior to RNA extraction that is at the time of sample collection, putting the

191 nasopharyngeal/ oropharyngeal swabs into a common viral transport media(VTM). By doing

so the major bottle neck of RNA extraction can be removed. Pooling can also be done by

193 pooling the VTMs and doing common extraction for pooled VTM samples, here the

194 limitation is that in case a pool is positive individual sample collection has to be done at the

195 patient level. Pooling RNA samples are easy to do in a laboratory as any pool turns out to be

196 positive ID test can be performed and result can be declared.

197 It can be argued that the sample size of our study is small. However, in the interest of

198 widespread need to enhance testing facilities and capabilities, the present study makes a

199 seminal contribution. In future, mathematical and computational model can be applied to

200 design appropriate pool size and increase the pool size further for SARS-CoV-2 testing.

201 Pooling of samples is essentially important in monitoring the infection in cohesive groups

such as quarantine facilities, health care workers, community surveillance and diagnosing the

asymptomatic cases. The infection load in these groups may be low but even a single positive

204 case amongst such groups can activate the surveillance system and quarantine the affected

205 group and prevent the further spread in the community.

#### 206 CONCLUSION

As the COVID-19 pandemic spans across the globe, implementation of expanded testing in

larger population groups is the only way out. We recommend that testing the pooled samples,

if done properly, is reliable, and can take the fight of detecting COVID-19 to the next level at

the earliest.

211

## 212 CONFLICT OF INTEREST

213 None

214

## 215 FUNDING

## 217 CONTRIBUTORS

- EG, SKS conceived and designed the study. AP, AK, RA, KR, VM, MK, SD did the
- 219 literature search. Laboratory data were collected by AK. Data analysis was done by
- 220 GK. EG interpreted the data. EG, AP and SKS drafted and prepared the manuscript.

221

- 222 Acknowledgements: We acknowledge all the technical staff involved in testing and ICMR
- for providing the Primers and probes for the conduct of RT-q PCR.

#### 224 **REFERENCES**

- 1. Naming the coronavirus disease (COVID-19) and the virus that causes it.
- 226 https://www.who.int/emergencies/diseases/novel-coronavirus-2019/technical-

227		guidance/naming-the-coronavirus-disease-(covid-2019)-and-the-virus-that-causes-it.
228		Accessed 15 Mar 2020
229	2.	Zhu N, Zhang D, Wang W, et al (2020) A Novel Coronavirus from Patients with
230		Pneumonia in China, 2019. New England Journal of Medicine 382:727–733
231	3.	Novel Coronavirus (2019-nCoV) situation reports.
232		https://www.who.int/emergencies/diseases/novel-coronavirus-2019/situation-reports.
233		Accessed 13 Apr 2020
234	4.	Wuhan, China, faces such a shortage of coronavirus test-kits that people say getting one
235		is like "winning the lottery." https://www.businessinsider.fr/us/wuhan-coronavirus-
236		china-shortage-test-kits-lottery-2020-1. Accessed 10 Apr 2020
237	5.	Nguyen NT, Bish EK, Aprahamian H (2018) Sequential prevalence estimation with
238		pooling and continuous test outcomes. Stat Med 37:2391-2426
239	6.	Bilder CR, Tebbs JM (2012) Pooled-testing procedures for screening high volume
240		clinical specimens in heterogeneous populations. Stat Med 31:3261-3268
241	7.	Arnold ME, Slomka MJ, Coward VJ, Mahmood S, Raleigh PJ, Brown IH (2013)
242		Evaluation of the pooling of swabs for real-time PCR detection of low titre shedding of
243		low pathogenicity avian influenza in turkeys. Epidemiol Infect 141:1286–1297
244	8.	QIAsymphony DSP Virus/Pathogen Kit - QIAGEN Online Shop.
245		https://www.qiagen.com/us/products/discovery-and-translational-research/dna-rna-
246		purification/multianalyte-and-virus/qiasymphony-dsp-viruspathogen-
247		kits/#orderinginformation. Accessed 10 Apr 2020
248	9.	Tang Y-W, Schmitz JE, Persing DH, Stratton CW (2020) The Laboratory Diagnosis of
249		COVID-19 Infection: Current Issues and Challenges. Journal of Clinical Microbiology.

## 250 https://doi.org/10.1128/JCM.00512-20

251	10. Abel U, Schosser R, Süss J (1999) Estimating the prevalence of infectious agents using
252	pooled samples: Biometrical considerations. Zentralblatt für Bakteriologie 289:550–563
253	11. Taylor SM, Juliano JJ, Trottman PA, Griffin JB, Landis SH, Kitsa P, Tshefu AK,
254	Meshnick SR (2010) High-Throughput Pooling and Real-Time PCR-Based Strategy for
255	Malaria Detection. Journal of Clinical Microbiology 48:512–519
256	12. Thermo Fisher to produce millions of coronavirus diagnostic tests - STAT.
257	https://www.statnews.com/2020/03/14/thermo-fisher-to-produce-millions-of-
258	coronavirus-diagnostic-tests/. Accessed 10 Apr 2020
259	13. WHO Director-General's opening remarks at the media briefing on COVID-19 - 11
260	March 2020. https://www.who.int/dg/speeches/detail/who-director-general-s-opening-
261	remarks-at-the-media-briefing-on-covid-19 11-march-2020. Accessed 15 Mar 2020
262	14. WHO Director-General's opening remarks at the media briefing on COVID-19 - 13
263	March 2020. https://www.who.int/dg/speeches/detail/who-director-general-s-opening-
264	remarks-at-the-mission-briefing-on-covid-1913-march-2020. Accessed 15 Mar 2020

265	15. Peto J (2020	) Covid-19 mass	s testing facilities	could end the	e epidemic	rapidly. H	BMJ.
-----	------------------	-----------------	----------------------	---------------	------------	------------	------

- 266 https://doi.org/10.1136/bmj.m1163
- 267 16. Population, total India | Data.
- 268 https://data.worldbank.org/indicator/SP.POP.TOTL?locations=IN. Accessed 10 Apr
- 269 2020
- 270 17. Gupta N, Potdar V, Praharaj I, et al (2020) Laboratory preparedness for SARS-CoV-2
- testing in India: Harnessing a network of Virus Research & Diagnostic Laboratories.
- Indian J Med Res. https://doi.org/10.4103/ijmr.IJMR\_594\_20
- 18. Yelin I, Aharony N, Shaer-Tamar E, et al (2020) Evaluation of COVID-19 RT-qPCR
- test in multi-sample pools. medRxiv 2020.03.26.20039438

275

276

**277** Figures Legends

#### **278** Figure 1: Scheme for pooling the samples on a 96 well plate.

**279** Figure 2: Effect of number of positive samples in a pool.

55250

556

281

# 578<sup>2</sup> Figure 3: Performance of single positive sample pool



Figure 1: Scheme for pooling the samples on a 96 well plate





Figure 2: Effect of number of positive samples in a pool.





Figure 3: Performance of single positive sample pool

Gene	Oligonucleotide	Sequence
	E_ Sarbeco _F	ACAGGTACGTTAATAGTTAATAGCGT
E gene	E_Sarbeco_P1	FAM-ACACTAGCCATCCTTACTGCGCTTCG-BBQ
	E_Sarbeco_R	ATATTGCAGCAGTACGCACACA
RdRP gene	RdRp_SARSr-F	GTGARATGGTCATGTGTGGCGG
	RdRp_SARSr-P2	FAM-CAGGTGGAACCTCATCAGGAGATGC-BBQ
	RdRp_SARSr-R	CARATGTTAAASACACTATTAGCATA

Table 1: Primers and probes for the RT-qPCR

FAM: 6-carboxyfluorescein; BBQ: blackberry quencher.

 Table 2: Comparison between the mean Ct values of individual test and pool test in different

 pool combinations.

Pool of 8 RNA san	nples	E gene detection					
Pool combinations	Total Number of Pools	Not detected in number of pools	Mean of ID test Ct value (± SD)	Mean of Pool test Ct value (± SD)	P value		
1 positive + 7 negative	11	1	33.1 (±3.3)	35.6 (±3.3)	0.041		
2 positive + 6 negative	5	0	30.4 (±3.7)	33(±3.1)	0.134		
3 positive + 5 negative	5	0	34.2 (±3.7)	34.9(±2.6)	0.581		
4 positive + 4 negative	1	0	31.7 (± 4.6)	31.7	-		
8 negative 13		13		NA	-		
Total	35	14					

Ct : cycle threshold ,RNA :ribonucleic acid ,ID :individual

S. No.	Pool with 1 positive sample(n= 11)		Pool with 2 positive Sample(n=5)		Pool with 3 positive Sample(n= 5)		Pool with 4 positive sample(n= 1)	
		Pool test						
	ID Ct	Ct	ID Ct	Pool Ct	ID Ct	Pool Ct	ID Ct	Pool Ct
1.	29	32.13	28.9		39.82		31.72	
2.	38	39	36.1	32.3	29.12		30.73	
3.	33.1	34.78	31.7		30.3	34.09	26.69	
4.	33.8	35.7	28.5	32.4	39.1		37.81	31.72
5.	34.46	39.1	31.5		37.72			
6.	31.05	35.94	23.9	29.1	31.72	38.32		
7.	33.37	36.06	27.74		36.82			
8.	29.17	29.17	27.74	33.32	35.97			
9.	32.7	34.13	33.26		35.56	36.35		
10.	30.26	39.78	34.36	37.74	34.37			
11.	39.1	Negative			38			
12.					29.19	33.89		
13.					32.7			
14.					29.17			
15.					33.37	31.61		
Mean	33.1	35.6	30.4	33.0	34.2	34.9	31.7	31.7
SD	3.3	3.3	3.7	3.1	3.7	2.6	4.6	

 $Table \ 3: \ \textbf{Comparison of individual test Ct value with pool test Ct value in the positive samples pool.}$ 

Table 1: Primers and probes for the RT-qPCR

Gene	Oligonucleotide	Sequence
	E_ Sarbeco _F	ACAGGTACGTTAATAGTTAATAGCGT
E gene	E_Sarbeco_P1	FAM-ACACTAGCCATCCTTACTGCGCTTCG-BBQ
	E_Sarbeco_R	ATATTGCAGCAGTACGCACACA
RdRP gene	RdRp_SARSr-F	GTGARATGGTCATGTGTGGCGG
	RdRp_SARSr-P2	FAM-CAGGTGGAACCTCATCAGGAGATGC-BBQ
	RdRp_SARSr-R	CARATGTTAAASACACTATTAGCATA

FAM: 6-carboxyfluorescein; BBQ: blackberry quencher.

**Table 2:** Comparison between the mean Ct values of individual test and pool test in different

 pool combinations.

Pool of 8 RNA sam	nples	E gene detection				
Pool combinations	Total Number of Pools	Not detected in number of pools	Mean of ID test Ct value (± SD)	Mean of Pool test Ct value (± SD)	P value	
1 positive + 7 negative	11	1	33.1 (±3.3)	35.6 (±3.3)	0.041	
2 positive + 6 negative	5	0	30.4 (±3.7)	33(±3.1)	0.134	
3 positive + 5 negative	5	0	34.2 (±3.7)	34.9(±2.6)	0.581	
4 positive + 4 negative	1	0	31.7 (± 4.6)	31.7	-	
8 negative	13	13		NA	-	
Total 35		14				

Ct : cycle threshold ,RNA :ribonucleic acid ,ID :individual

S No	Pool with 1 positive sample(n= 11)		Pool with 2 Sample	Pool with 2 positive Sample(n=5)		Pool with 3 positive Sample(n= 5)		Pool with 4 positive sample(n= 1)	
5. NO.	Sample	Pool test	Jampie	.(11-5)	Jampie	(11- 5)	Sample(II= 1)		
	ID Ct	Ct	ID Ct	Pool Ct	ID Ct	Pool Ct	ID Ct	Pool Ct	
1.	29	32.13	28.9		39.82		31.72		
2.	38	39	36.1	32.3	29.12		30.73		
3.	33.1	34.78	31.7		30.3	34.09	26.69		
4.	33.8	35.7	28.5	32.4	39.1		37.81	31.72	
5.	34.46	39.1	31.5		37.72				
6.	31.05	35.94	23.9	29.1	31.72	38.32			
7.	33.37	36.06	27.74		36.82				
8.	29.17	29.17	27.74	33.32	35.97				
9.	32.7	34.13	33.26		35.56	36.35			
10.	30.26	39.78	34.36	37.74	34.37				
11.	39.1	Negative			38				
12.					29.19	33.89			
13.					32.7				
14.					29.17				
15.					33.37	31.61			
Mean	33.1	35.6	30.4	33.0	34.2	34.9	31.7	31.7	
SD	3.3	3.3	3.7	3.1	3.7	2.6	4.6		

 $Table \ 3: \mbox{Comparison of individual test Ct value with pool test Ct value in the positive samples pool.}$ 



Figure 1: Scheme for pooling the samples on a 96 well plate



Figure 2: Effect of number of positive samples in a pool



Figure 3: Performance of single positive sample pool