

Computed-tomography estimates of interaural mismatch in insertion depth and scalar location in bilateral cochlear-implant users

Matthew J. Goupell¹, Jack H. Noble^{2,3,4}, Sandeep A. Phatak⁵, Elizabeth Kolberg¹, Miranda Cleary¹, Olga A. Stakhovskaya^{1,5}, Kenneth K. Jensen⁵, Michael Hoa⁶, H. Jeffrey Kim⁶, and Joshua G. W. Bernstein⁵

1) Department of Hearing and Speech Sciences, University of Maryland, College Park, MD, United States.

2) Department of Hearing and Speech Sciences, Vanderbilt University Medical Center, Nashville, TN, United States.

3) Department of Electrical Engineering and Computer Science, Vanderbilt University, Nashville, TN, United States.

4) Department of Otolaryngology, Vanderbilt University Medical Center, Nashville, TN, United States.

5) National Military Audiology and Speech Pathology Center, Walter Reed National Military Medical Center, Bethesda, MD, United States.

6) Department of Otolaryngology-Head and Neck Surgery, Georgetown University Medical Center, Washington, DC, United States.

Corresponding Author:

Matthew J. Goupell

Department of Hearing and Speech Sciences

University of Maryland

College Park, MD 20742

(301) 405-8552

goupell@umd.edu

Running title: Interaural mismatch in bilateral cochlear-implant users

ABSTRACT

Hypothesis: We hypothesized that the bilateral cochlear-implant (BI-CI) users would have a range of interaural insertion-depth mismatch because of different physical placements or characteristics of the arrays, but less than half of electrodes would have less than 75° or 3 mm of interaural insertion-depth mismatch. We also hypothesized that interaural insertion-depth mismatch would be more prevalent nearer the apex, when electrodes were located outside of scala tympani (i.e., possible interaural scalar mismatch), and when the arrays were a mix of pre-curved and straight types.

Background: Brainstem neurons in the superior olivary complex are exquisitely sensitive to interaural differences, the cues to sound localization. These binaurally sensitive neurons rely on interaurally place-of-stimulation-matched inputs at the periphery. BI-CI users may have interaural differences in insertion depth and scalar location, causing interaural place-of-stimulation mismatch that impairs binaural abilities.

Methods: Insertion depths and scalar locations were calculated from temporal-bone computed-tomography (CT) scans of 107 BI-CI users (27 Advanced Bionics, 62 Cochlear, and 18 Med-El). Each subject had either both pre-curved, both straight, or one of each type of array (mixed).

Results: The median interaural insertion-depth mismatch was 23.4° or 1.3 mm. Relatively large interaural insertion-depth mismatch sufficient to disrupt binaural processing occurred for about 15% of electrode pairs [defined as $>75^\circ$ (13.0% of electrode pairs) or >3 mm (19.0% of electrode pairs)]. There was a significant three-way interaction of insertion depth, scalar location, and array type. Interaural insertion-depth mismatch was most prevalent when electrode pairs were more apically located, electrode pairs had interaural

scalar mismatch (i.e., one in Scala Tympani, one in Scala Vestibuli), and when the arrays were both pre-curved.

Conclusion: Large interaural insertion-depth mismatch can occur in BI-CI users. For new BI-CI users, improved surgical techniques to avoid interaural insertion-depth and scalar mismatch is recommended. For existing BI-CI users with interaural insertion-depth mismatch, interaural alignment of clinical frequency allocation tables by an audiologist might remediate any negative consequences to spatial-hearing abilities.

I. INTRODUCTION

The purpose of a single cochlear implant (CI) is to partially restore sound and speech perception to people with poor hearing, typically those with bilaterally moderate to profound hearing loss. Recently, bilateral CIs (BI-CIs) have become increasingly prevalent^{1,2}. The primary purpose of BI-CIs is to convey cues for auditory spatial perception. Although BI-CI users have better sound localization and speech recognition in the presence of competing sounds compared to unilateral CI users³⁻⁶, the spatial-hearing benefits that they receive from two ears are far less than those experience by normal-hearing (NH) listeners⁷⁻¹⁰, and they continue to struggle to communicate in noisy environments¹¹⁻¹³.

In typical auditory systems, binaural sensitivity is computed in brainstem neurons that receive tonotopically symmetric input from the auditory periphery^{14,15}. For BI-CIs to convey maximally useful binaural cues, it is imperative to minimize interaural place-of-stimulation mismatch^{14,16-20}. This can be partially achieved by matching the insertion depth (i.e., minimizing interaural insertion-depth mismatch) for the two CIs in a BI-CI user. In reality, perfectly matching insertion depths is difficult to achieve in complex CI surgeries due to a high prevalence of relatively shallow insertion depths^{21,22} and scalar translocations²³. If information about the interaural insertion-depth mismatch were available, the audiologist could potentially realign the ears using asymmetric frequency mappings in the two CIs. In addition, interaural place-of-stimulation mismatch may occur in the scalar location of electrodes (i.e., interaural scalar mismatch). The electrode array is commonly inserted in Scala Tympani (ST) through the round window and is intended to remain entirely in ST²⁴, without damaging the basilar membrane and Reissner's membrane.

It is also possible that the apical end of the electrode array can unintentionally puncture one or both membranes, producing electrodes in Scala Media (SM) or Scala Vestibuli (SV)²⁴⁻²⁶. Finally, it is sometimes necessary to insert the array into SM or SV. While interaural insertion-depth mismatch for BI-CI users has been examined in small-N studies using psychophysical^{16-20,27,28} and electrophysiological techniques¹⁸, the prevalence and characteristics of this mismatch have not been studied at the population level, and interaural scalar mismatch has not been examined at all.

This study used computed-tomography (CT) scans and computer-model rendering²⁹⁻³¹ to assess the extent to which interaural insertion-depth mismatch exists in a population of 107 BI-CI users. By characterizing the prevalence of interaural insertion-depth and scalar mismatch, and their dependence on array type and tonotopic location, the intent was to provide guidance to surgeons and audiologists in choosing array types and programming strategies to optimize BI-CI hearing outcomes. First, we expected that BI-CI users would have a range of interaural insertion-depth mismatch because of different physical placements of the arrays. We estimate that the population standard deviation in insertion angle is about 75° , based on the summary of the literature provided in Table 1 of Landsberger et al.²¹. If an individual BI-CI users two insertions were selected at random from this distribution, we would expect a mean absolute interaural mismatch of 86° , with 49% of the population having mismatch larger than 75° . However, the cochleae for a given individual are interaurally symmetric. We therefore hypothesized that the amount of interaural insertion-depth mismatch would be smaller than this estimate.

Second, we hypothesized that scalar translocation, which has been shown to result in loss of residual hearing³² and degradation of speech-in-noise performance³³, may be

prevalent in BI-CI users, particularly at the array's apical end. This should increase the prevalence of interaural scalar mismatch, particularly in cases when the arrays follow different intended trajectories through the cochlea (i.e., they are a mix of pre-curved and straight types), which should in turn increase the amount of interaural insertion-depth mismatch.

II. METHODS

Post-operative CT scan and image analysis²⁹⁻³⁴ was performed on 107 BI-CI users from Vanderbilt University and University of Maryland-College Park. The study population included 27 Advanced Bionics users (13 both pre-curved, 10 both straight, 4 mixed), 62 Cochlear Ltd. users (49 both pre-curved, 3 both straight, 10 mixed), and 18 Med-El users (all both straight).

To determine the intra-cochlear location of the electrodes, patient-specific cochlea shape was derived from the patient CT scans using a model that includes segmentation of the intra-cochlear anatomy²⁹⁻³⁴. This analysis quantifies the estimated electrode position in three dimensions. For the present analysis, we were interested in (1) the insertion angle based on a coordinate system defined by a line drawn between the round window and the modiolus and (2) the scalar location of the electrode (ST, SM, or SV)³⁵.

Interaural insertion-depth mismatch for number-matched electrode pairs was characterized in three ways. First, insertion depth was estimated in angular degrees around the cochlear spiral, with the modiolus at the center of the coordinate system and the round window defined as zero degrees (horizontal dashed lines in Fig. 1). Second, for comparison with previous estimates^{16,17,19,20,36,37} that binaural processing is tolerant to interaural

insertion-depth mismatch < 3 mm along an average basilar membrane (BM), insertion angles were translated into BM distances based on the Stakhovskaya³⁸ spiral-ganglion correction to the Greenwood³⁹ frequency-to-place map, based on a 35-mm cochlear length. For both angular and BM distance, interaural insertion-depth mismatch was calculated as the absolute value of the difference for a given number-matched electrode pair. Third, to provide an acoustic-equivalent measure, interaural mismatch was calculated in octaves based on standard center-frequency (CF) allocation tables for each manufacturer. This interaural mismatch was calculated as the ratio of the CF for a given electrode to the effective CF for an electrode at the same insertion depth in the other ear, obtained by linearly interpolating the electrode insertion angles and allocation tables (log-frequency).

Electrode scalar location was categorized as ST, SM, or SV based on the anatomical landmarks in the cochlea; ST and SV are shown as different colors in the leftmost and rightmost panels in Fig. 1.

For the analysis in this paper, we focused on ST and SV locations. Electrodes in SM were omitted from the analysis because there were relatively few (7.3%) and because for electrodes near SM, there are a number of different types of trauma that could be occurring, but without direct visualization of scalar structures the specific type is unclear. Some electrodes were also located outside of the cochlea. After removing electrodes pairs where at least one electrode was classified as SM or outside of the cochlea, there were remaining data from 105 subjects and 1766 electrode pairs.

Figure 1 shows CT scans from two example subjects, one with a small amount of interaural insertion-depth mismatch and no interaural scalar mismatch (top row) and one

with a large amount of interaural insertion-depth mismatch and an interaural scalar mismatch (bottom row).




Figure 1: Two example BI-CI users, with the left two columns showing the left ear and the right two columns showing the right ear. Insertion depth can be seen in the center panels. Scala location can be seen in the leftmost and rightmost panels; ST is shown in red and SV is shown in blue. The example BI-CI user in the top row (A) has a relatively small amount of interaural insertion-depth mismatch (center panels). The electrode arrays are both located in ST (leftmost and rightmost panels), and thus there is no interaural scalar mismatch. The example BI-CI user in the bottom row (B) has a relatively large amount of insertion-depth mismatch (the right-ear array is much deeper than the left-ear array) and there are numerous electrodes located outside of ST (i.e., some electrode pairs have an interaural scalar mismatch).

III. RESULTS




Figure 2: Distributions of interaural insertion-depth mismatch evaluated with respect to angle (degrees; panel A), distance (mm; panel B), and center frequency (octaves; panel C). Bars represent the probability distribution function (pdf) and the dotted curve shows the cumulative distribution function (cdf; i.e., the cumulative summation of the pdf). The vertical dashed lines in panels A and B depict an interaural insertion-depth mismatch of 75° and 3 mm, respectively. These represent a relatively large interaural insertion-depth mismatch that would degrade binaural perception as reported in the literature (see text).

Figure 2 shows probability distribution function (pdf; bars) and cumulative distribution function (cdf; dotted curves) for interaural insertion-depth mismatch expressed in degrees (panel A), mm (panel B), and octaves (panel C) calculated for interpolated electrode pairs with the same electrode number (degrees and mm) or interpolated electrode pairs the same insertion depth (octaves). The median interaural insertion-depth mismatch was 23.4°, 1.3 mm, or 0.27 octaves. Figure 2B includes a vertical line at 3 mm of interaural insertion-depth mismatch, the approximate value where binaural sensitivity statistically degrades in perception studies^{16,17,19,20,36,37}. Figure 2A includes a similar vertical line at 75°, which is the average change in insertion angle equivalent to a 3-mm shift along the basilar

membrane based on the Stakhovskaya³⁸ spiral-ganglion correction to the Greenwood³⁹ map. Interaural insertion-depth mismatch $>75^\circ$ occurred for 13.0% of electrode pairs and >3 mm occurred for 19.0% of electrode pairs.




Figure 3: Histogram of individual electrodes located in Scala Tympani (ST) and Scala Vestibuli (SV) as a function of insertion depth in degrees. The numbers (N) at the top of the plot show the number of electrodes at each depth, and the numbers inside the legend indicate the total number of electrodes in each category. Error bars show +1 standard deviation.

Figure 3 shows the percentage of individual electrodes located in ST or SV as a function of the insertion depth. There were 2815 electrodes located in ST and 869 electrodes located in SV. The percentage of scalar translocation increased with insertion depth. For depths $<180^\circ$, fewer than 20% of electrodes were in SV. This increased to 25–35% for depths between 240° – 420° and about 50% for depths $>420^\circ$.




Figure 4: Interaural insertion-depth mismatch as a function of insertion depth for the different categories of scalar location and array type. Solid lines show linear fits to the data from the linear mixed effects model. Dashed lines show the 95% confidence interval of the fits. Number of electrode pairs (the data points) and number of subjects contributing in each condition is reported in each panel. Arrows indicate conditions where the points with $>240^\circ$ of interaural insertion-depth mismatch occur, but were outside the plot range to better show the fits.

Figure 4 shows the absolute interaural insertion-depth mismatch as a function of the average insertion depth for each electrode pair, plotted separately for each combination of scalar location (rows) and array type (columns). Figures 4E and 4F had insufficient counts to assess the relationship between interaural insertion-depth mismatch and insertion depth, with only 1 or 2 subjects in each case. For the other categories, the fitted regression lines show a tendency for more interaural insertion-depth mismatch with greater insertion

depth (i.e., positive slopes) when at least one of the arrays was pre-curved (left and right columns).

These trends were analyzed with a linear mixed-effects (LME) model (R version 4.0.3) using the `buildmer` (v1.7.1)⁴⁰ and `lme4` (v1.1-26)⁴¹ packages. The best-fitting model included interaural insertion-depth mismatch as the dependent variable and three fixed effects: insertion depth (continuous variable), scalar location (categorical variable: ST-ST, SV-SV, and ST-SV), and array type (categorical variable: both pre-curved, both straight, and mixed). Subject and scalar location were significant random effects. For the categorical variables, the model was referenced to the ST-ST and both pre-curved categories.

The results of the main LME analysis are reported in Table I. There were significant main effects of insertion depth on interaural insertion-depth mismatch (more mismatch for more apically located electrodes) and scalar location (cases of interaural scalar mismatch ST-SV showed more mismatch than cases with no interaural scalar mismatch ST-ST). There were, however, also many significant two- and three-way interactions, suggesting that the relationship between interaural insertion-depth mismatch and insertion depth varied across array and scalar categories.

Table I: Results of the linear mixed-effects model. Significant effects and interactions are highlighted in bold.

Fixed Effects	Estimate	Std. Err.	df	t	p
Intercept	28.74	4.14	111.40	6.94	<0.001
Insertion Depth	0.04	0.01	1593.00	5.96	<0.001
Scala (SV-SV)	-4.18	8.58	29.79	-0.49	0.629
Scala (ST-SV)	-27.08	4.75	74.39	-5.71	<0.001
Array (Both ST)	0.32	7.33	112.10	0.04	0.965
Array (Mixed)	-11.70	9.74	114.50	-1.20	0.232
Insertion Depth × Scala (SV-SV)	0.06	0.02	1560.00	3.55	<0.001
Insertion Depth × Scala (ST-SV)	0.11	0.01	1541.00	9.96	<0.001
Insertion Depth × Both ST	17.39	56.67	441.80	0.31	0.759
Insertion Depth × Mixed	52.13	12.55	103.90	4.15	<0.001
Scala (SV-SV) × Both ST	-6.97	42.03	118.60	-0.17	0.869
Scala (ST-SV) × Both ST	48.77	11.83	79.57	4.12	<0.001
Scala (SV-SV) × Mixed	-0.01	0.01	1585.00	-1.02	0.309
Scala (ST-SV) × Mixed	0.02	0.02	1590.00	1.54	0.125
Insertion Depth × Scala (SV-SV) × Both ST	-0.04	0.11	1543.00	-0.34	0.733
Insertion Depth × Scala (ST-SV) × Both ST	-0.17	0.03	1137.00	-5.36	<0.001
Insertion Depth × Scala (SV-SV) × Mixed	0.16	0.14	1548.00	1.13	0.260
Insertion Depth × Scala (ST-SV) × Mixed	-0.14	0.03	1596.00	-4.54	<0.001
Random Effects	Variance				
Subject	942.6				
Scala (SV-SV)	957.4				
Scala (ST-SV)	546.4				
Residual	229.0				

To examine the three-way interactions, seven post-hoc models were run (the previous analysis and six additional models with different reference levels), one for each combination of scalar location and array type, excluding the two SV-SV combinations with insufficient counts/subjects. Bonferroni corrections were applied for seven analyses (criterion: $p < 0.0071 = 0.05/7$). Table II reports the slopes for the different scalar and array combinations. It also reports the difference in slopes of the fits, which was obtained by

comparing the interactions between insertion depth and different categorical conditions of the relevelled models.

Table II: Slopes for the different scalar and array categories (left columns) and a comparison of slopes across categories (right columns). Bolded p values are significant at the $p < 0.0071$ level. PC=pre-curved; ST=Scala Tympani; STR=straight; SV=Scala Vestibuli.

					Slope Comparison								
Slopes					ST-ST	ST-ST	ST-ST	SV-SV	SV-SV	SV-SV	ST-SV	ST-SV	ST-SV
Scala	Array	Intercept	Slope	p	Both PC	Both STR	Mixed	Both PC	Both STR	Mixed	Both PC	Both STR	Mixed
ST-ST	Both PC	28.740	0.036	<0.0001	--	0.309	0.125	0.0004	--	--	<0.0001	<0.0001	<0.0001
ST-ST	Both STR	29.050	0.025	0.005		--	0.038	0.733	--	--	<0.0001	0.040	0.374
ST-ST	Mixed	17.032	0.059	<0.0001			--	0.260	--	--	<0.0001	0.374	0.356
SV-SV	Both PC	24.554	0.098	<0.0001				--	--	--	0.009	0.205	0.038
SV-SV	Both STR	--	--	--					--	--	--	--	--
SV-SV	Mixed	--	--	--					--	--	--	--	--
ST-SV	Both PC	1.653	0.147	<0.0001							--	<0.0001	<0.0001
ST-SV	Both STR	54.105	-0.038	0.194								--	0.061
ST-SV	Mixed	38.721	0.033	0.173									--

Figure 4 and Table II show that interaural insertion-depth mismatch increased with increasing insertion depth for four conditions: all three cases with both-pre-curved cases (Fig. 4A, D, G), and the mixed ST-ST case (Fig. 4C). There was no effect of insertion depth in both-straight ST-ST condition (Fig. 4B). In this case, the slope of the regression line was shallow (0.025), meaning that the interaural insertion-depth mismatch increased by only 9° for a full 360° turn around the cochlea towards the apex. There was also no effect of insertion depth in the both-straight ST-SV (Fig. 4H) and mixed ST-SV cases (Fig. 4I), although the confidence intervals for these cases were large because of relatively few data points.

Comparing the slopes (right half of Table II) across the top row of Fig. 4 (ST-ST electrodes) shows no significant effect of array type, while comparing across the bottom row (ST-SV electrodes) shows significantly greater slopes for the both-pre-curved case than for the both-straight or mixed cases. Comparing the slopes down the left column of Fig. 4 (both pre-curved) shows that the slopes were significantly greater for the SV-SV

(Fig. 4D) and ST-SV (Fig. 4G) cases compared to the ST-ST case (Fig. 4A). There was a significantly greater slope for the both-pre-curved ST-SV condition (Fig. 4G) than all the other conditions, except for both-pre-curved SV-SV condition (Fig. 4D). There were no significant changes in slope as a function of scalar type down the middle (both-straight) and right columns (mixed arrays) of Fig. 4, although this might be due, at least in part to the relatively few counts for the both-straight ST-SV (Fig. 4H) and mixed ST-SV (Fig. 4I) cases.

To summarize, the significant three-way interaction reflected different dependencies of interaural insertion-depth mismatch on insertion depth. When both electrodes were in ST, there was a slight dependence of mismatch on insertion depth for all array-type categories (Fig. 4, top row). But when both arrays were pre-curved, there was a greater dependence of mismatch on insertion depth when at least one of the electrodes was outside of ST (Fig. 4, left column). The largest dependence on insertion depth occurred for the cases with two pre-curved arrays and an interaural scalar mismatch (ST-SV scalar locations; Fig. 4G). For these cases, the slope was 0.147, which means that the interaural insertion-depth mismatch increased by 52.9° for a full 360° turn around the cochlea towards the apex.

IV. GENERAL DISCUSSION

A. Summary of Results and Relationship to Previous Studies

Many factors contribute to relatively small binaural benefits in BI-CI users, including device-related and surgical factors⁴²⁻⁴⁵. Interaural place-of-stimulation mismatch can reduce binaural functioning in BI-CI users because the binaural differences are

computed at the level of the brainstem using frequency-matched inputs¹⁴. Previously, the prevalence of interaural place-of-stimulation mismatch was unknown as it has never been systematically analyzed in a large number of patients; most psychophysical or electrophysiological studies investigating interaural place-of-stimulation mismatch have involved relatively small sample sizes ($N \approx 10$)^{17-20,27,28,43,46-54}. Using a larger database of CT scans from $N=107$ BI-CI users, interaural place-of-stimulation mismatch (which likely is a combination of non-neural and neural factors) was estimated based on non-neural physical interaural insertion-depth mismatch (insertion depth differences of individual electrode pairs, measured in degrees).

Our main finding was that interaural insertion-depth mismatch was fairly common (Fig. 2), where the median interaural insertion-depth mismatch was 23.4° or 1.3 mm. From previous reports, we estimated that less half of electrode pairs would exhibit a relatively “large” amount of mismatch of 75° or 3 mm, the latter quantity has been shown to incur significant decrements in perceptual binaural sensitivity and fusion^{16,17,19,20,36,37,55}. We found this to be true, where 13% of electrode pairs had $>75^\circ$ and 19% of electrode pairs had >3 mm of interaural insertion-depth mismatch. It is important to note that the large-amount-of-mismatch value is derived from studies involving single-electrode stimulation in each ear; it has also been suggested that BI-CI users may be less tolerant to interaural insertion-depth mismatch using multi-electrode stimulation⁵⁴. Indeed, it is unclear what to judge as a disruptive amount of interaural insertion-depth mismatch without paired perceptual measurements.

Our second main finding was a significant three-way interaction that revealed larger interaural insertion-depth mismatch for increasing insertion depths, when there was at least

one pre-curved array (i.e., both pre-curved or mixed arrays) but particularly when there were two pre-curved arrays, and when the electrodes had an interaural scalar mismatch (Fig. 4, Tables I and II). While this broadly confirms the second hypothesis of the study, our initial hypothesis did not include the both-pre-curved condition, and hence this finding was notable. Consistent with previous reports²⁴, the probability of non-ST electrode location for a single array increased with the insertion depth in the current data set (Fig. 3). Pre-curved arrays also tended to have more mismatch (Fig. 4, Tables I and II). When there were two pre-curved arrays and translocation occurs (typically at the apex), this was the instance that caused the most interaural insertion-depth mismatch.

These results might account for some of the dependence of binaural sensitivity on tonotopic place that is sometimes reported in the literature⁶, in particular the tendency for the poorest ITD discrimination performance to occur near the apical end of the array^{43,44,56,57}. This may be a result of sensitivity being degraded by translocations⁵⁸, and interaural insertion-depth or scalar mismatch (Figs. 3 and 4), all of which become more prevalent toward the apex. These occurrences might be inconsistent between studies with small study populations, explaining the mixed results in the literature.

B. Limitations and Future Directions

While we have shown interaural insertion-depth mismatch occurs readily in BI-CI users, this analysis reflects only the anatomical location and does not account for other factors that might affect binaural sensitivity. Peripheral biological factors such as neural degeneration, dead regions, and tissue scarring may alter the relative populations of peripheral neurons being stimulated⁵⁹⁻⁶¹. Furthermore, it is possible that binaural circuits

could adapt to address interaural insertion-depth mismatch⁶²⁻⁶⁴, which could affect binaural functioning and/or binaural fusion^{55,65}; however, little physiological evidence has supported such an idea to date¹⁸.

To more directly assess the relationship between physical and functional mismatch, future work should consider relating binaural sensitivity to CT measurements of intracochlear electrode location. Recent measurements in a smaller group of 20 BI-CI listeners and 23 SSD-CI listeners suggests that CT and ITD-based estimates of interaural insertion-depth mismatch are closely aligned⁶⁶. In contrast, perceptual^{18,65,66} or electrophysiological binaural sensitivity showed incongruence with pitch¹⁸. Another limitation was that while it included >100 BI-CI users, a larger sample would have provided more statistical power to examine some of the subgroups in this study (e.g., use of two straight arrays, or a mix of pre-curved and straight arrays, Fig. 4).

C. Clinical Implications

Interaural place-of-stimulation matching has the potential to improve binaural performance in BI-CI users. During implantation, improved surgical techniques to avoid interaural insertion-depth and scalar mismatch are recommended, for example, by using CT-based surgical guidance⁶⁷. Post-implantation, knowing the interaural cochlear locations that maximize binaural sensitivity could guide the audiological frequency-mapping process to optimize spatial hearing. While this might be facilitated by psychophysical and electrophysiological measurements of binaural performance^{16-18,27,28,55}, these measurements require ITD sensitivity, which is not possible for some CI users⁶⁸⁻⁷⁰, and can be laboriously long²⁸. Use of CT scans, on the other hand, are rapid and precise. This

approach acknowledges that the binaural hearing system requires coordination across the ears, rather than treating the two devices as separate systems.

When considering bilateral implantation, there are competing issues involved in the selection of an array type. On one hand, some studies suggest that a pre-curved array might provide an advantage for monaural speech-understanding outcomes and provide better channel independence^{71,72}, although translocations with such arrays appear to be a potential problem because of decreased performance⁵⁸. On the other hand, our results show that pre-curved arrays are also more likely to result in more interaural insertion-depth and scalar mismatch toward the apex, which might reduce binaural sensitivity, although frequency-mapping programming techniques might address this issue. In any case, it is important to qualify this interpretation because the observed significant differences between array types was small and it is unclear if these differences would substantially affect perception.

ACKNOWLEDGMENTS

We thank Danielle King, Kelly Miller, and Kristina Milvae for help with data analysis. Research reported in this publication was supported by the National Institute On Deafness And Other Communication Disorders of the National Institutes of Health under Award Number R01DC015798 (M.J.G. and J.G.B.W.), R01DC014037 (J.N.), and R01DC014462 (Dawant). The content is solely the responsibility of the authors and does not necessarily represent the official views of the National Institutes of Health. Portions of this work were presented at the 176th Meeting of the Acoustical Society of America, 42nd Midwinter Meeting of the Association for Research in Otolaryngology, and the 19th Conference on Implantable Auditory Prostheses.

The identification of specific products or scientific instrumentation is considered an integral part of the scientific endeavor and does not constitute endorsement or implied endorsement on the part of the authors, Department of Defense, or any component agency. The views expressed in this article are those of the authors and do not reflect the official policy of the Department of Army/Navy/Air Force, Department of Defense, or U.S. Government.

References:

1. Peters BR, Wyss J, Manrique M. Worldwide trends in bilateral cochlear implantation. *Laryngoscope*. 2010;120:S17-44.
2. Holder JT, Reynolds SM, Sunderhaus LW, Gifford RH. Current profile of adults presenting for preoperative cochlear implant evaluation. *Trends Hear*. 2018;22:1-16.
3. Reeder RM, Firszt JB, Holden LK, Strube MJ. A longitudinal study in adults with sequential bilateral cochlear implants: Time course for individual ear and bilateral performance. *J Speech Lang Hear Res*. 2014;57(3):1108-1126.
4. Buss E, Pillsbury HC, Buchman CA, et al. Multicenter U.S. bilateral MED-EL cochlear implantation study: Speech perception over the first year of use. *Ear Hear*. 2008;29(1):20-32.
5. Litovsky RY, Parkinson A, Arcaroli J. Spatial hearing and speech intelligibility in bilateral cochlear implant users. *Ear Hear*. 2009;30(4):419-431.
6. Litovsky RY, Parkinson A, Arcaroli J, Sammeth C. Simultaneous bilateral cochlear implantation in adults: A multicenter clinical study. *Ear Hear*. 2006;27(6):714-731.
7. Bernstein JGW, Goupell MJ, Schuchman G, Rivera A, Brungart DS. Having two ears facilitates the perceptual separation of concurrent talkers for bilateral and single-sided deaf cochlear implantees. *Ear Hear*. 2016;37(3):282-288.
8. Zeitler DM, Dorman MF, Natale SJ, Loiselle L, Yost WA, Gifford RH. Sound source localization and speech understanding in complex listening environments by single-sided deaf listeners after cochlear implantation. *Otol Neurotol*. 2015;36(9):1467-1471.

9. Loizou PC, Hu Y, Litovsky R, et al. Speech recognition by bilateral cochlear implant users in a cocktail-party setting. *J Acoust Soc Am*. 2009;125(1):372-383.
10. Goupell MJ, Kan A, Litovsky RY. Spatial attention in bilateral cochlear-implant users. *J Acoust Soc Am*. 2016;140(3):1652-1662.
11. Sladen DP, Nie Y, Berg K. Investigating speech recognition and listening effort with different device configurations in adult cochlear implant users. *Cochlear Implants Int*. 2018;19(3):119-130.
12. Hallberg LR, Hallberg U, Kramer SE. Self-reported hearing difficulties, communication strategies and psychological general well-being (quality of life) in patients with acquired hearing impairment. *Disabil Rehabil*. 2008;30(3):203-212.
13. Kramer SE, Kapteyn TS, Festen JM. The self-reported handicapping effect of hearing disabilities. *Audiology*. 1998;37(5):302-312.
14. Blanks DA, Roberts JM, Buss E, Hall JW, Fitzpatrick DC. Neural and behavioral sensitivity to interaural time differences using amplitude modulated tones with mismatched carrier frequencies. *J Assoc Res Otolaryngol*. 2007;8(3):393-408.
15. Yin TC, Kuwada S, Sujaku Y. Interaural time sensitivity of high-frequency neurons in the inferior colliculus. *J Acoust Soc Am*. 1984;76(5):1401-1410.
16. Kan A, Litovsky RY, Goupell MJ. Effects of interaural pitch-matching and auditory image centering on binaural sensitivity in cochlear-implant users. *Ear Hear*. 2015;36(3):e62-e68.
17. Kan A, Stoelb C, Litovsky RY, Goupell MJ. Effect of mismatched place-of-stimulation on binaural fusion and lateralization in bilateral cochlear-implant users. *J Acoust Soc Am*. 2013;134(4):2923-2936.

18. Hu H, Dietz M. Comparison of interaural electrode pairing methods for bilateral cochlear implants. *Trends Hear*. 2015;19:1-22.
19. Kan A, Goupell MJ, Litovsky RY. Effect of channel separation and interaural mismatch on fusion and lateralization in normal-hearing and cochlear-implant listeners. *J Acoust Soc Am*. 2019;146(2):1448-1463.
20. Goupell MJ. Interaural correlation-change discrimination in bilateral cochlear-implant users: Effects of interaural frequency mismatch, centering, and age of onset of deafness. *J Acoust Soc Am*. 2015;137(3):1282-1297.
21. Landsberger DM, Svrakic M, Roland JT, Jr., Svirsky M. The relationship between insertion angles, default frequency allocations, and spiral ganglion place pitch in cochlear implants. *Ear Hear*. 2015;36(5):e207-213.
22. Ketten DR, Skinner MW, Wang G, Vannier MW, Gates GA, Neely JG. In vivo measures of cochlear length and insertion depth of Nucleus cochlear implant electrode arrays. *Ann Otol Rhinol Laryngol Suppl*. 1998;175:1-16.
23. Riggs WJ, Dwyer RT, Holder JT, et al. Intracochlear electrocochleography: Influence of scalar position of the cochlear implant electrode on postinsertion results. *Otol Neurotol*. 2019;40(5):e503-e510.
24. Risi F. Considerations and rationale for cochlear implant electrode design - Past, present and future. *J Int Adv Otol*. 2018;14(3):382-391.
25. Gstoettner W, Franz P, Hamzavi J, Plenk H, Jr., Baumgartner W, Czerny C. Intracochlear position of cochlear implant electrodes. *Acta oto-laryngologica*. 1999;119(2):229-233.

26. Jwair S, Prins A, Wegner I, Stokroos RJ, Versnel H, Thomeer H. Scalar translocation comparison between lateral wall and perimodiolar cochlear implant arrays - A meta-analysis. *Laryngoscope*. 2020;00:1-9.
27. Long CJ, Eddington DK, Colburn HS, Rabinowitz WM. Binaural sensitivity as a function of interaural electrode position with a bilateral cochlear implant user. *J Acoust Soc Am*. 2003;114(3):1565-1574.
28. Poon BB, Eddington DK, Noel V, Colburn HS. Sensitivity to interaural time difference with bilateral cochlear implants: Development over time and effect of interaural electrode spacing. *J Acoust Soc Am*. 2009;126(2):806-815.
29. Noble JH, Dawant BM. Automatic graph-based localization of cochlear implant electrodes in CT. *Med Image Comput Comput Assist Interv*. 2015;9350:152-159.
30. Noble JH, Gifford RH, Labadie RF, Dawant BM. Statistical shape model segmentation and frequency mapping of cochlear implant stimulation targets in CT. *Med Image Comput Comput Assist Interv*. 2012;15(Pt 2):421-428.
31. Noble JH, Labadie RF, Dawant BM. Automatic classification of cochlear implant electrode cavity positioning. *Med Image Comput Comput Assist Interv*. 2018;11073:47-54.
32. Zhao Y, Chakravorti S, Labadie RF, Dawant BM, Noble JH. Automatic graph-based method for localization of cochlear implant electrode arrays in clinical CT with sub-voxel accuracy. *Med Image Anal*. 2019;52:1-12.
33. Noble JH, Labadie RF, Majdani O, Dawant BM. Automatic segmentation of intracochlear anatomy in conventional CT. *IEEE Trans Biomed Eng*. 2011;58(9):2625-2632.

34. Zhao Y, Dawant BM, Labadie RF, Noble JH. Automatic localization of closely spaced cochlear implant electrode arrays in clinical CTs. *Med Phys*. 2018;45(11):5030-5040.
35. Verbist BM, Skinner MW, Cohen LT, et al. Consensus panel on a cochlear coordinate system applicable in histologic, physiologic, and radiologic studies of the human cochlea. *Otol Neurotol*. 2010;31(5):722-730.
36. Goupell MJ, Stoelb C, Kan A, Litovsky RY. Effect of mismatched place-of-stimulation on the salience of binaural cues in conditions that simulate bilateral cochlear-implant listening. *J Acoust Soc Am*. 2013;133(4):2272-2287.
37. Goupell MJ, Stoelb CA, Kan A, Litovsky RY. The effect of simulated interaural frequency mismatch on speech understanding and spatial release from masking. *Ear Hear*. 2018;39(5):895-905.
38. Stakhovskaya O, Sridhar D, Bonham BH, Leake PA. Frequency map for the human cochlear spiral ganglion: Implications for cochlear implants. *J Assoc Res Otolaryngol*. 2007;8(2):220-233.
39. Greenwood DD. A cochlear frequency-position function for several species--29 years later. *J Acoust Soc Am*. 1990;87(6):2592-2605.
40. *Buildmer: Stepwise elimination and term reordering for mixed-effects regression* [computer program]. Version 1.7.12020.
41. Bates D, Maechler M, Bolker B, Walker S. Fitting linear mixed-effects models using lme4. *J Stat Softw*. 2015(67):1-48.
42. Kan A, Litovsky RY. Binaural hearing with electrical stimulation. *Hear Res*. 2015;322:127-137.

43. Litovsky RY, Goupell MJ, Godar S, et al. Studies on bilateral cochlear implants at the University of Wisconsin's Binaural Hearing and Speech Laboratory. *Journal of the American Academy of Audiology*. 2012;23(6):476-494.
44. Laback B, Egger K, Majdak P. Perception and coding of interaural time differences with bilateral cochlear implants. *Hear Res*. 2015;322:138-150.
45. Blamey P, Artieres F, Başkent D, et al. Factors affecting auditory performance of postlinguistically deaf adults using cochlear implants: An update with 2251 patients. *Audiol Neurootol*. 2013;18(1):36-47.
46. Steel MM, Papsin BC, Gordon KA. Binaural fusion and listening effort in children who use bilateral cochlear implants: A psychoacoustic and pupillometric study. *PLoS One*. 2015;10(2):1-29.
47. He S, Brown CJ, Abbas PJ. Effects of stimulation level and electrode pairing on the binaural interaction component of the electrically evoked auditory brain stem response. *Ear Hear*. 2010;31(4):457-470.
48. He S, Brown CJ, Abbas PJ. Preliminary results of the relationship between the binaural interaction component of the electrically evoked auditory brainstem response and interaural pitch comparisons in bilateral cochlear implant recipients. *Ear Hear*. 2012;33(1):57-68.
49. van Hoesel RJM. Exploring the benefits of bilateral cochlear implants. *Audiol Neurootol*. 2004;9(4):234-246.
50. van Hoesel RJM. Sensitivity to binaural timing in bilateral cochlear implant users. *J Acoust Soc Am*. 2007;121(4):2192-2206.

51. van Hoesel RJM, Clark GM. Psychophysical studies with two binaural cochlear implant subjects. *J Acoust Soc Am*. 1997;102(1):495-507.
52. van Hoesel RJM, Tyler RS. Speech perception, localization, and lateralization with bilateral cochlear implants. *J Acoust Soc Am*. 2003;113(3):1617-1630.
53. Gordon KA, Salloum C, Toor GS, van Hoesel R, Papsin BC. Binaural interactions develop in the auditory brainstem of children who are deaf: Effects of place and level of bilateral electrical stimulation. *J Neurosci*. 2012;32(12):4212-4223.
54. Williges B, Jurgens T, Hu H, Dietz M. Coherent coding of enhanced interaural cues improves sound localization in noise with bilateral cochlear implants. *Trends Hear*. 2018;22:1-18.
55. Reiss LAJ, Fowler JR, Hartling CL, Oh Y. Binaural pitch fusion in bilateral cochlear implant users. *Ear Hear*. 2018;39(2):390-397.
56. Best V, Laback B, Majdak P. Binaural interference in bilateral cochlear-implant listeners. *J Acoust Soc Am*. 2011;130(5):2939-2950.
57. Egger K, Majdak P, Laback B. Channel interaction and current level affect across-electrode integration of interaural time differences in bilateral cochlear-implant listeners. *J Assoc Res Otolaryngol*. 2016;17:55-67.
58. Chakravorti S, Noble JH, Gifford RH, et al. Further evidence of the relationship between cochlear implant electrode positioning and hearing outcomes. *Otol Neurotol*. 2019;40(5):617-624.
59. Bierer JA, Faulkner KF. Identifying cochlear implant channels with poor electrode-neuron interface: partial tripolar, single-channel thresholds and psychophysical tuning curves. *Ear Hear*. 2010;31(2):247-258.

60. DeVries L, Scheperle R, Bierer JA. Assessing the electrode-neuron interface with the electrically evoked compound action potential, electrode position, and behavioral thresholds. *J Assoc Res Otolaryngol*. 2016;17(3):237-252.
61. Long CJ, Holden TA, McClelland GH, et al. Examining the electro-neural interface of cochlear implant users using psychophysics, CT scans, and speech understanding. *J Assoc Res Otolaryngol*. 2014;15(2):293-304.
62. Reiss LA, Turner CW, Karsten SA, Gantz BJ. Plasticity in human pitch perception induced by tonotopically mismatched electro-acoustic stimulation. *Neuroscience*. 2014;256:43-52.
63. Reiss LA, Turner CW, Erenberg SR, Gantz BJ. Changes in pitch with a cochlear implant over time. *J Assoc Res Otolaryngol*. 2007;8(2):241-257.
64. Reiss LA, Ito RA, Eggleston JL, et al. Pitch adaptation patterns in bimodal cochlear implant users: Over time and after experience. *Ear Hear*. 2015;36(2):e23-34.
65. Staisloff HE, Aronoff JM. Comparing methods for pairing electrodes across ears with cochlear implants. *Ear Hear*. 2021.
66. Bernstein JGW, Jensen KK, Stakhovskaya OA, et al. Interaural place-of-stimulation mismatch estimates using CT scans and binaural perception, but not pitch, are consistent in cochlear-implant users. *medRxiv pre-print*. 2021.
67. Noble JH, Labadie RF, Gifford RH, Dawant BM. Image-guidance enables new methods for customizing cochlear implant stimulation strategies. *IEEE Trans Neural Syst Rehabil Eng*. 2013;21(5):820-829.

68. Ehlers E, Goupell MJ, Zheng Y, Godar SP, Litovsky RY. Binaural sensitivity in children who use bilateral cochlear implants. *J Acoust Soc Am*. 2017;141(6):4264-4277.
69. Litovsky RY, Jones GL, Agrawal S, van Hoesel R. Effect of age at onset of deafness on binaural sensitivity in electric hearing in humans. *J Acoust Soc Am*. 2010;127(1):400-414.
70. Salloum CA, Valero J, Wong DD, Papsin BC, van Hoesel R, Gordon KA. Lateralization of interimplant timing and level differences in children who use bilateral cochlear implants. *Ear Hear*. 2010;31(4):441-456.
71. Holder JT, Yawn RJ, Nassiri AM, et al. Matched cohort comparison indicates superiority of precurved electrode arrays. *Otol Neurotol*. 2019;40(9):1160-1166.
72. Berg KA, Noble JH, Dawant BM, Dwyer RT, Labadie RF, Gifford RH. Speech recognition as a function of the number of channels in perimodiolar electrode recipients. *J Acoust Soc Am*. 2019;145(3):1556-1564.

