

Standard pleural interventions are not high-risk aerosol generating procedures.

Arnold DT¹, Gregson FKA², Sheikh S², Hamilton FW^{3,4}, Welch H¹, Dipper A¹, Nava GW¹, AERATOR group*, Dodd JW^{1,3}, Clive AO¹, Bzdek BR², Reid JP², Maskell NA¹.

1. Academic Respiratory Unit, University of Bristol, BS10 5NB
2. Bristol Aerosol Research Centre, School of Chemistry, University of Bristol, BS8 1TS
3. MRC Integrative Epidemiology Unit, University of Bristol, BS8 2HW
4. Infection Sciences, North Bristol NHS Trust, BS10 5NB

*AERATOR group (in alphabetical order): The AERATOR group consists of (in alphabetical order):

Arnold, D; Brown, J; Bzdek, B; Davidson, A; Dodd, JW; Gormley M; Gregson, F; Hamilton, F; Maskell, N; Murray, J; Keller, J; Pickering, A.E; Reid, J; Sheikh, S; Shrimpton, A.

Corresponding author:

Dr David Arnold
NIHR Doctoral Research Fellow
Academic Respiratory Unit, University of Bristol
Level 2 Learning & Research Building
Southmead Hospital, Bristol BS10 5NB, UK

Word Count: 1000 words

Key words: COVID-19, Infection Control, Pleural Disease

Abstract

No evidence exists regarding the risk of aerosolisation from pleural procedures. This study used two discrete methodologies, in an environment with no background aerosol interference, to measure aerosol generation from 10 different pleural procedures (3 medical thorascopies, 3 indwelling pleural catheter insertions, 1 therapeutic thoracentesis, and 3 indwelling pleural catheter removals). The measurements indicated that, any aerosol production during these procedures was significantly lower than aerosols produced by the patient breathing or coughing. Pleural procedures should not be considered aerosol generating. We hope this study informs future iterations of guidelines on the appropriate use of PPE when performing these procedures.

Introduction

The coronavirus epidemic has focused attention on the risk of aerosol generating procedures (AGPs) in healthcare. SARS-CoV-2 has been isolated from pleural fluid which has the potential to infect staff if viraemic fluid were aerosolised during procedures^(1, 2). However, evidence for aerosol generation from pleural procedures is very limited. Current guidelines for appropriate use of personal protective equipment (PPE) while performing pleural procedures are based on expert opinion and application of the precautionary principle⁽³⁾. We set out to quantify the aerosol generation from pleural procedures compared to aerosol sampled during normal respiratory activities of breathing and coughing.

Methods

This study performed as part of the AERATOR study assessing the risk of aerosolised transmission of SARS-CoV-2 in healthcare. Ethical approval was granted by North-West Research Ethics Committee (Ref:20/NW/0393).

Aerosol measurements were recorded simultaneously using two devices: an Optical Particle Sizer (OPS, TSI Inc.model 3330, USA, sampling flow rate 1L.min⁻¹, samples 0.3-10µm diameter particles with a sampling period set as 1s) and an Aerodynamic Particle Sizer (APS, TSI Inc.model 3330, USA, sampling flow rate 1L.min⁻¹, sheath flow 4L.min⁻¹, samples 0.5-20µm diameter particles with a sampling period set as 1s). Technical specifications detailed in a previous publication (<https://www.medrxiv.org/content/10.1101/2021.01.29.21250552v1>), with aerosol sampled through a funnel 10cm from the operating site. The proximity of the sampling location is to ensure the source strength is measured without the convolved effects of dispersion and dilution at further distance from the source.

To reduce background aerosol concentration all procedures were performed in an ultra-clean laminar flow operating theatre (EXFLOW 32, Howarth Air Technology, UK) with high efficiency particulate air (HEPA) filtration and air supply rate of 1200m³.s⁻¹ (550-650 air changes/hr). The air flow is 0.2m.s⁻¹ at 1m above the floor below the laminar flow. To demonstrate that this airflow does not affect the sampling efficiency of aerosol generated under the laminar flow, the aerosol generated by voluntary coughing and breathing was sampled in the same position, 10 cm from a subject's face.

Given the different pleural procedures have common themes, the procedures were sub-classified into 5 different elements (see Table 1). We also measured any aerosol produced from 3 pleural fluid management systems.

Results

Ten patients requiring pleural procedures (3 medical thorascopies, 3 indwelling pleural catheter insertions (15.5Fr), 1 therapeutic aspiration (6Fr), 3 indwelling pleural catheter removals)) were recruited to the study, with further 2 further patients with chest tubes already in-situ for pneumothorax with ongoing air leak. The majority of patients were male (10/12) with a median age of 76 (IQR 72-79).

Figure 1 (logarithmic Y axis) shows the peak aerosol number concentration sampled during each procedure compared to peak number concentrations from coughing and mean number concentrations from breathing. For most procedures, the peak number concentration was of similar magnitude to or less than the mean aerosol number concentration measured during breathing from the same patients or from healthy volunteers from a previous study (<https://www.medrxiv.org/content/10.1101/2021.01.29.21250552v1>) and was significantly less than the peak number concentration detected from a cough. The mean concentration for all procedures is reported in Table 1 and is typically much less (up to two orders of magnitude) smaller than the mean concentration sampled when a subject is breathing. Again, it should be stressed that breathing is a sustained activity while coughs and these clinical interventions lead to transient events.

Figure 1: Bar chart showing the peak aerosol number concentration sampled by the APS and OPS methods during different procedural elements compared to mean aerosol number concentrations measured during breathing and peak aerosol number concentrations measured during cough.

Figure 2 illustrates the difference between the particle number concentration sampled during a single procedure compared to breathing or coughing for a patient undergoing a medical thoracoscopy.

Figure 2: Particle concentration over time during a single thoracoscopy compared to aerosol production from the patient breathing and coughing pre-procedure.

Table 1: The peak concentration sampled during a particular procedure of activity in this study (that reported in Fig. 1) and the mean particle concentration sampled throughout, for both the APS and OPS instruments.

	<i>n</i>		Peak concentration / cm-3		Mean concentration / cm-3	
	APS	OPS	APS	OPS	APS	OPS
Pleural Anaesthesia	8	10	0.023	0.036	9.52×10^{-4}	1.44×10^{-3}
Therapeutic Thoracentesis	3	5	0.04	0.036	1.10×10^{-3}	9.66×10^{-4}
Chest drain insertion	5	6	0.048	0.04	9.31×10^{-4}	1.04×10^{-3}
Open pleural space at thoracoscopy	2	3	0.03	0.04	1.19×10^{-4}	1.09×10^{-4}
Chest drain removal	3	3	0.1	0.018	7.89×10^{-3}	0.0124
IPC Aspiration	4	4	0.015	0.03	4.02×10^{-4}	5.07×10^{-4}
Underwater seal bottle	3	4	0.04	0.03	6.56×10^{-4}	9.84×10^{-4}
Thopaz device	2	2	0.06	0.09	1.59×10^{-3}	2.50×10^{-3}
Background	10	12	0.078	0.078	1.05×10^{-3}	1.37×10^{-3}
Cough	4	5	1.95	2.512	-	-
Breathing	5	6	-	-	0.276	0.223

Discussion

This study shows that percutaneous instrumentation of the pleura does not result in significant aerosol generation. Total aerosol production during these procedures was significantly below aerosols produced by breathing or coughing.

Current British Thoracic Society (BTS) guidelines recommend that “closed pleural procedures such as pleural aspirations and chest drain insertion can be undertaken in Level 1 PPE (surgical mask and visor, as well as gown and gloves)” whereas “open procedures such as thoracoscopy and IPC insertion, where pleural fluid may splash, should still be considered AGP. Therefore, Level 2 PPE should be worn (FFP3 mask, long sleeved gown, gloves, eye protection)”. On the basis of this evidence, pleural procedures are not aerosol generating and additional PPE (above that indicated for routine patient care) is not required, although eye protection should be worn given the risk of splash⁽⁴⁾.

It is well recognised that pleural procedures, especially those that generate a negative intrathoracic pressure (e.g. therapeutic thoracentesis) can induce a cough in participants. We would therefore recommend the patient be asked to wear a surgical facemask, which has been shown to significantly reduce aerosol produced during cough (<https://www.medrxiv.org/content/10.1101/2021.01.29.21250552v1>).

Pleural fluid management systems such as underwater seal chest tube bottles have also been seen as a source of aerosol generation with several studies advocating the use of antiviral filters. Duffy and colleagues assessed aerosol generation by bubbling air at different rates through an underwater seal bottle⁽⁴⁾, sampling a maximum aerosol concentration of particles (within the same size range to those studied here, 0.3–10 μ m) during the bubbling process of $\sim 4100\text{ft}^{-3}$, caused by atomization of the water. This equates to a peak concentration of $\sim 0.14\text{cm}^{-3}$, which is similar to the peak concentrations that we observed during the fluctuations of sampled aerosol concentration during pleural procedures. We show that the peak number concentration sampled during the pleural procedures was similar to that sampled during the background measurement, orders of magnitude smaller than that sampled during a cough and was never greater than the mean concentration sampled during a period of quiet breathing. However, given our sample size for underwater seal bottles is small ($n=3$) and the mitigating factors are simple we feel guidance should still encourage the use of viral filters or Thopaz devices until further evidence is gathered, especially in pneumothoraces with high air leaks.

In summary, using two methodologies to measure aerosol emission with no background aerosol interference, this study has shown that percutaneous pleural procedures are non-

aerosol generating. We hope this will inform future iterations of guidelines on the appropriate use of PPE when performing these procedures.

Acknowledgements

We would also like to thank all the patients and volunteers who took part in the AERATOR study.

Funding

The AERATOR study was funded by an NIHR/UKRI COVID-19 Rapid Rolling Call (Grant number COV0333).

Competing interests

All authors have completed the ICMJE uniform disclosure form at www.icmje.org/coi_disclosure.pdf and declare: no support from any organisation for the submitted work; no financial relationships with any organisations that might have an interest in the submitted work in the previous three years; no other relationships or activities that could appear to have influenced the submitted work.

Contributorship statement

DTA, FWH and NAM developed the study idea. FKAG, FWH, JWD, BRB and JPR developed the sampling design. DTA, SS, FWH, HW, AD, GN, NAM collected the primary aerosol data. FKAG, SS, BRB and JPR analysed the aerosol data. All authors were involved in the writing of the manuscript.

References

1. Malik MI, Fox N, Chopra A, Hughes HY, Washburn R, Huggins JT. Positive pleural fluid RT-PCR for virus detection in SARS-CoV-2 pneumonia. QJM. 2020;113(12):888-9.
2. Mei F, Bonifazi M, Menzo S, Di Marco Berardino A, Sediari M, Paolini L, et al. First Detection of SARS-CoV-2 by Real-Time Reverse Transcriptase-Polymerase Chain Reaction Assay in Pleural Fluid. Chest. 2020;158(4):e143-e6.
3. Halifax R, Wrightson J, Bibby A, Walker S, Stanton A, De Fonseca D, et al. Pleural services during the COVID-19 Pandemic <https://www.brit-thoracic.org.uk/document-library/quality-improvement/covid-19/pleural-services-during-covid-19-pandemic/> [accessed 11.03.2021]. 2020.
4. Sun S, Ravintharan N, Jassal K, Nandurkar R, Yesul K, Boo E, et al. COVID-19 pandemic consciousness: droplet contamination and aerosolization during pleural decompression. ANZ J Surg. 2020;90(12):2588-91.
5. Duffy C, Kidd A, Francis S, Tsim S, McNaughton L, Ferguson K, et al. Chest drain aerosol generation in COVID-19 and emission reduction using a simple anti-viral filter. BMJ Open Respir Res. 2020;7(1).



