1	
2	
3	
4	Validation of physical activity levels from shank-placed Axivity AX6
5	accelerometers
6	
7	
8	Fatima Gafoor ^{1*} , Matthew Ruder ² , Dylan Kobsar ^{1,2}
9	
10	
11	¹ School of Biomedical Engineering, McMaster University, Hamilton, Ontario, Canada
12	² Department of Kinesiology, McMaster University, Hamilton, Ontario, Canada
13	
14	
15	* Corresponding author
16	E-mail: gafoorf@mcmaster.ca (FG)
17	
18	
19	
20	
21	
22	

23 Abstract

24 This cross-sectional study aimed to identify and validate cut-points for measuring physical 25 activity using Axivity AX6 accelerometers positioned at the shank in older adults. Free-living 26 physical activity was assessed in 35 adults aged 55 and older, where each participant wore a shank 27 mounted Axivity and a waist mounted ActiGraph simultaneously for 72 hours. Optimized cut-28 points for each participant's Axivity data were determined using an optimization algorithm to align 29 with ActiGraph results. To assess validity between the physical activity assessments from the 30 optimized Axivity cut-points, a leave-one-out cross validation was conducted. Bland-Altman plots 31 with 95% limits of agreement, intraclass correlation coefficients (ICC), and mean differences were 32 used for comparing the systems. The results indicated good agreement between the two 33 accelerometers when classifying sedentary behaviour (ICC = 0.85) and light physical activity (ICC 34 = 0.80), and moderate agreement when classifying moderate physical activity (ICC = 0.67) and 35 vigorous physical activity (ICC = 0.70). Upon removal of a significant outlier, the agreement was 36 slightly improved for sedentary behaviour (ICC = 0.86) and light physical activity (ICC = 0.82). 37 but substantially improved for moderate physical activity (ICC = 0.81) and vigorous physical 38 activity (ICC = 0.96). Overall, the study successfully demonstrated the capability of the resultant 39 cut-point model to accurately classify physical activity using Axivity AX6 sensors placed at the 40 shank.

41 Introduction

42 Physical activity (PA) for the aging population has been identified as a protective factor 43 for various musculoskeletal disorders and noncommunicable diseases [1]. Even with the well-44 known benefits of PA, many older adults are representative of the most inactive portion of the

45 population [2]. Thus, an accurate assessment of PA can help to provide insight on the overall health 46 and rehabilitation of the aging population. Traditionally, PA data has been collected using self-47 reported measures such as questionnaires, activity logs, interviews, and diaries [3]. Common PA 48 self-report tools include the Physical Activity Scale for the Elderly (PASE), the International 49 Physical Activity Questionnaire (IPAQ), and the Short Questionnaire to Assess Health-Enhancing 50 Physical Activity (SQUASH) [4, 5, 6]. However, these measures are subject to recall and social 51 desirability biases [7]. Furthermore, many have difficulties distinguishing between different PA intensities in questionnaires [8, 9]. Consequently, these subjective measures lack accuracy, as they 52 53 often over- or underestimate the true levels of PA [10].

With the advent of wearable accelerometers, researchers can overcome these limitations and obtain objective measures of PA. These accelerometers can be worn outside of the lab and collect free-living data for multiple days, thereby providing a more comprehensive understanding of an individual's PA levels [7]. Among these devices, ActiGraph devices are widely used as the gold standard for clinical and research-based PA assessment, with over 20,000 papers published on its use in the past two decades [11].

The quantification of PA levels using these devices typically involves assessing time spent in different PA intensity ranges such as sedentary behaviour (SB), light physical activity (LPA), moderate physical activity (MPA) and vigorous physical activity (VPA). This is achieved by measuring the magnitude of three-dimensional acceleration signals, known as the resultant acceleration. The resultant accelerations are then used to determine activity counts, with higher accelerations resulting in greater the number of activity counts measured per minute [7]. Subsequently, cut-points are applied to these activity counts to classify the PA intensity,

delineating data into the different categories of SB, LPA, MPA, and VPA. These cut-points are
specific to the wear location of the device.

69 The ActiGraph GT9X Link, equipped with an inertial measurement unit (IMU) including 70 accelerometer, gyroscope, and magnetometer sensors, is one device commonly used for collecting 71 PA data. These devices can be worn at the wrist, waist, and the ankle but require different cut-72 point models for each site as the level of accelerations vary between segments [12]. Therefore, 73 deploying a sensor at any new location would require updating and validating a new set of cut-74 points for that placement site. While wrist, waist, and ankle placements offer sufficient flexibility 75 for assessing PA, recent advancements in sensor capabilities, including improved data quality, 76 higher frequency measurements, and enhanced data storage and battery life, have opened up many 77 new possibilities for free-living PA assessments.

78 Most notably, wearable sensors are being increasingly used to assess free-living gait. Free-79 living gait gives us the opportunity to gain insight into the motion of the limbs in an ecologically 80 valid, real-world setting, rather than a conventional gait lab. While conventional gait labs are 81 widely used in gait research, they are highly controlled settings and may be non-representative of 82 an individual's daily, functional gait [13]. Axivity sensors are an example of IMU sensors that 83 have been used in gait research surrounding physical function and disease state [14], but also 84 possess PA assessment capability given their potential for high frequency data collections with 85 enhanced data storage and battery life.

While there are a wide variety of placement sites used to acquire free-living gait data, a common placement to assess movement and impact accelerations near the knee is the proximal shank [15, 16, 17]. This placement site may offer insights into the progression of musculoskeletal and neurological diseases, such as knee osteoarthritis and Parkinson's through gait assessments

90 [18, 19]. Although this sensor placement can be ideal in assessing free-living gait, there are no
91 methods or cut-points available to determine PA from sensors worn at this wear location. As such,
92 obtaining concurrent PA data while measuring free-living gait would inefficiently require another
93 sensor placed at one of the previously validated sites.
94 Therefore, the purpose of this study is to develop and validate cut-points for a free-living

95 gait-capable sensor, the Axivity AX6, placed at the proximal shank by comparing it to the gold 96 standard waist-mounted ActiGraph GT9X. It was hypothesized that the agreement of SB and LPA 97 would be higher than the agreement of MPA and VPA. The outcomes compared between the 98 sensors were the percentage of time spent in the different PA intensities: SB, LPA, MPA, and VPA 99 for a total wear time of 72 hours.

100 Methods

101 **Participants**

102 Thirty-five adults over the age of 55 years (18F, age: 71 ± 9 years, height: 167.28 ± 9.79 cm, mass: 76.50 ± 15.10 kg, BMI: 27.00 ± 3.62 kg/m²) were recruited in this cross-sectional 103 104 validation study. The inclusion criteria included anyone over the age of 55 years that could 105 ambulate without any walking aids. Participants were recruited from the Physical Activity Centre 106 of Excellence (PACE) at McMaster University, a centre which encourages community exercise 107 and rehabilitation for older adults, from January 12, 2023 to March 23, 2023. All participants 108 provided their written informed consent prior to enrolling in this study. This study was approved 109 by the McMaster Research Ethics Board, Hamilton, ON.

110

111 **Procedures**

112 Participants were asked to report their age, sex, height, mass. Each subject reported if they 113 had received any lower limb injuries in the past year, had been diagnosed with osteoarthritis, and 114 if they had received any lower limb replacement surgeries. Two wearable sensors were used to 115 concurrently collect free-living PA data for 72 consecutive hours. The first of which was the gold 116 standard ActiGraph GT9X (ActiGraph LLC, FL, United States) which was worn by each 117 participant above the right iliac crest with a waistband and pouch which held the 118 device. Participants were encouraged to ensure the pouch sat above their right hip for the full 72 119 hours. Participants were instructed to only take off the waistband to shower or bath, and if 120 necessary, while sleeping, and wear it immediately upon waking and after their shower or bath. 121 The ActiGraph was initialized to start collecting $\pm 8g$ accelerometer data at 100 Hz simultaneously 122 with the Axivity for 72 hours through the ActiLife v6.13.4 software. The 'idle' sleep mode in the 123 ActiLife software was disabled for the ActiGraph sensors.

124 The sensor to be validated was the Axivity AX6 (Axivity Ltd, Newcastle upon Tyne, 125 United Kingdom) which was placed below the right knee, medial and inferior to the right tibial 126 tuberosity, of each participant using SIMPATCH adhesive patches. The Axivity accelerometer 127 was initialized through the Open Movement software (OMGui, version 1.0.0.44) and recorded $\pm 8g$ 128 accelerometer data at 100Hz, as well as $\pm 1000^{\circ}$ /s gyroscope data that was not analyzed in the 129 current study. Participants were shaved with an electric razor (Philips OneBlade, Amsterdam, 130 Netherlands) before the sensors were placed to ensure the sensors were both comfortable and 131 adhered to the skin well for the entirety of the 72 hours.

132

133 Data analysis

Following the retrieval of the sensors after 72 hours of wear, the data from the ActiGraphs were downloaded and processed through the ActiLife software. Each participants' data was categorized into SB, LPA, MPA, and VPA using the Freedson cut-points and 60-second epochs [20] which are available through the ActiLife software. The percentage of time spent in the different intensities was computed by the ActiLife software, based on the activity counts per minute. The time totalled to 72 hours for each participant.

140 Raw from the Axivity sensor were downloaded through the OpenMovement software 141 (OMGui) and saved in a binary CWA file. An open-source MATLAB (MathWorks, Natick, MA, 142 USA) script by Felix Liu [21], which uses methods published by Brønd, Andersen, and Arvidsson 143 [22] was used to generate ActiGraph equivalent activity counts per minute from Axivity 144 accelerometers. The activity counts for each participant were categorized into PA intensities using 145 optimized cut-points for each participant. However, using the same cut-point values applied to the 146 waist data would results in invalid data given the difference in acceleration signals across these sites [23]. Therefore, to optimize a new set of cut-points for the shank-placed Axivity sensor, an 147 148 optimization algorithm in MATLAB was created, using the Global Optimization Toolbox [24]. 149 Fig 1 outlines the optimization algorithm and resulting cross validation. In short, this optimization 150 algorithm was given raw Axivity acceleration data, an initial starting set of cut-points based on 151 pilot testing of data collected prior to this study, and the gold-standard results obtained from the 152 ActiGraph (i.e., time spent in SB, LPA, MPA, VPA). It then sequentially optimized the cut-points 153 for the Axivity data by first evaluating SB, then LPA, MPA, and finally VPA, resulting in three 154 values to distinguish SB, LPA, MPA, and VPA. In other words, the optimization algorithm was 155 given the raw data from the Axivity at the shank and asked to find the cut-points that best 156 approximated the results from the waist-mounted ActiGraph. This optimization occurred at the

individual level, resulting in 35 distinct sets of cut-points (i.e., one set for each participant). To generate a cross-validated set of cut-points that were blinded from each individual's result, a leaveone-out cross validation (LOOCV) was applied. This involved averaging the cut-points from 34 participants (i.e. all other participants) into a single set of cut-points that could then be applied to the data of the left-out subject. Consequently, the PA data generated for each subject was solely based on the optimized results from the other 34 subjects, ensuring complete blinding to the information of the excluded subject.

164

165 Fig 1. Flowchart of Leave-one-out cross validation analysis.

166 Statistical analysis

167 The agreement between the ActiGraph and the LOOCV Axivity results for each PA 168 intensity: SB, LPA, MPA, and VPA, were assessed using Bland-Altman plots with 95% limits of 169 agreement (LOA: mean difference of methods ± 1.96 SD) [25]. Additionally, intraclass correlation 170 coefficients (ICC; single measure, absolute agreement) with 95% confidence intervals (CIs) were 171 computed for each intensity to measure reliability of the Axivity. The ICC values were interpreted 172 as poor (ICC < 0.5), moderate (0.5 < ICC < 0.75), good (0.75 < ICC < 0.9), and excellent (ICC >173 0.9) [26]. The agreement between the ActiGraph and individually optimized Axivity results were 174 also computed, to serve as a comparison of the training data (i.e., individually optimized) versus 175 the generalizable, cross-validated results.

176

177 **Results**

Participant demographics are reported in Table 1. While 35 individuals were recruited forthe study, one participants' data was omitted from the analysis due to early removal of the Axivity

180 sensor. Results from the training model and the LOOCV analysis are shown in Table 2. Individual 181 results from the LOOCV analysis are shown in the Bland-Altman plots in Fig 2. The results from 182 the optimized cut-points training model matched closely with the ActiGraph results and all 183 intensities had excellent agreement, as expected given they are individually optimized. When 184 performing the LOOCV analysis, results showed the SB and LPA have good reliability, with mean 185 differences of 2.40% and 2.37%, respectively. Whereas the MPA and VPA have moderate

reliability and mean differences of 0.51% and 0.10% respectively.

Table 1. Participant Demographics.

Age	71 ± 9 years
Male	48.6 %
Height	$167.3 \pm 9.8 \text{ cm}$
Mass	76.5 ± 15.1 kg
BMI	$27.0 \pm 3.6 \text{ kg/m}^2$
Previous Lower Limb Surgeries	5.7%
Previous Lower Limb	11.4%
Replacements	
Diagnosed Osteoarthritis	31.4%
Mean Oxford Knee Score ^a	37.3 ± 7.1

^aThe Oxford Knee Score measures an individual's daily level of function, and how they have been impacted by pain in their knees. Participants only filled out the OKS questionnaire if they indicated they had osteoarthritis in either or both of their knees. Scores range from 0 to 48, with 48 representing maximal functioning knees [27]. A score ranging from 30 to 39 may indicate mild to

- 192 moderate knee osteoarthritis [28].
- 193

194**Table 2. Results.**

	Training Model			LOOCV Analysis		
PA Intensity	ICC	Mean Difference (%)	95% Limits of Agreemen	ICC	Mean Difference (%)	95% Limits of Agreemen
			t			t
Sedentary	0.99 (0.99,	0.01 ± 0.01	-0.02, 0.02	0.85 (0.73,	2.40 ± 1.84	-6.11, 5.85
	1.0)			0.92)		
Light	0.99 (0.99,	0.01 ± 0.01	-0.02, 0.02	0.80 (0.65,	2.37 ± 1.97	-5.93, 6.25
	1.0)			0.9)		

Moderate	0.98 (0.96,	0.05 ± 0.22	-0.03, 0.03	0.67 (0.44,	0.51 ± 0.70	-1.96, 1.82
	0.99)			0.82)		
Vigorous	0.96 (0.93,	0.04 ± 0.22	-0.02, 0.02	0.70 (0.48,	0.10 ± 0.73	-1.96, 1.30
	0.98)			0.84)		

195

196Fig 2. Bland-Altman plot of the time spent in different PA intensities in minutes, for197agreement between the Axivity and the ActiGraph in the LOOCV analysis.

198

199 After initial data analysis, a significant outlier was identified in the LOOCV analysis. 200 Outliers were defined as data points that deviated significantly from the remaining values of the 201 dataset, using a criterion of falling outside three standard deviations from the mean value. Based 202 on this criterion, one participant was considered an outlier in the LOOCV dataset. In order to 203 ensure the integrity of the results and subsequent cut-point models, a careful decision was made to 204 exclude this outlier from the analysis. By removing this outlier from the dataset, we aimed to 205 achieve a more representative dataset for further analysis. Following the outlier's removal, ICCs 206 for SB, LPA, and MPA showed good agreement, while the ICC for VPA showed excellent 207 agreement, as depicted in Table 3. Individual results from the LOOCV analysis with the outlier 208 removed are shown in the Bland-Altman plots in Fig 3.

209 **Table 3. Results with outlier removed.**

	Training Model			LOOCV Analysis		
PA	ICC	Mean	95%	ICC	Mean	95%
Intensity		Difference	Limits of		Difference	Limits of
		(%)	Agreement		(%)	Agreement
Sedentary	0.99 (0.99,	0.01 ± 0.01	-0.03, 0.02	0.86 (0.74,	2.34 ± 1.83	-5.88, 5.88
	1.0)			0.93)		,
Light	0.99 (0.99,	0.01 ± 0.01	-0.03, 0.03	0.82 (0.66,	2.28 ± 1.94	-5.92, 5.94
	1.0)			0.91)		,
Moderate	0.98 (0.96,	0.01 ± 0.01	-0.03, 0.03	0.81 (0.64,	0.42 ± 0.46	-1.20, 1.27
	0.99)			0.90)		

Vigorous	0.96 (0.93,	0.004 ±	-0.02, 0.02	0.96 (0.92,	0.10 ± 0.23	-0.53, 0.45
	0.98)	0.01		0.98)		

210

211

Fig 3. Bland-Altman plot of the time spent in different PA intensities in minutes, for agreement between the Axivity and the ActiGraph in the LOOCV analysis, with the outlier removed.

- 215
- Based on these results, the resultant cut-points developed for the Axivity accelerometer,

217 when placed below the knee, were computed as the average of the 34 participants and are shown

in Table 4. Also shown in Table 4 are the cut-points after the outlier was removed, which is the

219 more representative set of cut-points to use for accelerometers worn at the shank.

220 **Table 4. Resultant Cut-points.**

	Counts per Minute				
Intensity	Complete Dataset	Outlier Removed			
Sedentary	0 - 777	0 - 792			
Light	778 - 8415	793 - 8356			
Moderate	8416 - 18731	8357 - 18437			
Vigorous	18732 <	18438 <			

221

222 **Discussion**

The purpose of this study was to determine and validate cut-points to measure PA with Axivity sensors placed at the shank in older adults. Overall, the Axivity demonstrated good agreement with the gold standard for PA assessment, the ActiGraph, when assessing SB and LPA. When classifying MPA and VPA, the Axivity showed moderate agreement with the ActiGraph. However, with outliers accounted for, MPA and VPA showed good and excellent agreement, respectively. These data demonstrated that accelerometers placed just below the knee can effectively measure an older adult's PA. Overall, the resultant cut-points determined in this study

enable the assessment of PA in older adults at the proximal shank, either independently or inconjunction with research focused on free-living gait data from this site.

232 Our results supported our original hypothesis which predicted that SB and LPA would have 233 greater agreement than MPA and VPA with the ActiGraph, as the SB and LPA demonstrated ICCs 234 of 0.85 and 0.80, respectively. Whereas the MPA and VPA intensities demonstrated ICCs of 0.67 235 and 0.70, respectively. This may be because participants, being older adults, did not all perform 236 large amounts of MPA and VPA. The average percentage of time spent in VPA was 1.04 %, with 237 26 participants having no VPA. This likely reduced the samples of MPA and VPA collected and 238 could have affected the average of the resultant cut-points to identify MPA and VPA. These results 239 are consistent with previous studies that assessed PA between ActiGraph and other devices, 240 including the Axivity, which also found greatest agreement when classifying SB and LPA over 241 MPA, and VPA, such as the study done by Rowlands et al. [29], due to the relatively small amount 242 of time spent in moderate and vigorous PA levels [30].

243 With one outlier removed, these ICCs improved to 0.86, 0.82, 0.81, and 0.96 for SB, LPA, 244 MPA, and VPA, respectively. Upon evaluation, the recorded Axivity accelerations and 245 subsequently derived activity counts of the outlier were substantially higher in magnitude 246 compared to those of the remaining participants, despite the ActiGraph data of the outlier aligning 247 with that of the other participants. Consequently, the outlier's optimized cut-points were 248 significantly greater than the average cut-points of all subjects. Thus, when conducting the 249 LOOCV including the outlier's optimized cut-points, the model overestimated the time spent in 250 Specifically, the resultant cut-points including the outlier are higher for MPA and VPA intensities 251 when compared to the cut-points with the outlier removed. Higher recorded or erroneous 252 accelerations from one subject's Axivity can be due to many potential reasons, which can include

253 sensor movement as a result of being loosely attached and the introduction of artifacts and noise 254 into the accelerometer readings. Nevertheless, we cannot pin-point the exact cause as to why one 255 subject's Axivity sensor collected greater accelerations than others' Axivity sensors. The 256 optimized cut-points of the outlier affected all PA intensities, however, it had a limited impact on 257 SB and LPA data, which further highlights the quality of these outcomes in older adults compared 258 to MPA and VPA. Despite the outlier's removal improving the results to excellent validity, we 259 still caution the use of VPA data in older adults moving forward, given the limited amount of data 260 in these intensity levels.

261 With respect to the cut-points used for MPA and VPA, they are similar to cut-points found 262 in a study done by Rhudy et. al. [31] which identified cut-points for the ActiGraph GT9X at 263 varying wear-locations proximal to the shank, such as the ankle, foot, along with the wrist and hip 264 to measure moderate and vigorous PA. These cut-points are shown in Table 5 along with the cut-265 points from the shank developed from the results of the current study. These wear locations offer some opportunity for comparing cut-points at proximal locations, such as the ankle and foot. 266 267 Theoretically, the cut-points for the ankle and foot should be greater, as the impact accelerations 268 experienced would be greater than those at the shank. It is also expected that the cut-points for the 269 shank for both MPA and VPA are greater than those of the hip. This is mostly reflected in the 270 results from Rhudy et al. and the current study, as the cut-points for MPA and VPA measured at 271 the ankle are 14767 and 17818 counts per minute, respectively, and the cut-points for MPA and 272 VPA measured at the foot are 20575 and 22629 counts per minute, respectively. For the current 273 study, the cut-points for MPA and VPA measured at the shank are 8357 and 18438. As expected, 274 these shank cut-points are greater than the cut-points for the hip for MPA and VPA, which are 275 4978 and 6227 counts per minute. Furthermore, the cut-point for MPA at the shank is less than the

cut-points at the ankle and foot, as predicted. While the cut-point for VPA at the shank is less than the cut-point for the foot, it is greater than the cut-point at the ankle. This could be since the cutpoints for the foot and ankle in the aforementioned study were identified through a young, healthy population in a controlled setting, whereas the current study's population consisted solely of older adults in free-living conditions.

281

Table 5. Cut-points for MPA and VPA at foot, ankle, and hip wear locations from Rhudy et
 al. [31].

Wear Location	MPA cut-point	VPA cut-point
Foot	20575	22629
Ankle	14767	17818
Shank	8357	18438
Hip	4978	6227

²⁸⁴

This study has some limitations to be discussed. Firstly, participants removing the ActiGraph sensor when showering may have influenced the results, as the Axivity was worn for the full 72 hours. This likely would have minimal influence over the results with both accelerometers likely resulting in a SB classification, as showers are typically short in duration and most people won't move a significant amount while showering. However, if participants forgot to wear, or delayed, the placement of the ActiGraph after their shower, this could adversely affect agreement between two sensors.

Secondly, it is possible that individuals could have performed movement with their lower limbs, without moving their trunk, such as leg raises, which could have recorded more accelerations in the Axivity than the ActiGraph, if the hip-worn accelerometer was not experiencing movement. This would have impacted the final cut-points that were developed. Nevertheless, the current findings present a first step at classifying PA from shank-placed

accelerometers and identifying cut-points for this wear location, which can be widely used in gaitresearch in the elderly.

299

300 Conclusion

These results displayed a good agreement between the gold standard ActiGraph sensors and the Axivity sensors to classify PA and SB when worn at the shank. The resultant cut-points can be used for accelerometers, including the Axivity and ActiGraph, worn specifically just below the knee at the shank. The introduction of shank cut-points are a novel contribution that will aid in simplifying concurrent PA assessments that occur in older adults also completing gait assessments.

307 Acknowledgements

308 I would like to thank the Physical Activity Centre of Excellence at McMaster University for 309 supporting this study, as well as all the participants for their time and efforts, that made this study 310 possible.

311

312 **References**

- Langhammer B, Bergland A, Rydwik E. The importance of physical activity exercise among older people. BioMed Res Int. 2018;2018: 7856823. doi:10.1155/2018/7856823
- Watson KB, Carlson SA, Gunn JP, Galuska DA, O'Connor A, Greenlund KJ, et al. Physical inactivity among adults aged 50 years and older - United States, 2014. MMWR Morb Mortal Wkly Rep. 2016;65: 954–958. doi:10.15585/mmwr.mm6536a3
- Martins JC, Aguiar LT, Nadeau S, Scianni AA, Teixeira-Salmela LF, Faria CDC de M.
 Measurement properties of self-report physical activity assessment tools in stroke: a
 protocol for a systematic review. BMJ Open. 2017;7: e012655. doi:10.1136/bmjopen-2016 012655

- Washburn RA, McAuley E, Katula J, Mihalko SL, Boileau RA. The physical activity scale for the elderly (PASE): evidence for validity. J Clin Epidemiol. 1999;52: 643–651. doi:10.1016/s0895-4356(99)00049-9
- Scraig CL, Marshall AL, Sjöström M, Bauman AE, Booth ML, Ainsworth BE, et al.
 International physical activity questionnaire: 12-country reliability and validity. Med Sci
 Sports Exerc. 2003;35: 1381–1395. doi:10.1249/01.MSS.0000078924.61453.FB
- Wendel-Vos GCW, Schuit AJ, Saris WHM, Kromhout D. Reproducibility and relative validity of the short questionnaire to assess health-enhancing physical activity. J Clin Epidemiol. 2003;56: 1163–1169. doi:10.1016/s0895-4356(03)00220-8
- Sylvia LG, Bernstein EE, Hubbard JL, Keating L, Anderson EJ. A practical guide to measuring physical activity. J Acad Nutr Diet. 2014;114: 199–208. doi:10.1016/j.jand.2013.09.018
- Sallis JF, Saelens BE. Assessment of physical activity by self-report: status, limitations, and
 future directions. Res Q Exerc Sport. 2000;71: 1–14.
 doi:10.1080/02701367.2000.11082780
- 337 9. Timperio A, Salmon J, Crawford D. Validity and reliability of a physical activity recall
 338 instrument among overweight and non-overweight men and women. J Sci Med Sport.
 339 2003;6: 477-491. doi:10.1016/S1440-2440(03)80273-6
- Prince SA, Adamo KB, Hamel ME, Hardt J, Gorber SC, Tremblay M. A comparison of
 direct versus self-report measures for assessing physical activity in adults: a systematic
 review. Int J Behav Nutr Phys Act. 2008;5: 56. doi:10.1186/1479-5868-5-56
- 11. Neishabouri A, Nguyen J, Samuelsson J, Guthrie T, Biggs M, Wyatt J, et al. Quantification
 of acceleration as activity counts in ActiGraph wearable. Sci Rep. 2022;12: 11958.
 doi:10.1038/s41598-022-16003-x
- 346 12. Romero E, Neuman M, Warrington R. Kinetic energy harvester for body motion. 2009.
- 347 13. Shah VV, McNames J, Mancini M, Carlson-Kuhta P, Spain RI, Nutt JG, et al. Laboratory
 348 versus daily life gait characteristics in patients with multiple sclerosis, Parkinson's disease,
 349 and matched controls. J NeuroEngineering Rehabil. 2020;17: 159. doi:10.1186/s12984-020350 00781-4
- 14. Powell D, Nouredanesh M, Stuart S, Godfrey A. Investigating the AX6 inertial-based
 wearable for instrumented physical capability assessment of young adults in a low-resource
 setting. Smart Health. 2021;22: 100220. doi:10.1016/j.smhl.2021.100220
- 15. Li Y, Wang L. Human activity recognition based on residual network and BiLSTM.
 sensors. 2022;22: 635. doi:10.3390/s22020635
- 356

- 16. Li Q, Young M, Naing V, Donelan JM. Walking speed estimation using a shank-mounted
 inertial measurement unit. J Biomech. 2010;43: 1640–1643.
 doi:10.1016/j.jbiomech.2010.01.031
- 17. Celik Y, Stuart S, Woo WL, Godfrey A. Wearable inertial gait algorithms: impact of wear
 location and environment in healthy and Parkinson's populations. Sensors. 2021;21: 6476.
 doi:10.3390/s21196476
- 18. Tsukamoto H, Saito K, Saito H, Kijima H, Akagawa M, Komatsu A, et al. A novel
 classification of coronal plane knee joint instability using nine-axis inertial measurement
 units in patients with medial knee osteoarthritis. Sensors. 2023;23: 2797.
 doi:10.3390/s23052797
- 367 19. Morris R, Hickey A, Del Din S, Godfrey A, Lord S, Rochester L. A model of free-living
 368 gait: A factor analysis in Parkinson's disease. Gait Posture. 2017;52: 68–71.
 369 doi:10.1016/j.gaitpost.2016.11.024
- 370 20. Freedson PS, Melanson E, Sirard J. Calibration of the Computer Science and Applications,
 371 Inc. accelerometer. Med Sci Sports Exerc. 1998;30: 777.
- 372 21. Liu F. Axivity_AX3. 2022. Available: https://github.com/FelixLiu-SF/Axivity_AX3
- Brønd JC, Andersen LB, Arvidsson D. Generating ActiGraph counts from raw acceleration
 recorded by an alternative monitor. Med Sci Sports Exerc. 2017;49: 2351–2360.
 doi:10.1249/MSS.00000000001344
- 23. LaMunion SR, Bassett DRJ, Toth LP, Crouter SE. Effect of wear location on ActiGraph
 activity counts: 2286 Board #299 June 1 2: 00 PM 3: 30 PM. Med Sci Sports Exerc.
 2017;49: 643. doi:10.1249/01.mss.0000518693.36595.09
- 379 24. Optimization Toolbox version: 9.4 (R2022b), Natick, Massachusetts: The MathWorks Inc.;
 380 2022.
- Martin Bland J, Altman Douglas G. Statistical methods for assessing agreement between two
 methods of clinical measurement. The Lancet. 1986;327: 307–310. doi:10.1016/S0140 6736(86)90837-8
- Koo TK, Li MY. A guideline of selecting and reporting intraclass correlation coefficients for
 reliability research. J Chiropr Med. 2016;15: 155–163. doi:10.1016/j.jcm.2016.02.012
- 386 27. Murray DW, Fitzpatrick R, Rogers K, Pandit H, Beard DJ, Carr AJ, et al. The use of the
 387 Oxford hip and knee scores. J Bone Joint Surg Br. 2007;89-B: 1010–1014.
 388 doi:10.1302/0301-620X.89B8.19424
- 389 28. Oxford Knee Score | Auckland Surgery. [cited 13 Jul 2023]. Available:
- 390 https://www.aucklandsurgery.co.uk/navigator/oxford-knee-score/

- 29. Rowlands AV, Mirkes EM, Yates T, Clemes S, Davies M, Khunti K, et al. Accelerometer assessed physical activity in epidemiology: are monitors equivalent? Med Sci Sports Exerc.
 2018;50: 257. doi:10.1249/MSS.00000000001435
- 30. Pfister T, Matthews CE, Wang Q, Kopciuk KA, Courneya K, Friedenreich C. Comparison of
 two accelerometers for measuring physical activity and sedentary behaviour. BMJ Open
 Sport Exerc Med. 2017;3: e000227. doi:10.1136/bmjsem-2017-000227
- 397 31. Rhudy MB, Dreisbach SB, Moran MD, Ruggiero MJ, Veerabhadrappa P. Cut points of the
 398 Actigraph GT9X for moderate and vigorous intensity physical activity at four different
 399 wear locations. J Sports Sci. 2020;38: 503–510. doi:10.1080/02640414.2019.1707956

400

401 Supporting Information

- 402 S1 Table. Participant Axivity and ActiGraph Data. Axivity Data is missing from Participant
- 403 21 due to early removal of the Axivity AX6 sensor.

404

405



Individual Optimization Process

Fig1



Fig2

Percentage of Time in SB

Percentage of Time in LPA



Fig3