Assessing quadriceps strength in patellofemoral pain

- patients: A study on the reliability and validity of a low
 - cost load-cell for clinical practice
- 4 Germari Deysel, Mariaan van Aswegen, Mark Kramer*
- 5 Physical Activity, Sport, and Recreation (PhasRec) Research Focus Area, North-West
- 6 University, Potchefstroom, North West Province, South Africa
- 8 *Corresponding author

1

3

7

10

11

9 E-mail: <u>mark.kramer@nwu.ac.za</u> (MK)

Abstract

12

13

Background

- Patellofemoral pain (PFP) is a common knee complaint affecting diverse populations both
- acutely and chronically. Quadriceps muscle weakness is one possible aetiology, but current
- devices for measuring muscle strength (isokinetic dynamometer and hand-held
- 17 dynamometers) are frequently too expensive for practitioners, especially in under-resourced
- 18 settings. There is a need to evaluate a low-cost device to manage rehabilitation of people with
- 19 PFP.

20

Methods

- Isometric quadriceps strength of participants aged 18-35 years (total [n = 33], control group
- [n = 17] and PFP group [n = 16]) were evaluated on an isokinetic dynamometer and a low-
- 23 cost load cell at baseline and after an 8-week non-standardised intervention for validity
- 24 scores.

25 **Results**

- The load cell showed high absolute and relative reliability (intraclass correlation coefficient =
- 27 0.89-0.99; typical error of measurement = 3.9-10.4%). Clinically meaningful difference
- scores (12.2-45 Nm) were greater than the typical error of measurement, implying sufficient
- 29 sensitivity of the load cell to measure true changes in isometric quadricep strength. Strong to
- 30 very strong correlations were evident between the load cell and isokinetic dynamometer
- 31 torque measurements (r = 0.88-0.90, SEE = 0.05-0.07 Nm), but slope values ($\beta = 0.65-0.77$)
- 32 indicated that torque from the load cell was typically lower than that obtained from the
- isokinetic dynamometer. An average systematic bias of 16.3-28.8 Nm was evident in favour

- of the isokinetic dynamometer, with no statistically significant between-group differences
- being noted between the baseline and follow-up testing.

Conclusion

36

42

43

- 37 The load cell is a reliable instrument, sensitive enough to detect clinically meaningful
- differences in quadriceps strength in healthy individuals and those with PFP. The load cell
- 39 lacks validity and cannot replace isokinetic dynamometry. Given the low cost and excellent
- 40 reliability, the load cell can be a valuable tool to assess quadriceps muscle deficits and track
- 41 rehabilitation progress in people with PFP.

Introduction

- Patellofemoral pain (PFP) is a common knee complaint characterised by retropatellar pain
- 45 (pain behind the kneecap) or peripatellar pain (pain around the kneecap), aggravated by at
- least one knee loading activity during weight bearing on a flexed knee (e.g., squatting, stair
- 47 ambulation, jogging, running, hopping or jumping) [1]. Prevalence rates of PFP vary
- 48 according to age, gender, and activity levels with the literature showing rates of 28% in
- adolescents [2], 23% in adults [3], 15% in female adults [4] and 40% in recreational runners
- 50 [5]. People living with PFP often struggle with acute and chronic effects such as physical,
- 51 emotional and social problems during sports participation, activities related to work and
- 52 activities of daily living, which can persist several years [6–10]. Furthermore, PFP is
- considered a risk factor for the development of patellofemoral osteoarthritis [11], which has
- been attributed to reduced quadriceps strength relative to task-related loading of the
- patellofemoral joint [5,12,13]. It is therefore plausible to consider quadriceps strength as a
- 56 protective agent against patellofemoral osteoarthritis cartilage loss [14].

Given that people with PFP can be divided into different subgroups, depending on the individual aetiologies [15], it is important to determine whether quadriceps weakness is part of the cause for PFP in the individual, and to track whether progress is made when conducting an exercise program to improve quadriceps strength [15]. Presently, the 'gold standard' instrument for the evaluation of quadriceps strength is the isokinetic dynamometer (ID) [16,17], which is beyond the affordability of most clinicians (~R500k-R1.50mil). Although a more reasonable alternative to the ID exists, such as a handheld dynamometer (HHD) [17], the reliability of an HHD depends on the strength and aptitude of the practitioner [18,19]. The latter is especially true when assessing stronger individuals or muscle strength around larger joints such as the knee joint [17] which may have important clinical implications for both patients and practitioners. Furthermore, HHDs are still relatively expensive (~R20k) when considering starting a rehabilitation practice, especially in underresourced communities and clinical settings. Therefore, the purpose of this study was to: (i) establish the validity and reliability of a low-cost load cell (LC) (~R2k) for the assessment of isometric quadriceps muscle strength in people with PFP, (ii) evaluate the clinically meaningful difference (CMD) needed to enhance practitioner-based decision-making, and (iii) assess differences in isometric quadriceps strength with the use of a low-cost device between those with PFP and healthy control following an 8-week intervention.

Materials and Methods

Study design

57

58

59

60

61

62

63

64

65

66

67

68

69

70

71

72

73

74

75

76

77

This study used a repeated measures mixed study design consisting of a combination of between and within subject factors [20]. The between-subjects independent variable was group allocation (i.e., control group and PFP group), where group assignment was not

81 randomised (due to presence/absence of PFP). Within-subjects independent variables 82 included time (repeated measures) and extremity (injured [or non-dominant] and uninjured 83 [or dominant] limb). **Participants** 84 85 A total of 35 participants volunteered for the study, of which 17 were part of the control group (female [n=10] and male [n=7]) and 18 were part of the PFP group (female [n=15] and 86 87 male [n=3]. Given the requirements of the study design, a minimum total sample size of 24 88 was calculated based on a repeated measures analysis of variance (ANOVA) design that 89 incorporated (i) a within-between interaction, (ii) a moderate effect size (f = 0.25), (iii) a 90 type-I error rate of 5% ($\alpha = 0.05$), (iv) a type-II error rate of 20% ($\beta = 0.20$), (v) 2 groups 91 (control [n = 17] and PFP [n = 17]), (vi) 2 repeated measurements, (vii) an anticipated 92 dropout of 10%, and (viii) a minimum expected correlation of 0.50 among repeated 93 measurements [21]. Accounting for a potential drop-out of 20%, the minimum sample size 94 for adequate statistical power was 29 participants. Two participants were lost to follow-up, 95 both in the PFP group (female [n=1] and male [n=1]), resulting in a final sample size of 33 96 participants. 97 Participants were recruited by dispersing electronic flyers via social media platforms (e.g., 98 Facebook, WhatsApp), inviting prospective participants to contact the researcher for an 99 information letter which outlined the details of the study. Participants who were willing to 100 take part in the study were screened for eligibility (see inclusion and exclusion criteria below) 101 and received an informed consent form to read through and were given 72 hours to sign and 102 submit the form to an independent person affiliated with the study.

103

105

106

107

108

109

110

111

112

113

114

115

116

117

118

119

120

121

122

123

124

125

126

127

The inclusion criteria for the PFP group of this consisted of the following: (i) aged between 18 and 35 years, (ii) could be male or female, (iii) had to have retropatellar and/or peripatellar pain aggravated by at least one activity that loads the patellofemoral joint during weight bearing on a flexed knee, such as squatting, stair climbing, jogging/running, and hopping/jumping, and (iv) had to participate in some form of rehabilitation program. The control group had similar inclusion criteria, with the exception for points (iii) and (iv). The specific exclusion criteria for both the PFP and control group consisted of the following: participants should not (i) have had previous patellar dislocation or subluxation, (ii) have had previous injury or surgery to the knee, and (iii) have had recent (within the last 6 months) injury to the lower limbs (ex. Achilles tendinopathies, ankle sprain, etc.). All data collection occurred between 9th January 2023 to 22nd September 2023. Ethics approval was granted by the Human Research Ethics Committee of the university (NWU-00163-22-A1), and all participants completed the informed consent forms prior to participation. An independent researcher not affiliated with the study served as an independent witness and collected the signed consent forms. All ethical procedures conformed to the requirements of ethical conduct set forth in the Declaration of Helsinki. **Instruments** A general demographic questionnaire was used to obtain the participant's contact details, email address, age, sex, involved limb, dominant limb, previous injuries and PFP symptoms. The anterior knee pain scale (AKPS), also known as the Kujula patellofemoral scoring system, was completed electronically to capture knee-related pathologies where the total score out of 100 was captured. The AKPS was the chosen questionnaire, as it was developed specifically for evaluation of pain and disability in individuals with PFP [22], showed high test-retest reliability (intraclass correlation coefficient [ICC] = 0.95), and exhibited moderate responsiveness to clinical change, which implies that the score will likely reflect meaningful changes in a patient's condition over time [23]. Body mass and stature were measured to calculate body mass index (BMI) [24]. Body mass was measured with an electronic scale (Seca 874, Seca, Germany) to the nearest 0.1 kg and stature with a portable stadiometer (Holtain Ltd., U.K.) to the nearest 0.01 m. A cycle ergometer (Wattbike Pro, Wattbike Ltd, Nottingham, UK) was used for a ten-minute warm-up before the testing on the ID for optimal muscle performance and reduced risk of injury. The ID (Cybex II, CSMi, Stoughton, MA, USA) was used as the gold standard for evaluating isometric muscle strength and a LC (Crane & Hanging Scale, Micro Mini CS300, Border Scales & Labels) served as the low-cost alternative to measure isometric quadriceps strength. Strength was evaluated with the knee joint at 60° of knee flexion, where peak force of the quadriceps is usually generated [25]. A digital goniometer (EasyAngle, Meloq AB, Stockholm, Sweden) was used to measure the knee angle in all instances to ensure true validity and replicability.

Procedures

The demographic questionnaire was completed first, and only at the baseline testing. Thereafter the participants completed the AKPS questionnaire verbally, where the score was calculated out of 100 and recorded electronically. Stature and body mass were measured during barefoot standing and with minimal clothing. Lower leg length was measured from the lateral condyle of the femur to the lateral malleolus of the tibia to use as the lever length in the formula to convert force to torque for the LC measurements. Prior to testing participants completed a ten-minute warm-up on a cycle ergometer at lowest resistance and a comfortable speed (rating of perceived exertion [RPE] < 2 on the modified Borg scale). Participants completed baseline testing in two sessions separated by 24 hours, and then repeated the same testing eight weeks later. During the first session, isometric strength of the quadriceps muscles was measured first with the ID, and thereafter with the LC. During the second

154

155

156

157

158

159

160

161

162

163

164

165

166

167

168

169

170

171

172

173

174

175

176

177

session, only the testing on the LC was completed for the reliability analysis. Reliability analysis was done both at baseline testing and at follow-up testing conducted after 8 weeks to ensure a stronger reliability score. For the evaluation of isometric quadriceps strength on the ID, all variables related to the setup were recorded to replicate the exact position for follow-up testing. The dominant limb (defined as the preferred kicking limb) in the control group and the uninjured limb in the PFP group were measured first. The participants were placed in a seated position, stabilized with upper body straps and an upper leg strap just above the knee joint. The knee was positioned at a 60° angle, using the machine angle provided by the dynamometer. The lateral femoral epicondyle was aligned with the axis of rotation of the dynamometer, and the resistance pad was positioned anterior to the distal tibia just superior to the lateral and medial malleoli. A gravity correction was performed to account for any potential additional torque induced on the attachment. The participant completed a warm-up round consisting of three repetitions, with the instruction to do one repetition at approximately 25%, 50% and 75% of perceived maximum effort before commencing of the test to assist with familiarisation of the test. A 10 second rest period was allotted between trial repetitions. A one-minute rest period was granted after the warm-up, after which the participant completed repetitions of five second maximal extension contraction with ten seconds of rest between repetitions. Verbal encouragement was given throughout the whole procedure. Pain levels were monitored by asking the participant to rate pain on a scale of 1-10 throughout the procedure (where 1=pain free, 2= very mild, 3= discomforting, 4=tolerable, 5= very distressing, 6=intense, 7=very intense, 8=utterly horrible, 9=excruciating unbearable, 10=unimaginable unspeakable) [26], and participants were permitted to stop the test if pain levels were above bearable levels (≥ 4) . For the evaluation of isometric quadriceps torque using the LC, all positional measurements were recorded for replicability. Participants were seated on a standard chair, and the LC was attached to a strap that was attached to a fixed surface behind the chair. The strap was fastened to the participant's ankle just above the lateral and medial malleoli where the distance between the lateral malleolus and the start of the ankle strap was measured to more accurately calculate the lever arm length in the evaluation of torque from the LC. The straps were adjusted until the knee was at 60° of flexion, measured with a digital goniometer. The participant completed the same warm-up and testing protocol as with the ID. The tests were repeated in the second session for the reliability analysis, and all procedures were repeated after 8-weeks. The conversion of raw kilogram values from the LC were converted to torque values using the following equation:

Torque (N.m) = LC 'force' (kg)
$$\times$$
 9.81 (m.s⁻²) \times LA distance (m) [Equation 1]

Statistical analyses

 $CMD = TEM \times 1.96 \times \sqrt{2}$

All data were evaluated for normality using the Shapiro-Wilk test, with deviations from	
normality being accepted at p $<$ 0.05. All data are presented as mean \pm standard deviation	
(SD) unless otherwise stated. To determine the reliability of the LC, the intraclass correlation	
coefficient (ICC, two-way mixed effects, absolute agreement) between two measurements on	
separate occasions was used. The ICC values were interpreted as follows: poor: < 0.50;	
moderate: 0.50-0.75; good: 0.75-0.90; excellent: >0.90 [27]. Additional measures of	
reliability included: (i) typical error of measurement (TEM) (Equation 1), (ii) TEM%	
(Equation 3), and (iii) clinically meaningful difference (CMD) (Equation 4) [28]:	
$TEM = SD \sqrt{(1-ICC)}$	[Equation 2]
$TEM\% = TEM/mean \times 100$	[Equation 3]

[Equation 4]

The TEM% scores were qualitatively interpreted as: excellent:<5%; good: 5-10%; poor:

200

201

202

203

204

205

206

207

208

209

210

211

212

213

214

215

216

217

218

219

220

221

222

223

>10%. The CMD was used to evaluate the smallest detectable difference that would need to be overcome to conclude that a true change has occurred. A generalised linear model was used to evaluate mean differences between groups (2 levels: Con vs. PFP), and devices (2 levels: ID vs. LC) where each participant was treated as a random effect in the model. Posthoc analyses entailed the use of paired sample t-tests with a Holm correction to adjust for multiple comparisons. Standardised mean differences were calculated as Hedge's g, the magnitude of which was qualitatively interpreted as: trivial: <0.2; small: 0.2-0.6; moderate: 0.6-1.2; and large: > 1.2 [29]. The concurrent validity was determined by using linear regression where the Pearson correlation coefficient (r), coefficient of determination (r²), standard error of the estimate (SEE), and the slope of the regression lines were evaluated. The magnitude of the correlation coefficients were qualitatively interpreted as follows: negligible: 0.00-0.10; weak: 0.10-0.39; moderate: 0.40-0.69; strong: 0.70-0.89; and very strong: 0.90-1.00 [30]. Bland-Altman analyses were used to determine the systematic bias between the LC (reference measure) and the ID (criterion measure) [31]. For both the regression and Bland-Altman analyses, the point estimates were evaluated for potential outliers using Cook's distance where potential outliers (PO) were flagged when the Cook's distance exceeded a given threshold calculated as: 4/n (where n is the number of observations) [32]. All statistical analyses were completed using R [33]. **Results** The results pertaining to the relative (ICC_{3.1}) and absolute (TEM) reliability as well as CMD of the LC are highlighted in Figure 1. Generally, the LC shows excellent relative and absolute reliability both at baseline and following the 8-week training interval. The ICC point

225

226

227

228

229

230

231

232

233

234

235

236

237

238

239

240

241

242

243

244

245

246

estimates exhibit fairly narrow confidence intervals for the control group, but marginally longer intervals for the PFP group, indicating greater variability in torque measures. In all instances the CMD exceeded the TEM which implies that the LC had sufficient sensitivity to measure a true change in isometric quadricep strength (i.e. high signal-to-noise ratio). *** Figure 1 about here *** Simple interaction effects for both between-group and within-group differences in mean torque values are shown in Figure 2. Mean differences between devices and group are shown in Figure 2A, whereas the standardised effect size with 95% CI and corresponding uncertainty density distribution for the point estimates are shown in Figure 2B. *** Figure 2 about here *** For the concurrent validity, strong to very strong correlations were evident between the LC and ID torque measurements (r = 0.88-0.90, SEE = 0.05-0.07 Nm) (see Figure 3). Based on the slope analysis however, it is important to note that the torque values from the LC are typically lower than that obtained from the ID (slope = 0.65-0.77). *** Figure 3 about here *** The results from the Bland-Altman analysis confirm the measurement bias between the LC and the ID whereby an average systematic bias of 16.3-28.8 Nm is evident in favour of the ID (see Figure 4). *** Figure 4 about here ***

248

249

250

251

252

253

254

255

256

257

258

259

260

261

262

263

264

265266

267

268

269

270

Within-group differences in peak torque for each group and each device are shown in Figure 5. Although no statistically significant differences are evident, it is important to emphasise the inter-individual variability highlighted by the colour-gradients which indicate the magnitude of participant-specific torque improvements in quadriceps strength between time points. *** Figure 5 about here *** **Discussion** The present study yielded several novel findings. Firstly, the reliability and validity of a lowcost, commercially available LC were evaluated in comparison to a gold standard ID. The utility of such a device would have meaningful implications in rehabilitation services, especially in under-resourced practices. Secondly, we showed that the LC exhibited exceptional sensitivity for detection of clinically meaningful changes in isometric knee torque. Finally, we evaluated differences in isometric knee torque between a control and PFP group following an 8-week period showcasing substantial within-subject variability and constrained between-group differences. Reliability The LC showed excellent relative and absolute reliability in both the control and PFP groups (ICC>0.90 and TEM <10%) in all instances apart from the uninjured limb of the PFP group. The large ICC values obtained in the present study compare favourably with those from a separate study where belt-stabilisation was used with an HHD (ICC = 0.62 to 0.96) [34].

272

273

274

275

276

277

278

279

280

281

282

283

284

285

286

287

288

289

290 291

292

293

294

Similarly, the TEM% scores of the present study were substantially better those reported by Martins et al. [34] who recorded TEM% of 12% for a belt-stabilised HHD and Chamorro et al. [18] who yielded values ranging between 4-15% for measurements without stabilisation, both for knee extension strength. The ICC point estimates showed narrow confidence intervals in which the upper and lower limits fell within the margin defined as excellent in most instances. However, longer confidence intervals are observed in the PFP group than in the control group at the initial set of testing, indicating that the PFP group had greater variability in torque measures in the injured as well as the uninjured limbs compared to the control group. Torque variability for knee extensors is present in other knee injuries as well [35], and in this case the variability may be accounted for in the PFP group on the basis that PFP patients tend to have impaired quadriceps function often ascribed to impaired vastus medialis oblique firing [36,37] and reduced eccentric control [38]. However, the variability improved notably in the uninjured limb following the intervention period, possibly due to enhanced motor learning which tends to occur in injured populations [39]. It should be noted that the reliability of the LC, as with any device, is dependent on the set-up and therefor clinicians should be vigilant in following the set-up instructions set out in the procedures section and keep the set-up consistent when doing testing and re-testing in clinical settings, to ensure reliable results. Belt-stabilisation seems to be improving reliability in LCbased devices and are recommended to use by practitioners [25,34]. Clinically meaningful difference The CMD within the present study ranged between 12-45 Nm, implying that the differences in torque production by the quadriceps musculature between limbs or over the course of

rehabilitation can be detected with the LC. Previous research has shown that an HHD

exhibited comparable, although marginally lower CMD values of 17-27 Nm [40], with the ID yielding the lowest and most consistent CMD of approximately 27 Nm [41].

Group differences in torque production

295

296

297 298

299

300

301

302

303

304

305

306

307

308

309

310

311

312

313

314

315

316

317

318 319 An intriguing finding of the present study relates to the simple interaction effects regarding within-group (e.g., ID vs. LC) and between-group (e.g., Control vs. PFP) differences for quadriceps torque (see Figure 2A, 2B). The mean differences in torque production between devices and groups are classified as trivial-to-moderate, and typically range -2.28 Nm to 27.83 Nm (p = 0.113-0.999). In most instances the point estimates of the mean difference were not significantly different from zero largely due to the longer confidence intervals. Such a result generally implies, at least in principle, similarities in torque production between groups and between devices. The standardised effect sizes (together with the 95% CI) show that, despite a lack of statistical significance, the magnitude of the mean differences are likely to be meaningful, especially for the PFP group which consistently yielded higher torque values on the ID compared to the LC ($M_{diff} = 43.53$ Nm, p < 0.001). The interaction effects should be of interest to practitioners on the basis that the interchangeability of device measurements must be considered when interpreting results and making clinical inferences [41]. Given that, at least in some instances, practitioners might receive isokinetic results for a specific PFP patient or a referring clinician, these results should be verified when using the LC in the clinical setting so that more precise interpretations can be made when evaluating temporal changes in quadriceps strength. Moreover, the same device should be used by the same clinician to ensure adequate consistency in readings and insights of whether true changes have accrued.

Validity

The results of the preset study suggest that the concurrent validity of the LC, when compared to the ID, appears to be strong to very strong, although a few discrepancies are noteworthy. The torque values from the LC are typically lower than those from the ID, with the exception of torque values exceeding ~300 Nm (see Figure 3 and Figure 4). It is therefore reasonable to state that the LC might be of better use in injured and non-athletic populations that will most likely produce lower torque values compared to healthy, uninjured, or very athletic participants that might produce substantially higher torque values. The validity of the LC was greater than that of an HHD which yielded poorer scores even when stabilised with a belt, with correlations ranging between r = 0.3-0.8 [18,42]. The LC is not a perfectly valid tool given the bias in torque readings and moderate-to-large limits of agreement (LoA) and is therefore unlikely to replace the ID for absolute values. More specifically, there is evidence of a systematic bias in favour of the ID (see Figure 4), confirming that the LC typically measures lower torques than the ID on average. These results correspond with Martins et al. [34], who showed that knee extension tested with a belt-stabilised HHD, exhibited similar mean difference in torque production, with bias towards the ID. The HHD also shows, on average, wide LoA values (33.59%, CI_{95%} [23.91%, 43.26%]) for knee extension in other literature [18]. The wide LoA therefore support the notion that the LC is not valid enough to replace values obtained from the ID but given the excellent reliability would still be an exceptional tool for the evaluation of isometric quadriceps strength. It is also important to highlight that isokinetic norms should not be used to make clinical interpretations when comparing these to values derived from the LC. It would be important to develop independently generated normative data for LC-derived torque values to facilitate decisionmaking across different joints and population groups.

Temporal changes in torque

320

321

322

323

324

325

326

327

328

329

330

331

332

333

334

335

336

337

338

339

340

341

342

343 344 Finally, although the mean peak torque values did not change significantly from the baseline to follow-up testing ($M_{\rm diff}$ = 0.77-14.63 Nm, p = 0.531-0.969), it should be noted that there were substantial individual improvements both within- and between groups (see Figure 5). The variability in individual responses underscores the potential inadequacy of the intervention programmes followed by the groups which were beyond the control of the current study. Whether more focused and targeted interventions or prolonged rehabilitation timelines (> 8 weeks) would elicit more favourable outcomes would require further research, especially in those with PFP. It should however be noted that there were consistencies in the measured values between devices across time, implying that the LC indeed has exceptional utility as a measurement tool for evaluating isometric quadriceps strength.

Limitations

Given the strengths of the present study, it is also important to underscore some limitations. Firstly, participants could follow any rehabilitation program at any practitioner of their choice, and therefore could not control the details associated with targeted quadriceps strengthening as this was beyond the scope of the present study. Secondly, the ratios of males versus females in the groups differed, although consistency within groups were more important for this study than consistency between groups. Thirdly, participants were not divided into different groups according to the magnitude of quadriceps strength deficits. This latter point could provide more nuanced information on strength improvements over time in future studies.

Conclusions

The objectives of this study were to determine the validity and reliability of a low-cost LC for use in a clinical setting for measuring knee extension strength in those with PFP. The LC exhibited excellent reliability, and was deemed sensitive enough to detect clinically

meaningful differences over time in both healthy individuals as well as those with PFP. The reliability of the LC is dependent on the set-up of the individual, and therefore practitioners should take care to complete the set-up as described and ensure consistency across every testing session. The LC lacks validity and is therefore unlikely to be an adequate surrogate for isokinetic dynamometry. However, given the trade-off in (i) costing associated with the ID and the LC, (ii) the importance of evaluating and tracking changes in knee extension strength in those with PFP, and (iii) the excellent reliability of the LC, the utility of the LC as a viable assessment tool is advocated, especially in resource-restricted settings.

References

378

379 Crossley KM, Stefanik JJ, Selfe J, Collins NJ, Davis IS, Powers CM, et al. 2016 Patellofemoral pain 380 consensus statement from the 4th International Patellofemoral Pain Research Retreat, 381 Manchester. Part 1: Terminology, definitions, clinical examination, natural history, patellofemoral 382 osteoarthritis and patient-reported outcome measures. Br J Sports Med [Internet]. 383 2016;50(14):839-43. Available from: https://bjsm.bmj.com/content/50/14/831 384 Hall R, Barber Foss K, Hewett TE, Myer GD. Sport specialization's association with an increased 385 risk of developing anterior knee pain in adolescent female athletes. J Sport Rehabil. 386 2015;24(1):31-5. 387 Dey P, Callaghan M, Cook N, Sephton R, Sutton C, Hough E, et al. A questionnaire to identify 388 patellofemoral pain in the community: An exploration of measurement properties. BMC 389 Musculoskeletal Disorders. 2016;17(1):e237. 390 Roush JR, Bay RC. Prevalence of anterior knee pain in 18-35 year-old females. J Sports Phys Ther. 391 2012;7(4):396-401. 392 Kunene SH, Ramklass S, Taukobong NP. Anterior knee pain and its intrinsic risk factors among 393 runners in under-resourced communities in Ekurhuleni, Gauteng. S Afr J Physiother. 394 2018;74(1):e452. 395 Coburn SL, Barton CJ, Filbay SR, Hart HF, Rathleff MS, Crossley KM. Quality of life in individuals 396 with patellofemoral pain: A systematic review including meta-analysis. Physical Therapy in Sport 397 [Internet]. 2018;33:96-108. Available from: https://doi.org/10.1016/j.ptsp.2018.06.006 398 7. Collins NJ, Bierma-Zeinstra SMA, Crossley KM, Van Linschoten RL, Vicenzino B, Van Middelkoop 399 M. Prognostic factors for patellofemoral pain: a multicentre observational analysis. Br J Sports 400 Med [Internet]. 2013;47(4):227-33. Available from: 401 https://pubmed.ncbi.nlm.nih.gov/23242955/ 402 Reijnders L, Van de Groes SAW. The quality of life of patients with patellofemoral pain - a 403 systematic review. Acta Orthop Belg. 2020;86(4):678-87. 404 Tan SS, van Linschoten RL, van Middelkoop M, Koes BW, Bierma-Zeinstra SM, Koopmanschap 405 MA, et al. Cost-utility of exercise therapy in adolescents and young adults suffering from the 406 patellofemoral pain syndrome. Scand J Med Sci Sports. 2010;20(4):568-79. 407 10. Nimon G, Murray D, Sandow M, Goodfellow J. Natural History of Anterior Knee Pain: A 14- to 20-408 Year Follow-up of Nonoperative Management [Abstract]. Journal of Pediatric Orthopaedics 409 [Internet]. 1998 [cited 2023 Apr 17];18(1):118–22. Available from: 410 https://journals.lww.com/pedorthopaedics/Abstract/1998/01000/Natural History of Anterior 411 Knee Pain A 14 to.21.aspx 412 11. Thomas MJ, Wood L, Selfe J, Peat G. Anterior knee pain in younger adults as a precursor to 413 subsequent patellofemoral osteoarthritis: a systematic review. BMC Musculoskelet Disord 414 [Internet]. 2010 [cited 2023 Apr 17];11(1):e201. Available from: 415

https://www.ncbi.nlm.nih.gov/pmc/articles/PMC2944218/

- 416 12. Van Middelkoop M, Bennell KL, Callaghan MJ, Collins NJ, Conaghan PG, Crossley KM, et al.
- International patellofemoral osteoarthritis consortium: Consensus statement on the diagnosis,
- 418 burden, outcome measures, prognosis, risk factors and treatment. Semin Arthritis Rheum.
- 419 2018;47(5):666–75.
- 420 13. Boling MC, Padua DA, Marshall SW, Guskiewicz K, Pyne S, Beutler A. A Prospective Investigation
- 421 of Biomechanical Risk Factors for Patellofemoral Pain Syndrome The Joint Undertaking to
- 422 Monitor and Prevent ACL Injury (JUMP-ACL) Cohort. Am J Sports Med [Internet].
- 423 2009;37(11):2108–16. Available from: http://ajs.sagepub.com/supplemental/
- 424 14. Amin S, Baker K, Niu J, Clancy M, Goggins J, Guermazi A, et al. Quadriceps Strength and the Risk
- of Cartilage Loss and Symptom Progression in Knee Osteoarthritis. Arthritis Rheum [Internet].
- 426 2009 [cited 2023 Apr 17];60(1):189–98. Available from:
- 427 https://www.ncbi.nlm.nih.gov/pmc/articles/PMC3653642/
- 428 15. Willy RW, Hoglund LT, Barton CJ, Bolgla LA, Scalzitti DA, Logerstedt DS, et al. Patellofemoral pain.
- 429 J Orthop Sports Phys Ther. 2019;49(9):CPG1-95.
- 430 16. Janik F, Toulotte C, Seichepine AL, Masquelier B, Barbier F, Fabre C. Isometric Strength Database
- for Muscle Maximal Voluntary Endurance Field Tests: Normative Data. Sports Med [Internet].
- 432 2021;7(1):e47. Available from: /pmc/articles/PMC8273050/
- 433 17. Stark T, Walker B, Phillips JK, Fejer R, Beck R. Hand-held Dynamometry Correlation With the Gold
- 434 Standard Isokinetic Dynamometry: A Systematic Review. PM&R [Internet]. 2011;3(5):472–9.
- 435 Available from: https://onlinelibrary-wiley-
- 436 com.nwulib.nwu.ac.za/doi/full/10.1016/j.pmrj.2010.10.025
- 437 18. Chamorro C, Armijo-Olivo S, De La Fuente C, Fuentes J, Javier Chirosa L. Absolute reliability and
- 438 concurrent validity of hand held dynamometry and isokinetic dynamometry in the hip, knee and
- ankle joint: Systematic review and meta-analysis. Open Med [Internet]. 2017;12(1):359–75.
- 440 Available from: /pmc/articles/PMC5651404/
- 441 19. Deones VL, Wiley SC, Worrell T. Assessment of quadriceps muscle performance by a hand-held
- dynamometer and an isokinetic dynamometer. JOSPT [Internet]. 1994;20(6):296–301. Available
- from: https://pubmed.ncbi.nlm.nih.gov/7849749/
- 20. Christensen LB, Johnson RB, Turner LA. Research methods, design and analysis. 12th ed. Vol. 18.
- Harlow: Pearson Education Limited; 2015.
- 446 21. Faul F, Erdfelder E, Lang A, Buchner A. Statistics textbooks in the social. Behavior Research
- 447 Methods [Internet]. 2007;39(2):175–91. Available from:
- https://link.springer.com/content/pdf/10.3758/BF03193146.pdf
- 449 22. Howe TE, Dawson LJ, Syme G, Duncan L, Reid J. Systematic review Evaluation of outcome
- 450 measures for use in clinical practice for adults with musculoskeletal conditions of the knee: A
- 451 systematic review. Man Ther. 2012;17(2):100–18.
- 452 23. Watson CJ, Propps M, Ratner J, Zeigler DL, Horton P, Smith SS. Reliability and responsiveness of
- 453 the lower extremity functional scale and the anterior knee pain scale in patients with anterior
- knee pain. J Orthop Sports Phys Ther [Internet]. 2005;35(3):136–46. Available from:
- 455 https://pubmed.ncbi.nlm.nih.gov/15839307/

- 456 24. Hasan S. Quadriceps Femoris Strength Training: effect of Neuromuscular Electrical Stimulation Vs
- 457 Isometric Exercise in Osteoarthritis of Knee. Indian J Physiother Occup Therapy. 2015;9(3):129–
- 458 34.
- 459 25. McNamara S, Hevia EG, St. Louis R, Cho W, Lee S, Moyne M, et al. Isometric quadriceps strength
- 460 test device to improve the reliability of handheld dynamometry in patient with anterior cruciate
- ligament injury. In Minneapolis, MN, USA: ResearchGate; 2018. Available from:
- 462 http://asmedigitalcollection.asme.org/BIOMED/proceedings-
- 463 pdf/DMD2018/40789/V001T09A001/2788227/v001t09a001-dmd2018-6848.pdf
- 464 26. Aby F. Serotonergic neurons of the nucleus raphe Magnus in the control of nociceptive
- 465 transmission in the dorsal horn of the spinal cord: an optogenetic study in different
- pathophysiological contexts. [Nouvelle-Aquitaine, France]: University of Bordeaux; 2019.
- 467 27. Koo TK, Li MY. A Guideline of Selecting and Reporting Intraclass Correlation Coefficients for
- 468 Reliability Research. J Chiropr Med [Internet]. 2016;15(2):155–155. Available from:
- 469 /pmc/articles/PMC4913118/
- 470 28. Schwenk M, Gogulla S, Englert S, Czempik A, Hauer K. Test-retest reliability and minimal
- 471 detectable change of repeated sit-to-stand analysis using one body fixed sensor in geriatric
- 472 patients. Physiol Meas [Internet]. 2012 Oct [cited 2024 Jan 8];33(11):1931. Available from:
- 473 https://dx.doi.org/10.1088/0967-3334/33/11/1931
- 474 29. Hopkins WG, Marshall SW, Batterham AM, Hanin J. Progressive statistics for studies in sports
- 475 medicine and exercise science. Med Sci Sports Exerc. 2009 Jan;41(1):3–13.
- 476 30. Schober P, Boer C, Schwarte LA. Correlation Coefficients: Appropriate Use and Interpretation.
- 477 Anesthesia & Analgesia [Internet]. 2018 May [cited 2024 Jan 8];126(5):1763. Available from:
- 478 https://journals.lww.com/anesthesia-
- 479 analgesia/fulltext/2018/05000/correlation coefficients appropriate use and 50.aspx
- 480 31. Giavarina D. Understanding Bland Altman analysis. Biochem Med (Zagreb) [Internet]. 2015 Jun 5
- 481 [cited 2023 Dec 19];25(2):141–51. Available from:
- https://www.ncbi.nlm.nih.gov/pmc/articles/PMC4470095/
- 483 32. Hebbali A. rsquaredacademy/olsrr [Internet]. Rsquared Academy; 2023 [cited 2024 Jan 17].
- 484 Available from: https://github.com/rsquaredacademy/olsrr
- 485 33. R Core Team. R: The R Project for Statistical Computing [Internet]. Vienna, Austria: R Foundation
- for Statistical Computing; [cited 2024 Jan 8]. Available from: https://www.r-project.org/
- 487 34. Martins J, Rodrigues Da Silva J, Rodrigues Barbosa Da Silva; D´ Ebora Bevilaqua-Grossi M.
- 488 Reliability and Validity of the Belt-Stabilized Handheld Dynamometer in Hip-and Knee-Strength
- Tests. J Athl Train [Internet]. 2017;52(9):809–19. Available from: www.natajournals.org
- 490 35. Goetschius J, Hart JM. Knee-Extension Torque Variability and Subjective Knee Function in
- 491 Patients with a History of Anterior Cruciate Ligament Reconstruction. J Athl Train. 2016
- 492 Jan;51(1):22-7.
- 493 36. Van Tiggelen D, Cowan S, Coorevits P, Duvigneaud N, Witvrouw E. Delayed vastus medialis
- obliquus to vastus lateralis onset timing contributes to the development of patellofemoral pain
- in previously healthy men: a prospective study [Abstract]. Am J Sports Med. 2009;37(6):1099–

496 105.

498

499

500

501

502

503

504

505

506

507

508

509

510

511

512

513

514

515

516

517

518

519

520

521


37. Pal S, Draper CE, Fredericson M, Gold GE, Delp SL, Beaupre GS, et al. Patellar maltracking correlates with vastus medialis activation delay in patellofemoral pain patients. Am J Sports Med. 2011;39(3):590-8. 38. Eapen C, Nayak CD, Pazhyaottyil Zulfeeguer C. Effect of Eccentric Isotonic Quadriceps Muscle Exercises on Patellofemoral Pain Syndrome: An Exploratory Pilot Study. Asian J Sports Med [Internet]. 2011 [cited 2023 Jun 3];2(4):227–34. Available from: https://www.ncbi.nlm.nih.gov/pmc/articles/PMC3289225/ 39. Rodrigues-da-Silva JM, de Rezende MU, Spada TC, da Silva Francisco L, Greve JMD, Ciolac EG. Effects of Motor Learning on Clinical Isokinetic Test Performance in Knee Osteoarthritis Patients. Clinics (Sao Paulo) [Internet]. 2017 [cited 2024 Jan 12];72(4):202-6. Available from: https://www.ncbi.nlm.nih.gov/pmc/articles/PMC5401619/ 40. Knols RH, Aufdemkampe G, de Bruin ED, Uebelhart D, Aaronson NK. Hand-held dynamometry in patients with haematological malignancies: Measurement error in the clinical assessment of knee extension strength. BMC Musculoskelet Disord [Internet]. 2009 Mar 9 [cited 2024 Jan 12];10:31. Available from: https://www.ncbi.nlm.nih.gov/pmc/articles/PMC2662793/ 41. de Araujo Ribeiro Alvares JB, Rodrigues R, de Azevedo Franke R, da Silva BGC, Pinto RS, Vaz MA, et al. Inter-machine reliability of the Biodex and Cybex isokinetic dynamometers for knee flexor/extensor isometric, concentric and eccentric tests. Physical Therapy in Sport [Internet]. 2015 Feb 1 [cited 2024 Jan 17];16(1):59-65. Available from: https://www.sciencedirect.com/science/article/pii/S1466853X14000339 42. Hirano M, Katoh M, Gomi M, Arai S. Validity and reliability of isometric knee extension muscle strength measurements using a belt-stabilized hand-held dynamometer: a comparison with the measurement using an isokinetic dynamometer in a sitting posture. Journal of Physical Therapy Science. 2020;32(2):120-4.


Figure Captions

522

553

Figure 1: Between-group and within-group intra-class correlation coefficients. The ICC 523 values are shown with their 95% confidence intervals. Above each point estimate are shown 524 525 the precise ICC values, typical error of measurement (TEM%) and clinically meaningful difference (CMD) to provide greater context for the reliability score. Vertical dotted lines 526 527 represent thresholds for excellent (ICC > 0.90) and good (ICC: 0.70-0.90). Note: ND = non-528 dominant; Dom = dominant; Con = control group; PFP = patellofemoral pain group. 529 530 Figure 2: Absolute and relative within-group and between-group mean differences. Panel A: Mean differences with 95% CI; panel B: standardised mean effect size (Hedge's g) with 90% 531 532 CI (thick black line) and 95% CI (thin black line) as well as density estimates to highlight the uncertainty in the point estimate. Note: M_{diff} = mean difference; g = Hedge's g effect size; 533 534 Con = control group; PFP = patellofemoral pain group. 535 536 Figure 3: Linear regression between load-cell and Cybex for isometric knee extension torque. 537 PO = potential outlier; SEE = standard error of the estimate; ND = non-dominant; Dom = 538 dominant 539 540 Figure 4: Bland-Altman plots for the bias (95% CI) between instruments and the limits of agreement. Note: UL = upper limit; LL = lower limit; ND = non-dominant; Dom = dominant; 541 PO = potential outlier 542 543 544 Figure 5: Within-group differences in peak torque. Pre-post differences are shown for the 545 control group as evaluated by the LC (panel A) and ID (panel B). Pre-post differences are 546 shown for the PFP group as evaluated by the LC (panel C) and ID (panel D). Line segments 547 are coloured based on the percentage difference (Post-Pre/Pre×100) where improvements are 548 shifted towards orange, and decrements are shifted towards green. The magnitude of the 549 difference is captured by the intensity of the colour gradient (larger differences are darker, 550 smaller differences are lighter). Black lines indicate the group mean for a given time point. 551 Note: Mdiff = mean difference; g_{Hedges} = Hedge's g; Con = control group; PFP = patellofemoral pain group; LC = Load Cell 552





Α

