# 1 Enhancing Interventional Cardiology Training: A Porcine Heart-Based Coronary Intervention

- 2 **Simulator** Joerg Reifart, MD<sup>1</sup>; Paul Anthony Iaizzo PhD<sup>1</sup>.
- <sup>3</sup> <sup>1</sup> Visible Heart Laboratories, Department of Surgery, University of Minnesota, Minneapolis,
- 4 Minnesota, USA

# 5 Abbreviated Title: LOW-COST CORONARY INTERVENTION TRAINER

- 6 Data availability statement: Not applicable
- 7 Funding: None
- 8 Conflicts of interest: The authors have no potential conflicts of interest in regards to this
- 9 publication
- 10 Ethics approval statement: Not applicable
- 11 **Patient consent statement:** Not applicable
- 12 **Permission to reproduce material from other sources:** Not applicable
- 13 Clinical trial registration: Not applicable
- 14
- 15 Word count: 1866
- 16
- 17 Address for Correspondence:
- 18 Joerg Reifart MD, PhD
- 19 Visible Heart Laboratories, University of Minnesota
- 20 B172 Mayo, MMC195 | 420 Delaware Street S.E..
- 21 Minneapolis, MN 55455
- 22 Email: reifartpublications@gmail.com
- 23 Twitter/X handle: @opnaccesscardio

### 25 **ABSTRACT**:

#### 26 Introduction:

27 Access to simulators for interventional cardiology is currently limited. High acuity, low occurrence

28 procedures (HALO), such as coronary perforation or iatrogenic dissection, are not trained in currently

available simulators. We developed a cost-effective coronary intervention simulator designed to

30 enhance the training of both novice and experienced interventionalists.

### 31 Methods:

32 Porcine hearts from 6-month-old Yorkshire pigs (heart weight: 300-500g) were obtained from a large 33 animal research laboratory. Guide catheters were inserted into the coronary artery ostia and secured 34 with superglue. To maintain shape and rigidity, commercially available polyurethane insulation foam 35 was injected into the ventricles. The guide catheter was then connected to a Tuohy valve linked to a 3-36 way stopcock. One connection led to a pressure infusion bag filled with tap water (inflated at > 300 mmHg); the other was used for contrast injection. The heart was set on a radiolucent box with a fluid 37 38 collector underneath. Clinical scenarios were simulated using 3D-printed stenoses and occlusions, blood 39 clots for occlusive myocardial infarction, balloon oversizing for dissections, and needle trauma for 40 perforations.

#### 41 Results:

42 The simulator was used to practice coronary angiography, managing perforations, bifurcations,

43 dissections, and acute coronary occlusions.

Assembly, set-up, and simulation time until refilling the perfusion bag was required were 50, 25, and 30
 minutes, respectively. Intravascular imaging with Optical Coherence Tomography was successfully used

- to guide interventions. The simulator was frozen and reused more than three times without notable
- 47 deterioration.
- 48 **Conclusion**:
- 49 A wide range of clinical scenarios can be trained with our model. Its preparatory flexibility, including the
- ability to be frozen for on-demand training, enhances its utility. Limitations include the absence of
- 51 pulsatile flow, heart movement, and the inability to train guide catheter intubation.
- 52

# 54 INTRODUCTION:

55	Several studies showed the benefits of simulation of interventional or angiographic training with virtua
56	reality simulators to have new interventional fellows gain their first experience with coronary
57	procedures[1-3]. Hands-on alternatives are physical computer-based simulator systems, which have
58	learners more practically experience angiography and coronary interventions[4-7]. However, because
59	these systems are cost-prohibitive (only available online quote: \$55.380), this type of training is not
60	currently widely adopted[8].
61	Most interventional trainees take their first interventional steps, learning angiography and simple
62	procedures in patients under close supervision by their respective mentors. This may increase patient
63	risk, but it is also associated with high psychophysiologic stress experienced by the trainees[9].
64	Here, we present a low-cost system using porcine hearts that could be conjured from large animal
65	research facilities or possibly a slaughterhouse to create a training system. It allows novice
66	interventionalists to learn both what angulations to use in coronary angiography as well as coronary
67	interventions, including complication management training.
68	

### 70 METHODS:

#### 71 **HEART PREPARATION**:

Porcine hearts (fresh or frozen) from 6-month-old Yorkshire pigs (weight animal 60 – 80kg, weight of the
 hearts 300 -500g) that had been used for differing research projects in the Visible Heart Laboratories
 were used in this model.

75 Depending on the extent of dissection, guide catheters (6-8F JL4.0, EBU 3.75, or JR 4.0—previously used)

76 were placed into both coronary ostia (Figure 1, (C)). The aortic root can be trimmed in case of excess

aortic tissue or difficulty canulating the ostia. To prevent the catheters from exiting the ostia, they were

fixed in position with super glue.

79 We used commercially available polyurethane insulation foam (6.99\$) to fill the ventricles, atria, the

80 pulmonary artery, and the aorta. This is to prevent fluid from entering the ventricles, fix the guide

81 catheters in place, and to lend support to the organ, resulting in realistic coronary projections (Figure 1

82 (D)). Excess insulation foam was cut. Foam in hard-to-reach places or within vessels can be dissolved

using commercially available acetone. The guide catheter was then connected to a Tuohy valve and a

84 manifold, of which one port was connected to a pressure infusion bag filled with tap water (Figure 2;

85 Supplemental Digital Content 1 - Figure for fast refilling method). System assembly duration was timed

in minutes. The total cost for the used materials was \$42.56 (Table 1).

87

### 88 SYSTEM SET-UP:

The heart was placed on a support structure for which either a clear 28 Qt | 27l Storage Box with a cut-off 2 liter PET bottle was fixed to the bottom with epoxy (Figure 2) or alternatively two Polyethylene terephthalate (PETE) packages from Affinity Fusion Oxygenators (Supplemental Digital Content 2 -suppl

- 92 Figure). The support structure collects the fluid and contrast that circulates through the heart and
- 93 positions it at a height and angle similar to in-human coronary angiograms.
- 94 The pressure infusion system was inflated to pressures ~300mmHg, which allows for continuous
- 95 antegrade contrast clearance when valves are opened (standard). If feasible in the local catheterization
- 96 laboratory environment, a hose-based continuous perfusion set-up can be chosen to allow for
- 97 uninterrupted training (Supplemental Digital Content 3 Suppl Figure 3). We acquired residual unused
- 98 Iohexol from Catheterisation Laboratories and Radiology departments for intravenous contrast.
- 99 Alternatively, 10% Lugol's solution (lodine potassium iodide) can be used as an inexpensive alternative
- 100 contrast medium in this model. Lugol's solution, however, will stain the materials used.
- 101 System set-up duration was timed in minutes.
- 102

### 103 TRAINING SCENARIOS:

104 Different standard training scenarios apart from standard angiography angulations, balloon/stent

exchange practice, and handling of devices such as Optical Coherence Tomography (OCT) catheters for

106 intravascular imaging were simulated:

- Circular sutures around the artery produced Simple Type A lesions.
- Thrombotic acute occlusions were simulated by injecting previously collected coagulated
- porcine blood directly before guide catheter placement or through the guide catheters
  themselves.
- Complex lesions like chronic totally occluded coronary artery lesions or severely calcified lesions
   were modeled from diseased human coronary arteries and then 3-D printed. In short: Formalin fixed human hearts underwent iodine-enhanced micro-computed tomography scanning (North

114	Star Imaging X3000 Micro-CT scanner, Rogers, Minnesota) ) and were subsequently
115	segmented (Figure 3, MIMICS 26.0, Materialise, Plymouth, MI, USA), Figure 3, A). The
116	segmented coronary lesions were then 3-D printed at actual scale (Figure 3, B) and inserted into
117	the coronary arteries before fixing the guide catheters to the coronary ostia.
118	• Short, proximal, calcified stenoses were created by injection of Super-Glue mixed with Barium
119	sulfate powder into the artery before inserting the guide catheter.
120	• Dissections were created by inflating grossly oversized balloons (e.g., 4.5mm Balloon in a 2.5mm
121	diameter vessel).
122	• Simple lacerations with a needle or scalpel were used to simulate coronary perforations.
123	Duration, until the perfusing pressurized saline bag had to be refilled, was timed in minutes. Further
124	recommendations are in the supplementary Appendix.
125	
126	

# 128 **RESULTS**:

129	Initial assembly, including the polyurethane curing time of the coronary intervention trainer, was 50
130	minutes. Set-up in the catheterization laboratory was 25 minutes. Continuous antegrade perfusion on 11
131	pressure perfusion allowed for 30 minutes of uninterrupted simulation. High-pressure infusion through
132	the guiding catheter via the connected pressured fluid bag resulted in appropriate, non-collapsed vessel
133	diameters during interventions, with antegrade contrast clearance (see Videos, Supplemental Digital
134	Content 4 & 5). Optical coherence tomography (OCT) worked to acquire intravascular images without
135	the need for additional contrast injection.
136	Previously described scenarios were successfully simulated:
137	- Figure 4 : Angiographic and OCT image of the created acute thrombotic occlusion within the
138	simulator.
139	- Figure 5 & Video, Supplemental Digital Content 6: Coronary perforation on OCT and angiography
140	- Video, Supplemental Digital Content 7: Angiography of the 3D-printed CTO.
141	Other trainable coronary interventions and maneuvers are presented in Table 2.
142	The assembled hearts with guide catheters in place were frozen and used again more than three times,
143	with the main limitation to re-freezing and re-using being the accidental dislodgement of guiding
144	catheters.

# 146 **DISCUSSION**:

147	The coronary angiography and intervention trainer presented here could be valuable for novice and
148	advanced interventional cardiologists. As virtual reality-based angiography simulation training has
149	shown positive results, a physical system like the one we presented is likely to improve skills
150	surrounding coronary angiography angulations and interpretation[2]. Technical skills that other non-
151	digital PCI-focused trainers focused on can also be trained on our model[10,11]. Current non-digital
152	trainers tend to be more reductive and focus on single tasks such as balloon exchange or two-
153	dimensional wiring and interventions[10,11]. Currently available models do not meet the need to train
154	complex situations such as complications and other high acuity low occurrence (HALO) procedures.
155	Accordingly, interventional cardiologists may feel inappropriately prepared to deal with these events.
156	Simsiek et al. assessed the educational experience of interventional cardiologists in the US and Canada
157	and found that even after dedicated training, less than half of the fellows felt comfortable performing
158	various atherectomy techniques, implanting covered stents, fat embolization, or using embolic
159	protection devices[12].
160	Interventions on porcine hearts for research, training, and educational purposes, as demonstrated in the
161	Visible Heart Apparatus, are established as a valuable tool[13-16]Unfortunately, these methods are not
162	easily accessible to trainees.
163	We provided instructions to make these procedures more accessible. Compared to a computer-based
164	PCl simulator, which comes at a cost of ~\$55.000, we were able to compile a system for less than 50\$. In
165	contrast to the digitally simulated computer-based trainers, the model presented here allows any device
166	to be used or tested. Properly cleaned used wires, balloons, catheters, and microcatheters can be

167 trained and used for training to gain experience with these devices, as well as HALO procedures.

168 Furthermore, this model enables medical device prototype testing and the development or modification

169 of interventional techniques. Eventually, data sets from CT scans of patients with difficult coronary

170 lesions could be used to create 3D-printed lesions to incorporate into this model to have trainees

171 perform a previous case completed by a different operator or to train robotic PCI devices.

#### 172 **LIMITATIONS**:

- 173 While the simulator described here is reasonably realistic, some aspects lead to reduced fidelity of the
- system. As the guiding catheters are already in place, this system is not valuable for teaching catheter
- intubation or the difficulties in achieving and maintaining a good guide catheter position. There are no
- 176 ECG or pressure curves on display in this setup. Learning to recognize ischemia on ECG or dampened and
- 177 ventricularized pressure waveforms are important aspects of coronary intervention[17]. As the system is
- 178 not pulsatile, there is no heart movement. Therefore, especially RCA interventions, which can be more
- 179 difficult because of equipment movement, are not as realistic[18].
- 180 Porcine vessels are more compliant than diseased human coronary arteries and without coronary artery
- 181 disease. Therefore, the haptic feedback of balloon inflation and wire manipulation will not be the same.
- 182 Even though lesion characteristics can feasibly be modeled using 3-D printing, there is likely to be a
- difference. Similarly, the propensity to dissect or perforate differs between juvenile porcine arteries and
- 184 diseased human coronary arteries.

Acquiring an appropriate heart specimen may pose difficulties. Ideally, large animal hearts from research facilities are used. Most commercially available hearts for human consumption are dissected and cannot be used. Hearts used for school dissection projects can be formalin-fixed and may show more tissue rigidity than is ideal for this setting.

### 190 **CONCLUSION**:

- 191 We report on how to build and use a low-cost simulator for coronary angiography and percutaneous
- 192 coronary artery interventions. This model's reusability and versatility make it a valuable tool in training
- 193 interventional cardiologists. It could feasibly be used in any catheterization laboratory worldwide
- 194 without requiring specific equipment for assembly and set-up. Dedicated training units with more
- 195 specific preparation instructions and scenarios incorporating digital display simulation of arterial
- 196 pressures and ECG are promising next steps to increase the simulator's fidelity.

197

198

# 200 ACKNOWLEDGMENTS:

- 201 We thank the Visible Heart Laboratory staff and volunteers for their continuous engagement, which
- 202 makes this such an inspiring place to work. We also thank Erik Donley for his work on the support
- 203 structure.
- 204

# 205 **FINANCIAL DISCLOSURE SUMMARY:**

- 206 P.A.I. has a research contract and serves as an education consultant arrangement with Medtronic
- 207 (Minneapolis, MN, USA). These arrangements did not influence any part of the work presented here.

#### 208

### 209 **REFERENCES**

210 1. Voelker W, Petri N, Tönissen C, Störk S, Birkemeyer R, Kaiser E, et al. Does Simulation-Based 211 Training Improve Procedural Skills of Beginners in Interventional Cardiology?—A Stratified 212 Randomized Study. Journal of Interventional Cardiology 2016;29:75-82. 213 2. Popovic B, Pinelli S, Albuisson E, Metzdorf PA, Mourer B, Tran N, et al. The Simulation Training in 214 Coronary Angiography and Its Impact on Real Life Conduct in the Catheterization Laboratory. 215 American Journal of Cardiology 2019;123:1208-13. 216 3. Jensen UJ, Jensen J, Ahlberg G, Tornvall P. Virtual reality training in coronary angiography and its 217 transfer effect to real-life catheterisation lab. EuroIntervention 2016;11:1503-10. 218 4. Lee KS, Natarajan B, Wong WX, Yousman W, Koester S, Nyotowidjojo I, et al. A randomized 219 controlled trial of simulation training in teaching coronary angiographic views. BMC Med Educ 220 2022;22:644. 221 5. Schuetz M, Moenk S, Vollmer J, Kurz S, Mollnau H, Post F, et al. High Degree of Realism in 222 Teaching Percutaneous Coronary Interventions by Combining a Virtual Reality Trainer With a Full 223 Scale Patient Simulator. Simulation in Healthcare 2008;3:242-6. 224 Jensen UJ, Jensen J, Olivecrona GK, Ahlberg G, Tornvall P. Technical Skills Assessment in a 6. 225 Coronary Angiography Simulator for Construct Validation. Simulation in Healthcare 2013;8:324-226 8. 227 7. Räder SBEW, Abildgaard U, Jørgensen E, Bech B, Lönn L, Ringsted CV. Association Between 228 Endovascular Performance in a Simulated Setting and in the Catheterization Laboratory. 229 Simulation in Healthcare 2014;9:241-8. 230 8. Healthcare C. CAECathLabVR. GT Simulators Website: 231 https://www.gtsimulators.com/products/cathlabvr-simulator-base-unit Accessed April 29, 2024 232 2018. 233 9. Detling N, Smith A, Nishimura R, Keller S, Martinez M, Young W, et al. Psychophysiologic 234 responses of invasive cardiologists in an academic catheterization laboratory. American Heart 235 Journal 2006;151:522-8. 236 Petri N, Weber C, Maniuc O, Oder D, Lengenfelder B, Voelker W. A Skill Trainer for Balloon/Stent 10. 237 Exchange in Percutaneous Coronary Intervention—Concept and Evaluation. Simulation in 238 Healthcare 2022;17:249-55. 239 11. Nanto S. PCI TRAINER for BEGINNERS. Medialpha Web site: 240 https://medialpha.jp/book\_category/en Accessed May 2 2024 2011. 241 12. Simsek B, Kostantinis S, Karacsonyi J, Hakeem A, Prasad A, Prasad A, et al. Educational 242 Experience of Interventional Cardiology Fellows in the United States and Canada. JACC 243 Cardiovasc Interv 2023;16:247-57. 244 13. Valenzuela TF, Burzotta F, Iles TL, Lassen JF, laizzo PA. Assessment of single and double coronary 245 bifurcation stenting techniques using multimodal imaging and 3D modeling in reanimated swine 246 hearts using Visible Heart<sup>®</sup> methodologies. The International Journal of Cardiovascular Imaging 247 2021;37:2591-601. 248 14. Burzotta F, Cook B, laizzo PA, Singh J, Louvard Y, Latib A. Coronary bifurcations as you have 249 never seen them: the Visible Heart® Laboratory bifurcation programme. EuroIntervention 250 2015;11:V40-V3.

15. Cangemi S, Burzotta F, Bianchini F, DeVos A, Valenzuela T, Trani C, et al. Configuration of two stent coronary bifurcation techniques in explanted beating hearts: the MOBBEM study.
 EuroIntervention 2023;19:e423-e31.

- Bianchini F, Cangemi S, DeVos A, Bielecki M, Aurigemma C, Romagnoli E, et al. Multimodal
   Comparisons of Results Achieved by Different Side Branch Ballooning Techniques for Bifurcation
   Provisional Stenting. Circ Cardiovasc Interv 2023;16:e012908.
- 25717.Klein LW, Korpu D. Damped and Ventricularized Coronary Pressure Waveforms. J Invasive258Cardiol 2017;29:387-9.
- 18. Chen GC, Lu XM, Song YM, Gyawail L, Geng ZH, Song MB, et al. A 3-year experience of a simple,
   novel technique for accurate ostial/non-ostial coronary stenting: The buddy balloon anchor
   stent technique. Catheter Cardiovasc Interv 2018;92:1147-52.
- 262
- 263

# 264 Figures:

- Figure 1: (A) Porcine heart with aorta attached. (B) Heart after cutting the Aorta so the coronary ostia
- 266 can be accessed easily. (C) Guide catheters placed into the right and left coronary ostium and fixed in
- 267 place with Super-glue. (D) Injection of polyurethane foam into the left and right ventricle, leading to
- 268 expansion of all cavities and great vessels.

269

- 270 Figure 2: Coronary Intervention Simulator final setup. Fluoroscopy C-Arm on the left (blue box). Prepped
- 271 heart with guide catheters in place below the detector on the support structure putting it at the
- isocenter of the fluoroscopy machine and allowing to follow all clinically used coronary projections
- 273 (yellow box). Pressure infusion set on the right, as well as contrast solution following a standard
- 274 catheterization laboratory setup (red box).

276 Figure 3: (A) Digital 3-D model of a coronary chronically occluded artery segmented from a
--

- 277 resolution micro-computed tomography scan. (B) 3D print of the artery segment before introduction
- into the intervention simulator. Penny for scale.

279

Figure 4: OCT showing occlusive (A) and non-occlusive (B) white thrombus in the RCA that was saved

from a previous porcine case and then injected via the guide catheter. To train thrombus aspiration.

282

283 Figure 5: OCT of Ellis III coronary artery perforation before treatment.

284

#### 285 Supplemental Digital Content

Supplemental Digital Content 1: Supplemental Figure of our set up to quickly refill the infusion bag with tap water. A hose adapter was attached to a faucet. Then the drip counter of a IV Administration Set was cut in the lower third and then connected to the hose with a ½ inch connector.

289

290 Supplemental Digital Content 2: Supplemental Figure of how two Polyethylene terephthalate (PETE)

291 packages from Affinity Fusion Oxygenators were altered to be used as the support and fluid collection

- structure for the trainer. One package edge was cut off to fit into the other upside down.
- 293 Supplemental Digital Content 3: Figure of a tube, 3-way stopcock and connector setup that allows the
- 294 PCI trainer to be directly connected to a faucet-hose system for continuous perfusion.
- 295 Supplemental Video Content 1: wmv Video of an anterior-posterior projection of an angiogram of the
- 296 Left anterior descending artery (LAD) showing contrast being cleared from the vessel by the pressurized

fluid bag. Note small air bubble embolization.

299	Supplemental Video Content 2: avi Video of an OCT recording from within the LAD at the level of the
300	second diagonal branch showing the initiation perfusion via the guiding catheter with the notable
301	expansion of the vessel lumen as the artery is pressurized.
302	
303	Supplemental Video Content 3: wmv Video of Angiogram of a needle induced coronary artery
304	perforation Ellis III before treatment.
305	
306	Supplemental Video Content 4: wmv Video of the 3-D printed chronic total occlusion in the right
307	coronary artery with concurrent contralateral injection to show potential septal collaterals.

Component	Reusable	Cost per Unit	Source
Heart	Yes	\$10.50	Research lab, butcher
Tap water 1l	N/A	N/A	
Polyurethane foam	Yes	\$6.99	Hardware store
Super glue	Yes	\$3.64	Hardware store
Box	Yes	\$5.98	Hardware store
Empty 2l bottle	Yes	N/A	Recycle
Guide catheter & Pressure bag	No	N/A	Cath Lab
Iohexol/ Lugol's solution 500ml	No	\$25.95	Cath Lab/Chemical
			supplier
Total cost		42.56	

Table 1: Cost of Materials Used for single reusable PCI Trainer

Table 2: Trainable and non-trainable clinical situations

Trainable	Non-trainable
Wiring of tortuous vessels and side-branches	Ostial Interventions
Wiring septal/other collaterals for retrograde approach Precise placement of stents in bifurcated lesions	Bypass Graft interventions Fractional Flow Reserve (FFR)
Transcoronary ablation of septal hypertrophy (TASH) Embolization to occlude perforated, small vessels	Electrophysiologic Interventions Structural Interventions
Placement of covered stents to seal perforations Use of Atherectomy devices (rotational and orbital) Use of OCT and IVUS to guide PCI* Thrombus aspiration	







Exity preprint doi: https://doi.org/10.1101/2024.09.02.24312944; this version posted September 4, 2024. The copyright holder for this preprint (which was not certified by peer review) is the author/funder, who has granted medRxiv a license to display the preprint in perpetuity.



