

Diffuse optical tomography system for acute traumatic brain injury in the intensive care unit: a prospective study on healthy volunteers

Mario Forcione,^{a, b,*} Antonio Maria Chiarelli,^c David Perpetuini,^c Guy A. Perkins,^d Andrew R. Stevens,^{a, e} David J. Davies^{a, e} Antonio Belli^{a, e}

^a University of Birmingham, Neuroscience & Ophthalmology Research Group, Institute of Inflammation & Ageing, College of Medical and Dental Sciences, Edgbaston, Birmingham, UK, B15 2TT

^b University of Milan-Bicocca, School Medicine and Surgery, Piazza dell'Ateneo Nuovo, Milan, Italy, 20126

^c University 'G. D'Annunzio' of Chieti-Pescara, Institute for Advanced Biomedical Technologies (I.T.A.B.) and Department of Neuroscience, Imaging, and Clinical Sciences, Via Luigi Polacchi 13, Chieti, Italy, 66100

^d University of Birmingham, School of Computer Science, College of Engineering and Physical Sciences, Edgbaston, Birmingham, UK, B15 2TT

^e University Hospitals Birmingham NHS Foundation Trust, Neurosurgery Department, Mindelsohn Way, Birmingham, UK, B15 2GW

Abstract

Significance: Current systems for diffuse optical tomography (DOT) are unsuitable for clinical applications on acute traumatic brain injury (TBI) patients while in the intensive care unit (ICU).

Aim: To develop and test a method for DOT recordings suitable for TBI patients in the ICU. This method is based on measurements and co-registration using 3-D optical scans, and the acquisition of optical data using a custom-made helmet which would enable a multimodal (invasive and non-invasive) neuromonitoring.

Approach: Probe displacements compared to electromagnetic digitization co-registrations were assessed. The capacity to isolate and monitor, using functional near-infrared spectroscopy (fNIRS), the optical signal in the intracranial (ICT) and extracranial tissues (ECT) was tested on 23 healthy volunteers. Participants were scanned with a frequency-domain NIRS device (690 and 830 nm) during 5 Valsalva maneuvers (VM) in a simulated ICU environment.

Results: The results showed an average error in probe displacement of 5.5 mm, a sufficient capacity to isolate oxyhemoglobin O₂Hb ($p=6.4 \cdot 10^{-6}$) and total hemoglobin HbT ($p=2.8 \cdot 10^{-5}$) in the ICT from the ECT, and to follow the changes of hemoglobin in the ICT during the VM (O₂Hb, $p=9.2 \cdot 10^{-4}$; HbT, $p=1.0 \cdot 10^{-3}$).

Conclusions: The developed approach appears to be suitable for use on TBI patients in the ICU.

Keywords: diffuse optical tomography, functional near-infrared spectroscopy, Valsalva maneuver, traumatic brain injury, intensive care unit.

*Mario Forcione, E-mail: m.forcione@campus.unimib.it

1 Introduction

Traumatic brain injury (TBI) is a leading cause of injury-related death and disability, resulting in an estimated one million deaths per year and shortened life expectancy for survivors by an average of six years (1). Currently, there is no disease-modifying treatment for the primary injury sustained by the brain during trauma, and the treatment for severe TBI remains limited to the prevention of subsequent insults (e.g., intracranial hypertension, ischemia), known as secondary brain injury.

41 Pivotal components of established treatment strategies in TBI patients are continuous
42 neuromonitoring and adaptation of the physiological parameters accordingly to maintain intra-
43 cranial homeostasis.

44 Functional near-infrared spectroscopy (fNIRS) represents a non-invasive, continuous
45 neuromonitoring tool for tracking changes in brain oxygenation in severe TBI patients by
46 measuring changes of oxyhemoglobin (O₂Hb) and deoxyhemoglobin (HHb) concentrations (2-4).

47 Contrast-enhanced NIRS can also be employed to detect changes in optical density during a dye
48 passage (e.g., indocyanine green [ICG]), allowing one to assess cerebral perfusion and blood-brain
49 barrier (BBB) damage (5-7). So far, commercially available NIRS devices (e.g., INVOS 5100

50 Cerebral Oximeter; ISS Oxiplex TS) have not shown sufficient ability to detect episodes of
51 ischemia in cases of severe TBI compared to the invasive intracranial techniques (e.g., brain partial
52 oxygen pressure [PbtO₂]) (8-10). However, the validity of a direct comparison between (i) the level

53 of cerebral tissue saturation and its fluctuations measured by the optical device, and (ii) the
54 absolute values and their changes measured by the PbtO₂ monitor, can be impaired by the complex
55 pathogenesis of brain trauma (11). In contrast, the addition of fNIRS and contrast-enhanced NIRS

56 to currently used neuromonitoring techniques, including PbtO₂, could potentially lead to a more
57 precise assessment of the tissue status (5, 11). Thereby, the drive for "non-invasive equivalence"
58 to PbtO₂ should be dropped, and instead the hitherto untested abilities of the optical technique,

59 when included in a multimodal monitoring regime, to stratify injury severity and aid clinical
60 decision-making, could finally be explored.

61 Diffuse optical tomography (DOT) is an extension of NIRS imaging that uses a distributed high-
62 density array of NIRS probes to spatially map tissue optical properties on a co-registered structural
63 image (e.g., computerized tomography [CT], magnetic resonance imaging [MRI]) (12, 13).

64 Compared to commercially available NIRS devices, DOT offers several advantages in the TBI
65 monitoring by addressing the anatomical complexity of the TBI lesions (11). This advantage
66 suggests that the commercially available NIRS devices used in clinical trials to date have not fully
67 explored the capabilities of optical techniques and that DOT should be tested instead. Whereas
68 DOT has been successfully demonstrated as a neuromonitoring technique in infants admitted to
69 the neonatal intensive care unit (ICU), the authors are not aware of any studies which have
70 performed DOT measurements on adult severe TBI patients in the ICU (14). One of the difficulties
71 in performing such studies is related to the clinical environment in which the optical data should
72 be acquired (e.g., contamination of the signal from the ambient light, lack of space at the patient's
73 bedside, presence of an intracranial bolt for intracranial pressure [ICP] neuromonitoring).
74 Furthermore, the standard processes of probe digitization using an electromagnetic digitizer, or of
75 adding skin markers to the subject's structural imaging, would be impractical solutions to
76 implement DOT in the clinical practice of brain trauma care (15-17): the electromagnetic signal
77 used by common digitizers can be impaired by the presence of metal at patients' bedsides; the
78 optical analysis can only be performed *after* skin markers have been added. Since a CT scan is one
79 of the first neuroimaging assessments made on TBI patients upon hospital admission, it is usually
80 performed urgently (18). This makes it unlikely that skin markers could be positioned on TBI
81 patients, at the locations where the optodes will be placed, prior to their first scan. Furthermore, it
82 should be taken into consideration that the CT scan is crucial to the clinical decision process, and
83 the presence of skin markers may increase the risk of imaging artefacts which can impair accurate
84 radiological assessment. Therefore, a method that co-registers the optical data into a CT scan
85 without the aid of skin markers, or electromagnetic digitization, is needed to make DOT analysis
86 on TBI patients possible from the point of admission.

87 We devised a system to perform DOT recordings on acute TBI patients from the point of admission
88 into ICU, comprised of a custom-made helmet for optical data acquisition in a multimodal
89 neuromonitoring (e.g., intracranial probes), and a co-registration of the optical data on patients'
90 structural images acquired for clinical assessment (e.g., CT scans). On healthy volunteers, we
91 tested: (i) the accuracy of the novel co-registration compared to one based on electromagnetic
92 digitization; and (ii) the capacity to accurately detect changes in levels of Hb in the ICT and
93 separate signals from ICT and ECT, during VM.

94 **2 Materials and Methods**

95 *2.1 Participants*

96 23 healthy volunteers (17 males, ages ranging between 20 and 42 years old with a mean of 28
97 years) were recruited into the RECOS study (Repetitive Concussion in Sport) (IRAS ID: 216703;
98 Ref. REC: 17/EE/0275) performed at Centre for Human Brain Health, University of Birmingham,
99 UK. They had different hair colours and hair densities.

100 Participants took part in this study after providing their written, informed consent. The study was
101 conformed to the Declaration of Helsinki and was approved by the East of England-Essex
102 Research Ethics Committee on 22 September 2017.

103 *2.2 Equipment*

104 *2.2.1 Optical helmet*

105 A custom-built optical helmet was adapted from the description of Tan et al. to make it suitable
106 for optical recordings on TBI patients in the ICU (Figure 1) (19).




107
108 **Figure 1. Picture of the helmet. The design features an open area at the top of the skull to avoid interference**
109 **with invasive neuromonitoring devices.**

110 The helmet comprised multiple adjustable segments. It could be adjusted to fit a variety of head
111 shapes and sizes, including post-surgery (e.g., decompressive craniectomies; external ventricular
112 drains (EVD)). It was vital that the helmet design maintained patient comfort and safety (e.g., no
113 increase in ICP), whilst also being rigid enough to keep the probes still during the recording. The
114 helmet is designed to be affixed to the patient using disposable Velcro straps or medical bandages
115 under the chin. To reduce the risk of infection and to grant access for intracranial monitoring and
116 treatment, no slices were positioned in the vicinity of the Kocher's point, where the ICP bolt and
117 the external ventricular drain (EVD) are usually positioned (18).

118 The helmet comprised multiple black layers. The thicker part was made of two layers of black
119 foam (The Michaels Companies, Las Colinas, Irving, Texas, USA) to shield the probes from
120 ambient light as well as to keep the helmet lightweight. These two layers were glued together while
121 held in a curved shape, to model the helmet frame. The internal lining of the helmet was made of
122 a thin layer of polyurethane rubber, RenCast (Freeman manufacturing and Supply Company,
123 Avon, Ohio, USA), which came into contact with the scalp, along with the probes. Polyurethane

124 rubber has been used in previous studies on TBI patients in the ICU and, in our experience, it
125 increases the helmet's integrity whilst also being suitable for clinical environments, as it is non-
126 porous, soft and easy to clean (8, 20).

127 The layout for optical recording covered only the frontal lobes to facilitate the positioning of the
128 probes on TBI patients who have limited head movement due to being unconscious in a supine
129 position with their head being tilted-up, as per ICP control care, potentially wearing a hard cervical
130 collar, and being intubated or with a tracheostomy (18) (Figure 2, A).



131
132 **Figure 2. A: Illustration of the optical probes' positions on a structural MRI. The sources' positions are**
133 **illustrated as red dots and the detectors' positions as blue dots. B: Average light sensitivity of the probe**
134 **configuration used, computed with a FEM approach (21). The light sensitivity is displayed up to an**
135 **attenuation of 60 dB (1000 times), and it is overlaid on an MRI of a representative participant. C: Absolute**
136 **frequency of the SD distances of the channels across all subjects.**
137 Before making holes to house the probes, a "3-space" FastTrak 3-D digitizer (Polhemus,
138 Colchester, Vermont, USA) was used to electromagnetically digitize the helmet surface using the
139 nasion and the external acoustic meatuses (EAM) of an artificial head as fiducial points. Based on
140 the obtained digitization, a distribution of probes was selected to scan the broadest possible area
141 with the maximal number of SD distances between 25 and 35 mm, so as to maximise sensitivity


142 to the ICT (12) (Figure 2, C). A custom-made MATLAB graphical user interface, near-infrared
143 optode montage automated design (NOMAD) (<https://github.com/kylemath/nomad>), developed
144 by Mathewson et al., was used to evaluate the acceptability of a plotted configuration of SD
145 locations on the helmet, by ensuring the illumination of the frontal lobes of both hemispheres
146 whilst avoiding cross-talk between SD pairs in a time-multiplexing cycle of light-emission (22).
147 The average light sensitivity of the probe configuration used was computed with a FEM approach
148 (Figure 2, B) (21). Below 60 mm distances, the probes were deemed to be usable for fNIRS-DOT
149 analysis based on the data recorded during the VM. Holes to position the probes were punched
150 only where the NOMAD analysis recommended. The limited number of holes, compared to the
151 model described by Tan et al., was necessary to reduce the amount of ambient light that reaches
152 the probes, by ensuring that no hole went unutilized, as well as to allow data recording in a supine
153 position (19). The large diameter of the holes where the detectors were inserted allowed for the
154 removal of any hair between the detector and the skin, in order to maximise photon harvesting
155 from the scalp (23, 24). A plastic cone was inserted inside the hole along with each detector to
156 mechanically secure them to the helmet, and to reduce the signal contamination by ambient light
157 (24). The holes were labelled and color-coded to facilitate the positioning of the probes. Small
158 white dots, that could be easily seen on an optical 3-D scanner, and whose small size still
159 maintained a high level of pinpoint accuracy, were drawn on the helmet in the regions that were
160 to be exposed in a supine participant.

161 *2.2.2 Data acquisition*

162 *2.2.2.1 Digitization*

163 Participants were scanned with an optical 3-D scanner (Artec Leo, Artec 3D, Luxembourg,
164 European Union) while wearing the helmet and sitting on a chair. The quality of signal acquisition

165 was confirmed by a visual inspection on the 3-D scanner screen. On a 3-D image, the x - y - z
166 coordinates of the fiducials (i.e. nasion, EAM) and those of the white dots were measured using
167 complementary software (Artec Scanner, Artec 3D, Luxemburg, European Union) (Figure 3, A).



168
169 **Figure 3 A: Picture from Artec Leo reconstructed using Artec Scanner and measurement of the 3-D**
170 **coordinates of the nasion, EAM, and white dots. B: example electromagnetic digitization of fiducials (nasion**
171 **and eam), white dots and optodes with FastTrak.**
172 The fiducials, the holes for the probes, and the white dots were also digitized using the FastTrak
173 3-D digitizer (Figure 3, B). The nasion and the EAM were chosen as fiducials because, according
174 to our research experience, they can be identified relatively easily on an anatomical scan without
175 the use of skin markers.

176 2.2.2.2 *Optical recordings*


177 Optical data were recorded with a FD-NIRS system (Imagent; ISS Inc., Champaign, IL). Data were
178 collected from 15 detectors, each measuring light emitted by eight of the 16 total source locations
179 (120 dual-wavelength channels). Laser diodes delivered light, modulated at 110 MHz, for each
180 source location at 690 and 830 nm wavelength (max power: 10 mW, mean power: 1 mW). The
181 different diodes were switched on in a time-multiplexing manner. The light from the diodes was

182 transmitted to the scalp using thin optical fibres (diameter = 400 μm), whereas the back-scattered
183 light from the head was collected using detector fibre bundles (diameter = 3 mm) connected to
184 photomultiplier tubes (PMTs). Based on heterodyne detection, signal digitization, and fast Fourier
185 Transform, the system generated temporal modulations in the emitted light DC (average), AC
186 (amplitude), and phase. Optical data were sampled at 39.0625 Hz.

187 Coupling efficiency between the optodes and the skin, as well as signal noise, were monitored
188 before the beginning of each scan using a complementary software (BOXY, ISS Inc., Champaign,
189 IL) and adjustments were made to maximise the signal quality.

190 2.2.2.3 *Valsalva manoeuvre*

191 The VM is a forced attempt to exhale against a closed glottis. The detailed knowledge of the
192 expected hemodynamic changes during VM was used to validate the system presented, and so is
193 described hereafter. Based on the changes in blood pressure, the VM can be divided into four
194 phases (Figure 4) (25).



195
196
197
198

Figure 4. Representation of the physiological changes during the Valsalva manoeuvre's phases. The blue and grey backgrounds represent open and closed glottis, respectively. A: standard physiological state. B: Phase I. C: Phase II. D: Phase III. E: Phase IV.

199 During phase I, lasting approximately 2 seconds, an increase of intrathoracic pressure, due to the
200 forced exhalation attempt against the closed glottis, causes an increase of external pressure on the
201 main vessels (e.g., aorta, superior vena cava) with consequent increase of the arterial and venous
202 pressures (Figure 4 [B]) (25, 26). The increase of blood volume causes an increase in the levels of
203 total hemoglobin (HbT) and tissue saturation in the ICT and ECT (Figure 4 [B]) (27). Due to the
204 confinement of the brain into a non-expandable space, the increase of cerebral blood volume and
205 the incapacity for it to drain because of the high intrathoracic pressure is immediately followed by
206 a sharp rise in ICP (28-30). The high pressure exerted on the spinal compartment by the thorax is
207 another contributing factor in the increase of ICP (30). This simultaneous rise in ICP serves as a
208 protection mechanism for the brain and its blood vessels, which could otherwise be violently
209 damaged by the sudden increase of arterial pressure (29, 30). During phase II, the reduction of
210 end-diastolic volume, because of the reduction in venous return due to the high central venous
211 pressure, leads to a decrease of arterial pressure and an increase of HR as a compensatory
212 mechanism (Figure 4 [C]) (25). The reduction of blood supply, combined with the high brain
213 metabolism, causes a progressive, non-compensated consumption of tissue O₂Hb and a consequent
214 gradual decrease of ICT saturation (Figure 4 [C]) (27, 31, 32). On the other hand, a similar
215 reduction of tissue saturation is not present in the ECT due to its low level of metabolism (Figure
216 4 [C]) (32). During phase III, the glottis is opened, and this is followed by a drop of intrathoracic
217 pressure. The sudden lack of external pressure on the main vessels determines a high level of
218 venous return and a rapid decrease of arterial blood pressure and ICP (25, 33) (Figure 4 [D]). It is
219 worth noting that, at this stage, the level of arterial blood pressure is at its lowest point (33). The
220 large pool of Hb is removed from the ECT by the venous return (Figure 4 [D]). In ICT, the already
221 low tissue saturation further decreases due to the reduced arterial blood pressure (Figure 4 [D])

222 (27). In phase IV, there is a peak of arterial blood pressure and a subsequent increase of the O₂Hb
223 levels in the ECT and ICT (Figure 4 [E]) (25, 27).

224 The VM was considered a valid task to test the neuromonitoring assessment with fNIRS-DOT,
225 due to the substantial and opposite changes of tissue saturation in ICT and ECT. These two
226 different behaviours may give insight on the ability of the developed DOT approach to uncouple
227 the contributions from the ICT and ECT, and so test the capacity of the optical technique to record
228 the brain signal without interference from the ECT (34, 35). Furthermore, the VM simulates, on
229 healthy human subjects, a common post-brain trauma scenario, characterised by a decrease of brain
230 saturation and an increase of ICP (32, 36-39).

231 Participants were instructed to perform five VMs of 10 seconds each, preceded and followed by
232 three minutes of rest. 10 seconds before each VM, participants were asked to prepare for the
233 upcoming physical task taking long gasps as they were deemed suitable to generate intrathoracic
234 pressure during the VM. Participants were in a semi-supine position with their heads slightly tilted
235 by a cushion and in an illuminated environment. A clinician was observing the participants to
236 reduce the effects of adverse events (e.g., syncope) and to ensure that the VMs were performed
237 correctly. The VMs were monitored by measuring the expected HR increase using a pulse-
238 oximeter, Fingertip (ChoiceMMed, Yuquan Road, Haidian District, Beijing, PR China), and by
239 observing signs of physical effort (e.g., face redness, shakes for the strain, breath release at the end
240 of the VM). Participants could stop the experiment at any stage.

241 *2.2.3 Data analysis*

242 *2.2.3.1 Digitization*

243 The white dots from Artec Leo were co-registered on a structural image template through fiducial
244 alignment based on least square fitting. The optode locations were co-registered by combining the

245 locations of the white dots and the fiducials according to Artec Leo, with the locations of the white
246 dots and the optodes according to the FastTrak 3-D digitizer. This method was necessary to
247 mitigate against the potential for inaccurate optode localisation and, in extreme cases, optode
248 omission by Artec Leo, due to blurred images, large holder sizes (especially for the detectors), and
249 the difficulty of ascertaining the probes' depth within the helmet. On the other hand, the digitizer
250 is capable of measuring the coordinates of both the white dots and the optodes more reliably.
251 The root mean square error (RMSE) between the Artec Leo image, using the method described
252 above, and the FastTrak 3-D digitizer; and the number of white dots necessary to optimize this
253 approach were investigated. Finally, the co-registration of the optodes on a template was used for
254 the optical analysis.

255 *2.2.3.2 Optical data analysis*

256 The optical DC intensity data were transformed into optical density (OD), movement corrected,
257 and band-pass filtered between 0.01 and 0.5 Hz (zero-lag, 2nd order, Butterworth digital filter) (40,
258 41). For each wavelength and channel, in addition to the ten seconds of the VM trial, single-trial
259 OD was evaluated from 20 seconds prior up to 20 seconds after the VM onset. OD signal, of 50
260 seconds duration, was randomly chosen during each rest period to obtain a control measurement.
261 Average OD response for each wavelength, channel and subject was computed.

262 For each wavelength, only channels that provided a maximum average OD change below or equal
263 to 0.05 (~5% signal change) were deemed as usable. On average, 78 ± 10 and 72 ± 11 channels for
264 the 830 nm wavelength and the 690 nm wavelength, respectively, were used for further analysis.


265 Furthermore, a model of light propagation within head (forward model) and an inverse procedure
266 were performed using a structural MRI (MPRAGE sequence) as a template. The FEM applied to
267 the diffusion equation was used to estimate the forward model (42, 43). The FEM software
268 NIRFAST was used to model light propagation through the head and to compute Jacobian

269 (sensitivity) matrices of DC light intensity to absorption changes induced by Hb oscillations (21,
270 44). Figure 2 [B] reports an example of an average Jacobian (average optical sensitivity) for a
271 participant overlaid onto the anatomical MRI template. The average Jacobian is displayed up to an
272 attenuation of 60 dB (1000 times) from its maximum value. “Fine” meshes (maximum tetrahedral
273 volume 2 mm³) were generated for FEM using the MATLAB software iso2mesh (45). The
274 heterogeneous head models were based on the segmentation of the anatomical MRI. Segmentation
275 of the skull and scalp, CSF, white matter, and grey matter was performed using Statistical
276 Parametric Mapping (SPM) functions applied to the image (46, 47). Baseline optical properties
277 (μ_{α} , μ'_{s} and η) of the tissues at the relevant wavelength were taken from Tian et al. (48). An inverse
278 procedure based on energy minimisation of the solution was used to convert intensity changes on
279 individual channels to absorption changes in voxel space at each wavelength (44). The Lambert-
280 Beer Law was inverted to evaluate O₂Hb and HHb oscillations given absorption modulation at the
281 two wavelengths employed. The extinction coefficients of the two forms of Hb at the 690 nm and
282 830 nm wavelengths were extracted from Zijlstra et al. (49). Using the computed MRI
283 segmentation, the MRI was separated into ICT and ECT compartments. The ICT compartment
284 was evaluated by summing a smoothed version of white and grey matter segmentation. The ECT
285 compartment was obtained by subtracting a binary image of the head with the previously defined
286 ICT compartment. O₂Hb and HHb signals were evaluated for each subject by averaging the Hb
287 oscillations obtained in the different compartments, in a region where enough light sensitivity was
288 obtained (-60 dB of attenuation, Figure 2 [A]). Across subjects’ average responses of O₂Hb, HHb,
289 and HbT (HbT= O₂Hb +HHb) and associated statistics were evaluated.

290 3 Results

291 3.1 Co-registration

292 Figure 5 reports the RMSE between Artec Leo and FastTrak for the available white dots (51 in
293 total), as a function of the number of white dots used for co-registration between the two
294 modalities.




295
296 **Figure 5:** RMSE between Artec Leo and FastTrak for the available white dots (51 in total), as a function of the number
297 of white dots used for co-registration between the two modalities. The RMSE was computed using 10,000 iterations
298 for each number of white dots selected, where the white dots used were randomly chosen.
299 The RMSE reported in the figure was computed using 10,000 iterations for each number of white
300 dots selected, where the white dots used for co-registration in each iteration were randomly chosen
301 from all those available. The minimum number selected was 3. The plot in Figure 5 clearly shows

302 a plateau in the error between the two modalities with a number of white dots above ten. With this
303 number of dots, the average distance between the corresponding locations is within 5.5 mm.

304 3.2 Optical data

305 Five VMs were performed per participant, resulting in a mean increase in HR of 17 bpm (± 11)
306 ($t=5.17$, $df=21$, $p=3.9 \cdot 10^{-5}$).

307 Figure [A] reports the ICT and ECT compartments that were probed by the optical array, overlaid
308 onto the structural template image.



309
310 **Figure 6. A: Computed ICT and ECT compartments that were probed by the optical array overlaid on the**
311 **structural image. B: Across subjects' average modulations, and associated standard error, of O₂Hb, HHb**
312 **and HbT during the rest period and the VM in the two compartments. BH: breath-holding**
313

314 The compartments were computed based on the structural MRI employing the algorithm described.

315 Figure [B] reports the across subjects' average modulations, and associated standard error, of
316 O₂Hb, HHb, and HbT during the rest period and the VM in the two compartments.

317 A clear difference in fluctuations of levels of Hb during rest and the VM is visible. Moreover,
318 during the VM, decoupling in the Hb responses between the ICT and ECT compartments is
319 evident.

320 In both these two layers, the changes of O₂Hb, HHb, and HbT are in agreement with the expected
321 changes of tissue saturation during different VM phases. In the first few seconds of the task, the
322 levels of O₂Hb and HbT began to increase in the ECT and decrease in the ICT. Subsequently, the
323 levels of O₂Hb and HbT in the ECT remained high and low in the ICT. At the time of the breath
324 release (i.e. phase III), the levels of O₂Hb and HbT sharply decreased in the ECT. A reduction in
325 the ICT was less visible, probably because brain saturation was already compromised. In the
326 following seconds (i.e. phase IV), there was a sharp increase of O₂Hb and HbT in the ECT and
327 ICT, with a slight time-delay between the two. Having completed the VM, the parameters of O₂Hb
328 and HbT in the ECT and ICT gradually returned to the values of the resting state. Notice that
329 significant changes in O₂Hb and HbT, coupled with a rather stable HHb, are consistent with
330 modulations in the degree of perfusion of the tissue investigated, driven by changes in the pressure
331 gradient between the head and the body at a constant tissue oxygen consumption. This is in
332 agreement with previous studies on VM (50).

333 To test the statistical significance of the results, the peak changes in O₂Hb, HHb, and HbT during
334 the breath-holding (BH) phases of the VM (i.e., phases I and II) were extracted. A repeated-
335 measures analysis of variance (rmANOVA) was performed considering two within-subject factors
336 (i.e. Condition: Rest and VM; and Compartment: ICT and ECT). A significant effect was obtained
337 when considering O₂Hb for VM vs Rest ($F=6.04$, $df=1-21$, $p=0.023$) and ICT vs. ECT ($F=35.51$,
338 $df=1-21$, $p\sim 0$), and when considering HbT for VM vs Rest ($F=7.28$, $df=1-21$, $p=0.014$) and ICT
339 vs. ECT ($F=28.19$, $df=1-21$, $p\sim 0$). Importantly a highly significant interaction between factors

340 was obtained, indicating a synergic effect between condition and compartment for O₂Hb (F=35.56,
341 df=1-21, p≈0) and HbT (F=28.42, df=1-21, p≈0). When performing a post-hoc analysis, a
342 significant increase in O₂Hb and HbT was found in the ECT when comparing VM to Rest (O₂Hb:
343 t=5.10, df=21, p=4.7·10⁻⁵; HbT: t=4.78, df=21, p=9.9·10⁻⁴) whereas a significant decrease in O₂Hb
344 and HbT was obtained in the ICT (O₂Hb: t=-3.85, df=21, p=9.2·10⁻⁴; HbT: t=-3.74, df=21,
345 p=1.0·10⁻³).

346 Moreover, when comparing the ECT and the ICT during the VM, a significant difference was
347 obtained for both O₂Hb (t=5.96, df=21, p=6.4·10⁻⁶) and HbT (t=5.32, df=21, p=2.8·10⁻⁵).

348 **4 Discussion**

349 *4.1 Co-registration*

350 The developed approach yielded an average displacement error, expressed as RMSE, of 5.5 mm
351 (Figure 5), which importantly is well below the spatial resolution limit of DOT of ~15 mm (51).
352 This means that the Artec Leo images, through the method presented here, can be applied to obtain
353 an accurate co-registration of the optodes onto an anatomical template or a subject-specific
354 anatomical image, with similar performance to electromagnetic digitization.

355 It is important to stress that the method reported here relies upon measured distances between the
356 white dots and the optodes from a previously performed electromagnetic digitization. Hence, the
357 developed approach for the co-registration of optical data on TBI patients in the ICU would require
358 digitizing the white dots and the optical array on a phantom head or a healthy volunteer, with a
359 head circumference similar to that of the subject that is going to be measured. Using a digitized
360 helmet, the measurements performed on the TBI patient using an Artec Leo image would allow

361 clinicians to correctly locate the array onto the patient's CT scan or MRI, where the same fiducials
362 can be selected.

363 Remarkably, this method can be performed without additional CT scans beyond the one performed
364 upon admission for clinical assessment. The co-registration to the initial CT scan allows clinicians
365 to prioritise the optical analysis of the most affected and clinically significant regions. The optical
366 assessment of co-registered regions also allows clinicians to link the optical parameters and their
367 functional evolution, to known lesions on the CT scans and their anatomical evolution. It should
368 be mentioned that the ability to assess the optical properties of a brain lesion would depend on its
369 size, so lesions smaller than the margin of error between the two co-registrations are likely to be
370 overlooked, however, this problem is inherent to the limited spatial resolution of DOT (51). The
371 decoupling of the commencement of DOT recordings from a targeted anatomical image would be
372 useful either in the most severe cases of TBI, who for example can experience high ICP levels
373 when supine or have in place extracorporeal systems (e.g., hemodialysis), all of which can be
374 obstacles to performing further imaging, or in cases where clinicians are prioritising other patients
375 and a second anatomical imaging may be postponed.

376 The degree of error between the two techniques may differ slightly from the 5.5 mm RMSE
377 measured. The coordinates from Artec Leo can fail to exactly identify the digitized spots within
378 the area of the white dots, so the absolute value of RMSE could be partially due to the testing
379 analysis itself rather than the method implemented. However, this miscalculation can also affect
380 the optical analysis, as the optode co-registration by Artec Leo is based on the measured distances
381 between white dots and optodes by the electromagnetic digitization. The small sizes of the white
382 dots mitigate, to a greater or less extent, this error. The holders' width, especially for the detectors,

383 can be another cause of inaccuracy in the optical analysis, because there could be a mismatch
384 between the digitized spot within the holders and the real loci of the optodes.

385 Of note is that the analysis was performed on white dots that were uniformly distributed around
386 the optode locations. Thus, it is essential for future helmets to maintain the uniformity of white
387 dots in their design to accurately recover the optode locations on the anatomical image.

388 Although the results show a plateau of RMSE by using ten white dots, applications of this method
389 on TBI for research purposes led us to think that helmets should possess as many white dots as can
390 be fitted onto the exposed surface. This is because acquiring an optimal 3D image may not be
391 always possible in hospital settings and blurred images may impair the retrieval of all the white
392 dots. Moreover, the availability of white dots uniformly covering, not only the area around the
393 optodes, but the whole head surface, may allow to employ more complex algorithms that may
394 further reduce the optodes placement error (16, 17).

395 *4.2 Optical Analysis*

396 The fluctuations of O₂Hb, HHb, and HbT levels in the ICT due to the VM were separated in the
397 analysis from those in the ECT. In both layers, the results conformed with the expected changes
398 during the different phases of the VM. This consistency validates the optical system as being
399 capable of monitoring changes in brain Hb concentration through fNIRS-DOT analysis.

400 A decrease of cerebral saturation during VM was also reported in studies that used NIRS devices
401 (27, 32, 34, 52-54). However, these analyses were limited to measurement of the decrease of
402 cerebral saturation, and the ECT signal plausibly contaminated them by an unknown value. On the
403 contrary, the fNIRS-DOT analysis reported herein allowed for a comparison of the changes of
404 levels of Hb in the ICT and ECT, providing a better understanding of how the detected optical
405 signal relates to brain physiology. Similarly to other studies, we were unable to quantify the role

406 played by the ECT in the ICT signal, however the strong hemodynamic signal in the ICT leads us
407 to think that no matter the extent of this contamination, it would be insufficient to obscure the
408 tracking of the ICT signal during the VM (55). We believe that our test validated the aims of the
409 study: (i) to test the ability of ICT monitoring, and (ii) to separate the ICT-ECT signals.
410 These results are in agreement with a pilot study which was performed without the extensive use
411 of short channels, leading us to think that the co-registration process implemented in both studies
412 is a critical component in these optical results (56). This is in agreement with other studies that
413 showed that co-registering the optical signal with a structural image could improve the imaging
414 resolution and the localisation accuracy (57, 58). Furthermore, Clancy et al. specifically reported
415 an improvement in the accuracy of the signal detected during VM in simulated studies, by co-
416 registering the optical data with a structural image (59, 60). However, in a study on a healthy
417 volunteer, the co-registration of the optical signal from a high-density patch into an atlas did not
418 yield the same capacity to retrieve physiological changes of tissue saturation associated with the
419 VM as in the simulations (61). This limited capacity in an *in vivo* analysis may be explained to
420 some extent by a suboptimal co-registration process compared with the co-registration system
421 presented herein. Furthermore, as mentioned in the introduction, the co-registration of the optical
422 signal with a subject-specific image may be particularly significant for the assumption about the
423 baseline optical parameters in TBI patients to be valid. Layer thicknesses and optical properties
424 can vary significantly between individuals and across different areas of the head (62, 63). These
425 differences can be even more significant in TBI patients due to brain lesions (e.g., ischemia,
426 hemorrhage) or surgical interventions (e.g., decompressive craniectomy) (11). The co-registration
427 of the optical analysis into a subject-specific image takes into account the individual anatomy of
428 these layers and brain lesions.

429 *4.3 Limitations*

430 *4.3.1 Absence of arterial blood pressure monitoring*

431 Grading of the VM based on the changes in blood pressure was not possible. This may have
432 affected the results as the quality of the performance of the VM may have differed across
433 participants due to varying levels of commitment to the task and levels of cardiovascular fitness.
434 Furthermore, it is not possible to perform the VM in the supine position at the same level as in the
435 orthostatic one (26, 33). This further highlights the risk of different intensities of VM across
436 participants if the VM is not performed with sufficient strength.

437 The lack of blood pressure monitoring also limited the separation of the different phases of the
438 VM to an analysis of the expected changes of levels of Hb at different times without accurately
439 correlating them to the blood pressure changes.

440 *4.3.2 Absence of subject-specific structural image*

441 An MRI template was used in place of the subject-specific structural image for the co-registration
442 of the optical data. This reduced the accuracy of the fitting procedure of the probes on the
443 participants' scalps due to their different head shapes. Moreover, the approximation also affected
444 the optical data analysis, because the assumed optical properties of the different head layers were
445 probably not anatomically identical among subjects (e.g., layer thickness).

446 *4.3.3 Global hemodynamic response during Valsalva manoeuvre*

447 The cerebral perfusion and tissue saturation in TBI patients can vary significantly across the brain,
448 while the VM triggers a similar hemodynamic response across the whole brain (64, 65). Therefore,
449 the ability to separate the different hemodynamic across the brain of TBI patients could not be
450 determined from the results obtained.

451 *4.3.4 Limited field of view*

452 Although studies which investigated a broader coverage of the head using DOT have been
453 presented for clinical applications in non-ICU scenarios (e.g., Parkinson syndrome), we limited
454 the analysis to the frontal lobes to address the standard of care in ICU (44). In the future, new
455 studies with a broader field of acquisition should be implemented, while maintaining the standard
456 of care in acute TBI patients.

457 *4.3.5 Limited number of probes and wavelengths*

458 Our study was limited to 15 detectors and 16 sources at two wavelengths (690 and 830 nm) mainly
459 designed to measure O₂Hb and HHb levels. Future studies should apply HD (high density) DOT
460 to further improve the optical accuracy, as well as using different wavelengths to make the
461 recording more suitable for ICG measurement (6).

462 *4.3.6 Absence of phase-shift analysis*

463 Though a frequency-domain device was used, phase data were not implemented into the optical
464 analysis. From results on simulated and phantom studies, analysis of the phase-shift can,
465 theoretically, further increase imaging resolution and accuracy (66, 67). Future studies should
466 explore this potentiality.

467 **5 Conclusions**

468 The results suggest that our method is capable of addressing the foreseeable problems associated
469 with DOT neuromonitoring in TBI patients in the ICU. This procedure could enable fNIRS-DOT
470 and contrast-enhanced DOT recordings as part of a multimodal monitoring on TBI patients upon
471 hospital admission.

472 Future research should test other probe-position geometries up to HD-DOT, which may further
473 increase the capability of the optical analysis to separate the signal from the different head layers
474 and to monitor brain lesions. Clinical studies should be designed so that DOT measurements are
475 included in a multimodal monitoring.

476 **6 Conflict of interest**

477 None to disclose.

478 **7 Code, Data, and Materials Availability**

479 Code and anonymized optical data are available in Matlab format in FigShare
480 at <https://doi.org/10.6084/m9.figshare.25866682.v1>

481 **8 Acknowledgments**

482 This article presents independent research funded by the National Institute for Health Research
483 Surgical Reconstruction and Microbiology Research Centre (NIHR SRMRC), partnership between
484 University Hospitals Birmingham NHS Foundation Trust, the University of Birmingham, and the
485 Royal Centre for Defence Medicine. The views expressed are those of the authors and not
486 necessarily those of the NHS, the NIHR or the Department of Health.

487 Dr Mario Forcione thanks Mr Michael Orgill from Gill Learning (<https://www.gill-learning.co.uk>)
488 for his academic advice and proof reading.

489 **9 References**

- 490 1. A. I. R. Maas et al., "Traumatic brain injury: integrated approaches to improve prevention,
491 clinical care, and research," *Lancet Neurol* **16**(12), 987-1048 (2017).
- 492 2. D. J. Davies et al., "Near-Infrared Spectroscopy in the Monitoring of Adult Traumatic
493 Brain Injury: A Review," *Journal of neurotrauma* **32**(13), 933-941 (2015).

- 494 3. M. S. Green, S. Sehgal, and R. Tariq, "Near-Infrared Spectroscopy: The New Must Have
495 Tool in the Intensive Care Unit?," *Seminars in cardiothoracic and vascular anesthesia*
496 **20**(3), 213-224 (2016).
- 497 4. I. K. Haitsma, and A. I. Maas, "Monitoring cerebral oxygenation in traumatic brain injury,"
498 *Progress in brain research* **161**(207-216 (2007).
- 499 5. M. Forcione et al., "Cerebral perfusion and blood–brain barrier assessment in brain trauma
500 using contrast-enhanced near-infrared spectroscopy with indocyanine green: A review,"
501 *Journal of Cerebral Blood Flow & Metabolism* 0271678X20921973 (2020).
- 502 6. M. Forcione et al., "Dynamic contrast-enhanced near-infrared spectroscopy using
503 indocyanine green on moderate and severe traumatic brain injury: a prospective
504 observational study," *Quantitative Imaging in Medicine and Surgery* (2020).
- 505 7. M. Ganau et al., "Breakthrough in the assessment of cerebral perfusion and vascular
506 permeability after brain trauma through the adoption of dynamic indocyanin green-
507 enhanced near-infrared spectroscopy," *Quantitative Imaging in Medicine and Surgery*
508 **10**(11), 2081-2084 (2020).
- 509 8. D. J. Davies et al., "Cerebral Oxygenation in Traumatic Brain Injury: Can a Non-Invasive
510 Frequency Domain Near-Infrared Spectroscopy Device Detect Changes in Brain Tissue
511 Oxygen Tension as Well as the Established Invasive Monitor?," *Journal of neurotrauma*
512 **36**(7), 1175-1183 (2019).
- 513 9. S. R. Leal-Noval et al., "Invasive and noninvasive assessment of cerebral oxygenation in
514 patients with severe traumatic brain injury," *Intensive Care Med* **36**(8), 1309-1317 (2010).
- 515 10. W. Weigl et al., "Application of optical methods in the monitoring of traumatic brain
516 injury: A review," *J Cereb Blood Flow Metab* **36**(11), 1825-1843 (2016).
- 517 11. M. Forcione et al., "Mismatch between Tissue Partial Oxygen Pressure and Near-Infrared
518 Spectroscopy Neuromonitoring of Tissue Respiration in Acute Brain Trauma: The
519 Rationale for Implementing a Multimodal Monitoring Strategy," *International Journal of*
520 *Molecular Sciences* **22**(3), 1122 (2021).
- 521 12. I. J. Bigio, and S. Fantini, *Quantitative Biomedical Optics: Theory, Methods, and*
522 *Applications*, Cambridge University Press, Cambridge (2016).
- 523 13. M. D. Wheelock, J. P. Culver, and A. T. Eggebrecht, "High-density diffuse optical
524 tomography for imaging human brain function," *Review of Scientific Instruments* **90**(5),
525 051101 (2019).
- 526 14. T. Austin et al., "Three dimensional optical imaging of blood volume and oxygenation in
527 the neonatal brain," *NeuroImage* **31**(4), 1426-1433 (2006).
- 528 15. G. Giacalone et al., "Time-domain near-infrared spectroscopy in acute ischemic stroke
529 patients," *Neurophotonics* **6**(1), 015003 (2019).
- 530 16. C. Whalen et al., "Validation of a method for coregistering scalp recording locations with
531 3D structural MR images," *Hum Brain Mapp* **29**(11), 1288-1301 (2008).
- 532 17. A. M. Chiarelli et al., "Comparison of procedures for co-registering scalp-recording
533 locations to anatomical magnetic resonance images," *Journal of biomedical optics* **20**(1),
534 016009 (2015).
- 535 18. E. E. Moore, D. V. Feliciano, and K. L. Mattox, *Trauma, Eighth Edition*, McGraw-Hill
536 Education (2017).
- 537 19. C. H. Tan et al., "Mapping cerebral pulse pressure and arterial compliance over the adult
538 lifespan with optical imaging," *PLoS One* **12**(2), e0171305 (2017).

- 539 20. L. He et al., "Noninvasive continuous optical monitoring of absolute cerebral blood flow
540 in critically ill adults," *Neurophotonics* **5**(4), 045006 (2018).
- 541 21. H. Dehghani et al., "Near infrared optical tomography using NIRFAST: Algorithm for
542 numerical model and image reconstruction," *Communications in numerical methods in
543 engineering* **25**(6), 711-732 (2008).
- 544 22. K. E. Mathewson et al., "Dynamics of Alpha Control: Preparatory Suppression of Posterior
545 Alpha Oscillations by Frontal Modulators Revealed with Combined EEG and Event-
546 related Optical Signal," *Journal of cognitive neuroscience* **26**(10), 2400-2415 (2014).
- 547 23. C. S. Robertson et al., "Clinical evaluation of a portable near-infrared device for detection
548 of traumatic intracranial hematomas," *Journal of neurotrauma* **27**(9), 1597-1604 (2010).
- 549 24. M. Forcione et al., "Tomographic Task-Related Functional Near-Infrared Spectroscopy in
550 Acute Sport-Related Concussion: An Observational Case Study," *Int J Mol Sci* **21**(17),
551 (2020).
- 552 25. F. P. Tiecks et al., "Effects of the valsalva maneuver on cerebral circulation in healthy
553 adults. A transcranial Doppler Study," *Stroke* **26**(8), 1386-1392 (1995).
- 554 26. F. Pott et al., "Middle cerebral artery blood velocity during a valsalva maneuver in the
555 standing position," *Journal of applied physiology (Bethesda, Md. : 1985)* **88**(5), 1545-1550
556 (2000).
- 557 27. B. G. Perry et al., "Cerebral hemodynamics during graded Valsalva maneuvers," *Frontiers
558 in physiology* **5**(349) (2014).
- 559 28. M. H. Wilson, "Monro-Kellie 2.0: The dynamic vascular and venous pathophysiological
560 components of intracranial pressure," *J Cereb Blood Flow Metab* **36**(8), 1338-1350 (2016).
- 561 29. H. Prabhakar et al., "Intracranial pressure changes during Valsalva manoeuvre in patients
562 undergoing a neuroendoscopic procedure," *Minim Invasive Neurosurg* **50**(2), 98-101
563 (2007).
- 564 30. M. J. Haykowsky et al., "Resistance exercise, the Valsalva maneuver, and cerebrovascular
565 transmural pressure," *Med Sci Sports Exerc* **35**(1), 65-68 (2003).
- 566 31. J. C. Greenfield, Jr., J. C. Rembert, and G. T. Tindall, "Transient changes in cerebral
567 vascular resistance during the Valsalva maneuver in man," *Stroke* **15**(1), 76-79 (1984).
- 568 32. D. Davies et al., "Comparison of near infrared spectroscopy with functional MRI for
569 detection of physiological changes in the brain independent of superficial tissue," *The
570 Lancet* **387**(S34) (2016).
- 571 33. B. G. Perry et al., "The cerebrovascular response to graded Valsalva maneuvers while
572 standing," *Physiol Rep* **2**(2), e00233 (2014).
- 573 34. D. J. Davies et al., "The Valsalva maneuver: an indispensable physiological tool to
574 differentiate intra versus extracranial near-infrared signal," *Biomed. Opt. Express* **11**(4),
575 1712-1724 (2020).
- 576 35. J. Steinbrink et al., "Determining changes in NIR absorption using a layered model of the
577 human head," *Physics in Medicine and Biology* **46**(3), 879-896 (2001).
- 578 36. D. J. Davies, "Cerebral near infra-red spectroscopy in traumatic brain injury as a potential
579 independent monitoring modality and alternative to invasive tissue oxygen tension
580 sensors," in *School of Clinical and Experimental Medicine University of Birmingham,*
581 *University of Birmingham UBIRA E THESIS* (2017).
- 582 37. K. J. Lee et al., "Non-invasive detection of intracranial hypertension using a simplified
583 intracranial hemo- and hydro-dynamics model," *BioMedical Engineering OnLine* **14**(1),
584 51 (2015).

- 585 38. S. R. Mousavi et al., "Measurement of in vivo cerebral volumetric strain induced by the
586 Valsalva maneuver," *J Biomech* **47**(7), 1652-1657 (2014).
- 587 39. B. B. Ertl-Wagner et al., "Demonstration of periventricular brain motion during a Valsalva
588 maneuver: description of technique, evaluation in healthy volunteers and first results in
589 hydrocephalic patients," *European Radiology* **11**(10), 1998-2003 (2001).
- 590 40. A. M. Chiarelli et al., "A kurtosis-based wavelet algorithm for motion artifact correction
591 of fNIRS data," *NeuroImage* **112**(128-137) (2015).
- 592 41. D. Perpetuini et al., "A Motion Artifact Correction Procedure for fNIRS Signals Based on
593 Wavelet Transform and Infrared Thermography Video Tracking," *Sensors* **21**(15), 5117
594 (2021).
- 595 42. A. Ishimaru, "Diffusion of light in turbid material," *Appl. Opt.* **28**(12), 2210-2215 (1989).
- 596 43. K. D. Paulsen, and H. Jiang, "Spatially varying optical property reconstruction using a
597 finite element diffusion equation approximation," *Medical physics* **22**(6), 691-701 (1995).
- 598 44. A. T. Eggebrecht et al., "Mapping distributed brain function and networks with diffuse
599 optical tomography," *Nat Photonics* **8**(6), 448-454 (2014).
- 600 45. F. Qianqian, and D. A. Boas, "Tetrahedral mesh generation from volumetric binary and
601 grayscale images," *2009 IEEE International Symposium on Biomedical Imaging: From
602 Nano to Macro* 1142-1145 (2009).
- 603 46. K. J. Friston et al., "Statistical parametric maps in functional imaging: A general linear
604 approach," *Human Brain Mapping* **2**(4), 189-210 (1994).
- 605 47. W. Penny et al., *Statistical Parametric Mapping: The Analysis of Functional Brain Images*,
606 Elsevier: Amsterdam, The Netherlands (2011).
- 607 48. F. Tian, and H. Liu, "Depth-compensated diffuse optical tomography enhanced by general
608 linear model analysis and an anatomical atlas of human head," *NeuroImage* **85 Pt 1**(166-
609 180) (2014).
- 610 49. W. G. Zijlstra, A. Buursma, and W. P. Meeuwse-van der Roest, "Absorption spectra of
611 human fetal and adult oxyhemoglobin, de-oxyhemoglobin, carboxyhemoglobin, and
612 methemoglobin," *Clin Chem* **37**(9), 1633-1638 (1991).
- 613 50. A. Bluestone et al., "Three-dimensional optical tomography of hemodynamics in the
614 human head," *Opt Express* **9**(6), 272-286 (2001).
- 615 51. A. M. Chiarelli et al., "Combining energy and Laplacian regularization to accurately
616 retrieve the depth of brain activity of diffuse optical tomographic data," *Journal of
617 biomedical optics* **21**(3), 36008 (2016).
- 618 52. M. Clancy et al., "Comparison of Neurological NIRS signals during standing Valsalva
619 maneuvers, pre and post vasopressor injection," in *Diffuse Optical Imaging V H. Dehghani,*
620 *and P. Taroni, Eds., Spie-Int Soc Optical Engineering, Bellingham* (2015).
- 621 53. D. Canova et al., "Inconsistent detection of changes in cerebral blood volume by near
622 infrared spectroscopy in standard clinical tests," *Journal of applied physiology (Bethesda,*
623 *Md. : 1985)* **110**(6), 1646-1655 (2011).
- 624 54. R. Saager, and A. Berger, "Measurement of layer-like hemodynamic trends in scalp and
625 cortex: implications for physiological baseline suppression in functional near-infrared
626 spectroscopy," *Journal of biomedical optics* **13**(3), 034017 (2008).
- 627 55. N. Eleveld et al., "The Influence of Extracerebral Tissue on Continuous Wave Near-
628 Infrared Spectroscopy in Adults: A Systematic Review of In Vivo Studies," *J Clin Med*
629 **12**(8), (2023).

- 630 56. M. Forcione, "Neuromonitoring in mild, moderate, and severe acute brain trauma using
631 non-invasive diffuse optics," in *College of Medical and Dental Sciences*, p. 230, University
632 of Birmingham (2021).
- 633 57. D. A. Boas, and A. M. Dale, "Simulation study of magnetic resonance imaging-guided
634 cortically constrained diffuse optical tomography of human brain function," *Appl Opt*
635 **44**(10), 1957-1968 (2005).
- 636 58. D. A. Boas, A. M. Dale, and M. A. Franceschini, "Diffuse optical imaging of brain
637 activation: approaches to optimizing image sensitivity, resolution, and accuracy,"
638 *NeuroImage* **23 Suppl 1**(S275-288 (2004).
- 639 59. M. Clancy et al., "Monitoring the Injured Brain - Registered, patient specific atlas models
640 to improve accuracy of recovered brain saturation values," in *Diffuse Optical Imaging V*
641 H. Dehghani, and P. Taroni, Eds., Spie-Int Soc Optical Engineering, Bellingham (2015).
- 642 60. M. Clancy et al., "Improving the quantitative accuracy of cerebral oxygen saturation in
643 monitoring the injured brain using atlas based Near Infrared Spectroscopy models,"
644 *Journal of Biophotonics* **9**(8), 812-826 (2016).
- 645 61. M. Clancy, "Application and development of high-density functional near infrared
646 spectroscopy for traumatic brain injury," in *School of Chemistry*, University of
647 Birmingham, University of Birmingham UBIRA E THESES (2017).
- 648 62. L. Gagnon et al., "Short separation channel location impacts the performance of short
649 channel regression in NIRS," *NeuroImage* **59**(3), 2518-2528 (2012).
- 650 63. F. Scholkmann et al., "Absolute Values of Optical Properties (mu_a, mu_s, mu_{eff} and DPF)
651 of Human Head Tissue: Dependence on Head Region and Individual," *Adv Exp Med Biol*
652 **1072**(325-330 (2018).
- 653 64. J. P. Coles et al., "Effect of hyperventilation on cerebral blood flow in traumatic head
654 injury: clinical relevance and monitoring correlates," *Crit Care Med* **30**(9), 1950-1959
655 (2002).
- 656 65. W. B. Baker et al., "Continuous non-invasive optical monitoring of cerebral blood flow
657 and oxidative metabolism after acute brain injury," *Journal of Cerebral Blood Flow &*
658 *Metabolism* **39**(8), 1469-1485 (2019).
- 659 66. G. A. Perkins, A. T. Eggebrecht, and H. Dehghani, "Quantitative evaluation of frequency
660 domain measurements in high density diffuse optical tomography," *Journal of biomedical*
661 *optics* **26**(5), (2021).
- 662 67. G. A. Perkins, A. Eggebrecht, and H. Dehghani, "Multi-Modulated frequency domain high
663 densitydiffuse optical tomography," *Biomed. Opt. Express* **13**((2022).

664 **10 Author Biographies and Photographs**

665 **Mario Forcione** is a specialty registrar in the University of Milan-Bicocca in Anaesthesia and
666 Intensive Care and has an honorary contract with the University Hospital Birmingham NHS
667 Foundation Trust. In 2022, he received his Ph.D. from the University of Birmingham in clinical
668 applications of diffuse optics to brain trauma. In 2015, he received his BS in Medicine from La

669 Sapienza, University of Rome. His work relates to the application of diffuse optics to the care of
670 mild, moderate, and severe traumatic brain injury patients.

671
672 **Antonio Maria Chiarelli** is an Associate Professor in Applied Physics at the University of Chieti-
673 Pescara. He received his BS and MS in Physics Engineering at the Polytechnic of Milan and his
674 PhD in Neuroimaging at the University of Chieti. He worked as a Post Doctoral researcher at
675 Beckman Institute, IL, USA. His research focuses on neuroimaging with Magnetic Resonance
676 Imaging and Diffuse Optics for studying the human brain function and physiology.

677
678 **David Perpetuini** is an Assistant Professor in Biomedical Engineering at the University of Chieti-
679 Pescara. He graduated in Biomedical Engineering from the Marche Polytechnic University of
680 Ancona and he obtained his PhD in Neuroimaging at the University of Chieti. His research mainly
681 concerns the study of brain and autonomic nervous system activity through functional near-
682 infrared spectroscopy (fNIRS) and thermal imaging.

683
684 **Guy Perkins** completed his BSc in Physics in 2018 and then his PhD in 2024, both at the
685 University of Birmingham. During his PhD, he investigated the development of FD-fNIRS and
686 DOT for clinical applications. He is now a post-doctoral research fellow at the University of
687 Padova, investigating preterm infants' brain health in response to hyper/hypo glycemia.

688
689 **Andrew Stevens** is a Specialist Registrar in Neurosurgery, and academic researcher at the
690 University of Birmingham in the fields of traumatic brain and spinal cord injury. He graduated
691 from the University of Birmingham in 2016 with a Bachelors in Medicine and Surgery (MBChB),

692 following completion of a BMedSc in 2013. He is currently completing a PhD in neurotrauma

693 alongside his clinical practice.

694

695 Biographies and photographs for the other authors are not available.