

1 **Title**

2 Immunochromatographic test for the detection of SARS-CoV-2 in saliva

3

4 **Authors**

5 Katsuhito Kashiwagi¹, Yoshikazu Ishii², Kotaro Aoki², Shintaro Yagi^{3*}, Tadashi Maeda¹,

6 Taito Miyazaki¹, Sadako Yoshizawa^{1,2}, Katsumi Aoyagi³, Kazuhiro Tateda²

7

8 **Affiliations**

9 1) Department of General Medicine and Emergency Care, Toho University Omori Medical
10 Center, Tokyo, Japan. 2) Department of Microbiology and Infectious Diseases, Toho
11 University School of Medicine, Tokyo, Japan. 3) Fujirebio Inc., Tokyo, Japan.

12

13 **Keywords**

14 COVID-19, SARS-CoV2, POCT, Antigen test, immunochromatography.

15

16 **Running title**

17 Evaluation of the rapid SARS-CoV-2 N antigen test

18

19 **Corresponding Author:**

20 Shintaro Yagi, Ph.D.

21 Research and Development Division

22 Fujirebio Inc.

23 51 Komiya-machi, Hachioji, Tokyo, Japan

24 TEL +81-42-645-4755

25 e-mail: shintaro.yagi@miraca.com

26

1

2 **Abstract**

3 We evaluated the rapid immunochromatographic test for SARS-CoV-2 antigen
4 detection using 16 saliva specimens collected from 6 COVID-19 hospitalized patients,
5 and detected N-antigen in 4 of 7 RT-PCR positive specimens. The POCT antigen test
6 using saliva is highly considered to be a game-changer for COVID-19 diagnosis.

7

8 **Introduction**

9 COVID-19 caused by SARS-CoV-2 infection is diagnosed by nucleic acid amplification
10 test, as reverse-transcription polymerase chain reaction (RT-PCR), for viral genome RNA
11 in nasopharyngeal swab specimens. The collection procedure of nasopharyngeal
12 specimens poses a risk of secondary infection to health-care workers. The risk could be
13 reduced if saliva specimens could be used for diagnosis of COVID-19. Recently, several
14 reports indicated that SARS-CoV-2 was detected in saliva specimens of equally or higher
15 sensitivity than the nasopharyngeal specimens[1–3].

16 We recently developed a rapid antigen detection kit, ESPLINE[®] SARS-CoV-2 test, for
17 detection of the viral N antigens in nasopharyngeal swabs (manuscript in preparation).
18 The ESPLINE test is an enzyme-immunoassay based on immunochromatographic
19 technology using monoclonal antibodies specific to SARS-CoV and CoV-2 N antigens.
20 In this report, we evaluated the saliva specimens for diagnosis of COVID-19 by RT-PCR
21 and the point of care testing (POCT) antigen test.

22

23 **Materials and methods**

24 Sets of saliva and nasopharyngeal specimens were collected within several days from 6
25 COVID-19 patients hospitalized in Omori Hospital, who had been diagnosed by RT-PCR
26 test. Informed consent was obtained from all participants in the study. The study
27 protocol was approved by the Ethics Committee of Faculty of Medicine, Toho University
28 (No. A20028_A20020_A20014_A19099). Saliva was collected by passive drool method
29 according to previous report[4]. Ct values were obtained by RT-PCR test using N2 probe
30 according to the manual provided by National Institute of Infectious Diseases[5].

1 SARS-CoV-2 antigen was assessed using ESPLINE kit according to the manufacturer's
2 instruction (Fujirebio Inc., Tokyo). The procedure is briefly described as follows: specimen
3 collected by the Swab in the kit was treated with the sample treatment solution.
4 Approximately 20 μ L of the treated sample was applied to the immunochromatography
5 cassette, and was developed for 30 minutes. After incubation, test and reference lines
6 were visually assessed.

7

8 **Results**

9 SARS-CoV-2 RNA was detected in 13 of 16 (81.3%). On the other hand, RT-PCR detected
10 RNA in only 7 of 16 saliva specimens (43.8%). No significant difference in RNA copies of
11 nasopharyngeal specimens between positive and negative RT-PCR results of saliva
12 specimens was found. When 9 of the 13 RT-PCR positive nasopharyngeal specimens
13 were examined, SARS-CoV-2 antigen in 7 specimens (77.8% in sensitivity) was detected
14 by the ESPLINE test. The antigen was detected in 4 of 7 RT-PCR positive saliva specimens
15 (47.8%)(Figure 1). As shown in previous studies of SARS-CoV-2 antigen tests, sensitivity
16 of the antigen seems to be correlated with RNA copies concentrations in the specimens⁵
17 (Manuscript in preparation).

18

19 **Discussion**

20 Sensitivity of SARS-CoV-2 RT-PCR in saliva specimens was lower than those reported
21 previously[1,2,4,6–8]. Among several procedures of saliva collection which have been
22 applied, we chose the passive drool method to reduce the burden of the patients and
23 the secondary risk to healthcare workers. However, collection procedure should be
24 examined for further studies. In addition, symptoms, periods after the onset of
25 symptoms and infection could affect the sensitivity of the RNA detection in saliva. These
26 factors would also affect the sensitivity of antigen detection using the ESPLINE kit.

27 This study was carried out using specimens collected from hospitalized patients. The
28 mechanism and hazard of secondary infection caused by asymptomatic carriers and
29 super spreaders through biological materials from them have been discussed[9]. To
30 assess the secondary infection hazards, it would be important to measure infectious

1 virus particles in saliva and nasal fluid. Low prevalence of RNA in saliva specimens from
2 the hospitalized patients might provide clues to assessment of the secondary infection
3 hazards. To address this issue, mass-screening using POCT tests would be a powerful
4 tool.

5 This is the first report of SARS-CoV-2 antigen detection in saliva using a POCT antigen
6 test. The combination of saliva specimen and rapid antigen detection test could reduce
7 burden of medical settings by decreasing not only the risk of secondary infection, but
8 also time of diagnosis and cost of the specialized expensive equipment. It could be
9 useful for COVID-19 screening to prevent COVID-19 epidemic by identifying persons
10 who will spread COVID-19 through saliva.

11

12 **Funding**

13 A part of this research was supported by the Japan Agency for Medical Research and
14 Development under Grant Number JP19fk0108113.

15 **Conflict of Interest**

16 SY and KaA are an employee, and the director of Fujirebio, Inc., respectively.

17 **Acknowledgments**

18

19 **Author contributions**

20 YI, SY, KaA and KT contributed to design this study. KK, TMa, TM, and SaY contributed to
21 collect specimens, KoA and YI contributed to conduct and perform RT-PCR and antigen
22 tests, YI, KoA, SY and KaA contributed to analyze data, and SY, YI, KoA and KT
23 contributed to prepare this manuscript.

24

25 **Reference**

26

27 1. To KK-W, Tsang OT-Y, Leung W-S, et al. Temporal profiles of viral load in posterior
28 oropharyngeal saliva samples and serum antibody responses during infection by
29 SARS-CoV-2: an observational cohort study. *Lancet Infect Dis* **2020**;

- 1 2. Wyllie AL, Fournier J, Casanovas-Massana A, et al. Saliva is more sensitive for
2 SARS-CoV-2 detection in COVID-19 patients than nasopharyngeal swabs. Medrxiv
3 **2020**; :2020.04.16.20067835.
- 4 3. Pasomsub E, Watcharananan SP, Boonyawat K, et al. Saliva Sample as a Non-Invasive
5 Specimen for the Diagnosis of Coronavirus Disease-2019 (COVID-19): a Cross-Sectional
6 Study. Medrxiv **2020**; :2020.04.17.20070045.
- 7 4. To KK-W, Tsang OT, Yip CC-Y, et al. Consistent detection of 2019 novel coronavirus in
8 saliva. Clin Infect Dis **2020**;
- 9 5. Shinomiya KSNNSMMTTKKSH. Manual for the Detection of Pathogen 2019 - nCoV
10 Ver.2.6. 2020.
- 11 6. Williams E, Bond K, Zhang B, Putland M, Williamson DA. Saliva as a non-invasive
12 specimen for detection of SARS-CoV-2. J Clin Microbiol **2020**;
- 13 7. Zhong F, Liang Y, Xu J, et al. Continuously High Detection Sensitivity of Saliva, Viral
14 Shedding in Salivary Glands and High Viral Load in Patients with COVID-19. Ssrn
15 Electron J **2020**;
- 16 8. Chen L, Zhao J, Peng J, et al. Detection of 2019-nCoV in Saliva and Characterization of
17 Oral Symptoms in COVID-19 Patients. Ssrn Electron J **2020**;
- 18 9. Xu R, Cui B, Duan X, Zhang P, Zhou X, Yuan Q. Saliva: potential diagnostic value and
19 transmission of 2019-nCoV. Int J Oral Sci **2020**; 12:11.

20

21 **Legend of Figure 1.**

22 Detection of SARS-CoV-2 N antigen in nasopharyngeal swab and saliva specimens.
23 Closed circles indicate N antigen positive specimens. Among 16 nasopharyngeal swabs,
24 4 samples without Ag assay data are not plotted.

25

Figure 1

