

# Supplementary table 1

<u>Cross-reactivity Panel</u>	N501Y	SC2-S	S-DEL	SC2-N
Species				
Influenza A H1N1 pdm09 ( <i>A/California/60/2008</i> )	Negative	Negative	Negative	Negative
Influenza B Victoria ( <i>B/Brisbane/60/2008</i> )	Negative	Negative	Negative	Negative
RSV A	Negative	Negative	Negative	Negative
MERS Coronavirus	Negative	Negative	Negative	Negative
hCoV - OC43	Negative	Negative	Negative	Negative
Human Metapneumovirus (Subtype A)	Negative	Negative	Negative	Negative
Parainfluenzavirus 2	Negative	Negative	Negative	Negative
Parainfluenzavirus 3	Negative	Negative	Negative	Negative
Rhinovirus	Negative	Negative	Negative	Negative
hCoV - HKU1 (clinical sample)	Negative	Negative	Negative	Negative
hCoV - NL63 (clinical sample)	Negative	Negative	Negative	Negative
Adenovirus (clinical sample)	Negative	Negative	Negative	Negative
Boca-virus (clinical sample)	Negative	Negative	Negative	Negative
Mycoplasma pneumoniae (clinical sample)	Negative	Negative	Negative	Negative
Pneumocystis jirovecii (clinical sample)	Negative	Negative	Negative	Negative

## Supplementary table 2

SARS-CoV-2 "wild-type" (Accuplex Verification Panel)				
Step	cp/ml	SC2-S: pos/rep	SC2-N: pos/rep	Overall: pos/rep
1	500.00	8/8	8/8	8/8
2	250.00	8/8	8/8	8/8
3	125.00	8/8	8/8	8/8
4	62.50	8/8	8/8	8/8
5	31.25	8/8	8/8	8/8
6	15.63	8/8	8/8	8/8
7	7.81	6/8	7/8	8/8
8	3.91	4/8	5/8	6/8
9	1.95	4/8	2/8	5/8

SARS-CoV-2 B.1.1.7 lineage (clinical sample)					
Step	cp/ml	N501Y: pos/rep	S-Del: pos/rep	SC2-N: pos/rep	Overall: pos/rep
1	5561.00	8/8	8/8	8/8	8/8
2	2780.50	8/8	8/8	8/8	8/8
3	1390.25	8/8	8/8	8/8	8/8
4	695.13	8/8	8/8	8/8	8/8
5	347.56	8/8	8/8	8/8	8/8
6	173.78	8/8	8/8	8/8	8/8
7	86.89	2/8	7/8	8/8	8/8
8	43.45	1/8	7/8	8/8	8/8
9	21.72	0/8	2/8	7/8	7/8
10	10.86	0/8	1/8	6/8	6/8