

Pool raw data


B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T	U	V	W	X	Y	Z	AA	AB	AC	AD	AE	AF	AG	AH	AI	AJ	AK	AL	AM	AN	AO	AP		
Pool name	Original test date	Plate number	Sample	Claw	Cl ₂	CL	CL	Result	cut-off N	cut-off 9990	Original test date	Plate number	Sample	Claw	Cl ₂	CL	CL	Result	cut-off N	cut-off 9990	Original test date	Plate number	Sample	Claw	Cl ₂	CL	CL	Result	cut-off N	cut-off 9990	Test date	Plate number	Sample	Claw	Cl ₂	CL	CL	Result	cut-off N	cut-off 9990		
1N				29.40	40.66	N/A	N/A	Negative	35.59	34.67				28.76	37.60	N/A	N/A	Negative	35.81	35.08				28.98	N/A	N/A	N/A	Negative	35.81	35.08				29.40	37.25	N/A	N/A	Negative	36.08	35.95		
2N				29.02	40.24	37.58	N/A	Negative	35.59	34.67				29.04	38.01	42.29	39.06	Negative	35.81	35.08				30.07	N/A	N/A	43.27	44.51	Negative	35.81	35.08				29.66	N/A	N/A	42.87	42.20	Negative	36.08	35.95
3N				29.13	N/A	N/A	N/A	Negative	35.59	34.67				31.13	N/A	N/A	N/A	Negative	35.81	35.08				29.72	N/A	N/A	N/A	Negative	35.81	35.08				29.64	37.33	N/A	N/A	36.76	Negative	36.08	35.95	
4N				31.06	N/A	N/A	N/A	Negative	35.59	34.67				30.86	N/A	N/A	42.50	N/A	Negative	35.81	35.08				29.47	41.01	N/A	N/A	Negative	35.81	35.08				30.95	N/A	N/A	44.93	44.80	Negative	36.08	35.95
5N				30.59	N/A	N/A	N/A	Negative	35.59	34.67				30.76	N/A	N/A	N/A	38.68	Negative	35.81	35.08				29.38	N/A	N/A	N/A	Negative	35.81	35.08				29.65	37.90	N/A	N/A	N/A	Negative	36.08	35.95
6N				28.84	N/A	N/A	N/A	Negative	35.59	34.67				31.01	N/A	N/A	N/A	N/A	Negative	35.81	35.08				29.99	41.86	N/A	N/A	Negative	35.81	35.08				30.67	N/A	N/A	N/A	Negative	36.08	35.95	
7N				30.09	N/A	N/A	N/A	Negative	35.59	34.67				29.25	N/A	N/A	N/A	N/A	Negative	35.81	35.08				29.25	N/A	N/A	N/A	Negative	35.81	35.08				31.04	37.52	N/A	N/A	N/A	Negative	36.08	35.95
8N				29.50	N/A	N/A	N/A	Negative	35.59	34.67				30.12	N/A	N/A	N/A	N/A	Negative	35.81	35.08				29.48	N/A	N/A	N/A	Negative	35.81	35.08				29.84	36.22	N/A	N/A	38.27	Negative	36.08	35.95
9N				30.05	N/A	N/A	N/A	Negative	35.59	34.67				31.92	N/A	N/A	N/A	N/A	Negative	35.81	35.08				29.22	N/A	N/A	N/A	Negative	35.81	35.08				29.79	N/A	N/A	N/A	Negative	36.10	34.84	
10N				29.97	N/A	N/A	N/A	Negative	35.59	34.67				31.13	N/A	N/A	N/A	N/A	Negative	35.81	35.08				29.72	N/A	N/A	N/A	Negative	35.59	34.67				31.14	N/A	N/A	N/A	Negative	36.10	34.84	
11N				29.37	N/A	N/A	43.88	Negative	35.60	35.06				30.20	N/A	N/A	N/A	N/A	Negative	35.60	35.06				31.13	N/A	N/A	N/A	Negative	35.60	35.06				29.83	N/A	N/A	N/A	40.28	Negative	36.15	35.56
12N				29.96	N/A	N/A	N/A	Negative	35.60	35.06				30.01	40.05	N/A	N/A	N/A	Negative	35.60	35.06				29.75	N/A	N/A	N/A	Negative	35.60	35.06				29.48	N/A	N/A	N/A	Negative	36.15	35.56	
13N				31.63	N/A	N/A	N/A	Negative	35.60	35.06				30.16	N/A	N/A	N/A	N/A	Negative	35.60	35.06				30.14	N/A	N/A	N/A	Negative	35.60	35.06				30.46	N/A	N/A	N/A	Negative	36.15	35.56	
14N				29.81	N/A	N/A	N/A	Negative	35.60	35.06				30.22	N/A	N/A	N/A	N/A	Negative	35.60	35.06				30.18	44.99	N/A	N/A	Negative	35.60	35.06				30.11	N/A	N/A	N/A	Negative	36.15	35.56	
15N				29.87	N/A	N/A	N/A	Negative	35.60	35.06				29.46	N/A	N/A	37.69	N/A	Negative	35.60	35.06				28.86	N/A	N/A	40.91	Negative	35.60	35.06				29.32	N/A	N/A	N/A	Negative	36.15	35.56	
16N				29.23	N/A	N/A	N/A	Negative	35.60	35.06				26.20	N/A	N/A	43.98	N/A	Negative	35.60	35.06				30.26	N/A	N/A	N/A	Negative	35.60	35.06				29.88	N/A	N/A	N/A	Negative	36.15	35.56	
17N				29.97	N/A	N/A	N/A	Negative	35.60	35.06				30.48	36.31	N/A	N/A	N/A	Negative	35.60	35.06				30.37	N/A	N/A	40.47	Negative	35.60	35.06				29.78	N/A	N/A	N/A	Negative	36.15	35.56	
18N				30.02	N/A	N/A	N/A	Negative	35.60	35.06				29.88	N/A	N/A	N/A	N/A	Negative	35.60	35.06				29.55	N/A	N/A	N/A	Negative	35.60	35.06				29.60	N/A	N/A	N/A	Negative	36.15	35.56	
19N				29.86	N/A	N/A	N/A	Negative	35.60	35.06				29.63	N/A	N/A	N/A	N/A	Negative	35.60	35.06				30.03	N/A	N/A	N/A	Negative	35.52	34.96				30.03	N/A	N/A	N/A	Negative	36.15	35.56	
20N				29.95	N/A	N/A	N/A	Negative	35.60	35.06				30.24	38.69	N/A	N/A	N/A	Negative	35.60	35.06				32.18	N/A	N/A	N/A	Negative	35.52	34.96				30.12	N/A	N/A	N/A	Negative	36.15	35.56	
1P				29.62	13.29	14.94	15.31	Positive	35.87	36.67				29.20	N/A	N/A	N/A	N/A	Negative	35.81	35.08				34.10	N/A	N/A	N/A	Negative	35.81	35.08				29.66	15.21	16.69	16.89	Positive	36.08	35.95	
2P				30.67	17.24	17.69	17.77	Positive	35.87	36.62				28.06	N/A	N/A	N/A	N/A	Negative	35.81	35.08				33.04	N/A	N/A	N/A	Negative	35.81	35.08				29.63	18.22	18.52	18.65	Positive	36.08	35.95	
3P				30.61	18.55	20.04	20.17	Positive	35.87	36.62				29.21	N/A	N/A	44.93	N/A	Negative	35.81	35.08				29.10	N/A	N/A	N/A	Negative	35.81	35.08				31.34	21.62	22.81	23.21	Positive	36.08	35.95	
4P				31.02	22.28	22.27	22.07	Positive	36.09	35.73				29.94	N/A	N/A	N/A	N/A	Negative	35.81	35.08				29.38	N/A	N/A	N/A	Negative	35.81	35.08				31.78	23.63	23.67	23.57	Positive	36.08	35.95	
5P				31.22	17.15	18.22	18.54	Positive	36.09	35.73				28.76	37.60	N/A	N/A	N/A	Negative	35.81	35.08				28.98	N/A	N/A	N/A	Negative	35.81	35.08				30.01	18.84	19.88	20.23	Positive	36.08	35.95	
6P				29.15	17.18	18.54	18.88	Positive	36.09	35.73				29.04	38.01	42.29	39.06	Negative	35.81	35.08				30.07	N/A	43.27	44.51	Negative	35.81	35.08				29.64	18.71	19.63	20.10	Positive	36.08	35.95		
7P				29.90	17.05	18.17	18.53	Positive	35.97	36.32				30.11	N/A	N/A	N/A	N/A	Negative	35.81	35.08				28.95	N/A	N/A	N/A	Negative	35.81	35.08				30.74	18.52	20.06	20.94	Positive	36.08	35.95	
8P				32.00	22.00	23.00	23.30	Positive	35.47	36.52				29.05	N/A	N/A	41.95	N/A	Negative	35.81	35.08				28.93	N/A	N/A	44.70	Negative	35.81	35.08				30.63	23.69	24.80	25.28	Positive	36.08	35.95	
9P				29.13	16.78	18.44	19.05	Positive	36.14	36.22				29.36	N/A	N/A	N/A	N/A	Negative	35.81	35.08				30.07	43.32	N/A	N/A	Negative	35.81	35.08				29.30	18.04	19.19	20.01	Positive	36.08	35.95	
10P				29.02	24.75	25.29	25.51	Positive	35.53	35.84				30.27	N/A	N/A	N/A	N/A	Negative	35.81	35.08				29.38	N/A	N/A	38.52	Negative	35.81	35.08				29.36	24.88	25.64	26.38	Positive	36.08	35.95	
11P				30.89	15.28	17.27	17.78	Positive	35.88	34.58				31.13	N/A	N/A	N/A	N/A	Negative	35.81	35.08				29.72	N/A	N/A	N/A	Negative	35.81	35.08				30.88	16.00	17.49	18.67	Positive	36.08	35.95	
12P				30.88	17.31	18.76	19.30	Positive	35.40	35.38				31.13	N/A	N/A	41.61	N/A	Negative	35.81	35.08				28.55	N/A	40.18	41.41	Negative	35.81	35.08				31.38	17.91	19.03	19.69	Positive	36.08	35.95	
13P				27.15	22.54	22.26	21.56	Positive	35.40	35.38				30.86	N/A	42.50	N/A	N/A	Negative	35.81	35.08				29.47	41.01	N/A	N/A	Negative	35.81	35.08				28.31	23.54	23.50	23.32	Positive	36.08	35.95	
14P				31.85	15.86	17.44	17.86	Positive	35.40	35.38				30.76	N/A	N/A	38.68	N/A	Negative	35.81	35.08				29.38	N/A	N/A	N/A	Negative	35.81	35.08				30.42	16.65	17.69	17.86	Positive	36.08	35.95	
15P				30.25	15.27	16.31	16.54	Positive	35.40	35.38				31.01	N/A	N/A	N/A	N/A	Negative	35.81	35.08				28.99	41.86	N/A	N/A	Negative	35.81	35.08				30.23	17.07	18.10	18.75	Positive	36.08	35.95	
16P				31.57	20.15	21.46	22.22	Positive	35.40	35.38				31.47	N/A	N/A	N/A	N/A	Negative	35.81	35.08				29.25	N/A	N/A	N/A	Negative	35.81	35.08				31.58	20.74	22.05	22.97	Positive	36.08	35.95	
17P				29.94	20.48	21.76	22.11	Positive	35.40	35.38				30.12	N/A	N/A	N/A	N/A	Negative	35.81	35.08				30.17	21.70	22.80	23.48	Positive	35.81	35.08				30.17	21.70	22.80	23.48	Positive	36.08	35.95	
18P				30.03	17.27	18.81	19.38	Positive	35.40	35.38				31.92	N/A	N/A	N/A	N/A	Negative	35.81	35.08				29.22	N/A	N/A	N/A	Negative	35.81	35.08				30.09	17.77	19.30	20.21	Positive	36.08	35.95	
19P				29.23	16.83	17.94	17.89	Positive	35.40	35.38				31.07	N/A	N/A	N/A	N/A	Negative	35.81	35.08				29.46	N/A	N/A	N/A	Negative	35.59												

Positives

A	B	C	D	E	F	G	H	I	J	K	L	M
	Ct _{RPP}			Ct _N			Ct _E			Ct _S		
Pool name	Individual	Pool	Ct difference	Individual	Pool	Ct difference	Individual	Pool	Ct difference	Individual	Pool	Ct difference
1P	29.62	29.66	0.04	13.39	15.21	1.81	14.94	16.09	1.14	15.31	16.89	1.58
2P	30.67	29.83	-0.83	17.24	18.22	0.98	17.99	18.52	0.53	17.77	18.65	0.88
3P	30.61	31.34	0.73	18.55	21.62	3.07	20.04	22.81	2.77	20.17	23.21	3.04
4P	31.02	31.78	0.76	22.28	23.63	1.35	22.27	23.67	1.40	22.07	23.57	1.50
5P	31.22	30.01	-1.21	17.15	18.84	1.69	18.22	19.88	1.66	18.54	20.23	1.69
6P	29.15	29.64	0.49	17.18	18.71	1.53	18.54	19.63	1.09	18.88	20.10	1.22
7P	29.99	30.74	0.75	17.05	18.52	1.47	19.17	20.06	0.89	19.53	20.94	1.42
8P	32.00	30.63	-1.37	22.00	23.69	1.69	23.00	24.80	1.80	23.30	25.28	1.98
9P	29.13	29.30	0.17	16.78	18.04	1.26	18.44	19.19	0.75	19.05	20.01	0.96
10P	29.02	29.36	0.34	24.75	24.88	0.13	25.29	25.64	0.35	25.51	26.38	0.87
11P	30.89	30.88	-0.01	15.28	16.00	0.71	17.27	17.49	0.22	17.78	18.67	0.89
12P	30.88	31.38	0.50	17.33	17.91	0.58	18.76	19.03	0.27	19.30	19.89	0.59
13P	27.15	28.31	1.16	22.54	23.54	1.00	22.26	23.50	1.24	21.56	23.32	1.76
14P	31.85	30.42	-1.43	15.96	16.05	0.09	17.44	17.09	-0.34	17.86	17.86	0.00
15P	30.25	30.23	-0.02	15.27	17.07	1.80	16.31	18.10	1.79	16.54	18.75	2.21
16P	31.57	31.58	0.01	20.15	20.74	0.59	21.46	22.05	0.59	22.22	22.97	0.75
17P	29.94	30.17	0.24	20.48	21.70	1.21	21.76	22.80	1.04	22.11	23.48	1.37
18P	30.03	30.09	0.06	17.27	17.77	0.50	18.81	19.30	0.49	19.38	20.21	0.83
19P	29.23	30.47	1.24	16.83	17.97	1.13	17.94	19.10	1.16	17.89	19.44	1.55
20P	30.28	30.77	0.49	19.54	21.28	1.73	20.68	22.08	1.40	20.97	22.53	1.56
21P	31.28	31.06	-0.23	16.16	18.31	2.15	17.94	20.25	2.31	18.29	20.70	2.42
22P	30.00	31.15	1.15	29.20	30.92	1.72	28.80	30.70	1.90	29.00	30.94	1.94

Regression

A	B	C	D	E	F	G
	Ct _N		Ct _E		Ct _S	
Pool name	Individual	Pool	Individual	Pool	Individual	Pool
1P	13.39	15.21	14.94	16.09	15.31	16.89
2P	17.24	18.22	17.99	18.52	17.77	18.65
3P	18.55	21.62	20.04	22.81	20.17	23.21
4P	22.28	23.63	22.27	23.67	22.07	23.57
5P	17.15	18.84	18.22	19.88	18.54	20.23
6P	17.18	18.71	18.54	19.63	18.88	20.10
7P	17.05	18.52	19.17	20.06	19.53	20.94
8P	22.00	23.69	23.00	24.80	23.30	25.28
9P	16.78	18.04	18.44	19.19	19.05	20.01
10P	24.75	24.88	25.29	25.64	25.51	26.38
11P	15.28	16.00	17.27	17.49	17.78	18.67
12P	17.33	17.91	18.76	19.03	19.30	19.89
13P	22.54	23.54	22.26	23.50	21.56	23.32
14P	15.96	16.05	17.44	17.09	17.86	17.86
15P	15.27	17.07	16.31	18.10	16.54	18.75
16P	20.15	20.74	21.46	22.05	22.22	22.97
17P	20.48	21.70	21.76	22.80	22.11	23.48
18P	17.27	17.77	18.81	19.30	19.38	20.21
19P	16.83	17.97	17.94	19.10	17.89	19.44
20P	19.54	21.28	20.68	22.08	20.97	22.53
21P	16.16	18.31	17.94	20.25	18.29	20.70
22P	29.20	30.92	28.80	30.70	29.00	30.94



	Pool			Individual		
	LOD	95% CI low	95% CI up	LOD	95% CI low	95% CI upper
N	37.29	36.91	37.67	36.17	35.79	36.56
E	37.29	36.91	37.67	35.62	35.25	35.98
S	37.29	36.91	37.67	35.59	35.22	35.96

For a pool which is at the limit of detection, the individual sample would have had a Ct around 36.17 for N2. This means that during the pooling procedure samples with Cts between 36.17 and 37.29 would have been detected using individual testing but will not be due to pooling.

