

Supplementary Table II

<i>N</i> gene	<i>S</i> gene	<i>E</i> gene	<i>RPP30</i> gene	Status	Result	Action
NEG	NEG	NEG	NEG	Invalid	N/A	Repeat test. If the repeat results remain invalid, a new sample would need to be collected.
NEG	NEG	NEG	POS	Valid	SARS-CoV-2 no detected	Report results.
Only one SARS-CoV-2 target gene is POS			POS	Valid	SARS-CoV-2 inconclusive	Repeat test. If the repeat result remains inconclusive, additional confirmation testing should be conducted if clinically indicated.
Two or more SARS-CoV-2 target gene are POS			POS	Valid	SARS-CoV-2 positive	Report results.