

Supplementary Table III

cut-off value for N: 37.29, 95% CI: [36.91, 37.67]							
Viral RNA copies/reaction	Replicate	C _T _N	N Conclusion	C _T _E	E Conclusion	C _T _S	S Conclusion
300 copies	1	31.36	Positive	31.45	Positive	31.39	Positive
300 copies	2	31.25	Positive	31.15	Positive	31.79	Positive
300 copies	3	31.02	Positive	31.23	Positive	31.59	Positive
300 copies	4	31.32	Positive	31.55	Positive	31.55	Positive
300 copies	5	31.53	Positive	31.25	Positive	31.41	Positive
300 copies	6	31.28	Positive	30.74	Positive	31.34	Positive
300 copies	7	31.13	Positive	30.97	Positive	31.35	Positive
300 copies	8	31.04	Positive	31.03	Positive	31.31	Positive
300 copies	9	31.19	Positive	31.31	Positive	31.43	Positive
300 copies	10	31.31	Positive	31.45	Positive	31.63	Positive
300 copies	11	31.44	Positive	31.09	Positive	30.98	Positive
300 copies	12	31	Positive	30.73	Positive	30.46	Positive
300 copies	13	30.92	Positive	30.75	Positive	31.1	Positive
300 copies	14	31.09	Positive	31.04	Positive	30.74	Positive
300 copies	15	31.38	Positive	30.94	Positive	31.09	Positive
300 copies	16	31.2	Positive	30.57	Positive	30.53	Positive
300 copies	17	31.15	Positive	30.94	Positive	31.03	Positive
300 copies	18	31.1	Positive	30.79	Positive	30.97	Positive
300 copies	19	31.47	Positive	31.04	Positive	31.06	Positive
300 copies	20	31.36	Positive	31.26	Positive	30.97	Positive
	SD	0.17		0.27		0.36	
	% Positive		20/20		20/20		20/20