

Supplementary Table IV

cut-off value for N: 37.7747, 95% CI: [37.04, 38.51]							
Viral RNA copies/reaction	Replicate	C <sub>T</sub> _N	N Conclusion	C <sub>T</sub> _E	E Conclusion	C <sub>T</sub> _S	S Conclusion
20 copies	1	35	Positive	33.6	Positive	33.11	Positive
20 copies	2	33.39	Positive	33.72	Positive	33.42	Positive
20 copies	3	33.92	Positive	33.4	Positive	33.07	Positive
20 copies	4	33.38	Positive	33.09	Positive	32.59	Positive
20 copies	5	33.87	Positive	33.6	Positive	34.2	Positive
20 copies	6	33.2	Positive	33.18	Positive	33.43	Positive
20 copies	7	33.21	Positive	33.47	Positive	32.95	Positive
20 copies	8	33.81	Positive	33.8	Positive	34.51	Positive
20 copies	9	34.14	Positive	33.69	Positive	33.28	Positive
20 copies	10	34.04	Positive	34.41	Positive	33.67	Positive
20 copies	11	33.75	Positive	32.96	Positive	33.25	Positive
20 copies	12	34.06	Positive	33.83	Positive	33.2	Positive
20 copies	13	33.67	Positive	33.63	Positive	33.18	Positive
20 copies	14	33.68	Positive	33	Positive	33.27	Positive
20 copies	15	34.35	Positive	33.74	Positive	33.69	Positive
20 copies	16	33.67	Positive	33.35	Positive	33.32	Positive
20 copies	17	33.85	Positive	33.85	Positive	33.02	Positive
20 copies	18	33.71	Positive	33.61	Positive	32.5	Positive
20 copies	19	35.33	Positive	33.44	Positive	32.94	Positive
20 copies	20	33.66	Positive	34.27	Positive	32.82	Positive
	<b>SD</b>	<b>0.5298607</b>		<b>0.3733293</b>		<b>0.4812363</b>	
	<b>% Positive</b>		<b>20/20</b>		<b>20/20</b>		<b>20/20</b>