

Supplementary Table V

cut-off value for N: 37.29, 95% CI: [36.91, 37.67]							
Viral RNA copies/reaction	Replicate	C _T _N	N Conclusion	C _T _E	E Conclusion	C _T _S	S Conclusion
1.3 copies	1	37.17	Negative	N/A	Negative	N/A	Negative
1.3 copies	2	N/A	Negative	37.14	Negative	N/A	Negative
1.3 copies	3	N/A	Negative	N/A	Negative	37.49	Negative
1.3 copies	4	N/A	Negative	N/A	Negative	40.63	Negative
1.3 copies	5	N/A	Negative	N/A	Negative	N/A	Negative
1.3 copies	6	37.64	Negative	37.14	Negative	N/A	Negative
1.3 copies	7	N/A	Negative	37.23	Negative	N/A	Negative
1.3 copies	8	N/A	Negative	N/A	Negative	N/A	Negative
1.3 copies	9	N/A	Negative	N/A	Negative	37.5	Negative
1.3 copies	10	N/A	Negative	37.29	Negative	N/A	Negative
1.3 copies	11	N/A	Negative	N/A	Negative	N/A	Negative
1.3 copies	12	37.07	Negative	N/A	Negative	N/A	Negative
1.3 copies	13	N/A	Negative	N/A	Negative	36.6	Negative
1.3 copies	14	N/A	Negative	N/A	Negative	38.38	Negative
1.3 copies	15	N/A	Negative	37.42	Negative	N/A	Negative
1.3 copies	16	N/A	Negative	N/A	Negative	N/A	Negative
1.3 copies	17	37.35	Negative	37.25	Negative	37.55	Negative
1.3 copies	18	N/A	Negative	N/A	Negative	N/A	Negative
1.3 copies	19	N/A	Negative	N/A	Negative	44.65	Negative
1.3 copies	20	N/A	Negative	42.59	Negative	N/A	Negative
	SD	0.25		2.02		2.8	
	% Positive		0/20		0/20		0/20