

## Towards adaptive deep brain stimulation: clinical and technical notes on a novel commercial device for chronic brain sensing

Yohann Thenaisie<sup>a,b,\*</sup>, Chiara Palmisano<sup>c,\*</sup>, Andrea Canessa<sup>c,d</sup>, Bart J. Keulen<sup>e,f</sup>, Philipp Capetian<sup>c</sup>, Mayte Castro Jiménez<sup>g</sup>, Julien F. Bally<sup>g</sup>, Elena Manferlotti<sup>c,h</sup>, Laura Beccaria<sup>c</sup>, Rodi Zutt<sup>e</sup>, Grégoire Courtine<sup>b,i,l</sup>, Jocelyne Bloch<sup>b,i</sup>, Niels A. van der Gaag<sup>m</sup>, Carel F. Hoffmann<sup>m</sup>, Eduardo Martin Moraud<sup>a,b,#</sup>, Ioannis U. Isaias<sup>c,#</sup> and M. Fiorella Contarino<sup>e,n,#,§</sup>

<sup>§</sup> Corresponding Author M.F. Contarino, Els Borst-Eilersplein 275-2545AA, The Hague, The Netherlands. *E-mail address:* m.f.contarino@lumc.nl

**Supplementary Table 1. Location of maximum beta peak power and best clinical contact on Parkinson's disease patients**

Subject	Side	Indefinite streaming at rest		Stimulation with best clinical outcome	
		Contact pairs with highest beta power	Recording <sup>a</sup>	Stimulation contact	Stimulation parameters
CH1	L	0-2	IS (69s)	1-	1.7mA, 130Hz, 60μs
	R	0-2		1-	1.6mA, 130Hz, 60μs
CH2	L	1-3	IS (128s)	0-	1.7mA, 130Hz, 60μs
	R	1-3		1-	1.6mA, 130Hz, 60μs
CH3	L	0-3	IS (121s)	3-	3.0mA, 130Hz, 60μs
	R	0-3		3-	1.9mA, 130Hz, 60μs
CH4	L	1-3	IS (113s)	2-3- IL <sup>b</sup>	2.0mA, 180Hz, 90μs
	R	No beta <sup>d</sup>		2-3- IL <sup>b</sup>	2.1mA, 180Hz, 90μs
CH6	L	1-2	IS (90s)	1-	1.4mA, 130Hz, 60μs
	R	1-3		1-	1.4mA, 130Hz, 60μs
CH7	L	0-3	IS (107s)	1-	4.7mA, 140Hz, 90μs
	R	0-3		3-	2.9mA, 140Hz, 90μs
CH8	L	0-3	IS (105s)	2-	1.0mA, 245Hz, 60μs
	R	0-2		3-	3.2mA, 245Hz, 60μs
CH9	L	1-2	IS (111s)	3-	2.0mA, 130Hz, 60μs
	R	1-3		2-	2.0mA, 130Hz, 60μs
NL1 <sup>c</sup>	L	1-3	Survey	1-	4.3mA, 130Hz, 60μs
	R	2-3		3-	1.2mA, 130Hz, 60μs
NL2	L	No beta <sup>d</sup>	Survey	0-1+	2.8mA, 185Hz, 60μs
	R	No beta <sup>d</sup>		1+2-3-	3.0mA, 185Hz, 60μs,
PW1 <sup>c</sup>	L	0-3	Survey	2+3-	5.0 mA, 180Hz, 60μs
	R	1-3		2-	2.6mA, 180Hz, 60μs

<sup>a</sup> Data from 21 s of Survey or from a variable duration of Survey Indefinite Streaming (IS) recordings.

<sup>b</sup> Interleaved stimulation (IL).

<sup>c</sup> Survey were recorded on four (NL1) or two (PW1) sessions. Only the recording exhibiting the clearest beta peaks was considered for analysis.

<sup>d</sup> In three STN (two patients), no clear beta peak could be identified in any of the contact pairs.