

Supplementary Data

Supplementary Table 1: SARS-CoV-2 RT-dPCR assay linearity study on Clarity Plus™ with three independent experimental runs in duplicates, using a series of expected N1 concentrations ranging from 0.02 to 30,000 copies/μl

Expected Concentration (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 1 ^b (Copies/μl)	Mean Measured Concentration 1 (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 2 ^b (Copies/μl)	Mean Measured Concentration 2 (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 3 ^b (Copies/μl)	Mean Measured Concentration 3 (Copies/μl)
30000	41002	41064	0.99849	20954.04	22272.25	40527	40604	0.99810	20218.76	21418.38	44561	44627	0.99852	21020.77	20045.35
	44962	44992	0.99933	23590.46		41011	41048	0.99910	22618.00		46409	46535	0.99729	19069.93	
20000	42681	42811	0.99696	18700.05	18924.60	39414	39502	0.99777	19699.26	18887.03	46383	46515	0.99716	18918.48	18498.95
	42281	42393	0.99736	19149.16		41084	41236	0.99631	18074.79		43850	44012	0.99632	18079.42	
10000	41225	42829	0.96255	10595.85	10502.90	44826	46930	0.95517	10015.54	10048.33	45598	48068	0.94861	9575.48	9935.72
	43934	45749	0.96033	10409.95		43744	45754	0.95607	10081.11		43792	45669	0.95890	10295.95	
2000	21044	50136	0.41974	1755.73	1774.03	20235	46456	0.43557	1844.98	1827.02	21075	46281	0.45537	1960.16	1888.77
	20350	47738	0.42629	1792.33		20274	47231	0.42925	1809.06		20340	47223	0.43072	1817.38	
200	2791	45392	0.06149	204.70	205.88	2758	46164	0.05974	198.72	198.63	2772	47031	0.05894	195.96	2204.38
	2900	46642	0.06218	207.07		2838	47545	0.05969	198.54		2967	46475	0.06384	212.81	
20	296	43103	0.00687	22.23	22.39	298	43256	0.00689	22.30	21.55	302	42674	0.00708	22.91	22.16
	304	43612	0.00697	22.56		250	38888	0.00642	20.80		285	43102	0.00661	21.40	
2	22	42285	0.00052	1.68	1.79	30	45023	0.00067	2.15	2.48	32	49529	0.00065	2.08	1.75
	27	45717	0.00059	1.91		39	44662	0.00087	2.82		22	50111	0.00044	1.42	
1	13	43257	0.00030	0.97	0.97	16	45871	0.00035	1.13	1.32	14	42618	0.00033	1.06	1.06
	14	46659	0.00030	0.97		19	40482	0.00047	1.51		16	48865	0.00033	1.06	
0.2	3	47065	0.00006	0.21	0.31	1	44159	0.00002	0.07	0.11	3	45494	0.00007	0.21	0.21
	6	47428	0.00013	0.41		2	42523	0.00005	0.15		3	47429	0.00006	0.20	
0.02	3	45334	0.00007	0.21	0.17	3	41621	0.00007	0.23	0.15	1	45595	0.00002	0.07	0.10
	2	46210	0.00004	0.14		1	44579	0.00002	0.07		2	49740	0.00004	0.13	

The results represent individual positive signals, total partition number, ratio, and measured concentration of two replicates within independent experimental runs.

^aRatio was calculated by the formula (Number of Positive Signals / Total Partition Number)

^bMeasured Concentration (Copies/μl) is calculated by the formula: $-\ln(1 - \text{Ratio}) / \text{Partition Volume} \times 1000$. Partition volume used is 0.31nL.

Supplementary Data

Supplementary Table 2: SARS-CoV-2 RT-dPCR assay linearity study on Clarity Plus™ with three independent experimental runs in duplicates, using a series of expected ORFlab concentrations ranging from 0.02 to 30,000 copies/μl

Expected Concentration (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 1 ^b (Copies/μl)	Mean Measured Concentration 1 (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 2 ^b (Copies/μl)	Mean Measured Concentration 2 (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 3 ^b (Copies/μl)	Mean Measured Concentration 3 (Copies/μl)
30000	41034	41064	0.99927	23295.77	20274.11	40577	40604	0.99934	23599.31	22866.27	44570	44627	0.99872	21493.69	21324.76
	44778	44992	0.99524	17252.46		41005	41048	0.99895	22133.22		46469	46535	0.99858	21155.82	
20000	42566	42811	0.99428	16655.78	17067.08	39334	39502	0.99575	17613.36	17527.18	46195	46515	0.99312	16061.96	16326.37
	42205	42393	0.99557	17478.38		41051	41236	0.99551	17441.00		43755	44012	0.99416	16590.78	
10000	40203	42829	0.93869	9005.66	9020.12	43757	46930	0.93239	8690.26	8711.51	44570	48068	0.92723	8452.98	8579.86
	42969	45749	0.93923	9034.58		42701	45754	0.93327	8732.76		42597	45669	0.93273	8706.74	
2000	20021	50136	0.39933	1644.24	1662.72	19707	46456	0.42421	1780.67	1696.34	18825	46281	0.40675	1684.34	1684.34
	19390	47738	0.40618	1681.20		18576	47231	0.39330	1612.01		19208	47223	0.40675	1684.33	
200	2419	45392	0.05329	176.66	175.54	2457	46164	0.05322	176.43	172.23	2191	47031	0.04659	153.89	158.37
	2455	46642	0.05263	174.42		2413	47545	0.05075	168.02		2288	46475	0.04923	162.85	
20	262	43103	0.00608	19.67	20.30	250	43256	0.00578	18.70	17.66	239	42674	0.00560	18.12	17.09
	282	43612	0.00647	20.93		200	38888	0.00514	16.63		214	43102	0.00497	16.06	
2	17	42285	0.00040	1.30	1.43	29	45023	0.00064	2.08	2.05	17	49529	0.00034	1.11	1.30
	22	45717	0.00048	1.55		28	44662	0.00063	2.02		23	50111	0.00046	1.48	
1	16	43257	0.00037	1.19	0.97	6	45871	0.00013	0.42	0.81	13	42618	0.00031	0.98	1.09
	11	46659	0.00024	0.76		15	40482	0.00037	1.20		18	48865	0.00037	1.19	
0.2	3	47065	0.00006	0.21	0.28	1	44159	0.00002	0.07	0.19	2	45494	0.00004	0.14	0.58
	5	47428	0.00011	0.34		4	42523	0.00009	0.30		15	47429	0.00032	1.02	
0.02	1	45334	0.00002	0.07	0.18	2	41621	0.00005	0.16	0.19	6	45595	0.00013	0.42	0.21
	4	46210	0.00009	0.28		3	44579	0.00007	0.22		0	49740	0.00000	0.00	

The results represent individual positive signals, total partition number, ratio, and measured concentration of two replicates within independent experimental runs.

^aRatio was calculated by the formula (Number of Positive Signals / Total Partition Number)

^bMeasured Concentration (Copies/μl) is calculated by the formula: $-\ln(1 - \text{Ratio}) / \text{Partition Volume} \times 1000$. Partition volume used is 0.31nL.

Supplementary Data

Supplementary Table 3: Limit of Quantification (LOQ) in SARS-CoV-2 RT-dPCR assay quantified using Clarity Plus™. Three independent runs in duplicates were performed with a series of expected N1 concentrations ranging from 0.125 to 1 Copies/μl

Expected Concentration (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 1 ^b (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 2 ^b (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 3 ^b (Copies/μl)
1	10	41267	0.00024	0.78	12	39938	0.00030	0.97	12	43297	0.00028	0.89
	16	44536	0.00036	1.16	14	41669	0.00034	1.08	15	47542	0.00032	1.02
0.5	6	41436	0.00014	0.47	7	45073	0.00016	0.50	7	43668	0.00016	0.52
	7	44060	0.00016	0.51	7	43235	0.00016	0.52	6	44058	0.00014	0.44
0.25	4	46397	0.00009	0.28	1	41806	0.00002	0.08	5	43130	0.00012	0.37
	3	42857	0.00007	0.23	2	43240	0.00005	0.15	1	44437	0.00002	0.07
0.125	3	42761	0.00007	0.23	3	42145	0.00007	0.23	3	44287	0.00007	0.22
	3	43267	0.00007	0.22	2	41699	0.00005	0.15	2	44112	0.00005	0.15

The results represent individual positive signals, total partition number, ratio, and measured concentration of two replicates within independent experimental runs.

^aRatio was calculated by the formula (Number of Positive Signals / Total Partition Number)

^bMeasured Concentration (Copies/μl) is calculated by the formula: $-\ln(1 - \text{Ratio}) / \text{Partition Volume} \times 1000$. Partition volume used is 0.31nL

Supplementary Data

Supplementary Table 4: Limit of Quantification (LOQ) in SARS-CoV-2 RT-dPCR assay quantified using Clarity Plus™. Three independent runs in duplicates were performed with a series of expected ORF1ab concentrations ranging from 0.125 to 1 Copies/μl

Expected Concentration (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 1 ^b (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 2 ^b (Copies/μl)	Number of Positive Partition	Total Partition Number	Ratio ^a	Measured Concentration 3 ^b (Copies/μl)
1	15	41267	0.00036	1.17	8	39938	0.00020	0.65	17	43297	0.00039	1.27
	10	44536	0.00022	0.72	15	41669	0.00036	1.16	5	47542	0.00011	0.34
0.5	5	41436	0.00012	0.39	6	45073	0.00013	0.43	8	43668	0.00018	0.59
	7	44060	0.00016	0.51	11	43235	0.00025	0.82	0	44058	0.00000	0.00
0.25	4	46397	0.00009	0.28	2	41806	0.00005	0.15	2	43130	0.00005	0.15
	8	42857	0.00019	0.60	2	43240	0.00005	0.15	5	44437	0.00011	0.36
0.125	2	42761	0.00005	0.15	2	42145	0.00005	0.15	4	44287	0.00009	0.29
	5	43267	0.00012	0.37	4	41699	0.00010	0.31	7	44112	0.00016	0.51

The results represent individual positive signals, total partition number, ratio, and measured concentration of two replicates within independent experimental runs.

^aRatio was calculated by the formula (Number of Positive Signals / Total Partition Number)

^bMeasured Concentration (Copies/μl) is calculated by the formula: $-\ln(1 - \text{Ratio}) / \text{Partition Volume} \times 1000$. Partition volume used is 0.31nL