

SUPPLEMENTAL TABLES**Supplemental Table 1.** Ct values in Roche cobas SARS-CoV-2 RNA test on standard dilutions before and after inactivation by beta-propiolactone.

Standard		A-series		B-series	
Viral state	Dilution	ORFa/b	E gene	ORFa/b	E gene
native	10	21.95	22.10	21.78	21.96
native	100	25.05	25.13	24.91	25.18
native	1000	27.87	28.20	28.09	28.32
inactivated	10	23.54	24.55	23.77	24.67
inactivated	100	26.93	27.87	26.95	27.86
inactivated	1000	30.15	30.99	30.04	30.95

Supplemental Table 2a. Ct values on WHO and BioQ SARS-CoV-2 standard dilution series in cobas PCR assay

	series A		series B	
WHO 20/146 IU/mL	ORF Ct1	E Ct2	ORF Ct1	E Ct2
100000	27.63	27.78	27.31	27.38
30000	29.16	29.26	29.43	29.61
10000	30.71	30.82	30.39	30.61
3000	32.20	32.61	33.07	33.05
1000	32.94	33.37	33.27	33.77
BioQ inactivated RNA copies/mL	ORF Ct1	E Ct2	ORF Ct1	E Ct2
33784	26.94	27.75	26.54	27.30
11249	28.74	29.60	28.78	29.68
3378	30.34	31.25	30.09	31.02
1125	31.75	32.75	31.92	32.56
338	33.17	34.03	33.26	34.21

Supplemental Table 2b. Potency of BioQ inactivated standard against WHO 20/146 standard in cobas SARS-CoV-2 PCR assay

Target	parameter	Value (95% CI)
ORF	delta Ct	2.10 (1.78-2.42)
	IU/copy	4.29 (3.44-5.36)
E	delta Ct	1.42 (1.13-1.72)
	IU/copy	2.68 (2.19-3.29)

Supplemental Table 3. Estimated assay conversion times relative to the most sensitive assay, i.e. cobas PCR (Supplemental table 3a) and relative to an assumed infectivity threshold of 1000 NAT detectable RNA copies/mL equivalent to 1 TCID₅₀ (Supplemental table 3b) during ramp up phase of viremia of B.1 (Wuhan) type SARS-CoV-2 virus and the B.1.617.2 (delta) variant assuming a 10-fold and 10,000 fold daily rise of viral load respectively.

Supplemental table 3a

assay	50% LOD copies/mL [^]	B.1 (Wuhan) type		B1.617.2 (delta) variant
		days	hours	hours
cobas PCR	1.8	0	0	0
Aptima TMA	6.6	0.56	13.4	3.3
cobas MP6 PCR	11.1	0.78	18.7	4.7
SAMBA	15.0	0.91	21.8	5.4
LAMP	22.8	1.09	26.2	6.6
1 TCID ₅₀ /mL	1000	2.74	65.9	16.5
Fluorecare antigen	50000	4.44	106.5	26.6
Panbio antigen	75000	4.61	110.7	27.7
Roche antigen	100000	4.74	113.7	28.4

[^]50% NAT LODs or ± antigen LODs on native standard before inactivation

Supplemental table 3b

assay	50% LOD copies/mL [^]	B.1 (Wuhan) type		B1.617.2 (delta) variant
		days	hours	hours
cobas PCR	1.8	-2.74	-65.9	-16.5
Aptima TMA	6.6	-2.19	-52.5	-13.1
cobas MP6 PCR	11.1	-1.97	-47.2	-11.8
SAMBA	15.0	-1.84	-44.1	-11.0
LAMP	22.8	-1.65	-39.6	-9.9
1 TCID ₅₀ /mL	1000	0.00	0.0	0.0
Fluorecare antigen	50000	1.69	40.6	10.2
Panbio antigen	75000	1.87	44.8	11.2
Roche antigen	100000	1.99	47.8	12.0

[^]50% NAT LODs or ± antigen LODs on native standard before inactivation