

Supplementary Material Captions

Table S1: SARS-CoV-2 Antigen Diagnostic Tests with Emergency Use Authorization.

List of diagnostic tests for the detection of SARS-CoV-2 with FDA Emergency Use Authorizations as of 31May22. Organized first by target antigen, then by company name in alphabetical order. A majority of the tests target the N-protein and are designed as lateral flow assays.

Table S2: SARS-CoV-2 Variants and Current SARS-CoV-2 Interagency Group (SIG) Classification.

List of SARS-CoV-2 Variants of Concern/Variants Being Monitored (both WHO label and PANGO lineage) with SIG classification as of 31May22. Includes location and time each variant was first detected.

Table S3: Cross-Reactivity of High-Priority Pathogens Commonly Found in the Respiratory Tract Flora.

High-priority pathogens that are commonly found in the respiratory tract flora were tested for cross-reactivity with the Q-LAAD assay for detection of SARS-CoV-2.

Table S4: Interference of High-Priority Pathogens Commonly Found in the Respiratory Tract Flora.

High-priority pathogens that are commonly found in the respiratory tract flora were tested for interference with the Q-LAAD assay for detection of SARS-CoV-2.

Table S5: Interference of Endogenous Substances.

Endogenous substances commonly present in clinical samples from the respiratory tract were tested for interference with the Q-LAAD assay for detection of SARS-CoV-2.

Table S1: SARS-CoV-2 Antigen Diagnostic Tests with Emergency Use Authorization as of 22Sep2022¹

Company	Name of Diagnostic Test	Attributes ²	Date EUA Issued	Antigen Targeted	Authorized Settings ³
Celltrion USA, Inc.	Sampinute COVID-19 Antigen MIA	Magnetic Force-assisted Electrochemical Sandwich IA (MESIA), ST	10/23/2020	S-protein RBD	H, M
	Celltrion DiaTrust COVID-19 Ag Home Test	LF, VR, OTC, SS, MT	10/21/2021		N-protein
	Celltrion DiaTrust COVID-19 Ag Rapid Test	LF, VR, SS, MT	04/16/2021	S-protein RBD	H, M, W
Abbott Diagnostics Scarborough, Inc.	BinaxNOW COVID-19 Ag Card	LF, VR, ST	08/26/2020	N-protein	H, M, W
	BinaxNOW COVID-19 Ag Card Home Test	LF, VR, OTC, TPS, SS, ST	12/16/2020		Home, H, M, W
	BinaxNOW COVID-19 Ag 2 Card	LF, VR, Non-prescription Testing, SS, ST	03/31/2021		H, M, W
	BinaxNOW COVID-19 Antigen Self Test	LF, VR, OTC, SS, ST	03/31/2021		Home, H, M, W
Access Bio, Inc.	CareStart COVID-19 Antigen Home Test	LF, VR, OTC, SS, ST	08/02/2021	Home, H, M, W	Home, H, M, W
	CareStart COVID-19 Antigen test	LF, VR, SS, ST	10/08/2020		H, M, W
ACON Laboratories, Inc	Flowflex COVID-19 Antigen Home Test	LF, VR, OTC, Screening, ST	10/04/2021	Home, H, M, W	Home, H, M, W
ANP Technologies, Inc	NIDS COVID-19 Antigen Rapid Test Kit	LF, VR, SS, ST	09/24/2021		H, M, W
Becton, Dickinson and Company (BD)	BD Veritor System for Rapid Detection of SARS-CoV-2	Chromatographic Digital IA, IR, SS, ST	07/02/2020		H, M, W
	BD Veritor At-Home COVID-19 Test	LF, Digital Read, OTC, SS, ST	8/24/2021	Home, H, M, W	Home, H, M, W
DiaSorin, Inc.	LIAISON SARS-CoV-2 Ag	CLIA, ST	03/26/2021		H, M
Ellume Limited	ellume.lab COVID Antigen Test	LF, Fluorescence, IR, ST	07/08/2021		H, M, W
Ellume Limited	Ellume COVID-19 Home Test	LF, Fluorescence, IR, OTC, SS, ST	12/15/2020	Home, H, M, W	Home, H, M, W
Qorvo Biotechnologies, LLC.	Omnia SARS-CoV-2 Antigen Test	Bulk Acoustic Wave Biosensor, IR, ST	04/13/2021		H, M
Genabio Diagnostics, Inc.	Genabio COVID-19 Rapid Self-Test Kit	LF, VR, OTC, SS, ST	07/08/2022		Home, H, M, W
GenBody Inc.	GenBody COVID-19 Ag	LF, VR, SS, ST	07/13/2021	H, M, W	H, M, W
iHealth Labs, Inc.	iHealth COVID-19 Antigen Rapid Test Pro	LF, VR, SS, ST	01/14/2022		H, M, W
	iHealth COVID-19 Antigen Rapid Test	LF, VR, OTC, SS, ST	11/05/2021		Home, H, M, W
InBios International Inc.	SCoV-2 Ag Detect Rapid Self-Test	LF, VR, OTC, SS, ST	11/22/2021	Home, H, M, W	Home, H, M, W
	SCoV-2 Ag Detect Rapid Test	LF, VR, SS, ST	05/06/2021		H, M, W
Luminostics, Inc.	Clip COVID Rapid Antigen Test	LF immunoluminescent assay, IR, ST	12/07/2020		H, M, W
LumiraDx UK Ltd.	LumiraDx SARS-CoV-2 Ag Test	Microfluidic Immunofluorescence Assay, IR, Screening, ST	08/18/2020	H, M, W	H, M, W
Maxim Biomedical, Inc.	MaximBio ClearDetect COVID-19 Antigen Home Test	LF, VR, OTC, SS, ST	01/19/2022		Home, H, M, W
Nano-Ditech Corp.	Nano-Check COVID-19 Antigen Test	LF, VR, SS, ST	12/06/2021		H, M, W
Oceanit Foundry LLC	ASSURE-100 Rapid COVID-19 Test	LF, VR, ST	2/28/2022	H, M, W	H, M, W
OraSure Technologies, Inc.	IntelliSwab COVID-19 Rapid Test Rx	LF, VR, PHT, ST	06/04/2021		Home, H, M, W
	IntelliSwab COVID-19 Rapid Test	LF, VR, OTC, SS, ST	06/04/2021		Home, H, M, W
	IntelliSwab COVID-19 Rapid Test Pro	LF, VR, SS, ST	06/04/2021		H, M, W
Ortho Clinical Diagnostics, Inc.	VITROS Immunodiagnostic Products SARS-CoV-2 Antigen Reagent Pack	Chemiluminescence IA, IR, ST	01/11/2021	H, M	H, M
OSANG LLC	OHC COVID-19 Antigen Self Test	LF, VR, OTC, Serial Testing, ST	04/06/2022		Home, H, M, W
PHASE Scientific International, Ltd.	INDICAID COVID-19 Rapid Antigen Test	LF, VR, SS, ST	07/28/2021		H, M, W
	INDICAID COVID-19 Rapid Antigen At-Home Test	LF, VR, OTC, SS, ST	3/16/2022	Home, H, M, W	Home, H, M, W
QIAGEN GmbH	QIAreach SARS-CoV-2 Antigen	Digital LF, Fluorescence, IR, ST	08/05/2021		H, M
Quidel Corporation	QuickVue SARS Antigen Test	LF, VR, SS, ST	12/18/2020		H, M, W
	QuickVue At-Home OTC COVID-19 Test	LF, VR, OTC, SS, ST	03/31/2021		Home, H, M, W
	Sofia SARS Antigen FIA	LF, Fluorescence, IR, SS, ST	05/08/2020		H, M, W
	QuickVue At-Home COVID-19 Test	LF, VR, PHT, ST	03/01/2021		Home, H, M, W
Salofa Oy	Sienna-Clarity COVID-19 Antigen Rapid Test Cassette	LF, VR, ST	05/20/2021	H, M, W	H, M, W
SD Biosensor, Inc.	Pilot COVID-19 At-Home Test	LF, VR, OTC, SS, ST	12/24/2021		Home, H, M, W
Siemens Healthcare Diagnostics, Inc.	Atellica IM SARS-CoV-2 Antigen	CLIA, Single Target	3/11/2022		H, M
	ADVIA Centaur SARS-CoV-2 Antigen	CLIA, Single Target	3/11/2022	H, M	H, M
Siemens Healthineers	CLINITEST Rapid COVID-19 Antigen Self-Test	LF, VR, OTC, SS, ST	12/29/2021		Home, H, M, W
Watmind USA	Speedy Swab Rapid COVID-19 Antigen Self-Test	LF, VR, OTC, SS, ST	07/08/2022		Home, H, M, W
Xiamen Boson Biotech Co., Ltd.	Rapid SARS-CoV-2 Antigen Test Card	LF, VR, OTC, Serial Testing, ST	04/06/2022	Home, H, M, W	Home, H, M, W
Xtrava Health	SPERA COVID-19 Ag Test	LF, VR, ST	10/12/2021		H, M, W
Becton, Dickinson and Company (BD)	BD Veritor System for Rapid Detection of SARS-CoV-2 & Flu A+B	Chromatographic Digital IA, IR, Multi-analyte, ST	03/24/2021	N-protein Influenza A Influenza B	H, M, W
Princeton BioMeditech Corp.	Status COVID-19/Flu A&B	LF, VR, Multi-analyte, ST	02/04/2021		H, M, W
Quidel Corporation	Sofia 2 Flu + SARS Antigen FIA	LF, Fluorescence, IR, Multi-Analyte, ST	10/02/2020		H, M, W

¹ U.S. Food and Drug Administration. In Vitro Diagnostics EUAs - Antigen Diagnostic Tests for SARS-CoV-2. <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/in-vitro-diagnostics-euas-antigen-diagnostic-tests-sars-cov-2>. Accessed September 22, 2022.

² Lateral Flow (LF), Visual Read (VR), Single Target (ST), Serial Screening (SS), Over-the-Counter Home Testing (OTC), Telehealth Proctor Supervised (TPS), Instrument Read (IR), Immunoassay (IA), Prescription Home Testing (PHT), Multiple Targets (MT)

³ H - Labs certified under Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet requirements to perform high complexity tests.

M - Labs certified under CLIA, 42 U.S.C. §263a, that meet requirements to perform moderate complexity tests.

W - Patient care settings operating under a CLIA Certificate of Waiver.

Table S2: SARS-CoV-2 Variants and Current SARS-CoV-2 Interagency Group (SIG) Classification¹

WHO Label	PANGO Lineage	Location and Time First Detected	SIG Classification (as of 22Sep2022)
Alpha	B.1.1.7 and Q lineages	UK, Late 2020	Variant Being Monitored
Beta	B.1.351 and descendant lineages	South Africa, Late 2020	
Gamma	P.1 and descendant lineages	Brazil, Late 2020	
Delta	B.1.617.2 and AY lineages	India, Late 2020	
Omicron	B.1.1.529, BA.1, BA1.1, BA.2, BA.3, BA.4, and BA.5 lineages	South Africa, Nov 2021	Variant of Concern

¹ SARS-CoV-2 Variant Classifications and Definitions. <https://www.cdc.gov/coronavirus/2019-ncov/variants/variant-classifications.html>. Accessed September 22, 2022.

Table S3: Cross-Reactivity of High-Priority Pathogens Commonly Found in the Respiratory Tract Flora

Organism	Strain	Source	Identifier	Concentration tested (CFU/mL and PFU/mL)	Number Detected (3 Replicates)	Cross-Reactivity
Human coronavirus	229E	Isolate	ATCC Cat# VR-740	10 ⁵	0/3	No
Human coronavirus	OC43	Isolate	ATCC Cat# VR-1558	10 ⁵	0/3	No
Rhinovirus	B632	Isolate	ATCC Cat# VR-1645	10 ⁵	0/3	No
<i>Haemophilus influenzae</i>	Rd [KW20]	Isolate	ATCC Cat# 51907	10 ⁶	0/3	No
<i>Streptococcus pneumoniae</i>	262 [CIP 104340] (Klein) Chester	Isolate	ATCC Cat# 49619	10 ⁶	0/3	No
<i>Streptococcus pyogenes</i>	Bruno [CIP 104226]	Isolate	ATCC Cat# 19615	10 ⁶	0/3	No
<i>Candida albicans</i>	CBS 562 [572, CCRC 20512, CECT 1002, DBVPG 6133, IFO 1385, IGC 3436, JCM 1542, NCYC 597, NRRL Y-12983]	Isolate	ATCC Cat# 18804	10 ⁶	0/3	No
Pooled human nasal swab	N/A	N/A	N/A	N/A	0/3	No
Adenovirus 5	Adenoid 75	Isolate	ATCC Cat# VR-1516	10 ⁵	0/3	No
Parainfluenza virus 3	ATCC-2011-5	Isolate	ATCC Cat# VR-1782	10 ⁵	0/3	No
Parainfluenza virus 1	C35	Isolate	ATCC Cat# VR-94	10 ⁵	0/3	No
Influenza A (H1N1)	A/WS/33	Isolate	ATCC Cat# VR-1520	10 ⁵	0/3	No
Influenza B	B/Florida/78/2015	Isolate	ATCC Cat# VR-1931	10 ⁵	0/3	No
Enterovirus	H	Isolate	ATCC Cat# VR-1432	10 ⁵	0/3	No
Respiratory syncytial virus	Long	Isolate	ATCC Cat# VR-26	10 ⁵	0/3	No
<i>Bordetella pertussis</i>	18323 [NCTC 10739]	Isolate	ATCC Cat# 9797	10 ⁶	0/3	No
<i>Mycoplasma pneumoniae</i>	Eaton Agent [NCTC 10119]	Isolate	ATCC Cat# 15531	10 ⁶	0/3	No
<i>Staphylococcus aureus</i>	NCTC 8532 [IAM 12544, R. Hugh 2605]	Isolate	ATCC Cat# 12600	10 ⁶	0/3	No
<i>Staphylococcus epidermidis</i>	FDA strain PCI 1200	Isolate	ATCC Cat# CRM-12228	10 ⁶	0/3	No
**Human coronavirus	NL63 (heat-inactivated)	Isolate	ZeptoMetrix Cat# 0810228CFHI	1.70 ⁵ TCID 50/mL	0/3	No
Parainfluenza virus 2	Greer	Isolate	ATCC Cat# VR-92	10 ⁵	0/3	No
Parainfluenza virus 4b	19503	Isolate	BEI Cat# 3238	10 ⁵	0/3	No
Influenza A H3N2	A/Hong Kong/8/68	Isolate	ATCC Cat# VR-1679	10 ⁵	0/3	No
**MERS-CoV	EMC/2012	Gamma-irradiated	BEI Cat# NR-50549	10 ⁵	0/3	No
*SARS-CoV	Urbani	Gamma-irradiated	BEI Cat# NR-9323	10 ⁵	0/3	No
Human Metapneumovirus	TN/83-1211	Isolate	BEI Cat# NR-22227	10 ⁵	0/3	No
<i>Chlamydiaceae, Chlamydia</i>	AR-39	Isolate	ATCC Cat# 53592	10 ⁶	0/3	No
<i>Legionella pneumophila</i>	Concord 3 [NCTC 11985]	Isolate	ATCC Cat# 33152	10 ⁶	0/3	No

Table S4: Interference of High-Priority Pathogens Commonly Found in the Respiratory Tract Flora

Organism (Identifier)	Pool	Concentration Tested (CFU/mL and PFU/mL)	Number Detected (3 Replicates)	Interference
<i>Mycoplasma pneumoniae</i> (ATCC 15531)	1	1.00E+06	3/3	No
<i>Streptococcus pneumoniae</i> (ATCC 49619)		1.00E+06		
<i>Haemophilus influenzae</i> (ATCC 51907)		1.00E+06		
Adenovirus 5 (ATCC VR-1516)		1.00E+05		
<i>Bordetella pertussis</i> (ATCC 9797)		1.00E+06		
<i>Streptococcus pyogenes</i> ATCC (19615)	2	1.00E+06	3/3	No
Rhinovirus (ATCC VR-1645)		1.00E+05		
Human coronavirus OC43 (ATCC VR-1558)		1.00E+05		
<i>Candida albicans</i> (ATCC 18804)		1.00E+06		
<i>Staphylococcus aureus</i> (ATCC 12600)		1.00E+06		
Human coronavirus 229E (ATCC VR-740)	3	1.00E+05	3/3	No
Respiratory syncytial virus (ATCC VR-26)		1.00E+05		
Parainfluenza virus 1 (ATCC VR-94)		1.00E+05		
<i>Staphylococcus epidermidis</i> (ATCC CRM-12228)		1.00E+06		
SARS-CoV (BEI NR-9323)		1.00E+05		
Parainfluenza virus 3 (ATCC VR-1782)	4	1.00E+05	3/3	No
Influenza A (H1N1) (ATCC VR-1520)		1.00E+05		
Influenza B (ATCC VR-1931)		1.00E+05		
Enterovirus (ATCC VR-1432)		1.00E+05		
MERS-CoV (BEI NR-50549)		1.00E+05		
Influenza A (H3N2) (ATCC VR-1679)	5	1.00E+05	3/3	No
<i>Chlamydiaceae</i> (ATCC 53592)		1.00E+06		
<i>Legionella pneumophila</i> (ATCC 35096)		1.00E+06		
Coronavirus NL63 (Zeptometrix 0810228CFHI)		1.00E+05		
Human Metapneumovirus (BEI NR-22227)		1.00E+05		
Human parainfluenza 2 (ATCC VR-92)	6	1.00E+05	3/3	No
Parainfluenza 4B (BEI NR-3238)		1.00E+05		
Pooled Nasal Wash		N/A		

Table S5: Cross-Reactivity and Interference of Endogenous Substances

Potential interfering substances	Active Ingredient	(Recommended) Concentration Tested	Number Detected (3 Replicates)	Cross-Reactivity At Tested Concentration	Number Detected (3 Replicates)	Interference
Afrin	Oxymetazoline	15% v/v	0/3	No	3/3	No
Azithromycin	Azithromycin	250 ug/mL	0/3	No	3/3	No
Blood (Human)	Blood	4%	3/3	Yes	3/3	No
Chloraseptic	Menthol	1.5 mg/mL	0/3	No	3/3	No
Clathromycin	Clathromycin	1 mg/mL	0/3	No	3/3	No
Flonase	Fluticasone Propionate	5% v/v	0/3	No	3/3	No
Homeopathic	Alkalol	1:10 dilution	0/3	No	3/3	No
Mucin	Purified mucin protein	0.5% w/v	2/3	Yes	3/3	No
Mupirocin	Mupirocin	10 mg/mL	2/3	Yes	3/3	No
Nasal Drops	Phenylephrine hydrochloride	15% v/v	1/3	Yes	3/3	No
Nasal Spray	Cromolyn	15% v/v	0/3	No	3/3	No
Nelimed	Hyaluronic Acid	5% v/v	0/3	No	3/3	No
Phenol Throat Spray	Phenol	15% v/v	0/3	No	3/3	No
Saline Nasal Spray	Saline	5% v/v	0/3	No	3/3	No
Tamiflu	Oseltamivir	5 mg/mL	1/3	Yes	3/3	No
Tobramycin	Tobramycin	4 mg/mL	0/3	No	3/3	No
Zicam	Oxymetazoline hydrochloride	5% v/v	3/3	Yes	3/3	No