

SUPPLEMENTARY TABLE 1:

Virus	n	Co-infection	Variant/subtype
SARS-CoV-2 obtained between February 2020 to June 2022	56	0	Wuhan n=10; Alpha n=12; Delta n=12; Omicron n= 22 (sub-variants BA.1 n=7; BA.2 n=6; BA.2.12.1 n=3; BA.4 n=1; BA.5 n=5)
IAV from 2016 to 2019	84	RSV n = 4 (RSVA n= 3; RSVB n = 1)	H1N1 n=82 H3N2 n=2
IBV from 2016 to 2019	5	0	Victoria n=4 Yamagata n=1
RSV from 2016 to 2022	28	SARS-CoV-2 n= 2	RSVA n=29 RSVB n=3
None (negative) from 2019 to 2022	126	0	-

Legend: SARS-CoV-2: severe acute respiratory syndrome coronavirus 2, IAV: Influenza A; IBV: Influenza B; RSVA/B: respiratory syncytial virus

SUPPLEMENTAL TABLE 2: Detection of samples per target for each assay:

	SOC			FAO			TP†			PC			GX			
	+	-	total	+	-	total	+	-	total	+	-	total	+	-	total	
SARS-CoV-2	+	59	0	59	-	-	-	57	2	59	56	3	59	57	2	59
	-	0	180	180	-	-	-	0	180	180	1	179	180	0	165	165
	total	59	180	239	-	-	-	57	182	239	57	182	239	57	167	224
IAV	+	84	0	84	48	36	84	-	-	-	84	0	84	53*	0	53
	-	0	179	179	0	0	0	-	-	-	1	178	179	0	179	179
	total	84	179	263	48	36	84	-	-	-	85	178	263	53	179	232
IBV	+	5	0	5	-	-	-	-	-	-	5	0	5	5	0	5
	-	0	222	222	-	-	-	-	-	-	3	219	222	0	203	203
	total	5	222	227	-	-	-	-	-	-	8	219	227	5	203	208
Influenza†	+	89	0	89	-	-	-	89	0	89	-	-	-	-	-	-
	-	0	174	174	-	-	-	0	174	174	-	-	-	-	-	-
	total	89	174	263	-	-	-	89	174	236	-	-	-	-	-	-
RSV	+	32	0	32	-	-	-	31	1	32	32	0	32	28	4	32
	-	0	178	178	-	-	-	0	178	178	1	177	178	0	163	163
	total	32	178	210	-	-	-	31	179	210	33	177	210	28	167	195

Legend: SARS-CoV-2: severe acute respiratory syndrome coronavirus 2, IAV: Influenza A; IBV: Influenza B; RSVA/B: respiratory syncytial virus; SOC: standard of care (Berlin for Sars-CoV-2; CDC for IAV and IBV, Duplex VIDRL for RSV); FOA: Food and Agriculture Organization of United Nations recommended primers and probes developed by the Australian Center for Disease Prevention for the detection of M gene from avian influenza; GX: Xpert® Xpress SARS-CoV-2, Flu, RSV Kit; TP: TaqPath™ COVID-19, FluA/B, RSV Combo Kit; PC: PowerChek™ SARS-CoV-2, Influenza A&B, RSV Multiplex Real-time PCR Kit II. * For 31 specimens, the volume of sample available did not allow us to perform the GX test. † TaqPath does not provide separate results for IAV and IBV, as targets are combined in the same optical detection channel. Therefore, IAV and IBV samples were analyzed conjointly for comparison with SoC.

SUPPLEMENTAL TABLE 3: Discordant results among clinical specimens

Sample ID	Discordance type	SARS-CoV-2				Influenza				RSV				
		SOC	GX	TP	PC	SOC	GX	TP†	PC	SOC	GX*	TP*	PC*	
4	SARS-CoV-2	nd	nd	nd	+ (34)	nd	nd	nd	nd	+ (21)	+ (25)	+ (15)	+ (13)	
5		+ (38)	nd	nd	nd	nd	nd	nd	nd	+ (22)	+ (29)	+ (16)	+ (13)	
13		+ (39)	+ (37)	+ (33)	nd	nd	nd	nd	nd	+ (33)	+ (33)	+ (28)	+ (25)	
29	Influenza	nd	nd	nd	nd	nd IBV (30)	nd IBV (26)	Inf (32)	IAV (33) IBV (26)	nd	nd	nd	nd	
40		+ (33/37)	+ (35)	+ (32)	+ (28)	nd	nd	Inf (13)	nd	nd	nd	nd	nd	
126		nd	nd	nd	nd	IAV (20) nd	IAV (20/25) nd	Inf (11)	IAV (14) IBV (31)	nd	nd	nd	nd	
162		nd	nd	nd	nd	IAV (21) nd	IAV (22/28) nd	Inf (12)	IAV (17) IBV (31)	nd	nd	nd	nd	
221		nd	nd	nd	nd	nd	nd	nd	nd IBV (34)	nd	nd	nd	nd	
8	RSV	nd	nd	nd	nd	nd	nd	nd	nd	RSVA (30)	nd	+ (25)	+ (22)	
17		nd	nd	nd	nd	nd	nd	nd	nd	RSVA (26)	nd	+ (23)	+ (17)	
27		nd	nd	nd	nd	nd	nd	nd	nd	RSVA (30)	nd	+ (24)	+ (20)	
73		+ (35/36)	+ (32)	+ (24)	+ (26)	nd	nd	nd	nd	nd	nd	nd	+ (35)	nd
127		nd	nd	nd	nd	IAV (28)	IAV (27/29)	Inf (13)	IAV (22)	nd	nd	nd	nd	+ (33)
146		nd	nd	nd	nd	IAV (24)	IAV (25/28)	Inf (18)	IAV (19)	RSVA (39)	nd	nd	+ (38)	

Legend: SoC: standard of care (Berlin for Sars-CoV-2; CDC for IAV and IBV, Duplex VIDRL for RSV); GX: Xpert® Xpress SARS-CoV-2, Flu, RSV Kit; TP: TaqPath™ COVID-19, FluA/B, RSV Combo Kit; PC: PowerChek™ SARS-CoV-2, Influenza A&B, RSV Multiplex Real-time PCR Kit II; SARS-CoV-2: severe acute respiratory syndrome coronavirus 2, Flu: influenza, IAV: Influenza A; IBV: Influenza B; RSV/RSVA: respiratory syncytial virus; nd: not detected; † TP uses the same optical channel for the detection of IAV and IBV and therefore provides a combined result for influenza.* GX, TP and PC use the same optical channel for the detection of RSVA and RSVB and then provide a combined result for RSV.

