

**Supplemental Table 1.** Summary of comparator assays used in this study

Taxa	Target gene	Assay Type	Forward Primer Sequence	Reverse Primer Sequence	Probe Sequence	Reference
<i>Escherichia coli</i>	uidA	Taqman	CGGAAGCAACGCGTAAACTC	TGAGCGTCGCAGAACATTACA	CGCGTCCGATCACCTGCGTC	Silkie, et al., 2008
<i>Klebsiella pneumoniae</i>	khe	Taqman	GATGAAACGACCTGATTGCATTC	CCGGGCTGTCCGGATAAG	GCGAACTGGAAGGGCCCG	Hartman, et al., 2009
<i>Proteus mirabilis</i>	ureR	Taqman	CCATCAGATTATGTCATTCAA	GAGGAAAATGCAATTTATCTTTA	CACACCCTACCCAACATTCATTCA	Liu, et al., 2019
<i>Enterococcus faecalis</i>	16S	Taqman	CGCTTCTTTCCCTCCGAGT	GCCATGCGGCATAAATCG	CAATTGGAAAGAGGAGTGGCGGACG	Santo Domingo et al., 2003
<i>Staphylococcus aureus</i>	ebp5	Taqman	CCACATGCCTTAATAATG	GCGATTTTATTTCTTTGTAC	ATGCCATGCCTCAAATATCGC	Liu, et al., 2019
<i>Gardnerella vaginalis</i>	cpn60	Taqman	CGCATCTGCTAAGGATGTTG	CAGCAATCTTTTCGCCAACT	TGCAACTATTTCTGCAGCAGATCC	Menard et al., 2008
<i>Aerococcus urinae</i>	16S	Taqman	GACGGCTTTGCTGCTACTATCG	GCTATGCATCATTGCTTGGTAG	TGCATTAGCTCGTTGGTGGG	Biotia
<i>Prevotella species</i>	16S	SYBR green	GGTTCTGAGAGGAAGTCCCC	TCCTGCACGCTACTTGGCTG	NA	Stevenson and Weimer, 2007
<i>Candida species</i>	ITS	Sanger	TCCGTAGGTGAACCTGCGG	TCCTCCGCTTATTGATATGC	NA	White et al., 1990
<i>Anaerococcus vaginalis</i>	16S	Sanger	TGATTTCTTCGGAATGAAATTAAGTGGATTA	GGTCATTTTATCATGCGATATTTGACTTT	NA	Biotia
<i>Corynebacterium urealyticum</i>	16S	Sanger	CCTGCTTGCAGGGTACTCGA	CACCAACCACACTAAAGATTGGTC	NA	Biotia
<i>Enterococcus faecium</i>	16S	Sanger	TTCTTTTTCCACCGGAGCTT	AACCATGCGGTTTYGATTG	NA	Ryu et al., 2013
<i>Morganella morganii</i>	16S	Sanger	AGCTTGCTTCTGCTGACGAG	GAGGCCCGAAGGTCCCCCG	NA	Biotia
<i>Staphylococcus epidermidis</i>	16S	Sanger	GTTCAATAGTGAAAGACGGTTTTGCTGTC	GTTTTACGATCCGAAGACCTTCATCAC	NA	Biotia
<i>Staphylococcus saprophyticus</i>	16S	Sanger	TAAAGTGAAAGATGTTTTGCTATC	GTTTTACGAGCCGAAACCCTTCATCAC	NA	Biotia
<i>Streptococcus agalactiae</i>	16S	Sanger	KGTTTGGTGTTCACACTAGACTG	TTACCGTCACTTGGTAGATTTTCCA	NA	Biotia
<i>Streptococcus anginosus</i>	16S	Sanger	CGTAGTTTACTACCCGTATTCTGTGA	GTACCGTCACTGTGTGAAC	NA	Biotia

**Supplementary Table 2A.** Summary of Replicates Detected in the Reproducibility Studies

<b>Microbial Species</b>	<b>Replicates Detected</b>	<b>Qualitative Reproducibility</b>
<i>Enterococcus faecalis</i>	81/81	100%
<i>Escherichia coli</i>	81/81	100%
<i>Klebsiella pneumoniae</i>	81/81	100%
<i>Proteus mirabilis</i>	81/81	100%
<i>Staphylococcus aureus</i>	81/81	100%
<i>Candida albicans</i>	45/45	100%
<i>Candida auris</i>	45/45	100%
<i>Candida glabrata</i>	45/45	100%
<i>Candida krusei</i>	45/45	100%
<i>Candida parapsilosis</i>	45/45	100%
<i>Candida tropicalis</i>	45/45	100%
<i>Candida dubliniensis</i>	9/9	100%
<i>Candida guilliermondii</i>	9/9	100%
<i>Candida kefyr</i>	9/9	100%
<i>Candida lusitanae</i>	9/9	100%
<i>Negative</i>	42/42	100%

**Supplementary Table 2B. Reproducibility Studies Results by Concentration**

Analyte	Strain	Load	Day 1	Day 2	Day 3	Total
<i>Enterococcus faecalis</i>	URV-032-2	15,000	3/3	3/3	3/3	9/9
<i>Escherichia coli</i>	URV-012	15,000	3/3	3/3	3/3	9/9
<i>Klebsiella pneumoniae</i>	URV-155-1	15,000	3/3	3/3	3/3	9/9
<i>Proteus mirabilis</i>	URV-023	15,000	3/3	3/3	3/3	9/9
<i>Staphylococcus aureus</i>	ATCC-43300	15,000	3/3	3/3	3/3	9/9
<i>Enterococcus faecalis</i>	URV-032-2	20,000	3/3	3/3	3/3	9/9
<i>Escherichia coli</i>	URV-012	20,000	3/3	3/3	3/3	9/9
<i>Klebsiella pneumoniae</i>	URV-155-1	20,000	3/3	3/3	3/3	9/9
<i>Proteus mirabilis</i>	URV-023	20,000	3/3	3/3	3/3	9/9
<i>Staphylococcus aureus</i>	ATCC-43300	20,000	3/3	3/3	3/3	9/9
<i>Enterococcus faecalis</i>	URV-032-2	25,000	3/3	3/3	3/3	9/9
<i>Escherichia coli</i>	URV-012	25,000	3/3	3/3	3/3	9/9
<i>Klebsiella pneumoniae</i>	URV-155-1	25,000	3/3	3/3	3/3	9/9
<i>Proteus mirabilis</i>	URV-023	25,000	3/3	3/3	3/3	9/9
<i>Staphylococcus aureus</i>	ATCC-43300	25,000	3/3	3/3	3/3	9/9
<i>Enterococcus faecalis</i>	URV-220	25,000	3/3	3/3	3/3	9/9
<i>Escherichia coli</i>	URV-005-1	25,000	3/3	3/3	3/3	9/9
<i>Klebsiella pneumoniae</i>	URV-038	25,000	3/3	3/3	3/3	9/9
<i>Proteus mirabilis</i>	URV-010	25,000	3/3	3/3	3/3	9/9
<i>Staphylococcus aureus</i>	SJF-038	25,000	3/3	3/3	3/3	9/9
<i>Enterococcus faecalis</i>	URV-032-2	25,000	3/3	3/3	3/3	9/9
<i>Escherichia coli</i>	URV-012	25,000	3/3	3/3	3/3	9/9
<i>Klebsiella pneumoniae</i>	URV-036	25,000	3/3	3/3	3/3	9/9
<i>Proteus mirabilis</i>	URV-014	25,000	3/3	3/3	3/3	9/9
<i>Staphylococcus aureus</i>	SJF-044	25,000	3/3	3/3	3/3	9/9
<i>Enterococcus faecalis</i>	URV-007	25,000	3/3	3/3	3/3	9/9
<i>Escherichia coli</i>	URV-029	25,000	3/3	3/3	3/3	9/9
<i>Klebsiella pneumoniae</i>	URV-022	25,000	3/3	3/3	3/3	9/9
<i>Proteus mirabilis</i>	URV-040	25,000	3/3	3/3	3/3	9/9
<i>Staphylococcus aureus</i>	SJF-034	25,000	3/3	3/3	3/3	9/9
<i>Enterococcus faecalis</i>	URV-026	25,000	3/3	3/3	3/3	9/9
<i>Escherichia coli</i>	URV-006	25,000	3/3	3/3	3/3	9/9
<i>Klebsiella pneumoniae</i>	URV-031	25,000	3/3	3/3	3/3	9/9
<i>Proteus mirabilis</i>	URV-060	25,000	3/3	3/3	3/3	9/9
<i>Staphylococcus aureus</i>	SJF-035	25,000	3/3	3/3	3/3	9/9
<i>Enterococcus faecalis</i>	URV-261	25,000	3/3	3/3	3/3	9/9
<i>Escherichia coli</i>	URV-021	25,000	3/3	3/3	3/3	9/9
<i>Klebsiella pneumoniae</i>	URV-155-1	25,000	3/3	3/3	3/3	9/9
<i>Proteus mirabilis</i>	URV-016	25,000	3/3	3/3	3/3	9/9
<i>Staphylococcus aureus</i>	SJF-041	25,000	3/3	3/3	3/3	9/9
<i>Enterococcus faecalis</i>	URV-032-2	100,000	3/3	3/3	3/3	9/9
<i>Escherichia coli</i>	AR-104	100,000	3/3	3/3	3/3	9/9
<i>Klebsiella pneumoniae</i>	AR-557	100,000	3/3	3/3	3/3	9/9
<i>Proteus mirabilis</i>	AR-156	100,000	3/3	3/3	3/3	9/9
<i>Staphylococcus aureus</i>	SJF-035	100,000	3/3	3/3	3/3	9/9
<i>Candida albicans</i>	60193	25,000	3/3	3/3	3/3	9/9
<i>Candida auris</i>	MYA-5000	25,000	3/3	3/3	3/3	9/9
<i>Candida glabrata</i>	15126	25,000	3/3	3/3	3/3	9/9
<i>Candida krusei</i>	14243	25,000	3/3	3/3	3/3	9/9
<i>Candida parapsilosis</i>	22019	25,000	3/3	3/3	3/3	9/9
<i>Candida tropicalis</i>	1369	25,000	3/3	3/3	3/3	9/9

<i>Candida albicans</i>	10231	25,000	3/3	3/3	3/3	9/9
<i>Candida auris</i>	MYA-5001	25,000	3/3	3/3	3/3	9/9
<i>Candida glabrata</i>	2001	25,000	3/3	3/3	3/3	9/9
<i>Candida krusei</i>	34135	25,000	3/3	3/3	3/3	9/9
<i>Candida parapsilosis</i>	90018	25,000	3/3	3/3	3/3	9/9
<i>Candida tropicalis</i>	1369	25,000	3/3	3/3	3/3	9/9
<i>Candida albicans</i>	14053	25,000	3/3	3/3	3/3	9/9
<i>Candida auris</i>	MYA-5002	25,000	3/3	3/3	3/3	9/9
<i>Candida glabrata</i>	66032	25,000	3/3	3/3	3/3	9/9
<i>Candida krusei</i>	90878	25,000	3/3	3/3	3/3	9/9
<i>Candida parapsilosis</i>	SJF-019	25,000	3/3	3/3	3/3	9/9
<i>Candida tropicalis</i>	750	25,000	3/3	3/3	3/3	9/9
<i>Candida albicans</i>	90029	25,000	3/3	3/3	3/3	9/9
<i>Candida auris</i>	MYA-5003	25,000	3/3	3/3	3/3	9/9
<i>Candida glabrata</i>	90030	25,000	3/3	3/3	3/3	9/9
<i>Candida krusei</i>	6258	25,000	3/3	3/3	3/3	9/9
<i>Candida parapsilosis</i>	SJF-032	25,000	3/3	3/3	3/3	9/9
<i>Candida tropicalis</i>	13803	25,000	3/3	3/3	3/3	9/9
<i>Candida albicans</i>	URV-018	25,000	3/3	3/3	3/3	9/9
<i>Candida auris</i>	CDC B11903	25,000	3/3	3/3	3/3	9/9
<i>Candida glabrata</i>	MYA-2950	25,000	3/3	3/3	3/3	9/9
<i>Candida krusei</i>	URV-019	25,000	3/3	3/3	3/3	9/9
<i>Candida parapsilosis</i>	SJF-033	25,000	3/3	3/3	3/3	9/9
<i>Candida tropicalis</i>	66029	25,000	3/3	3/3	3/3	9/9
<i>Candida dubliniensis</i>	3949	25,000	3/3	3/3	3/3	9/9
<i>Candida guilliermondii</i>	6260	25,000	3/3	3/3	3/3	9/9
<i>Candida kefyr</i>	66028	25,000	3/3	3/3	3/3	9/9
<i>Candida lusitanae</i>	34449	25,000	3/3	3/3	3/3	9/9
Negative Samples	-	-	14/14	14/14	14/14	42/42

**Supplementary Table 3A.** Performance Characteristics of in silico Studies per Key Urogenital Taxa

Key pathogen	Method	TP	TN	FN	FP	Total	Sensitivity	Specificity
<i>Escherichia coli</i>	Clinical							
	Contrived							
	in silico	3	8239	0	24	8266	100.0%	99.7%
	Total							
<i>Klebsiella pneumoniae</i>	Clinical							
	Contrived							
	in silico	3	8250	0	13	8266	100.0%	99.8%
	Total							
<i>Proteus mirabilis</i>	Clinical							
	Contrived							
	in silico	3	8255	0	8	8266	100.0%	99.9%
	Total							
<i>Enterococcus faecalis</i>	Clinical							
	Contrived							
	in silico	3	8263	0	0	8266	100.0%	100.0%
	Total							
<i>Staphylococcus aureus</i>	Clinical							
	Contrived							
	in silico	3	8252	0	11	8266	100.0%	99.9%
	Total							
<i>Citrobacter species</i>	Clinical							
	Contrived							
	in silico	20	8246	0	0	8266	100.0%	100.0%
	Total							
<i>Enterobacter aerogenes</i>	Clinical							
	Contrived							
	in silico	3	8263	0	0	8266	100.0%	100.0%
	Total							
<i>Enterobacter cloacae</i>	Clinical							
	Contrived							
	in silico	3	8253	0	10	8266	100.0%	99.9%
	Total							
<i>Klebsiella variicola</i>	Clinical							
	Contrived							
	in silico	3	8260	0	3	8266	100.0%	100.0%
	Total							
<i>Klebsiella oxytoca</i>	Clinical							
	Contrived							
	in silico	3	8260	0	3	8266	100.0%	100.0%
	Total							
<i>Morganella morganii</i>	Clinical							
	Contrived							
	in silico	3	8262	0	1	8266	100.0%	100.0%
	Total							
<i>Proteus vulgaris</i>	Clinical							
	Contrived							
	in silico	3	8256	0	7	8266	100.0%	99.9%
	Total							
<i>Providencia rettgeri</i>	Clinical							
	Contrived							
	in silico	3	8260	0	3	8266	100.0%	100.0%
	Total							
<i>Providencia stuartii</i>	Clinical							
	Contrived							
	in silico	3	8260	0	3	8266	100.0%	100.0%
	Total							
<i>Raoultella ornithinolytica</i>	Clinical							
	Contrived							
	in silico	3	8260	0	3	8266	100.0%	100.0%
	Total							
<i>Serratia marcescens</i>	Clinical							
	Contrived							
	in silico	3	8245	0	18	8266	100.0%	99.8%
	Total							
<i>Acinetobacter baumannii</i>	Clinical							
	Contrived							
	in silico	3	8244	0	19	8266	100.0%	99.8%
	Total							

<i>Acinetobacter lwoffii</i>	Clinical							
	Contrived							
	in silico	3	8260	0	3	8266	100.0%	100.0%
	Total							
<i>Pseudomonas aeruginosa</i>	Clinical							
	Contrived							
	in silico	3	8260	0	3	8266	100.0%	100.0%
	Total							
<i>Streptrophomonas maltophilia</i>	Clinical							
	Contrived							
	in silico	3	8252	0	11	8266	100.0%	99.9%
	Total							
<i>Aerococcus species</i>	Clinical							
	Contrived							
	in silico	21	8242	3	0	8266	87.5%	100.0%
	Total							
Anginosus group Streptococci	Clinical							
	Contrived							
	in silico	5	8261	0	0	8266	100.0%	100.0%
	Total							
<i>Streptococcus agalactiae</i>	Clinical							
	Contrived							
	in silico	3	8263	0	0	8266	100.0%	100.0%
	Total							
<i>Enterococcus faecium</i>	Clinical							
	Contrived							
	in silico	3	8260	0	3	8266	100.0%	100.0%
	Total							
<i>Corynebacterium urealyticum</i>	Clinical							
	Contrived							
	in silico	3	8263	0	0	8266	100.0%	100.0%
	Total							
<i>Staphylococcus epidermidis</i>	Clinical							
	Contrived							
	in silico	3	8263	0	0	8266	100.0%	100.0%
	Total							
<i>Staphylococcus lugdunensis</i>	Clinical							
	Contrived							
	in silico	3	8263	0	0	8266	100.0%	100.0%
	Total							
<i>Staphylococcus saprophyticus</i>	Clinical							
	Contrived							
	in silico	3	8263	0	0	8266	100.0%	100.0%
	Total							
Mitis group Streptococci	Clinical							
	Contrived							
	in silico	3	8262	0	1	8266	100.0%	100.0%
	Total							
<i>Anaerococcus vaginalis</i>	Clinical							
	Contrived							
	in silico	3	8263	0	0	8266	100.0%	100.0%
	Total							
<i>Bacteroides fragilis</i>	Clinical							
	Contrived							
	in silico	3	8263	0	0	8266	100.0%	100.0%
	Total							
<i>Prevotella species</i>	Clinical							
	Contrived							
	in silico	9	8257	0	0	8266	100.0%	100.0%
	Total							
<i>Gardnerella vaginalis</i>	Clinical							
	Contrived							
	in silico	3	8263	0	0	8266	100.0%	100.0%
	Total							
Negative Organisms*	Contrived							
	in silico	7977	289	0	0	8266	100.0%	100.0%
	Total							

Supplementary Table 3B. Performance Characteristics of in silico Studies per Key Fungal Urogenital Taxa

Key pathogen	Method	TP	TN	FN	FP	Total	Sensitivity	Specificity
<i>Candida albicans</i>	Clinical							
	Contrived							
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total							
<i>Candida auris</i>	Clinical							
	Contrived							
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total							
<i>Candida dubliniensis</i>	Clinical							
	Contrived							
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total							
<i>Candida glabrata</i>	Clinical							
	Contrived							
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total							
<i>Candida guilliermondii</i>	Clinical							
	Contrived							
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total							
<i>Candida kefyr</i>	Clinical							
	Contrived							
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total							
<i>Candida krusei</i>	Clinical							
	Contrived							
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total							
<i>Candida lusitanae</i>	Clinical							
	Contrived							
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total							
<i>Candida parapsilopsis</i>	Clinical							
	Contrived							
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total							
<i>Candida tropicalis</i>	Clinical							
	Contrived							
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total							
Negative Organisms*	Contrived							
	<i>in silico</i>	4052	90	0	0	4142	100.0%	100.0%
	Total							

**Supplementary Table 4.** Summary of *in silico* study cross reactive species

Analyte	# of samples	Cross reactive species
<i>Acinetobacter baumannii</i>	19	<i>Acinetobacter genomosp</i>
<i>Acinetobacter lwoffii</i>	3	<i>Acinetobacter pseudolwoffii</i>
<i>Enterobacter cloacae</i>	10	<i>Enterobacter chengduensis</i>
<i>Enterococcus faecium</i>	3	<i>Enterococcus lactis</i>
<i>Escherichia coli</i>	48	<i>Escherichia marmotae</i>
<i>Klebsiella oxytoca</i>	3	<i>Klebsiella grimontii</i>
<i>Klebsiella pneumoniae</i>	13	<i>Klebsiella quasipneumoniae</i>
<i>Klebsiella variicola</i>	3	<i>Klebsiella quasivariicola</i>
<i>Morganella morganii</i>	1	<i>Morganella psychrotolerans</i>
<i>Proteus mirabilis</i>	8	<i>Proteus penneri</i>
<i>Proteus vulgaris</i>	7	<i>Proteus faecis</i>
<i>Providencia rettgeri</i>	3	<i>Providencia huaxiensis</i>
<i>Providencia stuartii</i>	3	<i>Providencia thailandensis</i>
<i>Pseudomonas aeruginosa</i>	3	<i>Pseudomonas paraeruginosa</i>
<i>Raoultella ornithinolytica</i>	3	<i>Raoultella planticola</i>
<i>Serratia marcescens</i>	18	<i>Serratia nematodiphila</i>
<i>Staphylococcus aureus</i>	11	<i>Staphylococcus roterodami</i>
<i>Stenotrophomonas maltophilia</i>	11	<i>Stenotrophomonas indicatrix</i>
<i>Streptococcus mitis</i>	1	<i>Streptococcus pseudopneumoniae</i>
Simulated Reads Species	# of samples	Cross reactive species
<i>Proteus vulgaris</i>	2	<i>Proteus terrae</i>
<i>Serratia nematodiphila</i>	1	<i>Serratia ureilytica</i>
<i>Streptococcus pseudopneumoniae</i>	1	<i>Streptococcus pneumoniae</i>
<i>Staphylococcus singaporensis</i>	3	<i>Staphylococcus haemolyticus</i>



**Supplementary Table 5.** Comparator testing was used to resolve discrepancies between the BIOTIA-DX and culture. A combination of qPCR assays and Sanger sequencing were used for resolving such discrepancies. Assays are summarized in Supplementary Table 1.

qPCR Comparator Testing Results							
Sample	Analyte	Prediction Probability	Relative abundance	Biotia-ID Result	Comparator result	qPCR Ct value	Outcome
URV-104	<i>Enterococcus faecalis</i>	0.988	0.198	Detected	Detected	16.07	True Positive
URV-110	<i>Enterococcus faecalis</i>	0.999	0.135	Detected	Detected	15.79	True Positive
URV-118	<i>Enterococcus faecalis</i>	0.999	0.215	Detected	Detected	24.00	True Positive
URV-122	<i>Enterococcus faecalis</i>	0.972	0.081	Detected	Detected	16.76	True Positive
URV-127	<i>Enterococcus faecalis</i>	0.998	0.021	Detected	Detected	21.29	True Positive
URV-130	<i>Enterococcus faecalis</i>	0.972	0.179	Detected	Detected	19.99	True Positive
URV-132	<i>Enterococcus faecalis</i>	0.999	0.398	Detected	Detected	27.53	True Positive
URV-133	<i>Enterococcus faecalis</i>	0.962	0.019	Detected	Detected	25.07	True Positive
URV-134	<i>Enterococcus faecalis</i>	0.984	0.008	Detected	Detected	19.91	True Positive
URV-137	<i>Enterococcus faecalis</i>	0.999	0.120	Detected	Detected	24.43	True Positive
URV-139	<i>Enterococcus faecalis</i>	0.595	0.012	Detected	Not_detected	Undetermined	False Positive
URV-146	<i>Enterococcus faecalis</i>	0.999	0.041	Detected	Detected	17.30	True Positive
URV-157	<i>Enterococcus faecalis</i>	0.999	0.384	Detected	Detected	16.81	True Positive
URV-163	<i>Enterococcus faecalis</i>	0.999	0.739	Detected	Detected	13.77	True Positive
URV-170	<i>Enterococcus faecalis</i>	0.999	0.194	Detected	Detected	21.68	True Positive
URV-172	<i>Enterococcus faecalis</i>	0.999	0.279	Detected	Detected	18.98	True Positive
URV-176	<i>Enterococcus faecalis</i>	0.999	0.090	Detected	Detected	17.53	True Positive
URV-180	<i>Enterococcus faecalis</i>	0.999	0.575	Detected	Detected	21.81	True Positive
URV-181	<i>Enterococcus faecalis</i>	0.963	0.989	Detected	Detected	14.67	True Positive
URV-182	<i>Enterococcus faecalis</i>	0.999	0.155	Detected	Detected	16.58	True Positive
URV-184	<i>Enterococcus faecalis</i>	0.676	0.013	Detected	Detected	27.10	True Positive
URV-190	<i>Enterococcus faecalis</i>	0.999	0.061	Detected	Detected	19.73	True Positive
URV-220	<i>Enterococcus faecalis</i>	0.982	0.805	Detected	Detected	15.03	True Positive
URV-226	<i>Enterococcus faecalis</i>	0.999	0.663	Detected	Detected	14.43	True Positive
URV-227	<i>Enterococcus faecalis</i>	0.998	0.021	Detected	Detected	26.06	True Positive
URV-232	<i>Enterococcus faecalis</i>	0.999	0.305	Detected	Detected	21.34	True Positive
URV-248	<i>Enterococcus faecalis</i>	0.999	0.924	Detected	Detected	16.66	True Positive
URV-258	<i>Enterococcus faecalis</i>	0.115	0.001	Not detected	Detected	24.59	False Negative
URV-259	<i>Enterococcus faecalis</i>	0.402	0.013	Not detected	Detected	24.6	False Negative
URV-261	<i>Enterococcus faecalis</i>	0.999	0.895	Detected	Detected	16.26	True Positive
URV-272	<i>Enterococcus faecalis</i>	0.999	0.616	Detected	Detected	15.76	True Positive
URV-278	<i>Enterococcus faecalis</i>	0.999	0.663	Detected	Detected	16.33	True Positive
URV-280	<i>Enterococcus faecalis</i>	0.999	0.046	Detected	Detected	22.42	True Positive
URV-283	<i>Enterococcus faecalis</i>	0.999	0.756	Detected	Detected	22.47	True Positive
URV-288	<i>Enterococcus faecalis</i>	0.999	0.061	Detected	Detected	24.09	True Positive
URV-291	<i>Enterococcus faecalis</i>	0.999	0.133	Detected	Detected	17.96	True Positive
URV-111	<i>Escherichia coli</i>	0.999	0.115	Detected	Detected	15.96	True Positive
URV-113	<i>Escherichia coli</i>	0.391	0.025	Not detected	Detected	19.96	False Negative
URV-114	<i>Escherichia coli</i>	0.998	0.104	Detected	Detected	18.05	True Positive
URV-117	<i>Escherichia coli</i>	0.999	0.020	Detected	Detected	18.10	True Positive
URV-130	<i>Escherichia coli</i>	0.999	0.133	Detected	Detected	30.40	True Positive
URV-131	<i>Escherichia coli</i>	0.999	0.840	Detected	Detected	17.06	True Positive
URV-137	<i>Escherichia coli</i>	0.020	0.000	Not detected	Not_detected	Undetermined	True Negative
URV-139	<i>Escherichia coli</i>	0.000	0.000	Not detected	Not Detected	Undetermined	True Negative
URV-141	<i>Escherichia coli</i>	0.000	0.000	Not detected	Not Detected	Undetermined	True Negative
URV-173	<i>Escherichia coli</i>	0.814	0.043	Detected	Detected	19.11	True Positive
URV-178	<i>Escherichia coli</i>	0.971	0.098	Detected	Detected	18.50	True Positive
URV-257	<i>Escherichia coli</i>	0.999	0.027	Detected	Detected	21.83	True Positive
URV-258	<i>Escherichia coli</i>	0.999	0.713	Detected	Detected	11.46	True Positive
URV-268	<i>Escherichia coli</i>	0.801	0.041	Detected	Detected	22.08	True Positive
URV-280	<i>Escherichia coli</i>	0.997	0.675	Detected	Detected	19.01	True Positive
URV-282	<i>Escherichia coli</i>	0.999	0.057	Detected	Detected	26.26	True Positive
URV-289	<i>Escherichia coli</i>	0.999	0.677	Detected	Detected	14.22	True Positive

URV-110	<i>Klebsiella pneumoniae</i>	0.801	0.012	Detected	Detected	15.81	True Positive
URV-122	<i>Klebsiella pneumoniae</i>	0.999	0.159	Detected	Detected	14.59	True Positive
URV-132	<i>Klebsiella pneumoniae</i>	0.046	0.000	Detected	Detected	37.81	True Positive
URV-147	<i>Klebsiella pneumoniae</i>	0.999	0.071	Detected	Detected	26.79	True Positive
URV-169	<i>Klebsiella pneumoniae</i>	0.953	0.043	Detected	Detected	14.82	True Positive
URV-184	<i>Klebsiella pneumoniae</i>	0.999	0.140	Detected	Detected	21.88	True Positive
URV-255	<i>Klebsiella pneumoniae</i>	0.781	0.002	Detected	Detected	25.35	True Positive
URV-262	<i>Klebsiella pneumoniae</i>	0.955	0.003	Detected	Detected	23.69	True Positive
URV-264	<i>Klebsiella pneumoniae</i>	0.422	0.004	Not detected	Detected	21.42	False Negative
URV-280	<i>Klebsiella pneumoniae</i>	0.801	0.011	Detected	Detected	25.16	True Positive
URV-289	<i>Klebsiella pneumoniae</i>	0.813	0.016	Detected	Detected	18.48	True Positive
URV-187	<i>Proteus mirabilis</i>	0.895	0.008	Detected	Not Detected	Undetermined	False Positive
URV-192	<i>Proteus mirabilis</i>	0.999	0.369	Detected	Detected	25.17	True Positive
URV-228	<i>Proteus mirabilis</i>	0.999	0.069	Detected	Detected	21.74	True Positive
URV-231	<i>Proteus mirabilis</i>	0.998	0.147	Detected	Detected	31.71	True Positive
URV-268	<i>Proteus mirabilis</i>	0.998	0.075	Detected	Detected	34.07	True Positive
URV-291	<i>Staphylococcus aureus</i>	0.978	0.109	Detected	Not detected	Undetermined	False Positive
URV-118	<i>Aerococcus urinae</i>	0.968	0.001	Detected	Detected	28.17	True Positive
URV-130	<i>Aerococcus urinae</i>	0.662	0.001	Detected	Detected	23.52	True Positive
URV-139	<i>Aerococcus urinae</i>	0.594	0.001	Detected	Detected	30.94	True Positive
URV-187	<i>Aerococcus urinae</i>	0.986	0.005	Detected	Detected	26.56	True Positive
URV-284	<i>Aerococcus urinae</i>	0.986	0.002	Detected	Detected	26.70	True Positive
URV-285	<i>Aerococcus urinae</i>	0.802	0.001	Detected	Detected	23.11	True Positive
URV-287	<i>Aerococcus urinae</i>	0.939	0.002	Detected	Detected	24.04	True Positive
URV-291	<i>Aerococcus urinae</i>	0.967	0.001	Detected	Detected	19.91	True Positive
URV-108	<i>Gardnerella vaginalis</i>	0.684	0.002	Detected	Not Detected	Undetermined	False Positive
URV-114	<i>Gardnerella vaginalis</i>	0.930	0.028	Detected	Detected	18.83	True Positive
URV-125	<i>Gardnerella vaginalis</i>	0.948	0.142	Detected	Detected	36.87	True Positive
URV-128	<i>Gardnerella vaginalis</i>	0.696	0.029	Detected	Not Detected	Undetermined	False Positive
URV-136	<i>Gardnerella vaginalis</i>	0.859	0.003	Detected	Detected	26.86	True Positive
URV-140	<i>Gardnerella vaginalis</i>	0.934	0.084	Detected	Detected	22.10	True Positive
URV-163	<i>Gardnerella vaginalis</i>	0.875	0.030	Detected	Not Detected	Undetermined	False Positive
URV-167	<i>Gardnerella vaginalis</i>	0.964	0.004	Detected	23.927	23.927	True Positive
URV-175	<i>Gardnerella vaginalis</i>	0.658	0.084	Detected	Not Detected	Undetermined	False Positive
URV-181	<i>Gardnerella vaginalis</i>	0.803	0.002	Detected	21.28	21.28	True Positive
URV-191	<i>Gardnerella vaginalis</i>	0.738	0.098	Detected	17.23	17.23	True Positive
URV-223	<i>Gardnerella vaginalis</i>	0.970	0.013	Detected	31.97	31.97	True Positive
URV-233	<i>Gardnerella vaginalis</i>	0.944	0.083	Detected	15.60	15.60	True Positive
URV-256	<i>Gardnerella vaginalis</i>	0.840	0.545	Detected	20.08	20.08	True Positive
URV-263	<i>Gardnerella vaginalis</i>	0.994	0.011	Detected	28.93	28.93	True Positive
URV-278	<i>Gardnerella vaginalis</i>	0.638	0.031	Detected	Not Detected	Undetermined	False Positive
URV-286	<i>Gardnerella vaginalis</i>	0.608	0.178	Detected	Not Detected	Undetermined	False Positive
URV-104	<i>Prevotella species</i>	0.986	0.001	Detected	Detected	14.83	True Positive
URV-112	<i>Prevotella species</i>	0.998	0.012	Detected	Detected	15.96	True Positive
URV-163	<i>Prevotella species</i>	0.874	0.001	Detected	Detected	15.49	True Positive
URV-249	<i>Prevotella species</i>	0.998	0.029	Detected	Detected	16.78	True Positive
URV-259	<i>Prevotella species</i>	0.992	0.006	Detected	Detected	17.4	True Positive
URV-270	<i>Prevotella species</i>	0.994	0.005	Detected	Detected	17.79	True Positive
URV-271	<i>Prevotella species</i>	0.961	0.003	Detected	Detected	17.87	True Positive

#### Sanger Sequencing Comparator Results

Sample	Analyte	Prediction Probability	Relative abundance	Biotia-ID Result	Comparator result	% Identity	Comparator Agreement
URV-146	<i>Enterococcus faecium</i>	0.998	0.017	Detected	Detected	100%	True Positive
URV-225	<i>Enterococcus faecium</i>	0.961	0.624	Detected	Detected	100%	True Positive
URV-176	<i>Morganella morganii</i>	0.988	0.004	Detected	Detected	100%	True Positive
URV-272	<i>Morganella morganii</i>	0.998	0.013	Detected	Detected	100%	True Positive
URV-112	<i>aphylococcus epidermid</i>	0.828	0.001	Detected	Detected	100%	True Positive
URV-116	<i>aphylococcus epidermid</i>	0.984	0.005	Detected	Detected	99.50%	True Positive

URV-291	<i>phylococcus saprophytic</i>	0.976	0.008	Detected	Detected	99.25%	True Positive
URV-146	<i>streptococcus agalactiae</i>	0.999	0.122	Detected	Detected	100%	True Positive
URV-167	<i>streptococcus agalactiae</i>	0.949	0.004	Detected	Detected	100%	True Positive
URV-186	<i>streptococcus agalactiae</i>	0.930	0.003	Detected	Detected	100%	True Positive
URV-232	<i>streptococcus agalactiae</i>	0.985	0.006	Detected	Detected	100%	True Positive
URV-291	ginosus group Streptoco	0.565	0.002	Detected	Detected	100%	True Positive
URV-140	ginosus group Streptoco	0.827	0.017	Detected	Not Detected	NA	False Positive
URV-275	ginosus group Streptoco	0.515	0.002	Detected	Not Detected	NA	False Positive
URV-287	ginosus group Streptoco	0.538	0.004	Detected	Not Detected	NA	False Positive
URV-154	<i>rynebacterium urealyticu</i>	0.998	0.024	Detected	Detected	100%	True Positive
URV-227	<i>rynebacterium urealyticu</i>	0.963	0.002	Detected	Detected	100%	True Positive
URV-287	<i>rynebacterium urealyticu</i>	0.549	0.000	Detected	Detected	100%	True Positive
URV-288	<i>rynebacterium urealyticu</i>	0.645	0.001	Detected	Detected	100%	True Positive
URV-225	<i>Candida albicans</i>	0.980	0.042	Detected	Detected	100%	True Positive
URV-230	<i>Candida albicans</i>	0.980	0.314	Detected	Detected	100%	True Positive
URV-233	<i>Candida albicans</i>	0.986	0.087	Detected	Detected	99.8%	True Positive
URV-260	<i>Candida albicans</i>	0.943	0.019	Detected	Detected	100%	True Positive
URV-278	<i>Candida albicans</i>	0.996	0.081	Detected	Detected	97.50%	True Positive
SBU-103	<i>Candida albicans</i>	0.998	0.171	Detected	Detected	99.8%	True Positive
SBU-161	<i>Candida albicans</i>	0.598	0.017	Detected	Detected	99.8%	True Positive
SBU-176	<i>Candida albicans</i>	0.963	0.378	Detected	Detected	100.0%	True Positive
URV-148	<i>Candida glabrata</i>	0.982	0.679	Detected	Detected	99.9%	True Positive
URV-182	<i>Candida glabrata</i>	0.982	0.314	Detected	Detected	98.70%	True Positive
URV-262	<i>Candida glabrata</i>	0.982	0.661	Detected	Detected	99.9%	True Positive
URV-288	<i>Candida glabrata</i>	0.991	0.047	Detected	Detected	99.90%	True Positive
SBU-021	<i>Candida glabrata</i>	0.974	0.347	Detected	Detected	100.00%	True Positive
SBU-047	<i>Candida glabrata</i>	0.972	0.305	Detected	Detected	99.80%	True Positive
SBU-137	<i>Candida glabrata</i>	0.967	0.016	Detected	Detected	99.88%	True Positive
SBU-162	<i>Candida glabrata</i>	0.981	0.075	Detected	Detected	99.76%	True Positive
SBU-017	<i>Candida tropicalis</i>	0.733	0.108	Detected	Detected	99.80%	True Positive
SBU-018	<i>Candida tropicalis</i>	0.988	0.392	Detected	Detected	100.00%	True Positive
SBU-027	<i>Candida tropicalis</i>	0.977	0.269	Detected	Detected	99.80%	True Positive
SBU-074	<i>Candida tropicalis</i>	0.577	0.171	Detected	Detected	98.80%	True Positive
SBU-017	<i>Candida krusei</i>	0.696	0.058	Detected	Detected	100%	True Positive

**Supplementary Table 6A.** Performance Characteristics fo Key Bacterial Urogenital Taxa Reported with BIOTIA-ID

Key pathogen	Method	TP	TN	FN	FP	Total	Sensitivity	Specificity
<i>Escherichia coli</i>	Clinical	39	204	1	1	245	97.50%	99.51%
	Contrived	157	1096	0	13	1266	100.00%	98.83%
	<i>in silico</i>	3	8239	0	24	8266	100.00%	99.71%
	Total	199	9539	1	38	9777	99.50%	99.60%
<i>Klebsiella pneumoniae</i>	Clinical	28	214	2	1	245	93.33%	99.53%
	Contrived	166	1099	0	1	1266	100.00%	99.91%
	<i>in silico</i>	3	8250	0	13	8266	100.00%	99.84%
	Total	197	9563	2	15	9777	98.99%	99.84%
<i>Proteus mirabilis</i>	Clinical	22	222	0	1	245	100.00%	99.55%
	Contrived	163	1103	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8255	0	8	8266	100.00%	99.90%
	Total	188	9580	0	9	9777	100.00%	99.91%
<i>Enterococcus faecalis</i>	Clinical	61	181	2	1	245	96.83%	99.45%
	Contrived	147	1119	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8263	0	0	8266	100.00%	100.00%
	Total	211	9563	2	1	9777	99.06%	99.99%
<i>Staphylococcus aureus</i>	Clinical	2	242	0	1	245	100.00%	99.59%
	Contrived	172	1094	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8252	0	11	8266	100.00%	99.87%
	Total	177	9588	0	12	9777	100.00%	99.88%
<i>Citrobacter species</i>	Clinical	0	243	0	2	245	-	99.18%
	Contrived	2	1264	0	0	1266	100.00%	100.00%
	<i>in silico</i>	20	8246	0	0	8266	100.0%	100.0%
	Total	22	9753	0	2	9777	100.00%	99.98%
<i>Enterobacter aerogenes</i>	Clinical	0	244	1	0	245	0.00%	100.00%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8263	0	0	8266	100.00%	100.00%
	Total	4	9772	1	0	9777	80.00%	100.00%
<i>Enterobacter cloacae</i>	Clinical	0	245	0	0	245	-	100.00%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8253	0	10	8266	100.00%	99.88%
	Total	4	9763	0	10	9777	100.00%	99.90%
<i>Klebsiella oxytoca</i>	Clinical	0	244	0	1	245	-	99.59%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8260	0	3	8266	100.00%	99.96%
	Total	4	9769	0	4	9777	100.00%	99.96%
<i>Klebsiella variicola</i>	Clinical	1	242	0	2	245	100.00%	99.18%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8260	0	3	8266	100.00%	99.96%
	Total	5	9767	0	5	9777	100.00%	99.95%
<i>Morganella morganii</i>	Clinical	2	243	0	0	245	100.00%	100.00%
	Contrived	2	1264	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8262	0	1	8266	100.00%	99.99%
	Total	7	9769	0	1	9777	100.00%	99.99%
<i>Proteus vulgaris</i>	Clinical	0	244	1	0	245	0.00%	100.00%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8256	0	7	8266	100.00%	99.92%
	Total	4	9765	1	7	9777	80.00%	99.93%
<i>Providencia rettgeri</i>	Clinical	0	244	0	1	245	-	99.59%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8260	0	3	8266	100.00%	99.96%
	Total	4	9769	0	4	9777	100.00%	99.96%
<i>Providencia stuartii</i>	Clinical	0	244	0	1	245	-	99.59%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8260	0	3	8266	100.00%	99.96%
	Total	4	9769	0	4	9777	100.00%	99.96%
<i>Raoultella ornithinolytica</i>	Clinical	0	245	0	0	245	-	100.00%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8260	0	3	8266	100.00%	99.96%
	Total	4	9770	0	3	9777	100.00%	99.97%
<i>Serratia marcescens</i>	Clinical	0	245	0	0	245	-	100.00%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8245	0	18	8266	100.00%	99.78%
	Total	4	9755	0	18	9777	100.00%	99.82%
<i>Acinetobacter baumannii</i>	Clinical	0	244	0	1	245	-	99.59%
	Contrived	4	1262	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8244	0	19	8266	100.00%	99.77%
	Total	7	9750	0	20	9777	100.00%	99.80%
	Clinical	0	245	0	0	245	-	100.00%

<i>Acinetobacter lwoffii</i>	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8260	0	3	8266	100.0%	100.0%
	Total	4	9770	0	3	9777	100.00%	99.97%
<i>Pseudomonas aeruginosa</i>	Clinical	0	244	0	1	245	-	100%
	Contrived	7	1259	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8260	0	3	8266	100.0%	100.0%
	Total	10	9763	0	4	9777	100.00%	99.96%
<i>Strenotrophomonas maltophilia</i>	Clinical	0	244	0	1	245	-	100%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8252	0	11	8266	100.0%	100.0%
	Total	4	9761	0	12	9777	100.00%	99.88%
<i>Aerococcus</i> species	Clinical	8	237	0	0	245	100.00%	100.00%
	Contrived	2	1264	0	0	1266	100.00%	100.00%
	<i>in silico</i>	21	8242	3	0	8266	87.50%	100.00%
	Total	31	9743	3	0	9777	91.18%	100.00%
Anginosus group Streptococci	Clinical	1	235	0	9	245	100.00%	96.31%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	5	8261	0	0	8266	100.00%	100.00%
	Total	7	9761	0	9	9777	100.00%	99.91%
<i>Corynebacterium urealyticum</i>	Clinical	4	241	0	0	245	100.00%	100.00%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8263	0	0	8266	100.00%	100.00%
	Total	8	9769	0	0	9777	100.00%	100.00%
<i>Enterococcus faecium</i>	Clinical	2	241	0	2	245	100.00%	99.18%
	Contrived	1	1003	0	0	1004	100.00%	100.00%
	<i>in silico</i>	3	8260	0	3	8266	100.00%	99.96%
	Total	6	9504	0	5	9515	100.00%	99.95%
Mitis group Streptococci	Clinical	0	245	0	0	245	-	100.00%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8262	0	1	8266	100.00%	99.99%
	Total	4	9772	0	1	9777	100.00%	99.99%
<i>Staphylococcus epidermidis</i>	Clinical	2	242	0	1	245	100.00%	99.59%
	Contrived	5	1261	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8263	0	0	8266	100.00%	100.00%
	Total	10	9766	0	1	9777	100.00%	99.99%
<i>Staphylococcus lugdunensis</i>	Clinical	0	245	0	0	245	-	100.00%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8263	0	0	8266	100.00%	100.00%
	Total	4	9773	0	0	9777	100.00%	100.00%
<i>Staphylococcus saprophyticus</i>	Clinical	1	242	0	2	245	100.00%	99.18%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8263	0	0	8266	100.00%	100.00%
	Total	5	9770	0	2	9777	100.00%	99.98%
Other Staphylococci	Clinical	0	245	0	0	245	-	100.00%
	Contrived	9	1257	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8263	0	0	8266	100.00%	100.00%
	Total	12	9765	0	0	9777	100.00%	100.00%
<i>Streptococcus agalactiae</i>	Clinical	4	241	0	0	245	100.00%	100.00%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8263	0	0	8266	100.00%	100.00%
	Total	8	9769	0	0	9777	100.00%	100.00%
<i>Anaerococcus vaginalis</i>	Clinical	0	245	0	0	245	-	100.00%
	Contrived	1	1265	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8263	0	0	8266	100.00%	100.00%
	Total	4	9773	0	0	9777	100.00%	100.00%
<i>Bacteroides fragilis</i>	Clinical	0	245	0	0	245	-	100.00%
	Contrived	7	1259	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8263	0	0	8266	100.00%	100.00%
	Total	10	9767	0	0	9777	100.00%	100.00%
<i>Prevotella</i> species	Clinical	7	238	0	0	245	100.00%	100.00%
	Contrived	7	1259	0	0	1266	100.00%	100.00%
	<i>in silico</i>	9	8257	0	0	8266	100.00%	100.00%
	Total	23	9754	0	0	9777	100.00%	100.00%
<i>Gardnerella vaginalis</i>	Clinical	11	228	0	6	245	100.00%	97.44%
	Contrived	7	1259	0	0	1266	100.00%	100.00%
	<i>in silico</i>	3	8263	0	0	8266	100.00%	100.00%
	Total	21	9750	0	6	9777	100.00%	99.94%
Negative Organisms	Clinical	15	230	0	0	245	100.00%	100.00%
	Contrived	223	1230	0	0	1266	100.00%	100.00%
	<i>in silico</i>	7977	289	0	0	8266	100.00%	100.00%
	Total	8215	1749	0	0	9777	100.00%	100.00%

Supplementary Table 6B. Performance Characteristics fo Key Fungal Urogenital Taxa Reported with BIOTIA-ID

Key pathogen	Method	TP	TN	FN	FP	Total	Sensitivity	Specificity
<i>Candida albicans</i>	Clinical	8	28	0	0	36	100.0%	100.0%
	Contrived	132	795	1	0	928	99.2%	100.0%
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total	149	4956	1	0	5106	99.3%	100.0%
<i>Candida auris</i>	Clinical	0	36	0	0	36	-	100.0%
	Contrived	129	797	0	2	928	100.0%	99.7%
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total	138	4966	0	2	5106	100.0%	100.0%
<i>Candida dubliniensis</i>	Clinical	0	36	0	0	36	-	100.0%
	Contrived	10	914	0	0	924	100.0%	100.0%
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total	19	5083	0	0	5102	100.0%	100.0%
<i>Candida glabrata</i>	Clinical	8	28	0	0	36	100.0%	100.0%
	Contrived	135	793	0	0	928	100.0%	100.0%
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total	152	4954	0	0	5106	100.0%	100.0%
<i>Candida guilliermondii</i>	Clinical	0	36	0	0	36	-	100.0%
	Contrived	10	918	0	0	928	100.0%	100.0%
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total	19	5087	0	0	5106	100.0%	100.0%
<i>Candida kefyr</i>	Clinical	0	36	0	0	36	-	100.0%
	Contrived	10	918	0	0	928	100.0%	100.0%
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total	19	5087	0	0	5106	100.0%	100.0%
<i>Candida krusei</i>	Clinical	1	35	0	0	36	-	100.0%
	Contrived	137	791	0	0	928	100.0%	100.0%
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total	147	4959	0	0	5106	100.0%	100.0%
<i>Candida lusitanae</i>	Clinical	0	36	0	0	36	-	100.0%
	Contrived	10	882	0	0	928	100.0%	100.0%
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total	19	5051	0	0	5106	100.0%	100.0%
<i>Candida parapsilosis</i>	Clinical	0	36	0	0	36	-	100.0%
	Contrived	139	789	0	0	928	100.0%	100.0%
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total	148	4958	0	0	5106	100.0%	100.0%
<i>Candida tropicalis</i>	Clinical	4	32	0	0	36	100.0%	100.0%
	Contrived	134	794	0	0	928	100.0%	100.0%
	<i>in silico</i>	9	4133	0	0	4142	100.0%	100.0%
	Total	147	4959	0	0	5106	100.0%	100.0%
Negative Organisms	Clinical	9	191	0	0	200	100.0%	100.0%
	Contrived	79	849	0	0	928	100.0%	100.0%
	<i>in silico</i>	4052	90	0	0	4142	100.0%	100.0%
	Total	4140	1130	0	0	5270	100.0%	100.0%