

Table of contents

\\USER

Holmes

PCY

CMRR Holmes_Yale_McLean_Jan2022

Localizer
AAHScout
Localizer_aligned
T1_vNav_setter
T1_MEMPRAGE_GSP_vNavTrk
T2_vNav_setter
T2_SPACE_GSP_vNavTrk
SpinEchoFieldMap_AP
SpinEchoFieldMap_PA
rfMRI_REST_AP
rfMRI_REST_PA
Stroop_BI1_AP
Stroop_BI2_PA
Hammer_BI1_AP
rfMRI_REST_AP
rfMRI_REST_PA

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\Localizer

TA: 9.2 s PM: REF Voxel size: 1.2×1.2×5.0 mmPAT: Off Rel. SNR: 1.00 : fl

Properties

Prio recon	On
Load images to viewer	Off
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
FoV read	300 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	40.0 ms
TE	3.00 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize, Elliptical filter
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	40.0 ms
TE	3.00 ms
MTC	Off
Magn. preparation	None
Flip angle	15 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Off

Resolution - Common

FoV read	300 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	256
Phase resolution	75 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
FoV read	300 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	40.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	2
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	3

Geometry - AutoAlign

Position	L0.0 A20.0 H0.0 mm
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	L0.0 A20.0 H0.0
L	0.0 mm
A	20.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slice-sel.

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
-----------------	------

Physio - Signal1

TR	40.0 ms
Concatenations	1
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	300 mm
FoV phase	100.0 %
Phase resolution	75 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	Off
------------------	-----

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	15 deg
Measurements	1
Contrasts	1
TR	40.0 ms
TE	3.00 ms

Sequence - Part 1

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Bandwidth	260 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	None

Sequence - Part 2

RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
Allowed delay	0 s

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\AAHScout

TA: 0:14 PM: REF Voxel size: 1.6×1.6×1.6 mmPAT: 3 Rel. SNR: 1.00 : fl

Properties

Prio recon	On
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.15 ms
TE	1.37 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	3.15 ms
TE	1.37 ms
Flip angle	8 deg

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
Base resolution	160
Phase resolution	100 %
Slice resolution	69 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Trajectory	Cartesian

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	24
Accel. factor 3D	1

Resolution - iPAT

Reference scan mode	Integrated
---------------------	------------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.15 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Flip angle	8 deg
Measurements	1
Time to center	6.2 s

Inline - Inline

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	8 deg
Measurements	1
Contrasts	1
TR	3.15 ms
TE	1.37 ms

Sequence - Part 1

Introduction	On
Dimension	3D

Sequence - Part 1

Asymmetric echo	Weak
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	540 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Normal
Excitation	Non-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
------	-----

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\Localizer_aligned

TA: 0:21 PM: REF Voxel size: 1.2×1.2×5.0 mmPAT: Off Rel. SNR: 1.00 : fl

Properties

Prio recon	On
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	2
Slices	7
Dist. factor	200 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	300 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	104.0 ms
TE	3.00 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize, Elliptical filter
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	104.0 ms
TE	3.00 ms
MTC	Off
Magn. preparation	None
Flip angle	15 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Off

Resolution - Common

FoV read	300 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	256
Phase resolution	75 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	2
Slices	7
Dist. factor	200 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
FoV read	300 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	104.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	2
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	3

Geometry - AutoAlign

Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slice-sel.

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
-----------------	------

Physio - Signal1

TR	104.0 ms
Concatenations	1
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	300 mm
FoV phase	100.0 %
Phase resolution	75 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	Off
------------------	-----

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	15 deg
Measurements	1
Contrasts	1
TR	104.0 ms
TE	3.00 ms

Sequence - Part 1

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Bandwidth	260 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	None

Sequence - Part 2

RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
Allowed delay	0 s

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\T1_vNav_setter

TA: 0.3 s PM: FIX Voxel size: 8.0×8.0×8.0 mmRel. SNR: 1.00 : ABCD

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	L0.0 P0.0 H50.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	32
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
TR	11.0 ms
TE	4.8 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	BC

Contrast - Common

TR	11.0 ms
TE	4.8 ms
MTC	Off
Flip angle	2 deg
Fat suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
Base resolution	32
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	6/8
Interpolation	Off

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	L0.0 P0.0 H50.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	32
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
TR	11.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	L0.0 P0.0 H50.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 H50.0
R	0.0 mm
A	0.0 mm
H	50.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	50.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	11.0 ms
Concatenations	1

Sequence - Part 1

Introduction	Off
Dimension	3D
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.27 ms
Bandwidth	4882 Hz/Px

Sequence - Part 2

EPI factor	32
RF pulse type	Normal
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On

Sequence - Special

Protocol filename	MPRAGE
-------------------	--------

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\T1_MEMPRAGE_GSP_vNavTrk

TA: 2:23 PM: FIX Voxel size: 1.2×1.2×1.2 mmPAT: 4 Rel. SNR: 1.00 : tfl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	144
FoV read	230 mm
FoV phase	100.0 %
Slice thickness	1.20 mm
TR	2200.0 ms
TE 1	1.57 ms
TE 2	3.39 ms
TE 3	5.21 ms
TE 4	7.03 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	2200.0 ms
TE 1	1.57 ms
TE 2	3.39 ms
TE 3	5.21 ms
TE 4	7.03 ms
Magn. preparation	Non-sel. IR
TI	1100 ms
Flip angle	7.0 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	230 mm
FoV phase	100.0 %
Slice thickness	1.20 mm

Resolution - Common

Base resolution	192
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	4
Ref. lines PE	32
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	On
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	144
FoV read	230 mm
FoV phase	100.0 %
Slice thickness	1.20 mm
TR	2200.0 ms
Multi-slice mode	Single shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	16.00 deg
Initial Orientation	Sagittal

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm

System - Miscellaneous

MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L0.0 P0.0 H6.0 mm
! Orientation	T > C-16.0
! Rotation	0.00 deg
! A >> P	208 mm
! R >> L	208 mm
! F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	2200.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. IR
TI	1100 ms
Fat suppr.	None
Dark blood	Off
FoV read	230 mm
FoV phase	100.0 %
Phase resolution	100 %

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off

Inline - MIP

MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Inline - Maplt

Save original images	On
Maplt	None
Flip angle	7.0 deg
Measurements	1
Contrasts	4
TR	2200.0 ms
TE 1	1.57 ms
TE 2	3.39 ms
TE 3	5.21 ms
TE 4	7.03 ms

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Off
Contrasts	4
Flow comp. 1	No
Multi-slice mode	Single shot
Echo spacing	9.4 ms
Bandwidth 1	650 Hz/Px
Bandwidth 2	650 Hz/Px
Bandwidth 3	650 Hz/Px
Bandwidth 4	650 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
Turbo factor	144

Sequence - Special

Readout polarity	Positive
Nav. location	Before
Apply moco to	nav only
Remeasure	0 TRs
Reacq. threshold	0.50
Feedback delay	80 ms
Moco ref. image	Use Temp Ref
K-space streaming	None
ABCD navigator	0.0 ms
Add. grad time	0.0 ms
Apply freq to	parent and nav
Averaging	RMS

Sequence - Assistant

Mode	Off
------	-----

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\T2_vNav_setter

TA: 0.3 s PM: FIX Voxel size: 8.0×8.0×8.0 mmRel. SNR: 1.00 : ABCD

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	L0.0 P0.0 H50.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	32
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
TR	13.0 ms
TE	6.1 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	BC

Contrast - Common

TR	13.0 ms
TE	6.1 ms
MTC	Off
Flip angle	2 deg
Fat suppr.	Water excit. normal

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
Base resolution	32
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	6/8
Interpolation	Off

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	L0.0 P0.0 H50.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	32
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	8.00 mm
TR	13.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	L0.0 P0.0 H50.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 H50.0
R	0.0 mm
A	0.0 mm
H	50.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	Water excit. normal
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	5.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	13.0 ms
Concatenations	1

Sequence - Part 1

Introduction	Off
Dimension	3D
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.27 ms
Bandwidth	4882 Hz/Px

Sequence - Part 2

EPI factor	32
RF pulse type	Normal
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On

Sequence - Special

Protocol filename	T2-SPACE
-------------------	----------

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\T2_SPACE_GSP_vNavTrk

TA: 2:37 PM: FIX Voxel size: 1.2×1.2×1.2 mmPAT: 2 Rel. SNR: 1.00 : spc

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	144
FoV read	230 mm
FoV phase	100.0 %
Slice thickness	1.20 mm
TR	2800 ms
TE	326 ms
Averages	1.0
Concatenations	1
Filter	Raw filter, Prescan Normalize
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	2800 ms
TE	326 ms
MTC	Off
Magn. preparation	None
Fat suppr.	None
Blood suppr.	Off
Restore magn.	Off

Contrast - Dynamic

Averages	1.0
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	230 mm
FoV phase	100.0 %
Slice thickness	1.20 mm
Base resolution	192
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Allowed
Slice partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	On
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	On
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	144
FoV read	230 mm
FoV phase	100.0 %
Slice thickness	1.20 mm
TR	2800 ms
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	16.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Restore magn.	Off
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off

System - Miscellaneous

Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L0.0 P0.0 H6.0 mm
! Orientation	T > C-16.0
! Rotation	0.00 deg
! A >> P	208 mm
! R >> L	208 mm
! F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
Trigger delay	0 ms
TR	2800 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	230 mm
FoV phase	100.0 %
Phase resolution	100 %

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Flow comp.	No
Echo spacing	3.5 ms
Adiabatic-mode	Off
Bandwidth	651 Hz/Px

Sequence - Part 2

Echo train duration	693 ms
RF pulse type	Normal
Gradient mode	Fast
Excitation	Non-sel.
Flip angle mode	T2 var
Turbo factor	212

Sequence - Special

Include nav.	On
Apply moco to	nav only
Remeasure	0 TRs
Reacq. threshold	0.50
Feedback delay	80 ms
Moco ref. image	Use Temp Ref
K-space streaming	None
ABCD navigator	On
Apply freq to	parent and nav

Sequence - Assistant

Allowed delay	0 s
---------------	-----

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\SpinEchoFieldMap_AP

TA: 8.0 s PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
TE	66.00 ms
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	8000 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 H6.0
L	0.0 mm
P	0.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-16.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8000 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\SpinEchoFieldMap_PA

TA: 8.0 s PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
TE	66.00 ms
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	8000 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	8000 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 H6.0
L	0.0 mm
P	0.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-16.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8000 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\rfMRI_REST_AP

TA: 6:40 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	488
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 H6.0
L	0.0 mm
P	0.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-16.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off

System - Adjustments

Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	488
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	DICOM
Triggering scheme	Standard

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\rfMRI_REST_PA

TA: 6:40 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	488
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 H6.0
L	0.0 mm
P	0.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-16.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off

System - Adjustments

Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	488
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	DICOM
Triggering scheme	Standard

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\Stroop_BI1_AP

TA: 6:58 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	510
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 H6.0
L	0.0 mm
P	0.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-16.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off

System - Adjustments

Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	510
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	DICOM
Triggering scheme	Standard

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\Stroop_BI2_PA

TA: 6:58 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	510
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 H6.0
L	0.0 mm
P	0.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-16.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off

System - Adjustments

Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	510
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	DICOM
Triggering scheme	Standard

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\Hammer_BI1_AP

TA: 6:44 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	493
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 H6.0
L	0.0 mm
P	0.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-16.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off

System - Adjustments

Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	493
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	DICOM
Triggering scheme	Standard

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\rfMRI_REST_AP

TA: 6:40 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	488
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 H6.0
L	0.0 mm
P	0.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-16.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off

System - Adjustments

Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	488
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	DICOM
Triggering scheme	Standard

\\USER\Holmes\PCY\CMRR Holmes_Yale_McLean_Jan2022\rfMRI_REST_PA

TA: 6:40 PM: REF Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : epfd

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
TE	37.00 ms
Multi-band accel. factor	8
Filter	None
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	800 ms
TE	37.00 ms
MTC	Off
Magn. preparation	None
Flip angle	52 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	488
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
------------------	-----

Resolution - Filter Image

Prescan Normalize	Off
-------------------	-----

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	800 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 H6.0
L	0.0 mm
P	0.0 mm
H	6.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-16.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off

System - Adjustments

Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H6.0 mm
Orientation	T > C-16.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.260801 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	800 ms
Multi-band accel. factor	8

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	488
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable freq. update	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	DICOM
Triggering scheme	Standard