

Supplementary Figure 1. Comparison between Wadsworth Center MIA and MSH ELISA. Serum samples were evaluated by ELISA and sent to the Wadsworth Center for MIA analysis. Individual MFI values for Abs reactive with the Nucleocapsid antigen (*A*) and the RBD antigen (*B*) were compared to ELISA endpoint titers. The Pearson Correlation Coefficient for the ELISA versus the N protein: r2 = 0.29, p<0.0001; ELISA versus RBD protein: r2 = 0.54, p<0.0001.

Supplementary Table 1. Comparison between Wadsworth Center MIA and MSH ELISA.

2398 Tested		NYS SARS CoV MIA				
		Positive	Indeterminate	Negative		
MSH ELISA (RBD antigen)	Positive	1120	199	249		
itee untigen)	Negative	29	19	782		
	nt at 6 SD: 79.3%)				
Percent Agreeme	It at 0 5D. 77.57					
Ŭ	ent at 3 SD: 87.6%					
Ŭ						
Percent Agreeme		•	ARS CoV MIA			
Percent Agreeme		•	ARS CoV MIA Indeterminate	Negative		
Ŭ		NYS S		Negative 65		

Serum samples were evaluated by ELISA and sent to the Wadsworth Center for MIA analysis. The MIA results were primarily reported based upon the SARS CoV-1 Nucleocapsid protein only (*top*) with a subset of specimens reported with both the Nucleocapsid and SARS-CoV-2 RBD proteins (*bottom*) as target antigens.

Day of Results	Number tested	COVID-19: Antibody Titers				
		2880	960	320	<320	
1	24	7	6	8	6	
2	47	7	12	18	10	
3	92	18	23	34	17	
4	223	47	69	76	31	
5	162	37	50	63	12	
6	138	48	51	27	12	
7	468	149	148	127	44	
8	278	84	82	80	22	
9	225	71	81	56	17	
Total	1657	468	532	489	168	
Percent	100%	28.2%	32.1%	29.5%	10.1%	

Supplementary Table 2. SARS-CoV-2 Ab Titers in Convalescent Sera.

Serum specimens from COVID-19 convalescent patients were screened at a 1:50 dilution for SARS-CoV-2 RBD reactivity using ELISA. Specimens reactive in the screen were then further tested by ELISA over a range of dilutions. The indicated titers are inverse dilutions of endpoint reactivity. Specimens were collected and assayed each day and individual days' results from a representative sampling of days are shown here.